

APPENDIX

Assessment of Methods to Collect and Analyze Perfluoroalkyl and Polyfluoroalkyl Substances (PFASs) in Air, Dust and Soil

Final Report

Agreement Number 19RD019

Principal Investigator: Asa Bradman, PhD, MS
Center for Environmental Research and Children's Health
University of California - Berkeley
February 2021

Prepared for the California Air Resources Board (CARB)

Report Authors

Asa Bradman, PhD
Rosemary Castorina, PhD
Teja Pattabhiraman
Anuroop Nirula
Monice Wong, B.A.
Sion Calabretta

Center for Environmental Research and Children's Health
University of California - Berkeley
1995 University Avenue, Suite 265
Berkeley, CA 94704

APPENDIX

Table of Contents

US EPA. 1999. Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air. Second Edition. Compendium Method TO-9A. Determination of Polychlorinated, Polybrominated and Brominated/Chlorinated Dibenzo-p-Dioxins and Dibenzofurans in Ambient Air.

US EPA. 1999. Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air. Second Edition. Compendium Method TO-13A. Determination of Polycyclic Aromatic Hydrocarbons (PAHs) in Ambient Air Using Gas Chromatography/Mass Spectrometry (GC/MS).

US EPA. 2021. Other Test Method 45 (OTM-45) Measurement of Selected Per- and Polyfluorinated Alkyl Substances from Stationary Sources

**Compendium of Methods
for the Determination of
Toxic Organic Compounds
in Ambient Air**

Second Edition

Compendium Method TO-9A

**Determination Of Polychlorinated,
Polybrominated And
Brominated/Chlorinated
Dibenzo-p-Dioxins And Dibenzofurans In
Ambient Air**

**Center for Environmental Research Information
Office of Research and Development
U.S. Environmental Protection Agency
Cincinnati, OH 45268**

January 1999

Method TO-9A Acknowledgements

This Method was prepared for publication in the *Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air, Second Edition* (EPA/625/R-96/010b), which was prepared under Contract No. 68-C3-0315, WA No. 3-10, by Midwest Research Institute (MRI), as a subcontractor to Eastern Research Group, Inc. (ERG), and under the sponsorship of the U.S. Environmental Protection Agency (EPA). Justice A. Manning, John Burckle, and Scott Hedges, Center for Environmental Research Information (CERI), and Frank F. McElroy, National Exposure Research Laboratory (NERL), all in the EPA Office of Research and Development, were responsible for overseeing the preparation of this method. Additional support was provided by other members of the Compendia Workgroup, which include:

- John O. Burckle, U.S. EPA, ORD, Cincinnati, OH
- James L. Cheney, Corps of Engineers, Omaha, NB
- Michael Davis, U.S. EPA, Region 7, KC, KS
- Joseph B. Elkins Jr., U.S. EPA, OAQPS, RTP, NC
- Robert G. Lewis, U.S. EPA, NERL, RTP, NC
- Justice A. Manning, U.S. EPA, ORD, Cincinnati, OH
- William A. McClenny, U.S. EPA, NERL, RTP, NC
- Frank F. McElroy, U.S. EPA, NERL, RTP, NC
- Heidi Schultz, ERG, Lexington, MA
- William T. "Jerry" Winberry, Jr., EnviroTech Solutions, Cary, NC

Method TO-9 was originally published in March of 1989 as one of a series of peer reviewed methods in the second supplement to "*Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air*," EPA 600/4-89-018. In an effort to keep these methods consistent with current technology, Method TO-9 has been revised and updated as Method TO-9A in this Compendium to incorporate new or improved sampling and analytical technologies.

This Method is the result of the efforts of many individuals. Gratitude goes to each person involved in the preparation and review of this methodology.

Author(s)

- Bob Harless, U.S. EPA, NERL, RTP, NC
- William T. "Jerry" Winberry, Jr., EnviroTech Solutions, Cary, NC
- Gil Radolovich, Midwest Research Institute, KC, MO
- Mark Horrigan, Midwest Research Institute, KC, MO

Peer Reviewers

- Audrey E. Dupuy, U.S. EPA, NSTL Station, MS
- Greg Jungclaus, Midwest Research Institute, KC, MO
- Stan Sleva, TRC, RTP, NC
- Robert G. Lewis, U.S. EPA, NERL, RTP, NC
- Lauren Drees, U.S. EPA, NRMRL, Cincinnati, OH

Finally, recognition is given to Frances Beyer, Lynn Kaufman, Debbie Bond, Cathy Whitaker, and Kathy Johnson of Midwest Research Institute's Administrative Services staff whose dedication and persistence during the development of this manuscript has enabled its production.

DISCLAIMER

This Compendium has been subjected to the Agency's peer and administrative review, and it has been approved for publication as an EPA document. Mention of trade names or commercial products does not constitute endorsement or recommendation for use.

Method TO-9A

Determination Of Polychlorinated, Polybrominated And Brominated/Chlorinated Dibenzo-p-Dioxins And Dibenzofurans In Ambient Air

TABLE OF CONTENTS

	<u>Page</u>
1. Scope	9A-1
2. Summary of Method	9A-2
3. Significance	9A-3
4. Safety	9A-3
5. Applicable Documents	9A-4
5.1 ASTM Standards	9A-4
5.2 EPA Documents	9A-4
5.3 Other Documents	9A-5
6. Definitions	9A-5
7. Interferences And Contamination	9A-9
8. Apparatus	9A-9
8.1 High-Volume Sampler	9A-9
8.2 High-Volume Sampler Calibrator	9A-9
8.3 High Resolution Gas Chromatograph-High Resolution Mass Spectrometer-Data System (HRGC-HRMS-DS)	9A-10
9. Equipment And Materials	9A-10
9.1 Materials for Sample Collection	9A-10
9.2 Laboratory Equipment	9A-11
9.3 Reagents and Other Materials	9A-11
9.4 Calibration Solutions and Solutions of Standards Used in the Method	9A-12
10. Preparation Of PUF Sampling Cartridge	9A-12
10.1 Summary of Method	9A-12
10.2 Preparation of Sampling Cartridge	9A-13
10.3 Procedure for Certification of PUF Cartridge Assembly	9A-13
10.4 Deployment of Cartridges for Field Sampling	9A-14
11. Assembly, Calibration And Collection Using Sampling System	9A-14
11.1 Description of Sampling Apparatus	9A-14
11.2 Calibration of Sampling System	9A-15
11.3 Sample Collection	9A-21

TABLE OF CONTENTS (continued)

	<u>Page</u>
12. Sample Preparation	9A-23
12.1 Extraction Procedure for Quartz Fiber Filters and PUF Plugs	9A-23
12.2 Cleanup Procedures	9A-23
12.3 Glassware Cleanup Procedures	9A-25
13. HRGC-HRMS System Performance	9A-25
13.1 Operation of HRGC-HRMS	9A-25
13.2 Colum Performance	9A-26
13.3 SIM Cycle Time	9A-26
13.4 Peak Separation	9A-26
13.5 Initial Calibration	9A-26
13.6 Criteria Required for Initial Calibration	9A-27
13.7 Continuing Calibration	9A-28
14. HRGC-HRMS Analysis And Operating Parameters	9A-28
14.1 Sample Analysis	9A-28
14.2 Identification Criteria	9A-29
14.3 Quantification	9A-29
14.4 Calculations	9A-30
14.5 Method Detection Limits (MDLs)	9A-31
14.6 2,3,7,8-TCDD Toxic Equivalents	9A-31
15. Quality Assurance/Quality Control (QA/QC)	9A-32
16. Report Format	9A-33
17. References	9A-34

METHOD TO-9A

Determination Of Polychlorinated, Polybrominated And Brominated/Chlorinated Dibenzo-p-Dioxins And Dibenzofurans In Ambient Air

1. Scope

1.1 This document describes a sampling and analysis method for the quantitative determination of polyhalogenated dibenzo-p-dioxins and dibenzofurans (PHDDs/PHDFs) in ambient air, which include the polychlorinated dibenzo-p-dioxins and dibenzofurans (PCDDs/PCDFs), polybrominated dibenzo-p-dioxins and dibenzofurans (PBDDs/PBDFs), and bromo/chloro dibenzo-p-dioxins and dibenzofurans (BCDDs/BCDFs). The method uses a high volume air sampler equipped with a quartz-fiber filter and polyurethane foam (PUF) adsorbent for sampling 325 to 400 m³ ambient air in a 24-hour sampling period. Analytical procedures based on high resolution gas chromatography-high resolution mass spectrometry (HRGC-HRMS) are used for analysis of the sample.

1.2 The sampling and analysis method was evaluated using mixtures of PHDDs and PHDFs, including the 2,3,7,8-substituted congeners (1,2). It has been used extensively in the U.S. Environmental Protection Agency (EPA) ambient air monitoring studies (3,4) for determination of PCDDs and PCDFs.

1.3 The method provides accurate quantitative data for tetra- through octa-PCDDs/PCDFs (total concentrations for each isomeric series).

1.4 Specificity is attained for quantitative determination of the seventeen 2,3,7,8-substituted PCDDs/PCDFs and specific 2,3,7,8-substituted PBDD/PBDF and BCDD/BCDF congeners.

1.5 Minimum detection limits (MDLs) in the range of 0.01 to 0.2 picograms/meter³ (pg/m³) can be achieved for these compounds in ambient air.

1.6 Concentrations as low as 0.2 pg/m³ can be accurately quantified.

1.7 The method incorporates quality assurance/quality control (QA/QC) measures in sampling, analysis, and evaluation of data.

1.8 The analytical procedures also have been used for the quantitative determination of these types of compounds in sample matrices such as stack gas emissions, fly ash, soil, sediments, water, and fish and human tissue (5-9).

1.9 The method is similar to methods used by other EPA, industry, commercial, and academic laboratories for determining PCDDs and PCDFs in various sample matrices (10-25). This method is an update of the original EPA Compendium Method TO-9, originally published in 1989 (26).

1.10 The method does not separately quantify gaseous PHDDs and PHDFs and particulate-associated PHDDs and PHDFs because some of the compounds volatilize from the filter and are collected by the PUF adsorbent. For example, most of the OCDD is collected by the filter and most of the TCDDs are collected by the PUF during sampling. PCDDs/PCDFs may be distributed between the gaseous and particle-adsorbed phases in ambient air. Therefore, the filter and PUF are combined for extraction in this method.

1.11 The sampling and analysis method is very versatile and can be used to determine other brominated and brominated/chlorinated dioxins and furans in the future when more analytical standards become available for use in the method. A recent modification of the sample preparation procedure provides the capability required to determine PCDDs, PCDFs, PCBs, and PAHs in the same sample (27).

2. Summary of Method

2.1 Quartz-fiber filters and glass adsorbent cartridges are pre-cleaned with appropriate solvents and dried in a clean atmosphere. The PUF adsorbent plugs are subjected to 4-hour Soxhlet extraction using an oversized extractor to prevent distortion of the PUF plug. The PUF plugs are then air dried in a clean atmosphere and installed in the glass cartridges. A 50 microliter (μL) aliquot of a 16 picogram/microliter ($\text{pg}/\mu\text{L}$) solution of $^{37}\text{Cl}_4$ -2,3,7,8-TCDD is spiked to the PUF in the laboratory prior to field deployment. (Different amounts and additional $^{13}\text{C}_{12}$ -labeled standards such as $^{13}\text{C}_{12}$ -1,2,3,6,7,8-HxCDF may also be used if desired.) The cartridges are then wrapped in aluminum foil to protect from light, capped with Teflon® end caps, placed in a cleaned labeled shipping container, and tightly sealed with Teflon® tap until needed.

2.2 For sampling, the quartz-fiber filter and glass cartridge containing the PUF are installed in the high-volume air sampler.

2.3 The high-volume sampler is then immediately put into operation, usually for 24 hours, to sample 325 to 400 m^3 ambient air.

[Note: Significant losses were not detected when duplicate samplers were operated 7 days and sampled 2660 m^3 ambient air (1-4).]

2.4 The amount of ambient air sampled is recorded at the end of the sampling session. Sample recovery involves placing the filter on top of the PUF. The glass cartridge is then wrapped with the original aluminum foil, capped with Teflon® end caps, placed back into the original shipping container, identified, and shipped to the analytical laboratory for sample processing.

2.5 Sample preparation typically is performed on a "set" of 12 samples, which consists of 9 test samples, a field blank, a method blank, and a matrix spike.

2.6 The filter and PUF are combined for sample preparation, spiked with 9 $^{13}\text{C}_{12}$ -labeled PCDD/PCDF and 4 PBDD/PBDF internal standards (28), and Soxhlet extracted for 16 hours. The extract is subjected to an acid/base clean-up procedure followed by clean-up on micro columns of silica gel, alumina, and carbon. The extract is then spiked with 0.5 ng $^{13}\text{C}_{12}$ -1,2,3,4-TCDD (to determine extraction efficiencies achieved for the $^{13}\text{C}_{12}$ -labeled internal standards) and then concentrated to 10 μL for HRGC-HRMS analysis in a 1 mL conical vial.

2.7 The set of sample extracts is subjected to HRGC-HRMS selected ion monitoring (SIM) analysis using a 60-m DB-5 or 60-m SP-2331 fused silica capillary column to determine the sampler efficiency, extraction efficiency, and the concentrations or the MDLs achieved for the PHDDs/PHDFs (28). Defined identification criteria and QA/QC criteria and requirements are used in evaluating the analytical data. The analytical results along with the volume of air sampled are used to calculate the concentrations of the respective tetra- through octa-isomers, the concentrations of the 2,3,7,8-chlorine or -bromine substituted isomers, or the MDLs. The concentrations and/or

MDLs are reported in pg/m^3 . The EPA toxicity equivalence factors (TEFs) can be used to calculate the 2,3,7,8-TCDD toxicity equivalents (TEQs) concentrations, if desired (18).

3. Significance

3.1 The PHDDs and PHDFs may enter the environment by two routes: (1) manufacture, use and disposal of specific chemical products and by-products and (2) the emissions from combustion and incineration processes. Atmospheric transport is considered to be a major route for widespread dispersal of these compounds in stack gas emissions throughout the environment. The PCDDs/PCDFs are found as complex mixtures of all isomers in emissions from combustion sources. The isomer profiles of PCDDs/PCDFs found in ambient air are similar to those found in combustion sources. Isomer profiles of PCDDs/PCDFs related to chemical products and by-products are quite different in that only a few specific and characteristic isomers are detectable, which clearly indicate they are not from a combustion source.

3.2 The 2,3,7,8-substituted PCDDs/PCDFs are considered to be the most toxic isomers. Fortunately, they account for the smallest percentage of the total PCDD/PCDF concentrations found in stack gas emissions from combustion sources and in ambient air. The 2,3,7,8-tetrachlorodibenzo-p-dioxin (2,3,7,8-TCDD), 1 of 22 TCDD isomers and the most toxic member of PCDDs/PCDFs, is usually found as a very minor component in stack gas emissions (0.5 to 10 percent of total TCDD concentration) and is seldom found in ambient air samples. All of the 2,3,7,8-substituted PCDDs/PCDFs are retained in tissue of life-forms such as humans, fish, and wildlife, and the non 2,3,7,8-substituted PCDDs/PCDFs are rapidly metabolized and/or excreted.

3.3 Attention has been focused on determining PHDDs/PHDFs in ambient air only in recent years. The analyses are time-consuming, complex, difficult, and expensive. Extremely sensitive, specific, and efficient analytical procedures are required because the analysis must be performed for very low concentrations in the pg/m^3 and sub pg/m^3 range. The MDLs, likewise, must be in the range of 0.01 to 0.2 pg/m^3 for the results to have significant meaning for ambient air monitoring purposes. The background level of total PCDDs/PCDFs detected in ambient air is usually in the range of 0.5 to 3 pg/m^3 , and the PBDFs is in the range of 0.1 to 0.2 pg/m^3 (2,3,14). Because PCDDs/PCDFs, PBDDs/PBDFs, and BCDDs/BCDFs can be formed by thermal reactions, there has been an increasing interest in ambient air monitoring, especially in the vicinities of combustion and incineration processes such as municipal waste combustors and resource recovery facilities (19,20). PBDDs/PBDFs can be created thermally (22,23), and they may also be formed in certain chemical processes (21). BCDDs/BCDFs have been detected in ash from combustion/incineration processes (9). The sampling and analysis method described here can be used in monitoring studies to accurately determine the presence or absence of pg/m^3 and sub pg/m^3 levels of these compounds in ambient air (26,27).

4. Safety

4.1 The 2,3,7,8-TCDD and other 2,3,7,8-chlorine or bromine substituted isomers are toxic and can pose health hazards if handled improperly. Techniques for handling radioactive and infectious materials are applicable to 2,3,7,8-TCDD and the other PHDDs and PHDFs. Only highly trained individuals who are thoroughly versed in appropriate laboratory procedures and familiar with the hazards of 2,3,7,8-TCDD should handle these substances. A good laboratory practice involves routine physical examinations and blood checks of employees working with 2,3,7,8-TCDD. It is the responsibility of the laboratory personnel to ensure that safe handling procedures are employed.

4.2 The toxicity or carcinogenicity of the other penta-, hexa-, hepta-, and octa-PHDDs/PHDFs with chlorine or bromine atoms in positions 2,3,7,8 are known to have similar, but lower, toxicities. However, each compound should be treated as a potential health hazard and exposure to these compounds must be minimized.

4.3 While the procedure specifies benzene as the extraction solution, many laboratories have substituted toluene for benzene (28). This is due to the carcinogenic nature of benzene. The EPA is presently studying the replacement of benzene with toluene.

4.4 A laboratory should develop a strict safety program for working with these compounds, which would include safety and health protocols; work performed in well ventilated and controlled access laboratory; maintenance of current awareness file of OSHA regulations regarding the safe handling of chemicals specified in the method; protective equipment; safety training; isolated work area; waste handling and disposal procedures; decontamination procedures; and laboratory wipe tests. Other safety practices as described in EPA Method 613, Section 4, July 1982 version, EPA Method 1613 Revision A, April 1990, Office of Water and elsewhere (29,30).

5. Applicable Documents

5.1 ASTM Standards

- Method D1365 *Definitions of Terms Relating to Atmospheric Sampling and Analysis*.
- Method E260 *Recommended Practice for General Gas Chromatography Procedures*.
- Method E355 *Practice for Gas Chromatography Terms and Relationships*.

5.2 EPA Documents

- *Quality Assurance Handbook for Air Pollution Measurement Systems, Volume II*, U. S. Environmental Protection Agency, EPA 600/R-94-038b, May 1994.
- *Protocol for the Analysis of 2,3,7,8-Tetrachlorodibenzo-p-Dioxin by High Resolution Gas Chromatography-High Resolution Mass Spectrometry*, U. S. Environmental Protection Agency, EPA 600/40-86-004, January 1986.
- "Evaluation of an EPA High Volume Air Sampler for Polychlorinated Dibenzo-p-Dioxins and Polychlorinated Dibenzofurans," undated report by Battelle under Contract No. 68-02-4127, Project Officers Robert G. Lewis and Nancy K. Wilson, U. S. Environmental Protection Agency, Research Triangle Park, North Carolina.
- *Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air: Method TO-9, Second Supplement*, U. S. Environmental Protection Agency, EPA 600/4-89-018, March 1989.
- *Technical Assistance Document for Sampling and Analysis of Toxic Organic Compounds in Ambient Air*, U. S. Environmental Protection Agency, EPA 600/4-83-027, June 1983.
- "Analytical Procedures and Quality Assurance for Multimedia Analysis of Polychlorinated Dibenzo-p-Dioxins and Dibenzofurans by High Resolution Gas Chromatography - Low Resolution Mass Spectrometry," U. S. Environmental Protection Agency/OSW, SW-846, RCRA 8280 HRGC-LRMS, January 1987.
- "Analytical Procedures and Quality Assurance for Multimedia Analysis of Polychlorinated Dibenzo-p-Dioxins and Dibenzofurans by High Resolution Gas Chromatography - High Resolution Mass Spectrometry," U. S. Environmental Protection Agency/OSW, SW-846, RCRA 8290 HRGC-HRMS, June 1987.

- Harless, R., "Analytical Procedures and Quality Assurance Plan for the Determination of PCDDs and PCDFs Ambient Air near the Rutland, Vermont Municipal Incinerator," Final Report, U. S. Environmental Protection Agency, AREAL, RTP, NC, 1988.
- *Feasibility of Environmental Monitoring and Exposure Assessment for a Municipal Waste Combustor: Rutland, Vermont Pilot Study*, U. S. Environmental Protection Agency, EPA 600/8-91/007, March 1991.
- "Method 23, Determination of Polychlorinated Dibenzo-p-Dioxins (PCDDs) and Dibenzofurans (PCDFs) from Stationary Sources." *Federal Register*, Vol. 56, No. 30, February 13, 1991.
- *Method 1613 Tetra- through Octa-Chlorinated Dioxins and Furans by Isotope Dilution HRGC-HRMS*, U. S. Environmental Protection Agency, Office of Solid Waste, Washington, DC, April 1990.

5.3 Other Documents

- "Operating Procedures for Model PS-1 Sampler," Graseby/General Metal Works, Inc., Village of Cleves, OH 45002 (800-543-7412).
- "Chicago Air Quality: PCB Air Monitoring Plan, Phase 2," IEAP/APC/86-011, Illinois Environmental Protection Agency, Division of Air Pollution Control, April 1986.
- "Operating Procedures for the Thermo Environmental Semi-volatile Sampler," Thermo Environmental Instruments, Inc. (formerly Wedding and Associates), 8 West Forge Parkway, Franklin, MA 02038 (508-520-0430).

6. Definitions

[Note: Definitions used in this document and any user-prepared Standard Operating Procedures (SOPs) should be consistent with those used in ASTM D1356. All abbreviations and symbols are defined within this document at the point of first use.]

6.1 Polychlorinated dibenzo-p-dioxins (PCDDs) and polychlorinated dibenzofurans (PCDFs)—compounds that contain from 1 to 8 chlorine atoms, resulting in a total of 75 PCDDs and 135 PCDFs. The structures are shown in Figure 1. The numbers of isomers at different chlorination levels are shown in Table 1. The seventeen 2,3,7,8-substituted PCDDs/PCDFs are shown in Table 2.

6.2 Polybrominated dibenzo-p-dioxins (PBDDs) and polybrominated dibenzofurans (PBDFs)—compounds that have the same structure and contain from 1 to 8 bromine atoms, resulting in a total of 75 PBDDs and 135 PBDFs. The structures and isomers are the same as those of the PCDDs/PCDFs shown in Figure 1 and Tables 1 and 2.

6.3 Brominated/chlorinated dibenzo-p-dioxins (BCDDs) and brominated/chlorinated dibenzofurans (BCDFs)—compounds with the same structures and may contain from 1 to 8 chlorine and bromine atoms, resulting in 1550 BCDD congeners and 3050 BCDF congeners.

6.4 Polyhalogenated dibenzo-p-dioxins (PHDDs) and polyhalogenated dibenzofurans (PHDFs)—dibenzo-p-dioxins and dibenzofurans substituted with 1 or more halogen atoms.

6.5 Isomer—compounds having the same number and type of halogen atoms, but substituted in different positions. For example, 2,3,7,8-TCDD and 1,2,3,4-TCDD are isomers. Additionally, there are 22 isomers that constitute the homologues of TCDDs.

6.6 Isomeric group—a group of dibenzo-p-dioxins or dibenzofurans having the same number of halogen atoms. For example, the tetra-chlorinated dibenzo-p-dioxins.

6.7 Internal Standard—is an isotopically-labeled analog that is added to all samples, including method blanks (process and field) and quality control samples, before extraction. They are used along with response factors to measure the concentration of the analytes. Nine PCDD/PCDF and 4 PBDD/PBDF internal standards are used in this method. There is one for each of the chlorinated dioxin and furan isomeric groups with a degree of halogenation ranging from four to eight, with the exception of OCDF.

6.8 High-Resolution Calibration Solutions (see Table 3)—solutions in tridecane containing known amounts of 17 selected PCDDs and PCDFs, 9 internal standards ($^{13}\text{C}_{12}$ -labeled PCDDs/PCDFs), 2 field standards, 4 surrogate standards, and 1 recovery standard. The set of 5 solutions is used to determine the instrument response of the unlabeled analytes relative to the $^{13}\text{C}_{12}$ -labeled internal standards and of the $^{13}\text{C}_{12}$ -labeled internal standards relative to the surrogate, field and recovery standards. Different concentrations and other standards may be used, if desired. Criteria for acceptable calibration as outlined in Section 13.5 should be met in order to use the analyte relative response factors.

6.9 Sample Fortification Solutions (see Table 4)—solutions (in isooctane) containing the $^{13}\text{C}_{12}$ -labeled internal standards that are used to spike all samples, field blanks, and process blanks before extraction. Brominated standards used only when desired.

6.10 Recovery Standard Solution (see Table 5)—Recovery Standard Solution (see Table 5)—an isooctane solution containing the $^{13}\text{C}_{12}$ -1,2,3,4-TCDD ($^{13}\text{C}_{12}$ -2,3,7,8,9-HxDD optional) recovery standards that are added to the extract before final concentration for HRGC-HRMS analysis to determine the recovery efficiencies achieved for the $^{13}\text{C}_{12}$ -labeled internal standards.

6.11 Air Sampler Field Fortification Solution (see Table 6)—an isooctane solution containing the $^{37}\text{Cl}_4$ -2,3,7,8-TCDD standard that is spiked to the PUF plugs prior to shipping them to the field for air sampling.

6.12 Surrogate Standard Solution (see Table 7)—an isooctane solution containing 4 $^{13}\text{C}_{12}$ -labeled standards that may be spiked to the filter or PUF prior to air sampling, to the sample prior to extraction, or to the sample extract before cleanup or before HRGC-HRMS analysis to determine sampler efficiency method efficiency or for identification purposes (28). Other standards and different concentrations may be used, if desired.

6.13 Matrix Spike and Method Spike Solutions (see Table 8)—isooctane solutions of native (non-labeled) PCDDs and PCDFs and PBDDs and PBDFs that are spiked to a clean PUF prior to extraction.

6.14 Sample Set—consists of nine test samples, field blank, method blank, and matrix spiked with native PHDDs/PHDFs. Sample preparation, HRGC-HRMS analysis, and evaluation of data is performed on a sample set.

6.15 Lab Control Spike—standard that is prepared during sample preparation and that contains exactly the same amounts of all of the labeled and unlabeled standards that were used in extraction and cleanup of the sample set for HRGC-HRMS analysis.

6.16 Field Blank—consists of a sample cartridge containing PUF and filter that is spiked with the filed fortification solution, shipped to the field, installed on the sampler, and passively exposed at the sampling area (the sampler is not operated). It is then sealed and returned to the laboratory for extraction, cleanup, and HRGC-HRMS analysis. It is treated in exactly the same manner as a test sample. A field blank is processed with each sampling episode. The field blank represents the background contributions from passive exposure to ambient air, PUF, quartz fiber filter, glassware, and solvents.

6.17 Laboratory Method Blank—represents the background contributions from glassware, extraction and cleanup solvents. A Soxhlet extractor is spiked with a solution of $^{13}\text{C}_{12}$ -labeled internal standards, extracted, cleaned up, and analyzed by HRGC-HRMS in exactly the same manner as the test samples.

6.18 Solvent Blank—an aliquot of solvent (the amount used in the method) that is spiked with the $^{13}\text{C}_{12}$ -labeled internal standards and concentrated to 60 μL for HRGC-HRMS analysis. The analysis provides the background contributions from the specific solvent.

6.19 GC Column Performance Evaluation Solution (see Table 9)—a solution containing a mixture of selected PCDD/PCDF isomers, including the first and last chromatographic eluters for each isomeric group. Used to demonstrate continued acceptable performance of the capillary column and to define the PCDD/PCDF retention time windows. Also includes a mixture of tetradoxin isomers that elute closest to 2,3,7,8-TCDD.

6.20 QA/QC Audit Samples—samples of PUF that contain known amounts of unlabeled PCDDs and PCDFs. These samples are submitted as "blind" test samples to the analytical laboratory. The analytical results can then be used to determine and validate the laboratory's accuracy, precision and overall analytical capabilities for determination of PCDDs/PCDFs.

6.21 Relative Response Factor—response of the mass spectrometer to a known amount of an analyte relative to a known amount of a labeled internal standard.

6.22 Method Blank Contamination—the method blank should be free of interferences that affect the identification and quantification of PHDDs and PHDFs. A valid method blank is an analysis in which all internal standard signals are characterized by S/N ratio greater than 10:1 and the MDLs are adequate for the study. The set of samples must be extracted and analyzed again if a valid method blank cannot be achieved.

6.23 Sample Rerun—additional cleanup of the extract and reanalysis of the extract.

6.24 Extract Reanalysis—analysis by HRGC-HRMS of another aliquot of the final extract.

6.25 Mass Resolution Check—a standard method used to demonstrate a static HRMS resolving power of 10,000 or greater (10 percent valley definition).

6.26 Method Calibration Limits (MCLs)—for a given sample size, a final extract volume, and the lowest and highest calibration solutions, the lower and upper MCLs delineate the region of quantitation for which the HRGC-HRMS system was calibrated with standard solutions.

6.27 HRGC-HRMS Solvent Blank—a 1 or 2 μL aliquot of solvent that is analyzed for tetra- through octa-PCDDs and PCDFs following the analysis of a sample that contains high concentrations of these compounds.

An acceptable solvent blank analysis (free of PHDDs/PHDFs) should be achieved before continuing with analysis of the test samples.

6.28 Sampler Spike (SS)—a sampler that is spiked with known amounts of the air sampler field fortification solution (see Table 6) and the matrix spike solutions (see Table 8) prior to operating the sampler for 24 hours to sample 325-400 std m³ ambient air. The results achieved for this sample can be used to determine the efficiency, accuracy and overall capabilities of the sampling device and analytical method.

6.29 Collocated Samplers (CS)—two samplers installed close together at the same site that can be spiked with known amounts of the air sampler field fortification solution (see Table 6) prior to operating the samplers for 24 hours to sample 325-400 std m³ ambient air. The analytical results for these two samples can be used to determine and evaluate efficiency, accuracy, precision, and overall capabilities of the sampling device and analytical method.

6.30 Congener—a term which refers to any one particular member of the same chemical family. As an example, there are 75 congeners of chlorinated dibenzo-p-dioxins. A specific congener is denoted by unique chemical notations. For example, 2,4,8,9-tetrachlorodibenzofuran is referred to as 2,4,8,9-TCDF.

6.31 Homologue—a term which refers to a group of structurally related chemicals that have the same degree of chlorination. For example, there are eight homologues of CDDs, monochlorinated through octochlorinated. Notation for homologous classes is as follows:

Class	Acronym	
Dibenzo-p-dioxin	D	
Dibenzofuran	F	
No. of halogens	Acronym	Example
1	M	
2	D	2,4-DCDD
3	Tr	
4	T	1,4,7,8-TCDD
5	Pe	
6	Hx	
7	Hp	
8	O	
1 through 8	CDDs and CDFs	

7. Interferences And Contamination

7.1 Any compound having a similar mass and mass/charge (m/z) ratio eluting from the HRGC column within ± 2 seconds of the PHDD/PHDF of interest is a potential interference. Also, any compound eluting from the HRGC column in a very high concentration will decrease sensitivity in the retention time frame. Some commonly encountered interferences are compounds that are extracted along with the PCDDs and PCDFs or other PHDDs/PHDFs, e.g., polychlorinated biphenyls (PCBs), methoxybiphenyls, polychlorinated diphenylethers, polychlorinated naphthalenes, DDE, DDT, etc. The cleanup procedures are designed to eliminate the majority of these substances. The capillary column resolution and mass spectrometer resolving power are extremely helpful in segregating any remaining interferences from PCDDs and PCDFs. The severity of an interference

problem is usually dependent on the concentrations and the mass spectrometer and chromatographic resolutions. However, polychlorinated diphenylethers are extremely difficult to resolve from PCDFs because they elute in retention time windows of PCDFs, and their fragment ion resulting from the loss of 2 chlorine atoms is identical to that of the respective PCDF. For example, the molecular ions of hexachlorodiphenylethers must be monitored to confirm their presence or absence in the analysis for TCDFs. This requirement also applies to the other PCDFs and PBDFs.

7.2 Since very low levels of PCDDs and PCDFs must be determined, the elimination of interferences is essential. High purity reagents and solvents must be used, and all equipment must be scrupulously cleaned. All materials, such as PUF, filter solvents, etc., used in the procedures are monitored and analyzed frequently to ensure the absence of contamination. Cleanup procedures must be optimized and performed carefully to minimize the loss of analyte compounds during attempts to increase their concentrations relative to other sample components. The analytical results achieved for the field blank, method blank, and method spike in a "set" of samples is extremely important in evaluating and validating the analytical data achieved for the test samples.

8. Apparatus

[Note: This method was developed using the PS-1 semi-volatile sampler provided by General Metal Works, Village of Cleves, OH as a guideline. EPA has experience in use of this equipment during various field monitoring programs over the last several years. Other manufacturers' equipment should work as well. However, modifications to these procedures may be necessary if another commercially available sampler is selected.]

8.1 High-Volume Sampler (see Figure 2). Capable of pulling ambient air through the filter/adsorbent cartridge at a flow rate of approximately 8 standard cubic feet per minute (scfm) (0.225 std m³/min) to obtain a total sample volume of greater than 325 scm over a 24-hour period. Major manufacturers are:

- Tisch Environmental, Village of Cleves, OH
- Andersen Instruments Inc., 500 Technology Ct., Smyrna, GA
- Thermo Environmental Instruments, Inc., 8 West Forge Parkway, Franklin, MA

8.2 High-Volume Sampler Calibrator. Capable of providing multipoint resistance for the high-volume sampler. Major manufacturers are:

- Tisch Environmental, Village of Cleves, OH
- Andersen Instruments Inc., 500 Technology Ct., Smyrna, GA
- Thermo Environmental Instruments, Inc., 8 West Forge Parkway, Franklin, MA

8.3 High Resolution Gas Chromatograph-High Resolution Mass Spectrometer-Data System (HRGC-HRMS-DS)

8.3.1 The GC should be equipped for temperature programming and all of the required accessories, such as gases and syringes, should be available. The GC injection port should be designed for capillary columns. Splitless injection technique, on-column injections, or moving needle injectors may be used. It is important to use the same technique and injection volume at all times.

8.3.2 The HRGC-HRMS interface, if used, should be constructed of fused silica tubing or all glass or glass lined stainless steel and should be able to withstand temperatures up to 340°C. The interface should not degrade the separation of PHDD/PHDF isomers achieved by the capillary column. Active sites or cold spots in the interface can cause peak broadening and peak tailing. The capillary column should be fitted directly into the HRMS ion source to avoid these types of problems. Graphite ferrules can adsorb PHDDs/PHDFs and cause problems. Therefore, Vespel® or equivalent ferrules are recommended.

8.3.3 The HRMS system should be operated in the electron impact ionization mode. The static resolving power of the instrument should be maintained at 10,000 or greater (10% valley definition). The HRMS should be operated in the selected ion monitoring (SIM) mode with a total cycle time of one second or less. At a minimum, the ions listed in Tables 10, 11, and 12 for each of the select ion monitoring (SIM) descriptors should be monitored. It is important to use the same set of ions for both calibration and sample analysis.

8.3.4 The data system should provide for control of mass spectrometer, data acquisition, and data processing. The data system should have the capability to control and switch to different sets of ions (descriptors/mass menus shown in Tables 10, 11, and 12) at different times during the HRGC-HRMS SIM analysis. The SIM traces/displays of ion signals being monitored can be displayed on the terminal in real time and sorted for processing. Quantifications are reported based on computer generated peak areas. The data system should be able to provide hard copies of individual ion chromatograms for selected SIM time intervals, and it should have the capability to allow measurement of noise on the baseline. It should also have the capability to acquire mass-spectral peak profiles and provide hard copies of the peak profiles to demonstrate the required mass resolution.

8.3.5 HRGC columns, such as the DB-5 (28) and SP-2331 fused silica capillary columns, and the operating parameters known to produce acceptable results are shown in Tables 13 and 14. Other types of capillary columns may also be used as long as the performance requirements can be successfully demonstrated.

9. Equipment And Materials

9.1 Materials for Sample Collection (see Figure 3a)

9.1.1 Quartz fiber filter. 102 millimeter bindless quartz microfiber filter, Whatman International Ltd, QMA-4.

9.1.2 Polyurethane foam (PUF) plugs. 3-inch thick sheet stock polyurethane type (density 0.022 g/cm³). The PUF should be of the polyether type used for furniture upholstery, pillows, and mattresses. The PUF cylinders (plugs) should be slightly larger in diameter than the internal diameter of the cartridge. Sources of equipment are Tisch Environmental, Village of Cleves, OH; University Research Glassware, 116 S. Merritt Mill Road, Chapel Hill, NC; Thermo Environmental Instruments, Inc., 8 West Forge Parkway, Franklin, MA; Supelco, Supelco Park, Bellefonte, PA; and SKC Inc., 334 Valley View Road, Eighty Four, PA (see Figure 3b).

9.1.3 Teflon® end caps. For sample cartridge. Sources of equipment are Tisch Environmental, Village of Cleves, OH; and University Research Glassware, 116 S. Merritt Mill Road, Chapel Hill, NC (see Figure 3b).

9.1.4 Sample cartridge aluminum shipping containers. For sample cartridge shipping. Sources of equipment are Tisch Environmental, Village of Cleves, OH; and University Research Glassware, 116 S. Merritt Mill Road, Chapel Hill, NC (see Figure 3b).

9.1.5 Glass sample cartridge. For sample collection. Sources of equipment are Tisch Environmental, Village of Cleves, OH; Thermo Environmental Instruments, Inc., 8 West Forge, Parkway, Franklin, MA; and University Research Glassware, 116 S. Merritt Mill Road, Chapel Hill, NC (see Figure 3b).

9.2 Laboratory Equipment

- 9.2.1 Laboratory hoods.
- 9.2.2 Drying oven.
- 9.2.3 Rotary evaporator. With temperature-controlled water bath.
- 9.2.4 Balances.
- 9.2.5 Nitrogen evaporation apparatus.
- 9.2.6 Pipettes. Disposal Pasteur, 150-mm long x 5-mm i.d.
- 9.2.7 Soxhlet apparatus. 500-mL.
- 9.2.8 Glass funnels.
- 9.2.9 Desiccator.
- 9.2.10 Solvent reservoir. 125-mL, Kontes, 12.35-cm diameter.
- 9.2.11 Stainless steel spoons and spatulas.
- 9.2.12 Glass wool. Extracted with methylene chloride, stored in clean jar.
- 9.2.13 Laboratory refrigerator.
- 9.2.14 Chromatographic columns.
- 9.2.15 Perfluorokerosenes.

9.3 Reagents and Other Materials

- 9.3.1 Sulfuric acid. Ultrapure, ACS grade, specific gravity 1.84, acid silica.
- 9.3.2 Sodium hydroxide. Potassium hydroxide, reagent grade, base silica.
- 9.3.3 Sodium sulfate.
- 9.3.4 Anhydrous, reagent grade.
- 9.3.5 Glass wool. Silanized, extracted with methylene chloride and hexane, and dried.
- 9.3.6 Diethyl ether. High purity, glass distilled.
- 9.3.7 Isooctane. Burdick and Jackson, glass-distilled.
- 9.3.8 Hexane. Burdick and Jackson, glass-distilled.
- 9.3.9 Toluene. Burdick and Jackson, glass-distilled, or equivalent.
- 9.3.10 Methylene chloride. Burdock and Jackson, chromatographic grade, glass distilled.
- 9.3.11 Acetone. Burdick and Jackson, high purity, glass distilled.
- 9.3.12 Tridecane. Aldrich, high purity, glass distilled.
- 9.3.13 Isooctane. Burdick and Jackson, high purity, glass distilled.
- 9.3.14 Alumina. Acid, pre-extracted (16-21 hours) and activated.
- 9.3.15 Silica gel. High purity grade, type 60, 70-230 mesh; extracted in a Soxhlet apparatus with methylene chloride (see Section 8.18) for 16-24 hours (minimum of 3 cycles per hour) and activated by heating in a foil-covered glass container for 8 hours at 130°C.
- 9.3.16 18 percent Carbopack C/Celite 545.
- 9.3.17 Methanol. Burdick and Jackson, high purity, glass distilled.
- 9.3.18 Nonane. Aldrich, high purity, glass distilled.
- 9.3.19 Benzene. High purity, glass distilled.

9.4 Calibration Solutions and Solutions of Standards Used in the Method

9.4.1 HRGC-HRMS Calibration Solutions (see Table 3). Solutions containing $^{13}\text{C}_{12}$ -labeled and unlabeled PCDDs and PCDFs at known concentrations are used to calibrate the instrument. These standards can be obtained from various commercial sources such as Cambridge Isotope Laboratories, 50 Frontage Road, Andover, MA 01810, 508-749-8000.

9.4.2 Sample Fortification Solutions (see Table 4). An isooctane solution (or nonane solution) containing the $^{13}\text{C}_{12}$ -labeled PCDD/PCDF and PBDD/PBDF internal standards at the listed concentrations. The internal standards are spiked to all samples prior to extraction and are used to measure the concentration of the unlabeled native analytes and to determine MDLs.

9.4.3 Recovery Standard Spiking Solution (see Table 5). An isooctane solution containing $^{13}\text{C}_{12}$ -1,2,3,4-TCDD at a concentration of 10 pg/ μL . Additional recovery standards may be used if desired.

9.4.4 Sampler Field Fortification Solution (see Table 6). An isooctane solution containing 10 pg/ μL $^{37}\text{Cl}_4$ -2,3,7,8-TCDD.

9.4.5 Surrogate Standards Solution (see Table 7). An isooctane solution containing the four $^{13}\text{C}_{12}$ -labeled standards at a concentration of 100 pg/ μL .

9.4.6 Matrix/Method Spike Solution (see Table 8). An isooctane solution containing the unlabeled PCDDs/PCDFs and PBDDs/PBDFs at the concentrations listed.

[Note: All PHDD/PHDF solutions listed above should be stored in a refrigerator at less than or equal to 4°C in the dark. Exposure of the solutions to light should be minimized.]

9.4.7 Column Performance Evaluation Solutions (see Table 9). Isooctane solutions of first and last chromatographic eluting isomers for each isomeric group of tetra- through octa-CDDs/CDFs. Also includes a mixture of tetradioxin isomers that elute closest to 2,3,7,8-TCDD.

10. Preparation Of PUF Sampling Cartridge

10.1 Summary of Method

10.1.1 This part of the procedure discusses pertinent information regarding the preparation and cleaning of the filter, adsorbents, and filter/adsorbent cartridge assembly. The separate batches of filters and adsorbents are extracted with the appropriate solvent.

10.1.2 At least one PUF cartridge assembly and one filter from each batch, or 10 percent of the batch, whichever is greater, should be tested and certified before the batch is considered for field use.

10.1.3 Prior to sampling, the cartridges are spiked with surrogate compounds.

10.2 Preparation of Sampling Cartridge

10.2.1 Bake the quartz filters at 400°C for 5 hours before use.

10.2.2 Set aside the filters in a clean container for shipment to the field or prior to combining with the PUF glass cartridge assembly for certification prior to field deployment.

10.2.3 The PUF plugs are 6.0-cm diameter cylindrical plugs cut from 3-inch sheet stock and should fit, with slight compression, in the glass cartridge, supported by the wire screen (see Figure 2). During cutting, rotate the die at high speed (e.g., in a drill press) and continuously lubricate with deionized or distilled water. Pre-cleaned PUF plugs can be obtained from commercial sources (see Section 9.1.2).

10.2.4 For initial cleanup, place the PUF plugs in a Soxhlet apparatus and extract with acetone for 16 hours at approximately 4 cycles per hour. When cartridges are reused, use diethyl ether/hexane (5 to 10 percent volume/volume [v/v]) as the cleanup solvent.

[Note: A modified PUF cleanup procedure can be used to remove unknown interference components of the PUF blank. This method consists of rinsing 50 times with toluene, acetone, and diethyl ether/hexane (5 to 10 percent v/v), followed by Soxhlet extraction. The extracted PUF is placed in a vacuum oven connected to a water aspirator and dried at room temperature for approximately 2 to 4 hours (until no solvent odor is detected). The extract from the Soxhlet extraction procedure from each batch may be analyzed to determine initial cleanliness prior to certification.]

10.2.5 Fit a nickel or stainless steel screen (mesh size 200/200) to the bottom of a hexane-rinsed glass sampling cartridge to retain the PUF adsorbents, as illustrated in Figure 2. Place the Soxhlet-extracted, vacuum-dried PUF (2.5-cm thick by 6.5-cm diameter) on top of the screen in the glass sampling cartridge using polyester gloves.

10.2.6 Wrap the sampling cartridge with hexane-rinsed aluminum foil, cap with the Teflon® end caps, place in a cleaned labeled aluminum shipping container, and seal with Teflon® tape. Analyze at least 1 PUF plug from each batch of PUF plugs using the procedures described in Section 10.3, before the batch is considered acceptable for field use. A level of 2 to 20 pg for tetra-, penta-, and hexa- and 40 to 150 pg for hepta- and octa-CDDs similar to that occasionally detected in the method blank (background contamination) is considered to be acceptable. Background levels can be reduced further, if necessary. Cartridges are considered clean for up to 30 days from date of certification when stored in their sealed containers.

10.3 Procedure for Certification of PUF Cartridge Assembly

10.3.1 Extract 1 filter and PUF adsorbent cartridge by Soxhlet extraction and concentrate using a Kuderna-Danish (K-D) evaporator for each lot of filters and cartridges sent to the field.

10.3.2 Assemble the Soxhlet apparatus. Charge the Soxhlet apparatus with 300 mL of the extraction solvent (10 percent v/v diethyl ether/hexane) and reflux for 2 hours. Let the apparatus cool, disassemble it, and discard the used extraction solvent. Transfer the filter and PUF glass cartridge to the Soxhlet apparatus (the use of an extraction thimble is optional).

[Note: The filter and adsorbent assembly are tested together in order to reach detection limits, to minimize cost and to prevent misinterpretation of the data. Separate analyses of the filter and PUF would not yield useful information about the physical state of most of the PHDDs and PHDFs at the time of sampling due to evaporative losses from the filter during sampling.]

10.3.3 Add 300 mL of diethyl ether/hexane (10 percent v/v) to the Soxhlet apparatus. Reflux the sample for 18 hours at a rate of at least 3 cycles per hour. Allow to cool; then disassemble the apparatus.

10.3.4 Assemble a K-D concentrator by attaching a 10-mL concentrator tube to a 500-mL evaporative flask.

10.3.5 Transfer the extract by pouring it through a drying column containing about 10 cm of anhydrous granular sodium sulfate and collect the extract in the K-D concentrator. Rinse the Erlenmeyer flask and column with 20 to 30 mL of 10 percent diethylether/hexane to complete the quantitative transfer.

10.3.6 Add 1 or 2 clean boiling chips and attach a 3-ball Snyder column to the evaporative flask. Pre-wet the Snyder column by adding about 1 mL of the extraction solvent to the top of the column. Place the K-D apparatus on a hot water bath (50°C) so that the concentrator tube is partially immersed in the hot water, and the entire lower rounded surface of the flask is bathed with hot vapor. Adjust the vertical position of the apparatus

and the water temperature as required to complete the concentration in one hour. At the proper rate of distillation, the balls of the column will actively chatter but the chambers will not flood with condensed solvent. When the apparent volume of liquid reaches approximately 5 mL, remove the K-D apparatus from the water bath and allow it to drain and cool for at least 5 minutes. Remove the Snyder column and rinse the flask and its lower joint into the concentrator tube with 5 mL of hexane. A 5-mL syringe is recommended for this operation.

10.3.7 Concentrate the extract to 1 mL, cleanup the extract (see Section 12.2.2), and analyze the final extract using HRGC-HRMS.

10.3.8 The level of target compounds must be less than or equal to 2 to 20 pg for tetra-, penta-, and hexa- and 40 to 150 pg for hepta- and octa-CDDs for each pair of filter and adsorbent assembly analyzed is considered to be acceptable.

10.4 Deployment of Cartridges for Field Sampling

10.4.1 Prior to field deployment, add surrogate compounds (i.e., chemically inert compounds not expected to occur in an environmental sample) to the center bed of the PUF cartridge, using a microsyringe. The surrogate compounds (see Table 3) must be added to each cartridge assembly.

10.4.2 Use the recoveries of the surrogate compounds to monitor for unusual matrix effects and gross sampling processing errors. Evaluate surrogate recovery for acceptance by determining whether the measured concentration falls within the acceptance limits.

11. Assembly, Calibration And Collection Using Sampling System

[Note: This method was developed using the PS-1 semi-volatile sampler provided by General Metal Works, Village of Cleves, OH as a guideline. EPA has experience in use of this equipment during various field monitoring programs over the last several years. Other manufacturers' equipment should work as well. However, modifications to these procedures may be necessary if another commercially available sampler is selected.]

11.1 Description of Sampling Apparatus

The entire sampling system is diagrammed in Figure 1. This apparatus was developed to operate at a rate of 4 to 10 scfm (0.114 to 0.285 std m³/min) and is used by EPA for high-volume sampling of ambient air. The method write-up presents the use of this device.

The sampling module (see Figure 2) consists of a filter and a glass sampling cartridge containing the PUF utilized to concentrate dioxins/furans from the air. A field portable unit has been developed by EPA (see Figure 4).

11.2 Calibration of Sampling System

Each sampler should be calibrated (1) when new, (2) after major repairs or maintenance, (3) whenever any audit point deviates from the calibration curve by more than 7 percent, (4) before/after each sampling event, and (5) when a different sample collection media, other than that which the sampler was originally calibrated to, will be used for sampling.

11.2.1 Calibration of Orifice Transfer Standard. Calibrate the modified high volume air sampler in the field using a calibrated orifice flow rate transfer standard. Certify the orifice transfer standard in the laboratory against a positive displacement rootsmeter (see Figure 5). Once certified, the recertification is performed rather

infrequently if the orifice is protected from damage. Recertify the orifice transfer standard performed once per year utilizing a set of five multiple resistance plates.

[Note: The set of five multihole resistance plates are used to change the flow through the orifice so that several points can be obtained for the orifice calibration curve. The following procedure outlines the steps to calibrate the orifice transfer standard in the laboratory.]

11.2.1.1 Record the room temperature (T_1 in $^{\circ}\text{C}$) and barometric pressure (P_b in mm Hg) on the Orifice Calibration Data Sheet (see Figure 6). Calculate the room temperature in K (absolute temperature) and record on Orifice Calibration Data Sheet.

$$T_1 \text{ in K} = 273^{\circ} + T_1 \text{ in } ^{\circ}\text{C}$$

11.2.1.2 Set up laboratory orifice calibration equipment as illustrated in Figure 5. Check the oil level of the rootsmeter prior to starting. There are 3 oil level indicators, 1 at the clear plastic end and 2 site glasses, 1 at each end of the measuring chamber.

11.2.1.3 Check for leaks by clamping both manometer lines, blocking the orifice with cellophane tape, turning on the high volume motor, and noting any change in the rootsmeter's reading. If the rootsmeter's reading changes, there is a leak in the system. Eliminate the leak before proceeding. If the rootsmeter's reading remains constant, turn off the hi-vol motor, remove the cellophane tape, and unclamp both manometer lines.

11.2.1.4 Install the 5-hole resistance plate between the orifice and the filter adapter.

11.2.1.5 Turn manometer tubing connectors 1 turn counter-clockwise. Make sure all connectors are open.

11.2.1.6 Adjust both manometer midpoints by sliding their movable scales until the zero point corresponds with the meniscus. Gently shake or tap to remove any air bubbles and/or liquid remaining on tubing connectors. (If additional liquid is required for the water manometer, remove tubing connector and add clean water.)

11.2.1.7 Turn on the high volume motor and let it run for 5 minutes to set the motor brushes. Turn the motor off. Insure manometers are set to zero. Turn the high volume motor on.

11.2.1.8 Record the time, in minutes, required to pass a known volume of air (approximately 200 to 300 ft^3 of air for each resistance plate) through the rootsmeter by using the rootsmeter's digital volume dial and a stopwatch.

11.2.1.9 Record both manometer readings-orifice water manometer (ΔH) and rootsmeter mercury manometer (ΔP) on Orifice Calibration Data Sheet (see Figure 6).

[Note: ΔH is the sum of the difference from zero (0) of the two column heights.]

11.2.1.10 Turn off the high volume motor.

11.2.1.11 Replace the 5-hole resistance plate with the 7-hole resistance plate.

11.2.1.12 Repeat Sections 11.2.1.3 through 11.2.1.11.

11.2.1.13 Repeat for each resistance plate. Note results on Orifice Calibration Data Sheet (see Figure 6). Only a minute is needed for warm-up of the motor. Be sure to tighten the orifice enough to eliminate any leaks. Also check the gaskets for cracks.

[Note: The placement of the orifice prior to the rootsmeter causes the pressure at the inlet of the rootsmeter to be reduced below atmospheric conditions, thus causing the measured volume to be incorrect. The volume measured by the rootsmeter must be corrected.]

11.2.1.14 Correct the measured volumes on the Orifice Calibration Data Sheet:

$$V_{\text{std}} = V_m \left(\frac{P_a - \Delta P}{P_{\text{std}}} \right) \left(\frac{T_{\text{std}}}{T_a} \right)$$

where:

- V_{std} = standard volume, std m³
- V_m = actual volume measured by the rootsmeter, m³
- P_a = barometric pressure during calibration, mm Hg
- ΔP = differential pressure at inlet to volume meter, mm Hg
- P_{std} = 760 mm Hg
- T_a = ambient temperature during calibration, K.

11.2.1.15 Record standard volume on Orifice Calibration Data Sheet.

11.2.1.16 The standard flow rate as measured by the rootsmeter can now be calculated using the following formula:

$$Q_{\text{std}} = \frac{V_{\text{std}}}{\theta}$$

where:

- Q_{std} = standard volumetric flow rate, std m³/min
- θ = elapsed time, min

11.2.1.17 Record the standard flow rates to the nearest 0.01 std m³/min.

11.2.1.18 Calculate and record $\sqrt{\Delta H (P_1/P_{\text{std}})(298/T_1)}$ value for each standard flow rate.

11.2.1.19 Plot each $\sqrt{\Delta H (P_1/P_{\text{std}})(298/T_1)}$ value (y-axis) versus its associated standard flow rate (x-axis) on arithmetic graph paper and draw a line of best fit between the individual plotted points.

[Note: This graph will be used in the field to determine standard flow rate.]

11.2.2 Calibration of the High Volume Sampling System Utilizing Calibrated Orifice Transfer Standard

For this calibration procedure, the following conditions are assumed in the field:

- The sampler is equipped with an valve to control sample flow rate.
- The sample flow rate is determined by measuring the orifice pressure differential, using a magnehelic gauge.
- The sampler is designed to operate at a standardized volumetric flow rate of 8 ft³/min (0.225 m³/min), with an acceptable flow rate range within 10 percent of this value.
- The transfer standard for the flow rate calibration is an orifice device. The flow rate through the orifice is determined by the pressure drop caused by the orifice and is measured using a "U" tube water manometer or equivalent.
- The sampler and the orifice transfer standard are calibrated to standard volumetric flow rate units (scfm or scmm).

- An orifice transfer standard with calibration traceable to NIST is used.
- A "U" tube water manometer or equivalent, with a 0- to 16-inch range and a maximum scale division of 0.1 inch, will be used to measure the pressure in the orifice transfer standard.
- A magnehelic gauge or equivalent, with a 9- to 100-inch range and a minimum scale division of 2 inches for measurements of the differential pressure across the sampler's orifice is used.
- A thermometer capable of measuring temperature over the range of 32° to 122°F (0° to 50°C) to ±2°F (±1°C) and referenced annually to a calibrated mercury thermometer is used.
- A portable aneroid barometer (or equivalent) capable of measuring ambient barometric pressure between 500 and 800 mm Hg (19.5 and 31.5 in. Hg) to the nearest mm Hg and referenced annually to a barometer of known accuracy is used.
- Miscellaneous handtools, calibration data sheets or station log book, and wide duct tape are available.

11.2.2.1 Monitor the airflow through the sampling system with a venturi/Magnehelic assembly, as illustrated in Figure 7. Set up the calibration system as illustrated in Figure 7. Audit the field sampling system once per quarter using a flow rate transfer standard, as described in the EPA *High Volume-Sampling Method, 40 CFR 50, Appendix B*. Perform a single-point calibration before and after each sample collection, using the procedures described in Section 11.2.3.

11.2.2.2 Prior to initial multi-point calibration, place an empty glass cartridge in the sampling head and activate the sampling motor. Fully open the flow control valve and adjust the voltage variator so that a sample flow rate corresponding to 110 percent of the desired flow rate (typically 0.20 to 0.28 m³/min) is indicated on the Magnehelic gauge (based on the previously obtained multipoint calibration curve). Allow the motor to warm up for 10 minutes and then adjust the flow control valve to achieve the desired flow rate. Turn off the sampler. Record the ambient temperature and barometric pressure on the Field Calibration Data Sheet (see Figure 8).

11.2.2.3 Place the orifice transfer standard on the sampling head and attach a manometer to the tap on the transfer standard, as illustrated in Figure 7. Properly align the retaining rings with the filter holder and secure by tightening the three screw clamps. Connect the orifice transfer standard by way of the pressure tap to a manometer using a length of tubing. Set the zero level of the manometer or magnehelic. Attach the magnehelic gauge to the sampler venturi quick release connections. Adjust the zero (if needed) using the zero adjust screw on face of the gauge.

11.2.2.4 To leak test, block the orifice with a rubber stopper, wide duct tape, or other suitable means. Seal the pressure port with a rubber cap or similar device. Turn on the sampler.

Caution: Avoid running the sampler from too long a time with the orifice blocked. This precaution will reduce the chance that the motor will be overheated due to the lack of cooling air. Such overheating can shorten the life of the motor.

11.2.2.5 Gently rock the orifice transfer standard and listen for a whistling sound that would indicate a leak in the system. A leak-free system will not produce an upscale response on the sampler's magnehelic. Leaks are usually caused either by damaged or missing gaskets by cross-threading and/or not screwing sample cartridge together tightly. All leaks must be eliminated before proceeding with the calibration. When the sample is determined to be leak-free, turn off the sampler and unblock the orifice. Now remove the rubber stopper or plug from the calibrator orifice.

11.2.2.6 Turn the flow control valve to the fully open position and turn the sampler on. Adjust the flow control valve until a Magnehelic reading of approximately 70 in. is obtained. Allow the Magnehelic and manometer readings to stabilize and record these values on the Field Calibration Data Sheet (see Figure 8).

11.2.2.7 Record the manometer reading under Y1 and the Magnehelic reading under Y2 on the Field Calibration Data Sheet. For the first reading, the Magnehelic should still be at 70 inches as set above.

11.2.2.8 Set the magnehelic to 60 inches by using the sampler's flow control valve. Record the manometer (Y1) and Magnehelic (Y2) readings on the Field Calibration Data Sheet.

11.2.2.9 Repeat the above steps using Magnehelic settings of 50, 40, 30, 20, and 10 inches.

11.2.2.10 Turn the voltage variator to maximum power, open the flow control valve, and confirm that the Magnehelic reads at least 100 inches. Turn off the sampler and confirm that the magnehelic reads zero.

11.2.2.11 Read and record the following parameters on the Field Calibration Data Sheet. Record the following on the calibration data sheet:

Data, job number, and operator's signature;

- Sampler serial number;
- Ambient barometric pressure; and
- Ambient temperature.

11.2.2.12 Remove the "dummy" cartridge and replace with a sample cartridge.

11.2.2.13 Obtain the Manufacturer High Volume Orifice Calibration Certificate.

11.2.2.14 If not performed by the manufacturer, calculate values for each calibrator orifice static pressure (Column 6, inches of water) on the manufacturer's calibration certificate using the following equation:

$$\sqrt{\Delta H(P_a/760)(298/[T_a + 273])}$$

where:

P_a = the barometric pressure (mm Hg) at time of manufacturer calibration, mm Hg

T_a = temperature at time of calibration, °C

11.2.2.15 Perform a linear regression analysis using the values in Column 7 of the manufacturer High Volume Orifice Calibration Certificate for flow rate (Q_{STD}) as the "X" values and the calculated values as the Y values. From this relationship, determine the correlation (CC1), intercept (B1), and slope (M1) for the Orifice Transfer Standard.

11.2.2.16 Record these values on the Field Calibration Data Sheet (see Figure 8).

11.2.2.17 Using the Field Calibration Data Sheet values (see Figure 8), calculate the Orifice Manometer Calculated Values (Y3) for each orifice manometer reading using the following equation:

Y3 Calculation

$$Y3 = [Y1(P_a/760)(298/[T_a + 273])]^{1/2}$$

11.2.2.18 Record the values obtained in Column Y3 on the Field Calibration Data Sheet (see Figure 8).

11.2.2.19 Calculate the Sampler Magnehelic Calculate Values (Y4) using the following equation:

Y4 Calculation

$$Y4 = [Y2(P_a/760)(298/[T_a + 273])]^{1/2}$$

11.2.2.20 Record the value obtained in Column Y4 on the Field Calibration Data Sheet (see Figure 8).

11.2.2.21 Calculate the Orifice Flow Rate (X1) in scm, using the following equation:

X1 Calculation

$$X1 = \frac{Y3 - B1}{M1}$$

11.2.2.22 Record the values obtained in Column X1, on the Field Calibration Data Sheet (see Figure 8).

11.2.2.23 Perform a linear regression of the values in Column X1 (as X) and the values in Column Y4 (as Y). Record the relationship for correlation (CC2), intercept (B2), and slope (M2) on the Field Calibration Data Sheet.

11.2.2.24 Using the following equation, calculate a set point (SP) for the manometer to represent a desired flow rate:

$$\text{Set point (SP)} = [(\text{Expected } P_a)/(\text{Expected } T_a)(T_{\text{std}}/P_{\text{std}})][M2 (\text{Desired flow rate}) + B2]^2$$

where:

P_a = Expected atmospheric pressure (P_a), mm Hg

T_a = Expected atmospheric temperature (T_a), °C

M2 = Slope of developed relationship

B2 = Intercept of developed relationship

T_{std} = Temperature standard, 25°C

P_{std} = Pressure standard, 760 mm Hg

11.2.2.25 During monitoring, calculate a flow rate from the observed Magnehelic reading using the following equations:

$$Y5 = [\text{Average Magnehelic Reading } (\Delta H) (P_a/T_a)(T_{\text{std}}/P_{\text{std}})]^{1/4}$$

$$X2 = \frac{Y5 - B2}{M2}$$

where:

Y5 = Corrected Magnehelic reading

X2 = Instant calculated flow rate, scm

11.2.2.26 The relationship in calibration of a sampling system between Orifice Transfer Standard and flow rate through the sampler is illustrated in Figure 9.

11.2.3 Single-Point Audit of the High Volume Sampling System Utilizing Calibrated Orifice Transfer Standard

Single point calibration checks are required as follows:

- Prior to the start of each 24-hour test period.
- After each 24-hour test period. The post-test calibration check may serve as the pre-test calibration check for the next sampling period if the sampler is not moved.
- Prior to sampling after a sample is moved.

For samplers, perform a calibration check for the operational flow rate before each 24-hour sampling event and when required as outlined in the user quality assurance program. The purpose of this check is to track the sampler's calibration stability. Maintain a control chart presenting the percentage difference between a sampler's indicated and measured flow rates. This chart provides a quick reference of sampler flow-rate drift problems and is useful for tracking the performance of the sampler. Either the sampler log book or a data sheet will be used

to document flowcheck information. This information includes, but is not limited to, sampler and orifice transfer standard serial number, ambient temperature, pressure conditions, and collected flow-check data.

In this subsection, the following is assumed:

- The flow rate through a sampler is indicated by the orifice differential pressure;
- Samplers are designed to operate at an actual flow rate of 8 scfm, with a maximum acceptable flow-rate fluctuation range of ± 10 percent of this value;
- The transfer standard will be an orifice device equipped with a pressure tap. The pressure is measured using a manometer; and
- The orifice transfer standard's calibration relationship is in terms of standard volumetric flow rate (Q_{std}).

11.2.3.1 Perform a single point flow audit check before and after each sampling period utilizing the Calibrated Orifice Transfer Standard (see Section 11.2.1).

11.2.3.2 Prior to single point audit, place a "dummy" glass cartridge in the sampling head and activate the sampling motor. Fully open the flow control valve and adjust the voltage variator so that a sample flow rate corresponding to 110 percent of the desired flow rate (typically 0.19 to 0.28 m³/min) is indicated on the Magnehelic gauge (based on the previously obtained multipoint calibration curve). Allow the motor to warm up for 10 minutes and then adjust the flow control valve to achieve the desired flow rate. Turn off the sampler. Record the ambient temperature and barometric pressure on a Field Test Data Sheet (see Figure 10).

11.2.3.3 Place the flow rate transfer standard on the sampling head.

11.2.3.4 Properly align the retaining rings with the filter holder and secure by tightening the 3 screw clamps. Connect the flow rate transfer standard to the manometer using a length of tubing.

11.2.3.5 Using tubing, attach 1 manometer connector to the pressure tap of the transfer standard. Leave the other connector open to the atmosphere.

11.2.3.6 Adjust the manometer midpoint by sliding the movable scale until the zero point corresponds with the water meniscus. Gently shake or tap to remove any air bubbles and/or liquid remaining on tubing connectors. (If additional liquid is required, remove tubing connector and add clean water.)

11.2.3.7 Turn on high-volume motor and let run for 5 minutes.

11.2.3.8 Record the pressure differential indicated, ΔH , in inches of water, on the Field Test Data Sheet. Be sure stable ΔH has been established.

11.2.3.9 Record the observed Magnehelic gauge reading, in inches of water, on the Field Test Data Sheet. Be sure stable ΔM has been established.

11.2.3.10 Using previous established Orifice Transfer Standard curve, calculate Q_{xs} (see Section 11.2.2.23).

11.2.3.11 This flow should be within ± 10 percent of the sampler set point, normally, 8 ft³. If not, perform a new multipoint calibration of the sampler.

11.2.3.12 Remove Flow Rate Transfer Standard and dummy adsorbent cartridge.

11.3 Sample Collection

11.3.1 General Requirements

11.3.1.1 The sampler should be located in an unobstructed area, at least 2 meters from any obstacle to air flow. The exhaust hose should be stretched out in the downwind direction to prevent recycling of air into the sample head.

11.3.1.2 All cleaning and sample module loading and unloading should be conducted in a controlled environment, to minimize any chance of potential contamination.

11.3.1.3 When new or when using the sampler at a different location, all sample contact areas need to be cleared. Use triple rinses of reagent grade hexane or methylene chloride contained in Teflon® rinse bottles. Allow the solvents to evaporate before loading the PUF modules.

11.3.2 Preparing Cartridge for Sampling

11.3.2.1 Detach the lower chamber of the cleaned sample head. While wearing disposable, clean, lint-free nylon, or powder-free surgical gloves, remove a clean glass adsorbent module from its shipping container. Remove the Teflon® end caps. Replace the end caps in the sample container to be reused after the sample has been collected.

11.3.2.2 Insert the glass module into the lower chamber and tightly reattach the lower chambers to the module.

11.3.2.3 Using clean rinsed (with hexane) Teflon-tipped forceps, carefully place a clean conditioned fiber filter atop the filter holder and secure in place by clamping the filter holder ring over the filter. Place the aluminum protective cover on top of the cartridge head. Tighten the 3 screw clamps. Ensure that all module connections are tightly assembled. Place a small piece of aluminum foil on the ball-joint of the sample cartridge to protect from back-diffusion of semi-volatile into the cartridge during transporting to the site.

[Note: Failure to do so could result in air flow leaks at poorly sealed locations which could affect sample representativeness.]

11.3.2.4 Place in a carrying bag to take to the sampler.

11.3.3 Collection

11.3.3.1 After the sampling system has been assembled, perform a single point flow check as described in Sections 11.2.3.

11.3.3.2 With the empty sample module removed from the sampler, rinse all sample contact areas using reagent grade hexane in a Teflon® squeeze bottle. Allow the hexane to evaporate from the module before loading the samples.

11.3.3.3 With the sample cartridge removed from the sampler and the flow control valve fully open, turn the pump on and allow it to warm-up for approximately 5 minutes.

11.3.3.4 Attach a "dummy" sampling cartridge loaded with the exact same type of filter and PUF media to be used for sample collection.

11.3.3.5 Turn the sampler on and adjust the flow control valve to the desired flow as indicated by the Magnehelic gauge reading determined in Section 11.2.2.4. Once the flow is properly adjusted, take extreme care not to inadvertently alter its setting.

11.3.3.6 Turn the sampler off and remove both the "dummy" module. The sampler is now ready for field use.

11.3.3.7 Check the zero reading of the sampler Magnehelic. Record the ambient temperature, barometric pressure, elapsed time meter setting, sampler serial number, filter number, and PUF cartridge number on the Field Test Data Sheet (see Figure 10). Attach the loaded sampler cartridge to the sampler.

11.3.3.8 Place the voltage variator and flow control valve at the settings used in Section 11.3.2, and the power switch. Activate the elapsed time meter and record the start time. Adjust the flow (Magnehelic setting), if necessary, using the flow control valve.

11.3.3.9 Record the Magnehelic reading every 6 hours during the sampling period. Use the calibration factors (see Section 11.2.2.23) to calculate the desired flow rate. Record the ambient temperature, barometric pressure, and Magnehelic reading at the beginning and during sampling period.

11.3.4 Sample Recovery

11.3.4.1 At the end of the desired sampling period, turn the power off. Carefully remove the sampling head containing the filter and adsorbent cartridge to a clean area.

11.3.4.2 While wearing disposable lint free nylon or surgical gloves, remove the PUF cartridge from the lower module chamber and lay it on the retained aluminum foil in which the sample was originally wrapped.

11.3.4.3 Carefully remove the glass fiber filter from the upper chamber using clean Teflon®-tipped forceps.

11.3.4.4 Fold the filter in half twice (sample side inward) and place it in the glass cartridge atop the PUF.

11.3.4.5 Wrap the combined samples in the original hexane rinsed aluminum foil, attached Teflon® end caps and place them in their original aluminum sample container. Complete a sample label and affix it to the aluminum shipping container.

11.3.4.6 Chain-of-custody should be maintained for all samples. Store the containers at <4°C and protect from light to prevent possibly photo-decomposition of collected analytes. If the time span between sample collection and laboratory analysis is to exceed 24 hours, refrigerate sample.

11.3.4.7 Perform a final calculated sample flow check using the calibration orifice, as described in Section 11.3.2. If calibration deviates by more than 10 percent from the initial reading, mark the flow data for that sample as suspect and inspect and/or remove from service.

11.3.4.8 Return at least 1 field filter/PUF blank to the laboratory with each group of samples. Treat a field blank exactly as the sample except that no air is drawn through the filter/adsorbent cartridge assembly.

11.3.4.9 Ship and store samples under ice (<4°C) until receipt at the analytical laboratory, after which it should be refrigerated at less than or equal to 4°C. Extraction must be performed within seven days of sampling and analysis within 40 days after extraction.

12. Sample Preparation

12.1 Extraction Procedure for Quartz Fiber Filters and PUF Plugs

12.1.1 Take the glass sample cartridge containing the PUF plug and quartz fiber filter out of the shipping container and place it in a 43-mm x 123-mm Soxhlet extractor. Add 10 µL of ¹³C₁₂-labeled sample fortification solution (see Table 4) to the sample. Put the thimble into a 50 mm Soxhlet extractor fitted with a 500 mL boiling flask containing 275 mL of benzene.

[Note: While the procedure specifies benzene as the extraction solution, many laboratories have substituted toluene for benzene because of the carcinogenic nature of benzene (28). The EPA is presently studying the replacement of benzene with toluene.]

12.1.2 Place a small funnel in the top of the Soxhlet extractor, making sure that the top of the funnel is inside the thimble. Rinse the inside of the corresponding glass cylinder into the thimble using approximately 25 mL of benzene. Place the extractor on a heating mantel. Adjust the heat until the benzene drips at a rate of 2 drops per second and allow to flow for 16 hours. Allow the apparatus to cool.

12.1.3 Remove the extractor and place a 3-bulb Snyder column onto the flask containing the benzene extract. Place on a heating mantel and concentrate the benzene to 25 mL (do not let go to dryness). Add 100 mL of hexane and again concentrate to 25 mL. Add a second 100 mL portion of hexane and again concentrate to 25 mL.

12.1.4 Let cool and add 25 mL hexane. The extract is ready for acid/base cleanup at this point.

12.2 Cleanup Procedures

12.2.1 Acid/Base Cleanup. Transfer the hexane extract to a 250 mL separatory funnel with two 25-mL portions of hexane. Wash the combined hexane with 30 ml of 2 N potassium hydroxide. Allow layers to separate and discard the aqueous layer. Repeat until no color is visible in the aqueous layer, up to a maximum of 4 washes. Partition the extract against 50 ml of 5% sodium chloride solution. Discard the aqueous layer. Carefully add 50 mL of concentrated sulfuric acid. Shake vigorously for 1 minute, allow layers to separate, and discard the acid layer. Repeat the acid wash until no color is visible in the aqueous layer, up to a maximum of 4 washes. Partition the extract against 50 ml of 5% sodium chloride solution. Discard the aqueous layer. Transfer the hexane through a 42-mm x 160-mm filter funnel containing a plug of glass wool and 3-cm of sodium sulfate into a 250-mL Kuderna-Danish (KD) concentrator fitter with a 15-mL catch tube. Rinse the filter funnel with two 25 mL portions of hexane. Place a 3-bulb Snyder column on the KD concentrator and concentrate on a steam bath to 1-2 mL. The extract is ready for the alumina column cleanup at this point, but it can be sealed and stored in the dark, if necessary. An extract that contains obvious contamination, such as yellow or brown color, is subjected to the silica column cleanup prior to the alumina cleanup.

12.2.2 Silica Column Preparation. Gently tamp a plug of glass wool into the bottom of a 5.75-inch (14.6 cm) disposable Pasteur pipette. Pour prewashed 100-200 mesh Bio-Sil®A (silica gel) into the pipette until a height of 3.0 cm of silica gel is packed into the column. Top the silica gel with 0.5 cm of anhydrous granular sodium sulfate. Place columns in an oven set at 220°C. Store columns in the oven until ready for use, at least overnight. Remove only the columns needed and place them in a desiccator until they have equilibrated to room temperature. Use immediately.

12.2.3 Silica Column Cleanup. Position the silica column over the alumina column so the eluent will drip onto the alumina column. Transfer the 2 mL hexane extract from the Acid/Base Cleanup onto the silica column with two separate 0.5-mL portions of hexane. Elute the silica column with an additional 4.0 mL of hexane. Discard the silica column and proceed with the alumina column cleanup at the point where the column is washed with 6.0 mL of carbon tetrachloride.

12.2.4 Alumina Column Preparation. Gently tamp a plug of glass wool into the bottom of a 5.75-inch (14.6 cm) disposable Pasteur pipette. Pour WOELM neutral alumina into the pipette while tapping the column with a pencil or wooden dowel until a height of 4.5 cm of alumina is packed into the column. Top the alumina with a 0.5 cm of anhydrous granular sodium sulfate. Prewash the column with 3 mL dichloromethane. Allow the dichloromethane to drain from the column; then force the remaining dichloromethane from the column with a stream of dry nitrogen. Place prepared columns in an oven set at 225°C. Store columns in the oven until ready for use, at least overnight. Remove only columns needed and place them in a desiccator over anhydrous calcium sulfate until they have equilibrated to room temperature. Use immediately.

12.2.5 Alumina Column Cleanup. Prewet the alumina column with 1 mL of hexane. Transfer the 2 mL hexane extract from acid/base cleanup into the column. Elute the column with 6.0 mL of carbon tetrachloride and archive. Elute the column with 4.0 mL of dichloromethane and catch the eluate in a 12- mL distillation receiver. Add 3 µL tetradecane, place a micro-Snyder column on the receiver and evaporate the dichloromethane just to dryness by means of a hot water bath. Add 2 mL of hexane to the receiver and evaporate just to dryness. Add another 2-mL portion of hexane and evaporate to 0.5 mL. The extract is ready for the carbon column cleanup at this point.

12.2.6 Carbon Column Preparation. Weigh 9.5 g of Bio-Sil®A (100-200 mesh) silica gel, which has been previously heated to 225°C for 24 hours, into a 50-mL screw cap container. Weigh 0.50 g of Amoco PX-21 carbon onto the silica gel cap and shake vigorously for 1 hour. Just before use, rotate the container by hand for at least 1 minute. Break a glass graduated 2.0-mL disposal pipette at the 1.8 mL mark and fire polish the end. Place a small plug of glass wool in the pipette and pack it at the 0.0 mL mark using two small solid glass rods. Add 0.1 mL of Bio-Sil®A 100-200 mesh silica gel. If more than 1 column is to be made at a time, it is best to

add the silica gel to all the columns and then add the carbon-silica gel mixture to all columns. Add 0.40 mL of the carbon silica gel mixture to the column; the top of the mixture will be at the 0.55-mL mark on the pipette. Top the column with a small plug of glass wool.

12.2.7 Carbon Column Cleanup. Place the column in a suitable clamp with the silica gel plug up. Add approximately 0.5 mL of 50 percent benzene-methylene chloride (v/v) to the top of the column. Fit a 10 mL disposable pipette on the top of the carbon column with a short piece of extruded teflon tubing. Add an additional 9.5 mL of the 50 percent benzene-methylene chloride. When approximately 0.5 mL of this solvent remains, add 10 mL of toluene. After all the toluene has gone into the column, remove the 10-mL reservoir and add at least 2.0 mL of hexane to the column. When approximately 0.1 mL of the hexane is left on the top of the column, transfer the sample extract onto the column with a Pasteur pipette. Rinse the distillation receiver column that contained the extract with two separate 0.2 mL portions of hexane and transfer each rinse onto the column. Allow the top of each transfer layer to enter the glass wool before adding the next one. When the last of the transfer solvent enters the glass wool, add 0.5 mL of methylene chloride, replace the 10-mL reservoir, and add 4.5 mL of methylene chloride to it. When approximately 0.5 mL of this solvent remains, add 10 mL of 50 percent benzene-methylene chloride. When all this solvent has gone onto the column, remove the reservoir, take the column out of the holder and rinse each end with toluene, turn the column over, and put it back in the holder. All previous elution solvents are archived. Place a suitable receiver tube under the column and add 0.5 mL of toluene to the top of the column. Fit the 10 mL reservoir on the column and add 9.5 mL of toluene to it. When all toluene has eluted through the column and has been collected in the receiving tube, add 5 mL of tetradecane and concentrate to 0.5 mL using a stream of nitrogen and water bath maintained at 60°C. Transfer the toluene extract to a 2.0 mL graduated Chromoflex® tube with two 0.5-mL portions of benzene. Add 0.5 ng of ¹³C₁₂-1,2,3,4-TCDD and store the extracts in the dark at room temperature. Concentrate the extract to 30 µL using a stream of nitrogen at room temperature just prior to analysis or shipping. Transfer the extracts that are to be shipped to a 2 mm i.d. x 75 mm glass tube that has been fire sealed on one end with enough benzene to bring the total volume of the extract to 100 µL. Then fire seal other end of the tube.

12.3 Glassware Cleanup Procedures

In this procedure, take each piece of glassware through the cleaning separately except in the oven baking process. Wash the 100-mL round bottom flasks, the 250 mL separatory funnels, the KD concentrators, etc., that were used in the extraction procedures three times with hot tap water, two times with acetone and two times with hexane. Then bake this glassware in a forced air oven that is vented to the outside for 16 hours at 450°C. Clean the PFTE stopcocks as above except for the oven baking step. Rinse all glassware with acetone and hexane immediately before use.

13. HRGC-HRMS System Performance

13.1 Operation of HRGC-HRMS

Operate the HRMS in the electron impact (EI) ionization mode using the selected ion monitoring (SIM) detection technique. Achieve a static mass resolution of 10,000 (10% valley) before analysis of a set of samples is begun. Check the mass resolution at the beginning and at the end of each day. (Corrective actions should be implemented whenever the resolving power does not meet the requirement.) Chromatography time required for PCDDs and PCDFs may exceed the long-term stability of the mass spectrometer because the instrument is operated in the high-resolution mode and the mass drifts of a few ppm (e.g., 5 ppm in mass) can have adverse effects on the analytical results. Therefore, a mass-drift correction may be required. Use a lock-mass ion for the reference

compound perfluorokerosene (PFK) to tune the mass spectrometer. The selection of the SIM lock-mass ions of PFK shown in the descriptors (see Tables 10, 11 and 12) is dependent on the masses of the ions monitored within each descriptor. An acceptable lock-mass ion at any mass between the lightest and heaviest ion in each descriptor can be used to monitor and correct mass drifts. Adjust the level of the reference compound (PFK) metered inside the ion chamber during HRGC-HRMS analyses so that the amplitude of the most intense selected lock-mass ion signal is kept to a minimum. Under those conditions, sensitivity changes can be more effectively monitored. Excessive use of PFK or any reference substance will cause high background signals and contamination of the ion source, which will result in an increase in "downtime" required for instrument maintenance.

Tune the instrument to a mass resolution of 10,000 (10% valley) at m/z 292.9825 (PFK). By using the peak matching unit (manual or computer simulated) and the PFK reference peak, verify that the exact m/z 392.9761 (PFK) is within 3 parts per million (ppm) of the required value.

Document the instrument resolving power by recording the peak profile of the high mass reference signal (m/z 392.9761) obtained during the above peak matching calibration experiment by using the low mass PFK ion at m/z 292.9825 as a reference. The minimum resolving power of 10,000 should be demonstrated on the high mass ion while it is transmitted at a lower accelerating voltage than the low mass reference ion, which is transmitted at full voltage and full sensitivity. There will be little, if any, loss in sensitivity on the high mass ion if the source parameters are properly tuned and optimized. The format of the peak profile representation should allow for computer calculated and manual determination of the resolution, i.e., the horizontal axis should be a calibrated mass scale (amu or ppm per division). Detailed descriptions for mass resolution adjustments are usually found in the instrument operators manual or instructions.

13.2 Column Performance

After the HRMS parameters are optimized, analyze an aliquot of a column performance solution containing the first and last eluting compounds (see Table 9), or a solution containing all congeners, to determine and confirm SIM parameters, retention time windows, and HRGC resolution of the compounds. Adjustments can be made at this point, if necessary. Some PeCDFs elute in the TCDD retention time window when using the 60 m DB-5 column. The PeCDF masses can be included with the TCDD/TCDF masses in Descriptor 1. Include the PeCDD/PeCDF masses with the TCDD/TCDF masses when using the 60 m SP-2331 polar column. The HRGC-HRMS SIM parameters and retention time windows can be rapidly and efficiently determined and optimized by analysis of a window defining solution of PCDDs/PCDFs using one mass for each isomer for the complete analysis of tetra- through octa- compounds, as illustrated in Figure 11.

13.3 SIM Cycle Time

The total time for each SIM cycle should be 1 second or less for data acquisition, which includes the sum of the mass ion dwell times and ESA voltage reset times.

13.4 Peak Separation

Chromatographic peak separation between 2,3,7,8-TCDD and the co-eluting isomers should be resolved with a valley of 25% or more (see Figure 12).

13.5 Initial Calibration

After the HRGC-HRMS SIM operating conditions are optimized, perform an initial calibration using the 5 calibration solutions shown in Table 3. The quantification relationships of labeled and unlabeled standards are illustrated in Tables 15, 16, 17, and 18. Figures 13 through 22 represent the extracted ion current profiles (EICP) for specific masses for 2,3,7,8-TCDF, 2,3,7,8-TCDD and other 2,3,7,8-substituted PCDF/PCDD (along with their labeled standards) through OCDF and OCDD respectively.

[Note: Other solutions containing fewer or different congeners and at different concentrations may also be used for calibration purposes.]

Referring to Tables 10, 11, or 12, calculate (1) the relative response factors (RRFs) for each unlabeled PCDD/PCDF and PBDD/PBDF [RRF (I)] relative to their corresponding $^{13}\text{C}_{12}$ -labeled internal standard and (2) the RRFs for the $^{13}\text{C}_{12}$ -labeled PCDD/PCDF and PBDD/PBDF internal standards [RRF (II)] relative to $^{37}\text{Cl}_4$ -2,3,7,8-TCDD recovery standard using the following formulae:

$$\text{RRF(I)} = \frac{(A_x \times Q_{is})}{(Q_x \times A_{is})}$$

$$\text{RRF(II)} = \frac{(A_{is} \times Q_{rs})}{(Q_{is} \times A_{rs})}$$

where:

A_x = the sum of the integrated ion abundances of the quantitation ions (see Tables 10, 11 or 12) for unlabeled PCDDs/PCDFs, and PBDDs/PBDFs and BCDDs/BCDFs.

A_{is} = the sum of the integrated ion abundances of the quantitation ions for the $^{13}\text{C}_{12}$ -labeled internal standards (see Table 10, 11 or 12).

[Note: Other $^{13}\text{C}_{12}$ -labeled analytes may also be used as the recovery standard(s)]

A_{rs} = the integrated ion abundance for the quantitation ion of the $^{37}\text{Cl}_4$ -2,3,7,8-TCDD recovery standard.

Q_{is} = the quantity of the $^{13}\text{C}_{12}$ -labeled internal standard injected, pg.

Q_x = the quantity of the unlabeled PCDD/PCDF analyte injected, pg.

Q_{rs} = the quantity of the $^{37}\text{Cl}_4$ -2,3,7,8-TCDD injected, pg.

RRF(I) and RRF(II) = dimensionless quantities. The units used to express Q_{is} and Q_x must be the same.

[Note: $^{13}\text{C}_{12}$ -1,2,3,7,8-PeBDF is used to determine the response factor for the unlabeled 2,3,7,8-substituted, PeBDD, HxBDF and HxBDD.]

Calculate the average RRFs for the 5 concentration levels of unlabeled and $^{13}\text{C}_{12}$ -labeled PCDDs/PCDFs and PBDDs/PBDFs for the initial calibration using the following equation:

$$\overline{\text{RRF}} = \frac{\text{RRF1} + \text{RRF2} + \text{RRF3} + \text{RRF4} + \text{RRF5}}{5}$$

13.6 Criteria Required for Initial Calibration

The analytical data must satisfy certain criteria for acceptable calibration. The isotopic ratios must be within the acceptable range (see Tables 19 and 20). The percent relative standard deviation for the response factors should be less than the values presented in Table 21. The signal-to-noise ratio for the $^{13}\text{C}_{12}$ -labeled standards must be 10:1 or more and 5:1 or more for the unlabeled standards.

13.7 Continuing Calibration

Conduct an analysis at the beginning of each day to check and confirm the calibration using an aliquot of the calibration solution. This analysis should meet the isotopic ratios and signal to noise ratios of the criteria stated in Section 13.6 (see Table 21 for daily calibration percent difference criteria). It is good practice to confirm the calibration at the end of the day also. Calculate the daily calibration percent difference using the following equation.

$$\% \text{RRF} = \frac{\text{RRF}_{\text{cc}} - \overline{\text{RRF}}}{\overline{\text{RRF}}} \times 100$$

RRF_{cc} = the relative response factor for a specific analyte in the continuing calibration standard.

14. HRGC-HRMS Analysis And Operating Parameters

14.1 Sample Analysis

Sample Analysis. An aliquot of the sample extract is analyzed with the HRGC-HRMS system using the instrument parameters illustrated in Tables 13 and 14 and the SIM descriptors and masses shown in Tables 10, 11, and 12. A 30-m SE-54 fused silica capillary column is used to determine the concentrations of total tetra-, penta-, hexa-, hepta- and octa-CDDs/CDFs and/or to determine the minimum limits of detections (MLDs) for the compounds. If the tetra-, penta-, and hexa-CDDs/CDFs were detected in a sample and isomer specific analyses are required, then an aliquot of the sample extract is analyzed using the 60 m SP-2331 fused silica capillary column to provide a concentration for each 2,3,7,8-substituted PCDD/PCDF and concentrations for total PCDDs and PCDFs also.

[Note: Other capillary columns such as the DB-5, SE-30, and DB-225 may be used if the performance satisfies the specifications for resolution of PCDDs/PCDFs. The SE-54 column resolves the four HpCDF isomers, two HpCDD isomers, OCDF and OCDD for isomer specific analysis. It does not resolve the tetra-, penta-, and hexa-2,3,7,8-substituted isomers. The SE-54 column is used for the analysis of PBDDs and PBDFs.]

Isomer specificity for all 2,3,7,8-substituted PCDDs/PCDFs cannot be achieved on a single HRGC capillary column at this time. However, many types of HRGC capillary columns are available and can be used for these analyses after their resolution capabilities are confirmed to be adequate using appropriate standards.

Two HRGC columns shown in Table 13 have been used successfully since 1984 (27, 28). The 60-m DB-5 provides an efficient analysis for total concentrations of PCDDs/PCDFs, specific isomers (total tetra-, penta-, hexa-CDDs/CDFs, four HpCDF isomers, two HpCDD isomers, OCDD and OCDF), PBDDs/PBDFs, and/or determination of MDLs. The 60 m SP-2331 column provides demonstrated and confirmed resolution of 2,3,7,8-substituted tetra-, penta-, and hexa-PCDDs/PCDFs (14). The descriptors and masses shown in Tables 10, 11 and 12 must be modified to take into account the elution of some of the PeCDDs and PeCDFs in the tetra retention time window using the SP-2331 column.

14.2 Identification Criteria

Criteria used for identification of PCDDs and PCDFs in samples are as follows:

- The integrated ion abundance ratio $M/(M+2)$ or $(M+2)/(M+4)$ shall be within 15 percent of the theoretical value. The acceptable ion abundance ranges are shown in Tables 19 and 20.
- The ions monitored for a given analyte, shown in Tables 10, 11, and 12, shall reach their maximum within 2 seconds of each other.
- The retention time for the 2,3,7,8-substituted analytes must be within 3 seconds of the corresponding $^{13}\text{C}_{12}$ -labeled internal standard, surrogate, or alternate standard.
- The identification of 2,3,7,8-substituted isomers that do not have corresponding $^{13}\text{C}_{12}$ -labeled standards is done by comparison to the analysis of a standard that contains the specific congeners. Comparison of the relative retention time (RRT) of the analyte to the nearest internal standard with reference (i.e., within 0.005 RRT time units to the comparable RRTs found in the continuing calibration or literature).
- The signal-to-noise ratio for the monitored ions must be greater than 2.5.
- The analysis shall show the absence of polychlorinated diphenyl-ethers (PCDPEs). Any PCDPEs that co-elute (± 2 seconds) with peaks in the PCDF channels indicates a positive interference, especially if the intensity of the PCDPE peak is 10 percent or more of the PCDF.

Use the identification criteria in Section 14.2 to identify and quantify the PCDDs and PCDFs in the sample. Figure 23 illustrates a reconstructed EICP for an environmental sample, identifying the presence of 2,3,7,8-TCDF as referenced to the labeled standard.

14.3 Quantification

The peak areas of ions monitored for $^{13}\text{C}_{12}$ -labeled PCDDs/PCDFs and ^{37}Cl -2,3,7,8-TCDD, unlabeled PCDDs/PCDFs, and respective relative response factors are used for quantification. The $^{37}\text{Cl}_4$ -2,3,7,8-TCDD, spiked to extract prior to final concentration, and respective response factors are used to determine the sample extraction efficiencies achieved for the nine $^{13}\text{C}_{12}$ -labeled internal standards, which are spiked to the sample prior to extraction (% recovery). The $^{13}\text{C}_{12}$ -labeled PCDD/PCDF internal standards and response factors are used for quantification of unlabeled PCDDs/PCDFs and for determination of the minimum limits of detection with but one exception: $^{13}\text{C}_{12}$ -OCDD is used for OCDF. Each $^{13}\text{C}_{12}$ -labeled internal standard is used to quantify all of the PCDDs/PCDFs in its isomeric group. For example, $^{13}\text{C}_{12}$ -2,3,7,8-TCDD and the 2,3,7,8-TCDD response factor are used to quantify all of the 22 tetra-chlorinated isomers. The quantification relationships of these standards are shown in Tables 15, 16, 17, and 18. The $^{37}\text{Cl}_4$ -2,3,7,8-TCDD spiked to the filter of the sampler

prior to sample collection is used to determine the sampler retention efficiency, which also indicates the collection efficiency for the sampling period.

14.4 Calculations

14.4.1 Extraction Efficiency. Calculate the extraction efficiencies (percent recovery) of the 9 $^{13}\text{C}_{12}$ -labeled PCDD/PCDF or the 3 $^{13}\text{C}_{12}$ -labeled PBDD/PBDF internal standards measured in the extract using the formula:

$$\%R_{is} = \frac{[A_{is} \times Q_{rs} \times 100]}{[Q_{is} \times A_{rs} \times \text{RRF(II)}]}$$

where:

$\%R_{is}$ = percent recovery (extraction efficiency).

A_{is} = the sum of the integrated ion abundances of the quantitation ions (see Tables 10, 11 or 12) for the $^{13}\text{C}_{12}$ -labeled internal standard.

A_{rs} = the sum of the integrated ion abundances of the quantitation ions (see Table 10, 11 or 12) for the $^{37}\text{Cl}_4$ - or $^{13}\text{C}_{12}$ -labeled recovery standard; the selection of the recovery standard(s) depends on the type of homologues.

Q_{is} = quantity of the $^{13}\text{C}_{12}$ -labeled internal standard added to the sample before extraction, pg.

Q_{rs} = quantity of the $^{37}\text{Cl}_4$ - or $^{13}\text{C}_{12}$ -labeled recovery standard added to the sample extract before HRGC-HRMS analysis, pg.

RRF(II) = calculated mean relative response factor for the labeled internal standard relative to the appropriate labeled recovery standard.

14.4.2 Calculation of Concentration. Calculate the concentration of each 2,3,7,8-substituted PCDD/PCDF, other isomers or PBDD/PBDF that have met the criteria described in Sections 14.2 using the following formula:

$$C_x = \frac{[A_x \times Q_{is}]}{[A_{is} \times V_{std} \times \text{RRF(I)}]}$$

where:

C_x = concentration of unlabeled PCDD/PCDF, PBDD/PBDF or BCDD/BCDF congener(s), pg/m^3 .

A_x = the sum of the integrated ion abundances of the quantitation ions (see Table 11, 12 or 13) for the unlabeled PCDDs/PCDFs, or PBDDs/PBDFs or BCDFs.

A_{is} = the sum of the integrated ion abundances of the quantitation ions (see Table 11, 12 or 13) for the respective $^{13}\text{C}_{12}$ -labeled internal standard.

Q_{is} = quantity of the $^{13}\text{C}_{12}$ -labeled internal standard added to the sample before extraction, pg.

V_{std} = standard volume of air, std m^3 .

RRF(I) = calculated mean relative response factor for an unlabeled 2,3,7,8-substituted PCDD/PCDF obtained in Section 13.4.

14.5 Method Detection Limits (MDLs)

The ambient background levels of total PCDDs/PCDFs are usually found in the range of 0.3 to 2.9 pg/m³. Therefore, the MDLs required to generate meaningful data for ambient air should be in the range of 0.02 to 0.15 pg/m³ for tetra-, penta-, and hexa-CDDs/CDFs. Trace levels, 0.05 to 0.25 pg/m³, of HpCDDs and OCDD are usually detected in the method blank (background contamination).

An MDL is defined as the amount of an analyte required to produce a signal with a peak area at least 2.5 x the area of the background signal level measured at the retention time of interest. MDLs are calculated for total PHDDs/PHDFs and for each 2,3,7,8-substituted congener. The calculation method used is dependent upon the type of signal responses present in the analysis. For example:

- Absence of response signals of one or both quantitation ion signals at the retention time of the 2,3,7,8-substituted isomer or at the retention time of non 2,3,7,8-substituted isomers. The instrument noise level is measured at the analyte's expected retention time and multiplied by 2.5, inserted into the formula below and calculated and reported as not detected (ND) at the specific MDL.
- Response signals at the same retention time as the 2,3,7,8-substituted isomers or the other isomers that have a S/N ratio in excess of 2.5:1 but that do not satisfy the identification criteria described in 14.2 are calculated and reported as ND at the elevated MDL and discussed in the narrative that accompanies the analytical results. Calculate the MDLs using the following formula:

$$\text{MDL} = \frac{[2.5 \times A_x \times Q_{is}]}{[A_{is} \times V_{std} \times \overline{\text{RRF}}]}$$

where:

MDL = concentration of unlabeled PHDD/PHDF, pg/m³.

A_x = sum of integrated ion abundances of the quantitation ions (see Table 10, 11 or 12) for the unlabeled PHDDs/PHDFs which do not meet the identification criteria or 2.5 x area of noise level at the analyte's retention time.

A_{is} = sum of the integrated ion abundances of the quantitation ions (see Table 10, 11, or 12) for the ¹³C₁₂-labeled internal standards.

Q_{is} = quantity of the ¹³C₁₂-labeled internal standard spiked to the sample prior to extraction, pg.

V_{std} = standard volume of ambient air sampled, std m³.

$\overline{\text{RRF}}$ = mean relative response factor for the unlabeled PHDD/PHDF.

14.6 2,3,7,8-TCDD Toxic Equivalents

Calculate the 2,3,7,8-TCDD toxic equivalents of PCDDs and PCDFs present in a sample according to the method recommended by EPA and the Center for Disease Control (18). This method assigns a 2,3,7,8-TCDD toxicity equivalency factor (TEF) for each of the seventeen 2,3,7,8-substituted PCDDs/PCDFs (see Table 22). The 2,3,7,8-TCDD equivalent of the PCDDs and PCDFs present in the sample is calculated by the respective TEF factors times their concentration for each of the compounds listed in Table 22. The exclusion of the other isomeric groupings (mono-, di-, and tri-chlorinated dibenzodioxins and dibenzofurans) does not mean that they are non-toxic. Their toxicity, as known at this time, is much less than the toxicity of the compounds listed in Table 22. The above procedure for calculating the 2,3,7,8-TCDD toxic equivalents is not claimed to be based on a

thoroughly established scientific foundation. The procedure, rather, represents a "consensus recommendation on science policy." Similar methods are used throughout the world.

15. Quality Assurance/Quality Control (QA/QC)

15.1 Certified analytical standards were obtained from Cambridge Isotope Laboratories, 50 Frontage Road, Andover, MA 01810, 508-749-8000.

15.2 Criteria used for HRGC-HRMS initial and continuing calibration are defined in Sections 13.5 and 13.6.

15.3 Analytical criteria used for identification purposes are defined in Section 14.2.

15.4 All test samples, method blanks, field blanks, and laboratory control samples are spiked with $^{13}\text{C}_{12}$ -labeled internal standards prior to extraction.

15.5 Sample preparation and analysis and evaluation of data are performed on a set of 12 samples, which may consist of 9 test samples, field blank, method blank, fortified method blank, or a laboratory control sample.

15.6 Method evaluation studies were performed to determine and evaluate the overall method capabilities (1, 2).

15.7 The $^{13}\text{C}_{12}$ -1,2,3,4-TCDD solution is spiked to filters of all samplers, including field blanks, immediately prior to operation or is spiked to all PUF plugs prior to shipping them to the field for sampling to determine and document the sampling efficiency.

15.8 Minimum equipment calibration and accuracy requirements achieved are illustrated in Table 23.

15.9 QA/QC requirements for data:

<u>Criteria</u>	<u>Requirements</u>
The data shall satisfy all indicated identification criteria	Discussed in Section 14.2
Method efficiency achieved for $^{13}\text{C}_{12}$ -labeled tetra-, penta-, hexa-CDDs/CDFs and PBDDs/PBDFs	50 to 120%
Method efficiency achieved for $^{13}\text{C}_{12}$ -labeled HpCDD and OCDD	40 to 120%
Accuracy achieved for PHDDs and PHDFs in method spike at 0.25 to 2.0 pg/m^3 concentration range	70 to 130%
Precision achieved for duplicate method spikes or QA samples	$\pm 30\%$
Sampler efficiency achieved for $^{13}\text{C}_{12}$ -1,2,3,4-TCDD	50 to 120%
Method blank contamination	Free of contamination that would interfere with test sample results.
Method detection limit range for method blank and field blank (individual isomers)	0.02 to 0.25 pg/m^3

16. Report Format

The analytical results achieved for a set of 12 samples should be presented in a table such as the one shown in Table 24. The analytical results, analysis, QA/QC criteria, and requirements used to evaluate data are discussed in an accompanying analytical report. The validity of the data in regard to the data quality requirements and any qualification that may apply is explained in a clear and concise manner for the user's information.

17. References

1. Harless, R. L. et al., "Evaluation of Methodology for Determination of Polyhalogenated Dibenzo-p-Dioxins and Dibenzofurans in Ambient Air," in *Proceedings of the 1991 EPA/A&WMA International Symposium on Measurement of Toxic and Related Air Pollutants*, U. S. Environmental Protection Agency, Research Triangle Park, NC 27711, EPA-600/9-91-018, May 1991.
2. Harless, R. L., et al., "Evaluation of a Sampling and Analysis Method for Determination of Polyhalogenated Dibenzo-p-Dioxins and Dibenzofurans in Ambient Air," in *Proceedings of the 11th International Symposium on Chlorinated Dioxins and Related Compounds*, U. S. Environmental Protection Agency, Research Triangle Park, NC 27711, EPA-600/D-91-106; *Chemosphere*, Vol. 25, (7-10):1317-1322, Oct-Nov 1992.
3. Smith-Mullen, C., et al., *Feasibility of Environmental Monitoring and Exposure Assessment for a Municipal Waste Combustor, Rutland, Vermont Pilot Study*, U. S. Environmental Protection Agency, Research Triangle Park, NC 27711, EPA-600/8-91-007, March 1991.
4. Harless, R. L., et al., *Sampling and Analysis for Polychlorinated Dibenzo-p-Dioxins and Dibenzofurans in Ambient Air*, U. S. Environmental Protection Agency, Research Triangle Park, NC 27711, EPA-600/D-9-172, May 1990.
5. Harless, R. L. et al., *Analytical Procedures and Quality Assurance Plan for the Analysis of 2,3,7,8-TCDD in Tier 3-7 Samples of the U. S. EPA National Dioxin Study*, U. S. Environmental Protection Agency, Research Triangle Park, NC 27711, EPA-600/3-85-019, May 1986.
6. Harless, R. L. et al., *Analytical Procedures and Quality Assurance Plan for the Analysis of Tetra Through Octa Chlorinated Dibenzo-p-Dioxins and Dibenzofurans in Tier 4 Combustion and Incineration Processes*, U. S. Environmental Protection Agency, Research Triangle Park, NC 27711, Addendum to EPA-600/3-85-019, May 1986.
7. Albro, P.W., et al., "Methods for the Quantitative Determination of Multiple Specific Polychlorinated Dibenzo-p-Dioxins and Dibenzofuran Isomers in Human Adipose Tissue in the Parts-Per-Trillion Range. An Interlaboratory Study," *Anal. Chem.*, Vol.57:2717-2725, 1985.
8. O'Keefe, P. W., et al., "Interlaboratory Validation of PCDD and PCDF Concentrations Found in Municipal Incinerator Emissions," *Chemosphere*, Vol. 18:185-192, 1989.
9. Harless, R. L., et al., "Identification of Bromo/Chloro Dibenzo-p-Dioxins and Dibenzofurans in Ash Samples," *Chemosphere*, Vol. 18:201-208, 1989.
10. Lafleur, L.E., and Dodo, G. H., "An Interlaboratory Comparison of Analytical Procedures for the Measurement of PCDDs/PCDFs in Pulp and Paper Industry Solid Wastes," *Chemosphere*, Vol. 18:77-84, 1989.
11. Patterson, D. G. et al., "Levels of Polychlorinated Dibenzo-p-Dioxins and Dibenzofurans in Workers Exposed to 2,3,7,8-TCDD," *American Journal of Industrial Medicine*, Vol. 16:135-146, 1989.

12. Lamparski, L. L. and Nestrick, T. J. "Determination of Tetra-, Hexa-, Hepta- and Octa-chlorodibenzo-p-dioxin isomers in Particulate Samples at Parts-Per-Trillion Levels," *Anal. Chem.*, Vol. 52:2045-2054, 1980.
13. Rappe, C., "Analysis of Polychlorinated Dioxins and Furans," *Environ. Sci. Technol.*, Vol. 18:78A-90A, 1984.
14. Rappe, C., et al., "Identification of PCDDs and PCDFs in Urban Air," *Chemosphere*, Vol. 17:3-20, 1988.
15. Tondeur, Y., et al., "Method 8290: An Analytical Protocol for the Multimedia Characterization of Polychlorinated Dibenzo-p-Dioxins and Dibenzofurans by High Resolution Gas Chromatography/High Resolution Mass Spectrometry," *Chemosphere*, Vol. 18:119-131, 1989.
16. "Method 23, Method for Measurement of Polychlorinated Dibenzo-p-Dioxins and Dibenzofurans from Stationary Sources," *Federal Register*, Vol. 56(30):5758-5770, February 13, 1991.
17. "Method 1613: Tetra- through Octa-Chlorinated Dioxins and Furans by Isotope Dilution HRGC-HRMS," *Federal Register*, Vol. 56(26):5098-5122, February 7, 1991.
18. *Interim Procedures for Estimating Risks Associated with Exposures to Mixtures of Chlorinated Dibenzo-p-Dioxins and Dibenzofurans (CDDs/CDFs)*, U. S. Environmental Protection Agency, Research Triangle Park, NC 27711, EPA-625/3-89-016, March 1989.
19. Tiernan, T., et al., "PCDD/PCDF in the Ambient Air of a Metropolitan Area in the U.S.," *Chemosphere*, Vol. 19:541-546, 1989.
20. Hunt, G., "Measurement of PCDDs/PCDFs in Ambient Air," *J. Air Pollut. Control Assoc.*, Vol. 39:330-331, 1989.
21. "40 CFR Parts 707 and 766, Polyhalogenated Dibenzo-p-Dioxins and Dibenzofurans: Testing and Reporting Requirements: Final Rule," *Federal Register*, Vol. 52 (108):21412-21452, June 5, 1987.
22. Buser, H., "Polybrominated Dibenzo-p-Dioxins and Dibenzofurans: Thermal reaction products of polybrominated diphenyl ether flame retardants," *Environ. Sci. Technol.*, Vol. 20:404-408, 1988.
23. Sovocol, G. W., et al., "Analysis of Municipal Incinerator Fly Ash for Bromo and Bromo/Chloro Dioxins, Dibenzofurans, and Related Compounds," *Chemosphere*, Vol. 18:193-200, 1989.
24. Lewis, R. G., et al., "Modification and Evaluation of a High-Volume Air Sampler for Pesticides and Semivolatile Industrial Organic Chemicals," *Anal. Chem.*, Vol. 54:592-594, 1982.
25. Lewis, R. G., et al., "Evaluation of Polyurethane Foam for Sampling Pesticides, Polychlorinated Biphenyls and Polychlorinated Naphthalenes in Ambient Air," *Anal. Chem.*, Vol. 49:1668-1672, 1977.
26. Winberry, W. T., Jr., et al., *Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air, Second Supplement, Method TO-9*, U. S. Environmental Protection Agency, Research Triangle Park, NC 27711, EPA 600/4-89-018, March 1989.

27. "Analysis of Air Samples for PCDDs, PCDFs, PCBs, and PAHs in Support of the Great Lakes Deposition Project," Draft Report, Midwest Research Institute, 425 Volker Boulevard, Kansas City, MO, MRI Project No. 3103-A, April 1990.
28. Boggess, K.E., "Analysis of Air Samples for PCDDs, PCDFs, PCBs and PAHs in Support of the Great Lakes Deposition Project," Final Report, Midwest Research Institute, 425 Volker Boulevard, Kansas City, MO, MRI Project No. 3103-A, April 1993.
29. "Working with Carcinogens," NIOSH, Publication 77-206, August 1977.
30. "Safety in the Academic Chemistry Laboratories," ACS Committee on Chemical Safety, 1979.

TABLE 1. NUMBER OF POLYCHLORINATED DIBENZO-P-DIOXIN AND DIBENZOFURAN (PCDD/PCDF) CONGENERS

No. of Chlorine Atoms	No. of PCDD Isomers	No. of PCDF Isomers
1	2	4
2	10	16
3	14	28
4	22	38
5	14	28
6	10	16
7	2	4
8	1	1
Total	75	135

[Note: This also applies for the polybrominated dibenzo-p-dioxins and dibenzofurans (PBDDs/PBDFs).]

TABLE 2. LIST OF 2,3,7,8-CHLORINE SUBSTITUTED PCDD/PCDF CONGENERS

PCDDs	PCDFs
2,3,7,8-TCDD	2,3,7,8-TCDF
1,2,3,7,8-PeCDD	1,2,3,7,8-PeCDF
	2,3,4,7,8-PeCDF
1,2,3,4,7,8-HxCDD	1,2,3,4,7,8-HxCDF
1,2,3,6,7,8-HxCDD	1,2,3,6,7,8-HxCDF
1,2,3,7,8,9-HxCDD	1,2,3,7,8,9-HxCDF
	2,3,4,6,7,8-HxCDF
1,2,3,4,6,7,8-HpCDD	1,2,3,4,6,7,8-HpCDF
	1,2,3,4,7,8,9-HpCDF
1,2,3,4,6,7,8,9-OCDD	1,2,3,4,6,7,8,9-OCDF

TABLE 3. COMPOSITIONS OF THE INITIAL CALIBRATION SOLUTIONS OF LABELED AND UNLABELED PCDDS AND PCDFS

Compound Solution No.	Concentrations (pg/ μ L)				
	1	2	3	4	5
Unlabeled Analytes					
2,3,7,8-TCDD	0.5	1	5	50	100
2,3,7,8-TCDF	0.5	1	5	50	100
1,2,3,7,8-PeCDD	2.5	5	25	250	500
1,2,3,7,8-PeCDF	2.5	5	25	250	500
2,3,4,7,8-PeCDF	2.5	5	25	250	500
1,2,3,4,7,8-HxCDD	2.5	5	25	250	500
1,2,3,6,7,8-HxCDD	2.5	5	25	250	500
1,2,3,7,8,9-HxCDD	2.5	5	25	250	500
1,2,3,4,7,8-HxCDF	2.5	5	25	250	500
1,2,3,6,7,8-HxCDF	2.5	5	25	250	500
1,2,3,7,8,9-HxCDF	2.5	5	25	250	500
2,3,4,6,7,8-HxCDD	2.5	5	25	250	500
1,2,3,4,6,7,8-HpCDD	2.5	5	25	250	500
1,2,3,4,6,7,8-HpCDF	2.5	5	25	250	500
1,2,3,4,7,8,9-HpCDF	2.5	5	25	250	500
OCDD	5.0	10	50	500	1000
OCDF	5.0	10	50	500	1000
Internal Standards					
¹³ C ₁₂ -2,3,7,8-TCDD	100	100	100	100	100
¹³ C ₁₂ -1,2,3,7,8-PeCDD	100	100	100	100	100
¹³ C ₁₂ -1,2,3,6,7,8-HxCDD	100	100	100	100	100
¹³ C ₁₂ -1,2,3,4,6,7,8-HpCDD	100	100	100	100	100
¹³ C ₁₂ -OCDD	200	200	200	200	200
¹³ C ₁₂ -2,3,7,8-TCDF	100	100	100	100	100

TABLE 3. (continued)

Compound Solution No.	Concentrations (pg/ μ L)				
	1	2	3	4	5
$^{13}\text{C}_{12}$ -1,2,3,7,8-PeCDF	100	100	100	100	100
$^{13}\text{C}_{12}$ -1,2,3,4,7,8-HxCDF	100	100	100	100	100
$^{13}\text{C}_{12}$ -1,2,3,4,6,7,8-HpCDF	100	100	100	100	100
Surrogate Standards					
$^{13}\text{C}_{12}$ -2,3,4,7,8-PeCDF	60	80	100	120	140
$^{13}\text{C}_{12}$ -1,2,3,4,7,8-HxCDF	60	80	100	120	140
$^{13}\text{C}_{12}$ -1,2,3,6,7,8-HxCDF	60	80	100	120	140
$^{13}\text{C}_{12}$ -1,2,3,6,7,8,9-HpCDF	60	80	100	120	140
Field Standards					
$^{37}\text{Cl}_4$ -2,3,7,8-TCDD	100	100	100	100	100
$^{13}\text{C}_{12}$ -1,2,3,7,8,9-HxCDD	100	100	100	100	100
Recovery Standard					
$^{13}\text{C}_{12}$ -1,2,3,4-TCDD	50	50	50	50	50

[Note: Standards specified in EPA Method 1613 can also be used in this method.]

TABLE 4. COMPOSITION OF THE SAMPLE FORTIFICATION SOLUTIONS

Analyte	Concentration (pg/ μ L)
Chlorinated Internal Standards	
$^{13}\text{C}_{12}$ -2,3,7,8-TCDD	100
$^{13}\text{C}_{12}$ -1,2,3,7,8-PeCDD	100
$^{13}\text{C}_{12}$ -1,2,3,6,7,8-HxCDD	100
$^{13}\text{C}_{12}$ -1,2,3,4,6,7,8-HpCDD	100
$^{13}\text{C}_{12}$ -OCDD	100
$^{13}\text{C}_{12}$ -2,3,7,8-TCDF	100
$^{13}\text{C}_{12}$ -1,2,3,7,8-PeCDF	100
$^{13}\text{C}_{12}$ -1,2,3,4,7,8-HxCDF	100
$^{13}\text{C}_{12}$ -1,2,3,4,6,7,8-HpCDF	100
Brominated Internal Standards	
$^{13}\text{Cl}_{12}$ -2,3,7,8-TBDD	0.86
$^{13}\text{C}_{12}$ -2,3,7,8-TBDF	0.86
$^{13}\text{C}_{12}$ -1,2,3,7,8-PeBDF	0.86

TABLE 5. COMPOSITION OF RECOVERY STANDARD SOLUTION

Analyte	Concentration (pg/ μ L)
Recovery Standard	
$^{13}\text{C}_{12}$ -1,2,3,4-TCDD	10

TABLE 6. COMPOSITION OF AIR SAMPLER FIELD FORTIFICATION STANDARD SOLUTION

Analyte	Concentration (pg/ μ L)
Field Fortification Standard	
$^{37}\text{Cl}_4$ -2,3,7,8-TCDD	10

TABLE 7. COMPOSITION OF SURROGATE STANDARD SOLUTION

Analyte	Concentration (pg/ μ L)
Surrogate Standards	
$^{13}\text{C}_{12}$ -1,2,3,4,7,8-HxCDD	100
$^{13}\text{C}_{12}$ -2,3,4,7,8-PeCDF	100
$^{13}\text{C}_{12}$ -1,2,3,6,7,8-HxCDF	100
$^{13}\text{C}_{12}$ -1,2,3,4,7,8,9-HpCDF	100

TABLE 8. COMPOSITION OF MATRIX AND METHOD SPIKE AND METHOD SPIKE SOLUTIONS OF PCDDS/PCDFs AND PBDDS/PBDFs^a

Analyte	Concentration (pg/ μ L)	Analyte	Concentration (pg/ μ L)
Native PCDDs and PCDFs		Native PBDDs and PBDFs	
2,3,7,8-TCDD	1	2,3,7,8-TBDD	1
2,3,7,8-TCDF	1	2,3,7,8-TBDF	1
1,2,3,7,8-PeCDD	5	1,2,3,7,8-PeBDD	5
1,2,3,7,8-PeCDF	5	1,2,3,7,8-PeBDF	5
2,3,4,7,8-PeCDF	5	1,2,3,4,7,8-HxBDD	5
1,2,3,4,7,8-HxCDD	5	1,2,3,4,7,8-HxBDF	5
1,2,3,6,7,8-HxCDD	5		
1,2,3,7,8,9-HxCDD	5		
1,2,3,4,7,8-HxCDF	5		
1,2,3,6,7,8-HxCDF	5		
1,2,3,7,8,9-HxCDF	5		
2,3,4,6,7,8-HxCDF	5		
1,2,3,4,6,7,8-HpCDD	5		
1,2,3,4,6,7,8-HpCDF	5		
1,2,3,4,7,8,9-HpCDF	5		
OCDD	10		
OCDF	10		

^aSolutions at different concentrations and those containing different congeners may also be used.

TABLE 9. HRGC-HRMS COLUMN PERFORMANCE EVALUATION SOLUTIONS

Congener	First Eluted	Last Eluted
SE-54 Column GC Retention Time Window Defining Standard ^a		
TCDF	1,3,6,8-	1,2,8,9-
TCDD	1,3,6,8-	1,2,8,9-
PeCDF	1,3,4,6,8-	1,2,3,8,9-
PeCDD	1,2,4,7,9-	1,2,3,8,9-
HxCDF	1,2,3,4,6,8-	1,2,3,4,8,9-
HxCDD	1,2,4,6,7,9-	1,2,3,4,6,7-
HpCDF	1,2,3,4,6,7,8-	1,2,3,4,7,8,9-
HpCDD	1,2,3,4,6,7,9-	1,2,3,4,6,7,8-
OCDF	OCDF	
OCDD	OCDD	
SE-54 TCDD Isomer Specificity Test Standard ^b		
	1,2,3,4-TCDD	
	1,4,7,8-TCDD	2,3,7,8-TCDD
SP-2331 Column TCDF Isomer Specificity Test Standard ^c		
	2,3,4,7-TCDF	
	2,3,7,8-TCDF	
	1,2,3,9-TCDF	

^aA solution containing these congeners and the seventeen 2,3,7,8-substituted congeners may also be used for these purposes.

^bA solution containing the 1,2,3,4,-TCDD and 2,3,7,8-TCDD may also be used for this purpose.

^cSolution containing all tetra- through octa-congeners may also be used for these purposes.

TABLE 10. DESCRIPTORS, MASSES, M/Z TYPES, AND ELEMENTAL COMPOSITIONS OF THE PCDDS AND PCDFS

Descriptor Number	Accurate Mass	m/z Type	Elemental Composition	Compound ²	Primary m/z
1	292.9825	Lock	C ₇ F ₁₁	PFK	
	303.9016	M	C ₁₂ H ₄ ³⁵ Cl ₄ O	TCDF	Yes
	305.8987	M+2	C ₁₂ H ₄ ³⁵ Cl ₃ ³⁷ ClO	TCDF	
	315.9419	M	¹³ C ₁₂ H ₄ ³⁵ Cl ₄ O	TCDF ³	Yes
	317.9389	M+2	¹³ C ₁₂ H ₄ ³⁵ Cl ₃ ³⁷ ClO	TCDF ³	
	319.8965	M	C ₁₂ H ₄ ³⁵ Cl ₄ O ₂	TCDD	Yes
	321.8936	M+2	C ₁₂ H ₄ ³⁵ Cl ₃ ³⁷ ClO ₂	TCDD	
	327.8847	M	C ₁₂ H ₄ ³⁷ Cl ₄ O ₂	TCDD ⁴	
	330.9792	QC	C ₇ F ₁₃	PFK	
	331.9368	M	¹³ C ₁₂ H ₄ ³⁵ Cl ₄ O ₂	TCDD ³	Yes
	333.9339	M+2	¹³ C ₁₂ H ₄ ³⁵ Cl ₃ ³⁷ ClO ₂	TCDD ³	
	375.8364	M+2	C ₁₂ H ₄ ³⁵ Cl ₃ ³⁷ ClO	HxCDFE	
2	339.8597	M+2	C ₁₂ H ₃ ³⁵ Cl ₄ ³⁷ ClO	PeCDF	Yes
	341.8567	M+4	C ₁₂ H ₃ ³⁵ Cl ₃ ³⁷ Cl ₂ O	PeCDF	
	351.9000	M+2	¹³ C ₁₂ H ₃ ³⁵ Cl ₄ ³⁷ ClO	PeCDF ³	Yes
	353.8970	M+4	¹³ C ₁₂ H ₃ ³⁵ Cl ₃ ³⁷ Cl ₂ O	PeCDF ³	
	354.9792	Lock	C ₉ F ₁₃	PFK	
	355.8546	M+2	C ₁₂ H ₃ ³⁵ Cl ₄ ³⁷ ClO ₂	PeCDD	Yes
	357.8516	M+4	C ₁₂ H ₃ ³⁵ Cl ₃ ³⁷ Cl ₂ O ₂	PeCDD	
	367.8949	M+2	¹³ C ₁₂ H ₃ ³⁵ Cl ₄ ³⁷ ClO ₂	PeCDD ⁴	Yes
	369.8919	M+4	¹³ C ₁₂ H ₃ ³⁵ Cl ₃ ³⁷ Cl ₂ O ₂	PeCDD ⁴	
	409.7974	M+2	C ₁₂ H ₃ ³⁵ Cl ₆ ³⁷ ClO	HpCDFE	

TABLE 10. (continued)

Descriptor Number	Accurate Mass	m/z Type	Elemental Composition	Compound ²	Primary m/z
3	373.8208	M+2	C ₁₂ H ₂ ³⁵ Cl ₅ ³⁷ Cl O	HxCDF	Yes
	375.8178	M+4	C ₁₂ H ₂ ³⁵ Cl ₄ ³⁷ Cl ₂ O	HxCDF	
	383.8639	M	¹³ C ₁₂ H ₂ ³⁵ Cl ₆ O	HxCDF ³	Yes
	385.8610	M+2	¹³ C ₁₂ H ₂ ³⁵ Cl ₅ ³⁷ Cl O	HxCDF ³	
	389.8157	M+2	C ₁₂ H ₂ ³⁵ Cl ₅ ³⁷ Cl O ₂	HxCDD	Yes
	391.8127	M+4	C ₁₂ H ₂ ³⁵ Cl ₄ ³⁷ Cl ₂ O ₂	HxCDD	
	392.9760	Lock	C ₉ F ₁₅	PFK	
	401.8559	M+2	¹³ C ₁₂ H ₂ ³⁵ Cl ₅ ³⁷ Cl O ₂	HxCDD ³	Yes
	403.8529	M+4	¹³ C ₁₂ H ₂ ³⁵ Cl ₄ ³⁷ Cl ₂ O ₂	HxCDD ³	
	430.9729	QC	C ₉ F ₁₃	PFK	
	445.7555	M+4	C ₁₂ H ₂ ³⁵ Cl ₆ ³⁷ Cl ₂ O	OCDFE	
4	407.7818	M+2	C ₁₂ H ³⁵ Cl ₆ ³⁷ Cl O	H _p CDF	Yes
	409.7789	M+4	C ₁₂ H ³⁵ Cl ₅ ³⁷ Cl ₂ O	H _p CDF	
	417.8253	M	¹³ C ₁₂ H ³⁵ Cl ₇ O	H _p CDF ³	Yes
	419.8220	M+2	¹³ C ₁₂ H ³⁵ Cl ₆ ³⁷ Cl O	H _p CDF ³	
	423.7766	M+2	C ₁₂ H ³⁵ Cl ₆ ³⁷ Cl O ₂	H _p CDD	Yes
	425.7737	M+4	C ₁₂ H ³⁵ Cl ₅ ³⁷ Cl ₂ O ₂	H _p CDD	
	430.9729	Lock	C ₉ F ₁₇	PFK	
	435.8169	M+2	¹³ C ₁₂ H ³⁵ Cl ₆ ³⁷ Cl O ₂	H _p CDD ³	Yes
	437.8140	M+4	¹³ C ₁₂ H ³⁵ Cl ₅ ³⁷ Cl ₂ O ₂	H _p CDD ³	
	479.7165	M+4	C ₁₂ H ³⁵ Cl ₇ ³⁷ Cl ₂ O	NCDPE	

TABLE 10. (continued)

Descriptor Number	Accurate Mass	m/z Type	Elemental Composition	Compound ²	Primary m/z
5	441.7428	M+2	C ₁₂ ³⁵ Cl ₇ ³⁷ Cl O	OCDF	Yes
	442.9728	Lock	C ₁₀ F ₁₇	PFK	
	443.7399	M+4	C ₁₂ ³⁵ Cl ₆ ³⁷ Cl ₂ O	OCDF	
	457.7377	M+2	C ₁₂ ³⁵ Cl ₇ ³⁷ Cl O ₂	OCDD	Yes
	459.7348	M+4	C ₁₂ ³⁵ Cl ₆ ³⁷ Cl ₂ O ₂	OCDD	
	469.7779	M+2	¹³ C ₁₂ ³⁵ Cl ₇ ³⁷ Cl O ₂	OCDD ³	Yes
	471.7750	M+4	¹³ C ₁₂ ³⁵ Cl ₆ ³⁷ Cl ₂ O ₂	OCDD ³	
	513.6775	M+4	C ₁₂ ³⁵ Cl ₈ ³⁷ Cl ₂ O	DCDPE	

¹Nuclidic masses used:

H = 1.007825 C = 12.00000 ¹³C = 13.003355 F = 18.9984
 O = 15.994915 ³⁵Cl = 34.968853 ³⁷Cl = 36.965903

²Compound abbreviations:

Polychlorinated dibenzo-p-dioxins

TCDD = Tetrachlorodibenzo-p-dioxin
 PeCDD = Pentachlorodibenzo-p-dioxin
 HxCDD = Hexachlorodibenzo-p-dioxin
 HpCDD = Heptachlorodibenzo-p-dioxin
 OCDD = Octachlorodibenzo-p-dioxin

Polychlorinated dibenzofurans

TCDF = Tetrachlorodibenzofuran
 PeCDF = Pentachlorodibenzofuran
 HxCDF = Hexachlorodibenzofuran
 HpCDF = Heptachlorodibenzofuran

Polychlorinated diphenyl ethers

HxCDE = Hexachlorodiphenyl ether
 HpCDE = Heptachlorodiphenyl ether
 OCDE = Octachlorodiphenyl ether
 NCDPE = Nonachlorodiphenyl ether
 DCDPE = Decachlorodiphenyl ether

Lock mass and OC compound

PFK = Perfluorokerosene

³Labeled compound

⁴There is only one m/z for ³⁷Cl₄-2,3,7,8-TCDD (recovery standard).

TABLE 11. DESCRIPTORS, M/Z TYPES, EXACT MASSES AND ELEMENTAL COMPOSITIONS OF THE PBDDS AND PBDFO

Descriptor Number	Accurate Mass ¹	Ion Type	Elemental Composition	Compound ²
1	327.8847	M	C ₁₂ H ₄ ³⁷ Cl ₄ O ₂	TCDD ⁴
	330.9792	QC	C ₇ F ₁₃	PFK
	331.9368	M	C ₁₂ H ₄ ³⁵ Cl ₄ O ₂	TCDD ³
	333.9339	M+2	C ₁₂ H ₄ ³⁵ Cl ₃ ³⁷ ClO ₂	TCDD ³
2	417.825	M	¹³ C ₁₂ H ³⁵ Cl ₇ O	HpCDF ³
	419.822	M+2	¹³ C ₁₂ H ³⁵ Cl ₆ ³⁷ ClO	HpCDF ³
	466.973	QC		PFK
	481.698	M+2	C ₁₂ H ₄ ⁷⁹ Br ₃ ⁸¹ BrO	TBDF
	483.696	M+4	C ₁₂ H ₄ ⁷⁹ Br ₂ ⁸¹ Br ₂ O	TBDF
	485.694	M+6	C ₁₂ H ₄ ⁷⁹ Br ⁸¹ Br ₃ O	TBDF
	492.970	LOCK MASS		PFK
	493.738	M+2	¹³ C ₁₂ H ₄ ⁷⁹ Br ₃ ⁸¹ BrO	TBDF ³
	495.736	M+4	¹³ C ₁₂ H ₄ ⁷⁹ Br ₂ ⁸¹ Br ₂ O	TBDD ³
	497.692	M+2	C ₁₂ H ₄ ⁷⁹ Br ₃ ⁸¹ BrO ₂	TBDD
	499.690	M+4	C ₁₂ H ₄ ⁷⁹ Br ₂ ⁸¹ Br ₂ O ₂	TBDD
	501.689	M+6	C ₁₂ H ₄ ⁷⁹ Br ⁸¹ Br ₃ O	TBDD
	509.733	M+2	¹³ C ₁₂ H ₄ ⁷⁹ Br ₃ ⁸¹ BrO ₂	TBDD ³
	511.731	M+4	¹³ C ₁₂ H ₄ ⁷⁹ Br ₂ ⁸¹ Br ₂ O ₂	TBDD ³
	565.620	M+6	C ₁₂ H ₅ ⁷⁹ Br ₂ ⁸¹ Br ₃ O	PeBDPO
	643.530	M+6	C ₁₂ H ₄ ⁷⁹ Br ₃ ⁸¹ Br ₃ O	HxBDPO

TABLE 11. (continued)

Descriptor Number	Accurate Mass ¹	Ion Type	Elemental Composition	Compound ²
3	469.778	M+2	¹³ C ₁₂ ³⁵ Cl ₇ ³⁷ Cl O ₂	OCDD ³
	471.775	M+4	¹³ C ₁₂ ³⁵ Cl ₆ ³⁷ Cl O ₂	OCDD ³
	559.608	M+2	C ₁₂ H ₃ ⁷⁹ Br ₄ ⁸¹ Br O	PeBDF
	561.606	M+4	C ₁₂ H ₃ ⁷⁹ Br ₃ ⁸¹ Br ₂ O	PeBDF
	563.604	M+6	C ₁₂ H ₃ ⁷⁹ Br ₂ ⁸¹ Br ₃ O	PeBDF
	566.966	LOCK MASS		PFK
	573.646	M+4	¹³ C ₁₂ H ₃ ⁷⁹ Br ₃ ⁸¹ Br ₂ O	PeBDF ³
	575.644	M+6	¹³ C ₁₂ H ₃ ⁷⁹ Br ₂ ⁸¹ Br ₃ O	PeBDF ³
	575.603	M+2	C ₁₂ H ₃ ⁷⁹ Br ₄ ⁸¹ Br O ₂	PeBDD
	577.601	M+4	C ₁₂ H ₃ ⁷⁹ Br ₃ ³⁷ Br ₂ O ₂	PeBDD
	579.599	M+6	C ₁₂ H ₃ ⁷⁹ Br ₂ ⁸¹ Br ₃ O ₂	PeBDD
	589.641	M+4	¹³ C ₁₂ H ₃ ⁷⁹ Br ₃ ³⁷ Br ₂ O ₂	PeBDD ³
	591.639	M+6	¹³ C ₁₂ H ₃ ⁷⁹ Br ₃ ⁸¹ Br ₂ O ₂	PeBDD ³
	616.963	QC		PFK

TABLE 11. (continued)

Descriptor Number	Accurate Mass ¹	Ion Type	Elemental Composition	Compound ²
4	643.530	M+6	C ₁₂ H ₄ ⁷⁹ Br ₃ ⁸¹ Br ₃ O	HxBDO
	721.441	M+6	C ₁₂ H ₃ ⁷⁹ Br ₄ ⁸¹ Br ₃ O	HpBDPO
	616.963	QC		PFK
	639.517	M+4	C ₁₂ H ₂ ⁷⁹ Br ₄ ⁸¹ Br ₂ O	HxBDF
	641.514	M+6	C ₁₂ H ₂ ⁷⁹ Br ₃ ⁸¹ Br ₃ O	HxBDF
	643.512	M+8	C ₁₂ H ₂ ⁷⁹ Br ₂ ⁸¹ Br ₄ O	HxBDF
	655.511	M+4	C ₁₂ H ₂ ⁷⁹ Br ₄ ⁸¹ Br ₂ O ₂	HxBDD
	657.509	M+6	C ₁₂ H ₂ ⁷⁹ Br ₃ ⁸¹ Br ₃ O ₂	HxBDD
	659.507	M+8	C ₁₂ H ₂ ⁷⁹ Br ₂ ⁸¹ Br ₄ O ₂	HxBDD
	666.960	LOCK MASS		PFK
	721.441	M+6	C ₁₂ H ₃ ⁷⁹ Br ₄ ⁸¹ Br ₃ O	HpBDPO
	801.349	M+8	C ₁₂ H ₂ ⁷⁹ Br ₄ ⁸¹ Br ₄ O	OBDO

TABLE 11. (continued)

Descriptor Number	Accurate Mass ¹	Ion Type	Elemental Composition	Compound ²
5	717.427	M+4	C ₁₂ H ⁷⁹ Br ₅ ⁸¹ Br ₂ O	HpBDF
	719.425	M+6	C ₁₂ H ⁷⁹ Br ₄ ⁸¹ Br ₃ O	HpBDF
	721.423	M+8	C ₁₂ H ⁷⁹ Br ₃ ⁸¹ Br ₄ O	HpBDF
	733.422	M+4	C ₁₂ H ⁷⁹ Br ₅ ⁸¹ Br ₂ O ₂	HpBDD
	735.420	M+6	C ₁₂ H ⁷⁹ Br ₄ ⁸¹ Br ₃ O ₂	HpBDD
	737.418	M+4	C ₁₂ H ⁷⁹ Br ₃ ⁸¹ Br ₄ O ₂	HpBDD
	754.954	QC		PFK
	770.960	LOCK MASS ALTERNATE		HpTriazine
	801.349	M+8	C ₁₂ H ₂ ⁷⁹ Br ₄ ⁸¹ Br ₄ O	OBDPO
	816.951	LOCK MASS		PFK
	879.260	M+8	C ₁₂ H ⁷⁹ Br ₅ ⁸¹ Br ₄ O	NBDPO
	865.958	QC ALTERNATE		HpTriazine

¹Nuclidic masses used: H = 1.007825 C = 12.000000 ¹³C = 13.003355
O = 15.994915 ⁷⁹Br = 78.91834 ⁸¹Br = 80.91629
¹⁹F = 18.9984

²Compound abbreviations:

Polybrominated dibenzo-p-dioxins

TBDD = Tetrabromodibenzo-p-dioxin
PeBDD = Pentabromodibenzo-p-dioxin
HxBDD = Hexabromodibenzo-p-dioxin
HpBDD = Heptabromodibenzo-p-dioxin
OBDD = Octabromodibenzo-p-dioxin

Polybrominated dibenzofurans

TBDF = Tetrabromodibenzofuran
PeBDF = Pentabromodibenzofuran
HxBDF = Hexabromodibenzofuran
HpBDF = Heptabromodibenzofuran
OBDF = Octabromodibenzofuran

³Labeled Compound

⁴There is only one m/z for ³⁷Cl₄-2378-TCDD (recovery standard).

Polybrominated diphenyl ethers

HxBDPE = Hexabromodiphenyl ether
HpBDPE = Heptabromodiphenyl ether
OBDPE = Octabromodiphenyl ether
NBDPE = Nonabromodiphenyl ether
DBDPE = Decabromodiphenyl ether
PFK = Perfluorokerosene
HpTriazine = Tris-(perfluoroheptyl)-s-Triazine

TABLE 12. DESCRIPTORS, MASSES, M/Z TYPES, AND ELEMENTAL COMPOSITIONS OF THE BCDDS AND BCDFS

Descriptor Number	Accurate mass ¹	m/z Type	Elemental Composition	Compound ²	Primary m/z
1	315.942	M	¹³ C ₁₂ H ₄ ³⁵ Cl ₄ O	TCDF ⁴	
	317.939	M+2	C ₁₂ H ₄ ³⁵ Cl ₃ ³⁷ Cl O	TCDF ⁴	Yes
	327.885	M	C ₁₂ H ₄ ³⁵ Cl ₄ O ₂	TCDD ³	Yes
	330.979	Lock	C ₇ F ₁₃	PFK	
	331.937	M	¹³ C ₁₂ H ₄ ³⁵ Cl ₄ O ₂	TCDD ⁴	
	333.934	M+2	¹³ C ₁₂ H ₄ ³⁵ Cl ₃ ³⁷ Cl O ₂	TCDD ⁴	Yes
	347.851	M	C ₁₂ H ₄ ³⁵ Cl ₃ ⁷⁹ Br O	Br Cl ₃ DF	
	349.849	M+2	C ₁₂ H ₄ ³⁵ Cl ₂ ³⁷ Cl ⁷⁹ Br O	Br Cl ₃ DF	Yes
	363.846	M	C ₁₂ H ₄ ³⁵ Cl ₃ ⁷⁹ Br O ₂	Br Cl ₃ DD	
	365.844	M+2	C ₁₂ H ₄ ³⁵ Cl ₂ ³⁷ Cl ⁷⁹ Br O ₂	Br Cl ₃ DD	Yes

TABLE 12. (continued)

Descriptor Number	Accurate mass ¹	m/z Type	Elemental Composition	Compound ²	Primary m/z
2	351.900	M+2	¹³ C ₁₂ H ₃ ³⁵ Cl ₃ O	PeCDF ₄	
	353.897	M+4	¹³ C ₁₂ H ₃ ³⁵ Cl ₄ ³⁷ Cl O	PeCDF ⁴	
	354.979	Lock	C ₉ F ₃	PFK	
	367.895	M+2	¹³ C ₁₂ H ₃ ³⁵ Cl ₃ O ₂	PeCDD ⁴	Yes
	369.892	M+4	¹³ C ₁₂ H ₃ ³⁵ Cl ₄ ³⁷ Cl O ₂	PeCDD ⁴	
	381.812	M	C ₁₂ H ₃ ³⁵ Cl ₄ ⁷⁹ Br O	Br Cl ₄ DF	
	383.809	M+2	C ₁₂ H ₃ ³⁵ Cl ₃ ³⁷ Cl ⁷⁹ Br O	Br Cl ₄ DF	Yes
	397.807	M	C ₁₂ H ₃ ³⁵ Cl ₄ ⁷⁹ Br O ₂	Br Cl ₄ DD	
	399.804	M+2	C ₁₂ H ₃ ³⁵ Cl ₃ ³⁷ Cl ⁷⁹ Br O ₂	Br Cl ₄ DD	Yes

¹Nuclidic masses used:

H = 1.007825

O = 15.994915

F = 18.9984

C = 12.00000

³⁵Cl = 34.968853⁷⁹Br = 78.91834¹³C = 13.003355³⁷Cl = 36.965903⁸¹Br = 80.91629²Compound abbreviations:Polychlorinated dibenzo-p-dioxins

TCDD = Tetrachlorodibenzo-p-dioxin

PeCDD = Pentachlorodibenzo-p-dioxin

HxCDD = Hexachlorodibenzo-p-dioxin

HpCDD = Heptachlorodibenzo-p-dioxin

OCDD = Octachlorodibenzo-p-dioxin

Polychlorinated dibenzofurans

TCDF = Tetrachlorodibenzofuran

PeCDF = Pentachlorodibenzofuran

HxCDF = Hexachlorodibenzofuran

HpCDF = Heptachlorodibenzofuran

Brominated/Chlorinateddibenzo-p-dioxins and dibenzofuransBrCl₃DD = Bromotrichloro dibenzo-p-dioxinBrCl₄DD = Bromotetrachloro dibenzo-p-dioxinBrCl₃DF = Bromotrichloro dibenzofuranBrCl₄DF = Bromotetrachloro dibenzofuranLock mass and OC compound

PFK = Perfluorokerosene

³There is only one m/z for ³⁷Cl₄-2,3,7,8-TCDD (recovery standard).⁴Labeled compound

TABLE 13. HRGC OPERATING CONDITIONS

Column Type	DB-5	SE-54	SP-2331
Length (m)	60	30	60
i.d. (mm)	0.25	0.25	0.25
Film Thickness (μm)	0.25	0.25	0.20
Carrier Gas	Helium	Helium	Helium
Carrier Gas Flow (mL/min)	1-2	1-2	1-2
Injector temperature ($^{\circ}\text{C}$)	290	308	308
Injection Mode	Splitless	<--- Moving needle --->	
Initial Temperature ($^{\circ}\text{C}$)	200	170.0	150.0
Initial Time (min)	2	7.0	7.0
Rate 1 ($^{\circ}\text{C}/\text{min}$)	5	8.0	10.0
Temperature ($^{\circ}\text{C}$)	220		
Hold Time (min)	16		
Rate 2 (deg. C/min)	5		
Temperature ($^{\circ}\text{C}$)	235		
Hold Time (min)	7		
Rate 2 (deg. C/min)	5		
Final Temperature ($^{\circ}\text{C}$)	330	300.0	250.0
Hold Time (min)	5		

TABLE 14. HRMS OPERATING CONDITIONS

Electron impact ionization	25-70 eV
Mass resolution	>10,000 (10% Valley Definition)
Analysis	Selected ion monitoring (SIM)
Exact masses monitored	Masses shown in Tables 10, 11, 12

TABLE 15. UNLABELED AND LABELED
ANALYTE QUANTIFICATION RELATIONSHIPS

Analyte	Internal Standard Used During Quantification
2,3,7,8-TCDD	¹³ C ₁₂ -2,3,7,8-TCDD
Other TCDDs	¹³ C ₁₂ -2,3,7,8-TCDD
³⁷ Cl ₄ -2,3,7,8-TCDD	¹³ C ₁₂ -2,3,7,8-TCDD
1,2,3,7,8-PeCDD	¹³ C ₁₂ -1,2,3,7,8-PeCDD
Other PeCDDs	¹³ C ₁₂ -1,2,3,7,8-PeCDD
1,2,3,4,7,8-HxCDD	¹³ C ₁₂ -1,2,3,6,7,8-HxCDD
1,2,3,6,7,8-HxCDD	¹³ C ₁₂ -1,2,3,6,7,8-HxCDD
1,2,3,7,8,9-HxCDD	¹³ C ₁₂ -1,2,3,6,7,8-HxCDD
Other HxCDDs	¹³ C ₁₂ -1,2,3,6,7,8-HxCDD
1,2,3,4,6,7,8-HpCDD	¹³ C ₁₂ -1,2,3,4,6,7,8-HpCDD
Other HpCDDs	¹³ C ₁₂ -1,2,3,4,6,7,8-HpCDD
OCDD	¹³ C ₁₂ -OCDD
2,3,7,8-TCDF	¹³ C ₁₂ -2,3,7,8-TCDF
Other TCDFs	¹³ C ₁₂ -2,3,7,8-TCDF
1,2,3,7,8-PeCDF	¹³ C ₁₂ -1,2,3,7,8-PeCDF
2,3,4,7,8-PeCDF	¹³ C ₁₂ -1,2,3,7,8-PeCDF
Other PeCDFs	¹³ C ₁₂ -1,2,3,7,8-PeCDF
1,2,3,4,7,8-HxCDF	¹³ C ₁₂ -1,2,3,4,7,8-HxCDF
1,2,3,6,7,8-HxCDF	¹³ C ₁₂ -1,2,3,4,7,8-HxCDF
1,2,3,7,8,9-HxCDF	¹³ C ₁₂ -1,2,3,4,7,8-HxCDF
2,3,4,6,7,8-HxCDF	¹³ C ₁₂ -1,2,3,4,7,8-HxCDF
Other HxCDFs	¹³ C ₁₂ -1,2,3,4,7,8-HxCDF
1,2,3,4,6,7,8-HpCDF	¹³ C ₁₂ -1,2,3,4,6,7,8-HpCDF
1,2,3,4,7,8,9-HpCDF	¹³ C ₁₂ -1,2,3,4,6,7,8-HpCDF
Other HpCDFs	¹³ C ₁₂ -1,2,3,4,6,7,8-HpCDF
OCDF	¹³ C ₁₂ -OCDD

TABLE 16. INTERNAL STANDARDS QUANTIFICATION RELATIONSHIPS

Internal Standard	Standard Used During Percent Recovery Determination ^a
¹³ C ₁₂ -2,3,7,8-TCDD	¹³ C ₁₂ -1,2,3,4-TCDD
¹³ C ₁₂ -1,2,3,7,8-PeCDD	¹³ C ₁₂ -1,2,3,4-TCDD
¹³ C ₁₂ -1,2,3,6,7,8-HxCDD	¹³ C ₁₂ -1,2,3,7,8,9-HxCDD
¹³ C ₁₂ -1,2,3,4,6,7,8-HpCDD	¹³ C ₁₂ -1,2,3,7,8,9-HxCDD
¹³ C ₁₂ -OCDD	¹³ C ₁₂ -1,2,3,7,8,9-HxCDD
¹³ C ₁₂ -2,3,7,8-TCDF	¹³ C ₁₂ -1,2,3,4-TCDD
¹³ C ₁₂ -1,2,3,7,8-PeCDF	¹³ C ₁₂ -1,2,3,4-TCDD
¹³ C ₁₂ -1,2,3,4,7,8-HxCDF	¹³ C ₁₂ -1,2,3,7,8,9-HxCDD
¹³ C ₁₂ -1,2,3,4,6,7,8-HpCDF	¹³ C ₁₂ -1,2,3,7,8,9-HxCDD

^aSurrogate standards shown in Table 7 may also be used.

TABLE 17. SURROGATE/ALTERNATE STANDARDS QUANTIFICATION RELATIONSHIPS

Surrogate Standard	Standard Used During Percent Recovery Determination
¹³ C ₁₂ -2,3,4,7,8-PeCDF	¹³ C ₁₂ -1,2,3,7,8-PeCDF
¹³ C ₁₂ -1,2,3,4,7,8-HxCDD	¹³ C ₁₂ -1,2,3,6,7,8-HxCDD
¹³ C ₁₂ -1,2,3,6,7,8-HxCDF	¹³ C ₁₂ -1,2,3,4,7,8-HxCDF
¹³ C ₁₂ -1,2,3,4,7,8,9-HpCDF	¹³ C ₁₂ -1,2,3,4,6,7,8-HpCDF

[*Note: Other surrogate standards may be used instead*]

TABLE 18. QUANTIFICATION RELATIONSHIPS OF THE CARBON-LABELED STANDARDS AND THE ANALYTES

Analytes	Quantification Standard
2,3,7,8-TBDD	¹³ C ₁₂ -2,3,7,8-TBDD
2,3,7,8-TBDF	¹³ C ₁₂ -2,3,7,8-TBDF
1,2,3,7,8-PeBDD	¹³ C ₁₂ -1,2,3,7,8-PeBDD
1,2,3,7,8-PeBDF	¹³ C ₁₂ -1,2,3,7,8-PeBDF
2,3,4,7,8-PeBDF	¹³ C ₁₂ -1,2,3,7,8-PeBDF
1,2,3,4,7,8-HxBDD	¹³ C ₁₂ -1,2,3,7,8-PeBDD

[Note: 0.5 ng ³⁷Cl₁-2,3,7,8-TCDD spiked to the extract prior to final concentration to 60 μL was used to determine the method efficiency (% recovery of the ¹³C₁₂-labeled PBDDs/PBDFs).

- Additional 2,3,7,8-substituted PBDDs/PBDFs are now commercially available.
- Retention Index for the PBDDs/PBDFs were published by Sovocool, et al., *Chemosphere* 16, 221-114, 1987; and Donnelly, et al., *Biomedical Environmental Mass Spectrometry*, 14, pp. 465-472, 1987.]

TABLE 19. THEORETICAL ION ABUNDANCE RATIOS AND CONTROL LIMITS FOR PCDDS AND PCDFS

No. of Chlorine Atoms	m/z's Forming Ratio	Theoretical Ratio	Control Limits ¹	
			Lower	Upper
4 ²	M/M+2	0.77	0.65	0.89
5	M+2/M+4	1.55	1.32	1.78
6	M+2/M+4	1.24	1.05	1.43
6 ³	M/M+2	0.51	0.43	0.59
7	M+2/M+4	1.04	0.88	1.20
7 ⁴	M/M+2	0.44	0.37	0.51
8	M+2/M+4	0.89	0.76	1.02

¹Represent ± 15% windows around the theoretical ion abundance ratios.

²Does not apply to ³⁷Cl₁-2,3,7,8-TCDD (cleanup standard).

³Used for ¹³C₁₂-HxCDF only.

⁴Used for ¹³C₁₂-HpCDF only.

TABLE 20. THEORETICAL ION ABUNDANCE RATIOS AND CONTROL LIMITS FOR PBDDS AND PBDFS

Number of Bromine Atoms	Ion Type	Theoretical Ratio	Control Limits	
			Lower	Upper
4	M+2/M+4	0.68	0.54	0.82
4	M+4/M+6	1.52	1.22	1.82
5	M+2/M+4	0.51	0.41	0.61
5	M+4/M+6	1.02	0.82	1.22
6	M+4/M+6	0.77	0.62	0.92
6	M+6/M+8	1.36	1.09	1.63
7	M+4/M+6	0.61	0.49	0.73
7	M+6/M+8	1.02	0.82	1.22

TABLE 21. MINIMUM REQUIREMENTS FOR INITIAL AND DAILY CALIBRATION RESPONSE FACTORS

Compound	Relative Response Factors	
	Initial Calibration RSD	Daily Calibration % Difference
Unlabeled Analytes		
2,3,7,8-TCDD	25	25
2,3,7,8-TCDF	25	25
1,2,3,7,8-PeCDD	25	25
1,2,3,7,8-PeCDF	25	25
2,3,4,7,8-PeCDF	25	25
1,2,4,5,7,8-HxCDD	25	25
1,2,3,6,7,8-HxCDD	25	25
1,2,3,7,8,9-HxCDD	25	25
1,2,3,4,7,8-HxCDF	25	25
1,2,3,6,7,8-HxCDF	25	25
1,2,3,7,8,9-HxCDF	25	25
2,3,4,6,7,8-HxCDF	25	25
1,2,3,4,6,7,8-HpCDD	25	25
1,2,3,4,6,7,8-HpCDF	25	25
OCDD	25	25
OCDF	30	30
Internal Standards		
¹³ C ₁₂ -2,3,7,8-TCDD	25	25
¹³ C ₁₂ -1,2,3,7,8-PeCDD	30	30
¹³ C ₁₂ -1,2,3,6,7,8-HxCDD	25	25
¹³ C ₁₂ -1,2,3,4,6,7,8-HpCDD	30	30

TABLE 21. (continued)

Compound	Relative Response Factors	
	Initial Calibration RSD	Daily Calibration % Difference
¹³ C ₁₂ -OCDD	30	30
¹³ C ₁₂ -2,3,7,8-TCDF	30	30
¹³ C ₁₂ -1,2,3,7,8-PeCDF	30	30
¹³ C ₁₂ -1,2,3,4,7,8-HxCDF	30	30
¹³ C ₁₂ -1,2,3,4,6,7,8-HpCDF	30	30
Surrogate Standards		
³⁷ Cl ₄ -2,3,7,8-TCDD	25	25
¹³ C ₁₂ -2,3,4,7,8-PeCDF	25	25
¹³ C ₁₂ -1,2,3,4,7,8-HxCDD	25	25
¹³ C ₁₂ -1,2,3,4,7,8-HxCDF	25	25
¹³ C ₁₂ -1,2,3,4,7,8,9-HpCDF	25	25

TABLE 22. 2,3,7,8-TCDD EQUIVALENT FACTORS (TEFS)¹
FOR THE POLYCHLORINATED DIBENZODIOXINS
AND POLYCHLORINATED DIBENZOFURANS

Number	Compound	TEF
1	2,3,7,8-TCDD	1.00
2	1,2,3,7,8-PeCDD	0.50
3	1,2,3,4,7,8-HxCDD	0.1
4	1,2,3,6,7,8-HxCDD	0.1
5	1,2,3,7,8,9-HxCDD	0.1
6	1,2,3,4,6,7,8-HpCDD	0.01
7	OCDD	0.001
8	2,3,4,7,8-TCDF	0.10
9	1,2,3,7,8-PeCDF	0.05
10	2,3,4,7,8-PeCDF	0.5
11	1,2,3,4,7,8-HxCDF	0.1
12	1,2,3,6,7,8-HxCDF	0.1
13	1,2,3,7,8,9-HxCDF	0.1
14	2,3,4,6,7,8-HxCDF	0.1
15	1,2,3,4,6,7,8-HpCDF	0.01
16	1,2,3,4,7,8,9-HpCDF	0.01
17	OCDF	0.001

¹Interim procedures for Estimating Risks associated with Exposures to mixtures of Chlorinated Dibenzo-p-Dioxins and Dibenzofurans (CDDs/CDFs), WPA-625/3-89-016, March 1989.

[Note: The same TEFs are assigned to the PBDDs/PBDFs and BCDDs/BCDFs.]

TABLE 23. MINIMUM SAMPLING EQUIPMENT CALIBRATION AND ACCURACY REQUIREMENTS

Equipment	Acceptance limits	Frequency and method of measurement	Action if requirements are not met
<u>Sampler</u>	Indicated flow rate = true flow rate $\pm 10\%$.	Calibrate with certified transfer standard on receipt, after maintenance on sampler, and any time audits or flow checks deviate more than $\pm 10\%$ from the indicated flow rate or $\pm 10\%$ from the design flow rate.	Recalibrate
<u>Associated equipment</u>			
Sampler on/off timer	± 30 min/24 hour	Check at purchase and routinely on sample-recovery days	Adjust or replace
Elapsed-time meter	± 30 min/24 hour	Compare with a standard time-piece of known accuracy at receipt and at 6-month intervals	Adjust or replace
Flowrate transfer standard (orifice device)	Check at receipt for visual damage	Recalibrate annually against positive displacement standard volume meter	Adopt new calibration curve

TABLE 24. FORMAT FOR TABLE OF ANALYTICAL RESULTS

IDENTIFICATION					
AIR SAMPLER EFFICIENCY (% RECOVERY)					
¹³ C ₁₂ -1,2,3,4,-TCDD					
METHOD EFFICIENCY (% RECOVERY)					
¹³ C ₁₂ -2,3,7,8-TCDF					
¹³ C ₁₂ -2,3,7,8-TCDD					
¹³ C ₁₂ -1,2,3,7,8-PeCDF					
¹³ C ₁₂ -1,2,3,7,8-PeCDD					
¹³ C ₁₂ -1,2,3,4,7,8-HxCDF					
¹³ C ₁₂ -1,2,3,6,7,8-HxCDD					
¹³ C ₁₂ -1,2,3,4,6,7,8-HpCDD					
¹³ C ₁₂ -OCDD					
CONCENTRATIONS DETECTED or MDL (pg/m ³)					
TCDDs (TOTAL) ¹					
2,3,7,8-TCDD					
PeCDDs (TOTAL)					
1,2,3,7,8-PeCDD					
HxCDDs (TOTAL)					
1,2,3,4,7,8-HxCDD					
1,2,3,6,7,8-HxCDD					
1,2,3,7,8,9-HxCDD					
HpCDDs (TOTAL)					
1,2,3,4,6,7,8-HpCDD					

TABLE 24. (continued)


IDENTIFICATION					
OCDD					
TCDFs (TOTAL)					
2,3,7,8-TCDF					
PeCDFs (TOTAL)					
1,2,3,7,8-PeCDF					
2,3,4,7,8-PeCDF					
HxCDFs (TOTAL)					
1,2,3,4,7,8-HxCDF					
1,2,3,6,7,8-HxCDF					
1,2,3,7,8,9-HxCDF					
2,3,4,6,7,8-HxCDF					
HpCDFs (TOTAL)					
1,2,3,4,6,7,8-HpCDF					
1,2,3,4,7,8,9-HpCDF					
OCDF					

¹(TOTAL) = All congeners, including the 2,3,7,8-substituted congeners.

ND = Not detected at specified minimum detection limit (MDL).

[Note: Please refer to text for discussion and qualification that must accompany the results.]

Figure 1. Dibenzo-p-dioxin and dibenzofuran structures.






Figure 2. Typical dioxins/furan high volume air sampler.




Figure 3a. Typical absorbent cartridge assembly for sampling dioxin/furans.





Figure 3b. Typical glass PUF cartridge (1) and shipping container (2) for use with hi-vol sampling systems.




Figure 4. Portable high volume air sampler developed by EPA.




Figure 5. Positive displacement rootsmeter used to calibrate orifice transfer standard.

COMPENDIUM METHOD TO-9A
ORIFICE CALIBRATION DATA SHEET

T₁ _____

Name _____

P₁ _____ mmHg

Date _____

Orifice No. _____

Rootsmer No. _____

Resistance Plants (No. of holes)	Air Volume Measured by Rootsmer V _m		Standard Volume, V _{std} (std m ³)	Time for Air Volume to Pass Through Rootsmer, θ (min)	Rootsmer Pressure Differential, ΔP (mm Hg)	Pressure Drop Across Orifice, ΔH (in. H ₂ O)	X-Axis Standard Flowrate, Q _{std} (std m ³ /min)	Y-axis $\sqrt{\Delta H(P_1/P_{std})(298/T_1)}$ value
	(R ³)	(m ³)						
5	200	5.66						
7	200	5.66						
10	300	8.50						
13	300	8.50						
18	300	8.50						

Factors: (R³)(0.02832 $\frac{m^3}{R^3}$) = m³ and (in. Hg) 25.4 ($\frac{mm\ Hg}{in.\ Hg}$) = mm Hg

Calculation Equations:

$$1. \quad V_{std} = V_m \left(\frac{P_1 - \Delta P}{P_{std}} \right) \left(\frac{T_{std}}{T_1} \right)$$

where:

$$T_{std} = 296^\circ K$$

$$P_{std} = 760.0\ mm\ Hg$$

$$2. \quad Q_{std} = \frac{V_{std}}{\theta}$$

Figure 6. Orifice calibration data sheet.




Figure 7. Field calibration configuration of the dioxin/furan sampler.

COMPENDIUM METHOD TO-9A
FIELD CALIBRATION DATA SHEET DIOXIN/FURAN SAMPLER CALIBRATION

Sampler ID: _____ Calibration Orifice ID: _____
 Sampler Location: _____ Job No.: _____
 High Volume Transfer Orifice Data:
 Correlation Coefficient (CC1): _____ Slope (M1): _____
 (CC2): _____ (M2): _____
 Intercept (B1): _____
 (B2): _____

Calibration Date: ____ Time: _____
 Calibration Ambient Temperature: ____°F ____°C CALIBRATOR'S SIGNATURE
 Calibration Ambient Barometric Pressure: ____ "Hg ____ mm Hg
 Calibration set point (SP): _____

SAMPLER CALIBRATION

Actual values from calibration		Calibrated values		
Orifice manometer, inches (Y1)	Monitor magnehelic, inches (Y2)	Orifice manometer (Y3)	Monitor magnehelic (Y4)	Calculated value orifice flow, scm (X1)
	70			
	60			
	50			
	40			
	30			
	20			
	10			

Definitions

- Y1 = Calibration orifice reading, in. H₂O
- Y2 = Monitor magnehelic reading, in. H₂O
- P_a = Barometric pressure actual, mm Hg
- B1 = Manufacturer's Calibration orifice Intercept
- M1 = Manufacturer's Calibration orifice manometer slope
- Y3 = Calculated value for orifice manometer
= $[Y1(P_a/760)(298/\{T_a + 273\})]^{1/2}$
- Y4 = Calculated value for magnehelic
= $[Y2(P_a/760)(298/\{T_a + 273\})]^{1/2}$
- X1 = Calculated value orifice flow, scm
= $\frac{Y3 - B1}{M1}$
- P_{std} = Barometric pressure standard, 760 mm Hg
- T_a = Temperature actual, °C
- T_{std} = Temperature standard, 25°C

Figure 8. Orifice transfer field calibration data sheet.




Figure 9. Relationship between orifice transfer standard and flow rate through sampler.

**COMPENDIUM METHOD TO-9A
FIELD TEST DATA SHEET
GENERAL INFORMATION**

Sampler I.D. No.: _____
 Lab PUF Sample No.: _____
 Sample location: _____

Operator: _____
 Other: _____

PUF Cartridge Certification Date: _____
 Date/Time PUF Cartridge Installed: _____
 Elapsed Timer: _____
 Start _____
 Stop _____
 Diff. _____

Barometric pressure ("Hg) Start _____ Stop _____
 Ambient Temperature (°F) _____
 Rain Yes _____ Yes _____
 No _____ No _____

Sampling

Sampling time
 Start _____
 Stop _____
 Diff. _____

M1 _____ B1 _____
 M2 _____ B2 _____

Audit flow check within ±10 of set point
 _____ Yes
 _____ No

TIME	TEMP	BAROMETRIC PRESSURE	MAGNEHELIC READING	CALCULATED FLOW RATE (scmm)	READ BY
Avg.					

Comments

Figure 10. Field test data sheet.




Figure 11. Chromatograms showing the window defining mix. First and last eluting PCDF and PCDD isomer in each group are shown above the respective internal quantification standard (IQS).




Figure 11. (continued)




Figure 11. (continued)




Figure 11. (continued)




Figure 12. HRGC-HRMS column performance mix showing acceptable separation of 2,3,7,8-TCDD from the adjacent isomers.




Figure 13. Extracted ion current profiles (EICP) for 2,3,7,8-TCDF and labeled standard.




Figure 14. Extracted ion current profiles (HLCP) for 2,3,7,8-TCDD and labeled standard.




Figure 15. Extracted ion current profiles (ELCP) for 2,3,7,8-substituted PeCDF and labeled standard.




Figure 16. Extracted ion current profiles (EICP) for 2,3,7,8-substituted PeCDD and labeled standard.




Figure 17. Extracted ion current profiles (EICP) for 2,3,7,8-substituted HxCDF and labeled standard.




Figure 18. Extracted ion current profiles (EICP) for 2,3,7,8-substituted HxCDD and labeled standard.




Figure 19. Extracted ion current profiles (EICP) for 2,3,7,8-substituted HpCDF and labeled standard.




Figure 20. Extracted ion current profiles (EICP) for 2,3,7,8-substituted HpCDD and labeled standard.




Figure 21. Extracted ion current profiles (EICP) for OCDF and labeled standard.




Figure 22. Extracted ion current profiles (EICP) for OCDD and labeled standard.




Figure 23. Extracted ion current profiles (EICP) for 2,3,7,8-TCDF and labeled standard in a complex environmental sample showing presence of other TCDF isomers.

This Page Intentionally Left Blank



**Compendium of Methods
for the Determination of
Toxic Organic Compounds
in Ambient Air**

Second Edition

Compendium Method TO-13A

**Determination of Polycyclic Aromatic
Hydrocarbons (PAHs) in Ambient Air Using Gas
Chromatography/Mass Spectrometry (GC/MS)**

**Center for Environmental Research Information
Office of Research and Development
U.S. Environmental Protection Agency
Cincinnati, OH 45268**

January 1999

Method TO-13A Acknowledgements

This Method was prepared for publication in the *Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air, Second Edition* (EPA/625/R-96/010b), under Contract No. 68-C3-0315, WA No. 3-10, by Midwest Research Institute (MRI), as a subcontractor to Eastern Research Group, Inc. (ERG), and under the sponsorship of the U.S. Environmental Protection Agency (EPA). Justice A. Manning, John O. Burckle, and Scott R. Hedges, Center for Environmental Research Information (CERI), and Frank F. McElroy, National Exposure Research Laboratory (NERL), all in the EPA Office of Research and Development (ORD), were responsible for overseeing the preparation of this method. Additional support was provided by other members of the Compendia Workgroup, which include:

- John O. Burckle, U.S. EPA, ORD, Cincinnati, OH
- James L. Cheney, Corps of Engineers, Omaha, NB
- Michael Davis, U.S. EPA, Region 7, KC, KS
- Joseph B. Elkins Jr., U.S. EPA, OAQPS, RTP, NC
- Robert G. Lewis, U.S. EPA, NERL, RTP, NC
- Justice A. Manning, U.S. EPA, ORD, Cincinnati, OH
- William A. McClenny, U.S. EPA, NERL, RTP, NC
- Frank F. McElroy, U.S. EPA, NERL, RTP, NC
- Heidi Schultz, ERG, Lexington, MA
- William T. "Jerry" Winberry, Jr., EnviroTech Solutions, Cary, NC

Method TO-13 was originally published in March of 1989 as one of a series of peer-reviewed methods in the second supplement to *Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air*, EPA 600/4-89-018. In an effort to keep these methods consistent with current technology, Method TO-13 has been revised and updated as Method TO-13A in this Compendium to incorporate new or improved sampling and analytical technologies.

This method is the result of the efforts of many individuals. Gratitude goes to each person involved in the preparation and review of this methodology.

Author(s)

- William T. "Jerry" Winberry, Jr., EnviroTech Solutions, Cary, NC
- Greg Jungclaus, Midwest Research Institute, Kansas City, MO

Peer Reviewers

- Nancy Wilson, U.S. EPA, NERL, RTP, NC
- Joan Bursey, ERG, Morrisville, NC
- Irene D. DeGraff, Supelco, Bellefonte, PA
- Jane Chuang, Battelle Laboratories, Cincinnati, OH
- Robert G. Lewis, U.S. EPA, NERL, RTP, NC
- Lauren Drees, U.S. EPA, NRMRL, Cincinnati, OH

Finally, recognition is given to Frances Beyer, Lynn Kaufman, Debbie Bond, Cathy Whitaker, and Kathy Johnson of Midwest Research Institute's Administrative Services staff whose dedication and persistence during the development of this manuscript has enabled its production.

DISCLAIMER

This Compendium has been subjected to the Agency's peer and administrative review, and it has been approved for publication as an EPA document. Mention of trade names or commercial products does not constitute endorsement or recommendation for use.

This page intentionally left blank

METHOD TO-13A

Determination of Polycyclic Aromatic Hydrocarbons (PAHs) in Ambient Air Using Gas Chromatography/Mass Spectrometry (GC/MS)

TABLE OF CONTENTS

	<u>Page</u>
1. Scope	13A-1
2. Summary of Method	13A-2
3. Significance	13A-3
4. Applicable Documents	13A-3
4.1 ASTM Standards	13A-3
4.2 EPA Documents	13A-3
4.3 Other Documents	13A-4
5. Definitions	13A-4
6. Limitations and Interferences	13A-5
6.1 Limitations	13A-5
6.2 Interferences	13A-6
7. Safety	13A-6
8. Apparatus	13A-7
8.1 Sampling	13A-7
8.2 Sample Clean-Up and Concentration	13A-8
8.3 Sample Analysis	13A-9
9. Equipment and Materials	13A-10
9.1 Materials for Sample Collection	13A-10
9.2 Sample Clean-up and Concentration	13A-11
9.3 GC/MS Sample Analysis	13A-11
10. Preparation of PUF Sampling Cartridge	13A-12
10.1 Summary of Method	13A-12
10.2 Preparation of Sampling Cartridge	13A-12
10.3 Procedure for Certification of PUF Cartridge Assembly	13A-13
10.4 Deployment of Cartridges for Field Sampling	13A-14
11. Assembly, Calibration, and Collection Using Sampling System	13A-15
11.1 Sampling Apparatus	13A-15
11.2 Calibration of Sampling System	13A-15
11.3 Sample Collection	13A-22

TABLE OF CONTENTS, CONTINUED

	<u>Page</u>
12. Sample Extraction, Concentration, and Cleanup	13A-23
12.1 Sample Identification	13A-23
12.2 Soxhlet Extraction and Concentration	13A-24
12.3 Sample Cleanup	13A-25
13. Gas Chromatography with Mass Spectrometry Detection	13A-26
13.1 General	13A-26
13.2 Calibration of GC/MS/DS	13A-26
13.3 GC/MS Instrument Operating Conditions	13A-29
13.4 Sample Analysis by GC/MS	13A-37
14. Quality Assurance/Quality Control (QA/QC)	13A-41
14.1 General System QA/QC	13A-41
14.2 Process, Field, and Solvent Blanks	13A-42
15. References	13A-42

METHOD TO-13A

Determination of Polycyclic Aromatic Hydrocarbons (PAHs) in Ambient Air Using Gas Chromatography/Mass Spectrometry (GC/MS)

1. Scope

1.1 Polycyclic aromatic hydrocarbons (PAHs) have received increased attention in recent years in air pollution studies because some of these compounds are highly carcinogenic or mutagenic. In particular, benzo[a]pyrene (B[a]P) has been identified as being highly carcinogenic. To understand the extent of human exposure to B[a]P and other PAHs, reliable sampling and analytical methods are necessary. This document describes a sampling and analysis procedure for common PAHs involving the use of a combination of quartz filter and sorbent cartridge with subsequent analysis by gas chromatography with mass spectrometry (GC/MS) detection. The analytical methods are modifications of EPA Test Method 610 and 625, *Methods for Organic Chemical Analysis of Municipal and Industrial Wastewater*, and Methods 8000, 8270, and 8310, *Test Methods for Evaluation of Solid Waste*.

1.2 Fluorescence methods were among the very first methods used for detection of B[a]P and other PAHs as carcinogenic constituents of coal tar (1-7). Fluorescence methods are capable of measuring subnanogram quantities of PAHs, but tend to be fairly non-selective. The normal spectra obtained are often intense and lack resolution. Efforts to overcome this difficulty led to the use of ultraviolet (UV) absorption spectroscopy (8) as the detection method coupled with pre-specified techniques involving liquid chromatography (LC) and thin layer chromatography (TLC) to isolate specific PAHs, particularly B[a]P. As with fluorescence spectroscopy, the individual spectra for various PAHs are unique, although portions of spectra for different compounds may be the same. As with fluorescence techniques, the possibility of spectral overlap requires complete separation of sample components to ensure accurate measurement of component levels. Hence, the use of UV absorption coupled with pre-speciation involving LC and TLC and fluorescence spectroscopy declined and was replaced with the more sensitive high performance liquid chromatography (HPLC) with UV/fluorescence detection (9) or highly sensitive and specific gas chromatography/mass spectrometry (GC/MS) for detection (10-11).

1.3 The choice of GC/MS as the recommended procedure for analysis of B[a]P and other PAHs was influenced by its sensitivity and selectivity, along with its ability to analyze complex samples.

1.4 The analytical methodology has consequently been defined, but the sampling procedures can reduce the validity of the analytical results. Recent studies (12-17) have indicated that non-volatile PAHs (vapor pressure $<10^{-8}$ mm Hg) may be trapped on the filter, but post-collection volatilization problems may distribute the PAHs downstream of the filter to the back-up sorbent. A wide variety of sorbents such as Tenax®, XAD-2® and polyurethane foam (PUF) have been used to sample common PAHs. All sorbents have demonstrated high collection efficiency for B[a]P in particular. In general, XAD-2® resin has a higher collection efficiency (18-21) for volatile PAHs than PUF, as well as a higher retention efficiency. PUF cartridges, however, are easier to handle in the field and maintain better flow characteristics during sampling. Likewise, PUF has demonstrated (22) its capability in sampling organochlorine pesticides, polychlorinated biphenyls (22), and polychlorinated dibenzo-p-dioxins (23). PUF also has demonstrated a lower recovery efficiency and storage capability for naphthalene than XAD-2®. There have been no significant losses of PAHs up to 30 days of storage at room temperature (23 °C) using XAD-2®. It also appears that XAD-2® resin has a higher collection efficiency for volatile PAHs than PUF, as well as a higher retention efficiency for both volatile and reactive PAHs.

Consequently, while the literature cites weaknesses and strengths of using either XAD-2® or PUF, this method includes the utilization of PUF as the primary sorbent.

1.5 This method includes the qualitative and quantitative analysis of the following PAHs (see Figure 1) specifically by utilizing PUF as the sorbent followed by GC/MS analysis:

Acenaphthene (low collection efficiency; see Section 6.1.3)	Coronene
Acenaphthylene (low collection efficiency; see Section 6.1.3)	Dibenz(a,h)anthracene
Anthracene	Fluoranthene
Benz(a)anthracene	Fluorene
Benzo(a)pyrene	Benzo(b)fluoranthene
Benzo(e)pyrene	Indeno(1,2,3-cd)pyrene
Benzo(g,h,i)perylene	Naphthalene (low collection efficiency; see Section 6.1.3)
Benzo(k)fluoranthene	Phenanthrene
Chrysene	Pyrene
	Perylene

The GC/MS method is applicable to the determination of PAHs compounds involving three member rings or higher. Naphthalene, acenaphthylene, and acenaphthene have only ~35 percent recovery when using PUF as the sorbent. Nitro-PAHs have *not* been fully evaluated using this procedure; therefore, they are not included in this method.

1.6 With optimization to reagent purity and analytical conditions, the detection limits for the GC/MS method range from 1 ng to 10 pg based on field experience.

2. Summary of Method

2.1 Filters and sorbent cartridges (containing PUF or XAD-2®) are cleaned in solvents and vacuum dried. The filters and sorbent cartridges are stored in screw-capped jars wrapped in aluminum foil (or otherwise protected from light) before careful installation on the sampler.

2.2 Approximately 300 m³ of air is drawn through the filter and sorbent cartridge using a high-volume flow rate air sampler or equivalent.

2.3 The amount of air sampled through the filter and sorbent cartridge is recorded, and the filter and cartridge are placed in an appropriately labeled container and shipped along with blank filter and sorbent cartridges to the analytical laboratory for analysis.

2.4 The filters and sorbent cartridge are extracted by Soxhlet extraction with appropriate solvent. The extract is concentrated by Kuderna-Danish (K-D) evaporator, followed by silica gel cleanup using column chromatography to remove potential interferences prior to analysis by GC/MS.

2.5 The eluent is further concentrated by K-D evaporation, then analyzed by GC/MS. The analytical system is verified to be operating properly and calibrated with five concentration calibration solutions.

2.6 A preliminary analysis of the sample extract is performed to check the system performance and to ensure that the samples are within the calibration range of the instrument. If the preliminary analysis indicates non-performance, then recalibrate the instrument, adjust the amount of the sample injected, adjust the calibration solution concentration, and adjust the data processing system to reflect observed retention times, etc.

2.7 The samples and the blanks are analyzed and used (along with the amount of air sampled) to calculate the concentration of PAHs in the air sample.

3. Significance

3.1 As discussed in Section 1, several documents have been published that describe sampling and analytical approaches for common PAHs. The attractive features of these methods have been combined in this procedure. Although this method has been validated in the laboratory, one must use caution when employing it for specific applications.

3.2 Because of the relatively low levels of common PAHs in the environment, the methodology suggest the use of high volume (0.22 m³/min) sampling technique to acquire sufficient sample for analysis. However, the volatility of certain PAHs prevents efficient collection on filter media alone. Consequently, this method utilizes both a filter and a backup sorbent cartridge, which provides for efficient collection of most PAHs involving three member rings or higher.

4. Applicable Documents

4.1 ASTM Standards

- **Method D1356** *Definitions of Terms Relating to Atmospheric Sampling and Analysis.*
- **Method 4861-94** *Standard Practice for Sampling and Analysis of Pesticides and Polychlorinated Biphenyl in Air*
- **Method E260** *Recommended Practice for General Gas Chromatography Procedures.*
- **Method E355** *Practice for Gas Chromatography Terms and Relationships.*
- **Method E682** *Practice for Liquid Chromatography Terms and Relationships.*

4.2 EPA Documents

- *Technical Assistance Document for Sampling and Analysis of Toxic Organic Compounds in Ambient Air*, U. S. Environmental Protection Agency, EPA-600/4-83-027, June 1983.
- *Quality Assurance Handbook for Air Pollution Measurement Systems*, U. S. Environmental Protection Agency, EPA-600/R-94-038b, May 1994.
- *Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air: Method TO-13, Second Supplement*, U. S. Environmental Protection Agency, EPA-600/4-89-018, March 1989.

4.3 Other Documents

- Existing Procedures (24-32).
- Ambient Air Studies (33-50).
- General Metal Works, Inc., "Operating Procedures for Model PS-1 Sampler," Village of Cleves, OH 45002 (800-543-7412).
- Illinois Environmental Protection Agency, Division of Air Quality, "Chicago Air Quality: PCB Air Monitoring Plan (Phase 2)," Chicago, IL, IEAP/APC/86/011, April 1986.
- Thermo Environmental, Inc. (formerly Wedding and Associates), "Operating Procedures for the Thermo Environmental Semi-Volatile Sampler," 8 West Forge Parkway, Franklin, MA 02038 (508-520-0430).
- American Chemical Society (ACS), "Sampling for Organic Chemicals in Air," *ACS Professional Book*, ACS, Washington, D.C., 1996.
- International Organization for Standardization (ISO), "Determination of Gas and Particle-Phase Polynuclear Aromatic Hydrocarbons in Ambient Air - Collected on Sorbent-Backed Filters with Gas Chromatographic/Mass Spectrometric Analysis," ISO/TC 146/SC 3/WG 17N, Case Postale 56, CH-1211, Genève 20, Switzerland.

5. Definitions

[Note: Definitions used in this document and in any user-prepared standard operating procedures (SOPs) should be consistent with ASTM Methods D1356, E260, and E255. All abbreviations and symbols are defined within this document at point of use.]

5.1 Retention time (RT)-time to elute a specific chemical from a chromatographic column. For a specific carrier gas flow rate, RT is measured from the time the chemical is injected into the gas stream until it appears at the detector.

5.2 Sampling efficiency (SE)-ability of the sampler to trap and retain PAHs. The %SE is the percentage of the analyte of interest collected and retained by the sampling medium when it is introduced into the air sampler and the sampler is operated under normal conditions for a period of time equal to or greater than that required for the intended use.

5.3 Dynamic retention efficiency-ability of the sampling medium to retain a given PAH that has been added to the sorbent trap in a spiking solution when air is drawn through the sampler under normal conditions for a period of time equal to or greater than that required for the intended use.

5.4 Polycyclic aromatic hydrocarbons (PAHs)-two or more fused aromatic rings.

5.5 Method detection limit (MDL)-the minimum concentration of a substance that can be measured and reported with confidence and that the value is above zero.

5.6 Kuderna-Danish apparatus-the Kuderna-Danish (K-D) apparatus is a system for concentrating materials dissolved in volatile solvents.

5.7 MS-SCAN-the GC is coupled to a mass spectrometer where the instrument is programmed to acquire all ion data.

5.8 Sublimation-the direct passage of a substance from the solid state to the gaseous state and back into the solid form without at any time appearing in the liquid state. Also applied to the conversion of solid to vapor without the later return to solid state, and to a conversion directly from the vapor phase to the solid state.

5.9 Surrogate standard-a chemically inert compound (not expected to occur in the environmental sample) that is added to each sample, blank, and matrix-spiked sample before extraction and analysis. The recovery of the surrogate standard is used to monitor unusual matrix effects, gross sample processing errors, etc. Surrogate recovery is evaluated for acceptance by determining whether the measured concentration falls within acceptable limits.

5.10 CAL-calibration standards are defined as five levels of calibration: CAL 1, CAL 2, CAL 3, CAL 4, and CAL 5. CAL 1 is the lowest concentration and CAL 5 is the highest concentration. CAL 3, which is the mid-level standard, is designated as the solution to be used for continuing calibrations.

5.11 Continuing calibration check-a solution of method analytes used to evaluate the mass spectrometer response over a period of time. A continuing calibration check (CCC) is performed once each 12-hour period. The CCC solution (CAL 3) is the standard of the calibration curve.

5.12 GC Response (A_x)-the peak area or height of analyte, x.

5.13 Internal standard (IS)-a compound added to a sample extract in known amounts and used to calibrate concentration measurements of other compounds that are sample components. The internal standard must be a compound that is not a sample component.

6. Limitations and Interferences

6.1 Limitations

6.1.1 PAHs span a broad spectrum of vapor pressures (e.g., from 1.1×10^{-2} kPa for naphthalene to 2×10^{-13} kPa for coronene at 25°C). PAHs that are frequently found in ambient air are listed in Table 1. Those with vapor pressures above approximately 10^{-8} kPa will be present in the ambient air substantially distributed between the gas and particulate phases. This method will permit the collection of both phases.

6.1.2 Particulate-phase PAHs will tend to be lost from the particle filter during sampling due to volatilization. Therefore, separate analysis of the filter will not reflect the concentrations of the PAHs originally associated with particles, nor will analysis of the sorbent provide an accurate measure of the gas phase. Consequently, this method calls for *extraction of the filter and sorbent together* to permit accurate measurement of total PAH air concentrations.

6.1.3 Naphthalene, acenaphthylene, and acenaphthene possess relatively high vapor pressures and may not be efficiently trapped by this method when using PUF as the sorbent. The sampling efficiency for naphthalene has been determined to be about 35 percent for PUF. The user is encouraged to use XAD-2® as the sorbent if these analytes are part of the target compound list (TCL).

6.2 Interferences

6.2.1 Method interferences may be caused by contaminants in solvents, reagents, glassware, and other sample processing hardware that result in discrete artifacts and/or elevated baselines in the detector profiles. All of these materials must be routinely demonstrated to be free from interferences under the conditions of the analysis by running laboratory reagent blanks.

6.2.2 Glassware must be scrupulously cleaned (51). All glassware should be cleaned as soon as possible after use by rinsing with the last solvent used in it and then high-purity acetone and hexane. These rinses should be followed by detergent washing with hot water and rinsing with copious amounts of tap water and several portions of reagent water. The glassware should then be drained dry and heated in a muffle furnace at 400°C for four hours. Volumetric glassware must not be heated in a muffle furnace; rather it should be solvent rinsed with acetone and spectrographic grade hexane. After drying and rinsing, glassware should be sealed and stored in a clean environment to prevent any accumulation of dust or other contaminants. Glassware should be stored inverted or capped with aluminum foil.

[Note: The glassware may be further cleaned by placing in a muffle furnace at 450°C for 8 hours to remove trace organics.]

6.2.3 The use of high purity water, reagents, and solvents helps to minimize interference problems. Purification of solvents by distillation in all-glass systems may be required.

6.2.4 Matrix interferences may be caused by contaminants that are coextracted from the sample. Additional clean-up by column chromatography may be required (see Section 12.3).

6.2.5 During sample transport and analysis, heat, ozone, NO₂, and ultraviolet (UV) light may cause sample degradation. Incandescent or UV-shielded fluorescent lighting in the laboratory should be used during analysis.

6.2.6 The extent of interferences that may be encountered using GC/MS techniques has not been fully assessed. Although GC conditions described allow for unique resolution of the specific PAH compounds covered by this method, other PAH compounds may interfere. The use of column chromatography for sample clean-up prior to GC analysis will eliminate most of these interferences. The analytical system must, however, be routinely demonstrated to be free of internal contaminants such as contaminated solvents, glassware, or other reagents which may lead to method interferences. A laboratory reagent blank should be analyzed for each reagent used to determine if reagents are contaminant-free.

6.2.7 Concern about sample degradation during sample transport and analysis was mentioned above. Heat, ozone, NO₂, and ultraviolet (UV) light also may cause sample degradation. These problems should be addressed as part of the user-prepared standard operating procedure (SOP) manual. Where possible, incandescent or UV-shielded fluorescent lighting should be used during analysis. During transport, field samples should be shipped back to the laboratory chilled (~4°C) using blue ice/dry ice.

7. Safety

7.1 The toxicity or carcinogenicity of each reagent used in this method has not been precisely defined; however, each chemical compound should be treated as a potential health hazard. From this viewpoint, exposure to these chemicals must be reduced to the lowest possible level by whatever means available. The laboratory is responsible for maintaining a current awareness file of Occupational Safety and Health Administration (OSHA) regulations regarding the safe handling of the chemicals specified in this method. A reference file of material safety data sheets (MSDSs) should also be made available to all personnel involved in the chemical analysis. Additional references to laboratory safety are available and are included in the reference list (52-54).

7.2 B[a]P has been tentatively classified as a known or suspected, human or mammalian carcinogen. Many of the other PAHs have been classified as carcinogens. Care must be exercised when working with these substances. This method does not purport to address all of the safety problems associated with its use. It is the responsibility of whomever uses this method to consult and establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use. The user should be thoroughly familiar with the chemical and physical properties of targeted substances (see Table 1 and Figure 1).

7.3 All PAHs should be treated as carcinogens. Neat compounds should be weighed in a glove box. Spent samples and unused standards are toxic waste and should be disposed according to regulations. Counter tops and equipment should be regularly checked with "black light" for fluorescence as an indicator of contamination.

7.4 The sampling configuration (filter and backup sorbent) and collection efficiency for target PAHs has been demonstrated to be greater than 95 percent (except for naphthalene, acenaphthylene and acenaphthene). Therefore, no field recovery evaluation will be required as part of this procedure.

[Note: Naphthalene, acenaphthylene and acenaphthene have demonstrated significant breakthrough using PUF cartridges, especially at summer ambient temperatures. If naphthalene, acenaphthylene and acenaphthene are target PAHs, the user may want to consider replacing the PUF with XAD-2® in order to minimize breakthrough during sampling.]

8. Apparatus

[Note: This method was developed using the PS-1 semi-volatile sampler provided by General Metal Works, Village of Cleves, OH as a guideline. EPA has experience in the use of this equipment during various field-monitoring programs over the last several years. Other manufacturers' equipment should work as well; however, modifications to these procedures may be necessary if another commercially available sampler is selected.]

8.1 Sampling

8.1.1 High-volume sampler (see Figure 2). Capable of pulling ambient air through the filter/sorbent cartridge at a flow rate of approximately 8 standard cubic feet per minute (scfm) (0.225 std m³/min) to obtain a total sample volume of greater than 300 m³ over a 24-hour period. Major manufacturers are:

- Tisch Environmental, Village of Cleves, OH
- Andersen Instruments Inc., 500 Technology Ct., Smyrna, GA
- Thermo Environmental Instruments, Inc., 8 West Forge Parkway, Franklin, MA

Recent EPA studies have concluded that sample volumes *less than* 300 m³ still collect enough PAHs on the filter/PUF for quantitation. The user is encouraged to investigate appropriate sample volume needed to meet project specific data quality objectives.

8.1.2 Sampling module (see Figure 3). Metal filter holder (Part 2) capable of holding a 102-mm circular particle filter supported by a 16-mesh stainless-steel screen and attaching to a metal cylinder (Part 1) capable of holding a 65-mm O.D. (60-mm I.D.) x 125-mm borosilicate glass sorbent cartridge containing PUF or XAD-2®. The filter holder is equipped with inert sealing gaskets (e.g., polytetrafluorethylene) placed on either side of the

filter. Likewise, inert, pliable gaskets (e.g., silicone rubber) are used to provide an air-tight seal at each end of the glass sorbent cartridge. The glass sorbent cartridge is indented 20 mm from the lower end to provide a support for a 16-mesh stainless-steel screen that holds the sorbent. The glass sorbent cartridge fits into Part 1, which is screwed onto Part 2 until the sorbent cartridge is sealed between the silicone gaskets. Major manufacturers are:

- Tisch Environmental, Village of Cleves, OH
- Andersen Instruments Inc., 500 Technology Ct., Smyrna, GA
- Thermo Environmental Instruments, Inc., 8 West Forge Parkway, Franklin, MA

8.1.3 High-volume sampler calibrator. Capable of providing multipoint resistance for the high-volume sampler. Major manufacturers are:

- Tisch Environmental, Village of Cleves, OH
- Andersen Instruments Inc., 500 Technology Ct., Smyrna, GA
- Thermo Environmental Instruments, Inc., 8 West Forge Parkway, Franklin, MA

8.1.4 Ice chest. To hold samples at 4°C or below during shipment to the laboratory after collection.

8.1.5 Data sheets. Used for each sample to record the location and sample time, duration of sample, starting time, and volume of air sampled.

8.2 Sample Clean-Up and Concentration (see Figure 4).

8.2.1 Soxhlet apparatus extractor (see Figure 4a). Capable of extracting filter and sorbent cartridges (5.75-cm x 12.5-cm length), 1,000 mL flask, and condenser, best source.

8.2.2 Pyrex glass tube furnace system. For activating silica gel at 180°C under purified nitrogen gas purge for an hour, with capability of raising temperature gradually, best source.

8.2.3 Glass vial. 40 mL, best source.

8.2.4 Erlenmeyer flask. 50 mL, best source.

[Note: Reuse of glassware should be minimized to avoid the risk of cross contamination. All glassware that is used must be scrupulously cleaned as soon as possible after use. Rinse glassware with the last solvent used in it and then with high-purity acetone and hexane. Wash with hot water containing detergent. Rinse with copious amounts of tap water and several portions of distilled water. Drain, dry, and heat in a muffle furnace at 400°C for 4 hours. Volumetric glassware must not be heated in a muffle furnace; rather, it should be rinsed with high-purity acetone and hexane. After the glassware is dry and cool, rinse it with hexane, and store it inverted or capped with solvent-rinsed aluminum foil in a clean environment.]

8.2.5 White cotton gloves. For handling cartridges and filters, best source.

8.2.6 Minivials. 2 mL, borosilicate glass, with conical reservoir and screw caps lined with Teflon®-faced silicone disks, and a vial holder, best source.

8.2.7 Teflon®-coated stainless steel spatulas and spoons. Best source.

8.2.8 Kuderna-Danish (K-D) apparatus (see Figure 4b). 500 mL evaporation flask (Kontes K-570001-500 or equivalent), 10 mL graduated concentrator tubes (Kontes K570050-1025 or equivalent) with ground-glass stoppers, 1 mL calibrated K-D concentration tubes, and 3-ball macro Snyder Column (Kontes K-570010500, K-50300-0121, and K-569001-219, or equivalent), best source.

8.2.9 Adsorption column for column chromatography (see Figure 4c). 1-cm x 10-cm with stands.

8.2.10 Glove box. For working with extremely toxic standards and reagents with explosion-proof hood for venting fumes from solvents, reagents, etc.

8.2.11 Vacuum oven. Vacuum drying oven system capable of maintaining a vacuum at 240 torr (flushed with nitrogen) overnight.

8.2.12 Concentrator tubes and a nitrogen evaporation apparatus with variable flow rate. Best source.

8.2.13 Laboratory refrigerator. Best source.

8.2.14 Boiling chips. Solvent extracted, 10/40 mesh silicon carbide or equivalent, best source.

8.2.15 Water bath. Heated, with concentric ring cover, capable of $\pm 5^\circ\text{C}$ temperature control, best source.

8.2.16 Nitrogen evaporation apparatus. Best source.

8.2.17 Glass wool. High grade, best source.

8.3 Sample Analysis

8.3.1 Gas Chromatography with Mass Spectrometry Detection Coupled with Data Processing System (GC/MS/DS). The gas chromatograph must be equipped for temperature programming, and all required accessories must be available, including syringes, gases, and a capillary column. The gas chromatograph injection port must be designed for capillary columns. The use of splitless injection techniques is recommended. On-column injection techniques can be used, but they may severely reduce column lifetime for nonchemically bonded columns. In this protocol, a 2 μL injection volume is used consistently to maximize auto sampler reproducibility. With some gas chromatograph injection ports, however, 1 μL injections may produce some improvement in precision and chromatographic separation. A 1 μL injection volume may be used if adequate sensitivity and precision can be achieved.

[Note: If 1 μL is used as the injection volume, the injection volumes for all extracts, blanks, calibration solutions and performance check samples must be 1 μL .]

All GC carrier gas lines must be constructed from stainless steel or copper tubing. Poly-tetrafluoroethylene (PTFE) thread sealants or flow controllers should only be used.

8.3.2 Gas chromatograph-mass spectrometer interface. The GC is usually coupled directly to the MS source. The interface may include a diverter valve for shunting the column effluent and isolating the mass spectrometer source. All components of the interface should be glass or glass-lined stainless steel. Glass can be deactivated by silanizing with dichlorodimethylsilane. The interface components should be compatible with 320 $^\circ\text{C}$ temperatures. Cold spots and/or active surfaces (adsorption sites) in the GC/MS interface can cause peak tailing and peak broadening. It is recommended that the GC column be fitted directly into the MS source. Graphite ferrules should be avoided in the gas chromatograph injection area since they may adsorb PAHs. Vespel[®] or equivalent ferrules are recommended.

8.3.3 Mass spectrometer. The MS should be operated in the full range data acquisition (SCAN) mode with a total cycle time (including voltage reset time) of one second or less (see Section 13.3.2). Operation of the MS in the SCAN mode allows monitoring of all ions, thus assisting with the identification of other PAHs beyond Compendium Method TO-13A target analyte list. In addition, operating in the SCAN mode assists the analyst with identification of possible interferences from non-target analytes due to accessibility of the complete mass spectrum in the investigative process. The MS must be capable of scanning from 35 to 500 amu every 1 sec or less, using 70 volts (nominal) electron energy in the electron impact (EI) ionization mode. The mass spectrometer must be capable of producing a mass spectrum for a 50 ng injection of decafluorotriphenyl phosphine (DFTPP) which meets all of the response criteria (see Section 13.3.3). To ensure sufficient precision of mass spectral data, the MS scan rate must allow acquisition of at least five scans while a sample compound elutes from the GC. The

GC/MS system must be in a room with atmosphere demonstrated to be free of all potential contaminants which will interfere with the analysis. The instrument must be vented outside the facility or to a trapping system which prevents the release of contaminants into the instrument room.

8.3.4 Data system. A dedicated computer data system is employed to control the rapid multiple ion monitoring process and to acquire the data. Quantification data (peak areas or peak heights) and multi-ion detector (MID) traces (displays of intensities of each m/z being monitored as a function of time) must be acquired during the analyses. Quantifications may be reported based upon computer generated peak areas or upon measured peak heights (chart recording). The detector zero setting must allow peak-to-peak measurement of the noise on the baseline. The computer should have software that allows searching the GC/MS data file for ions of a specific mass and plotting such ion abundances versus time or scan number. This type of plot is defined as Selected Ion Current Profile (SICP). The software used must allow integrating the abundance in any SICP between specified time or scan number limits. The data system should be capable of flagging all data files that have been edited manually by laboratory personnel.

8.3.5 Gas chromatograph column. A fused silica DB-5 column (30 m x 0.32 mm I.D.) crosslinked 5 percent phenyl methylsilicone, 1.0 μm film thickness is utilized to separate individual PAHs. Other columns may be used for determination of PAHs. Minimum acceptance criteria must be determined as per Section 13.3. At the beginning of each 12-hour period (after mass resolution has been demonstrated) during which sample extracts or concentration calibration solutions will be analyzed, column operating conditions must be attained for the required separation on the column to be used for samples.

8.3.6 Balance. Mettler balance or equivalent.

8.3.7 All required syringes, gases, and other pertinent supplies. To operate the GC/MS system.

8.3.8 Pipettes, micropipettes, syringes, burets, etc. Used to make calibration and spiking solutions, dilute samples if necessary, etc., including syringes for accurately measuring volumes such as 25 μL and 100 μL .

9. Equipment and Materials

9.1 Materials for Sample Collection (see Figure 3)

9.1.1 Quartz fiber filter. 102 millimeter binderless quartz microfiber filter, Whatman Inc., 6 Just Road, Fairfield, NJ 07004, Filter Type QMA-4.

9.1.2 Polyurethane foam (PUF) plugs (see Figure 5a). 3-inch thick sheet stock polyurethane type (density .022 g/cm^3). The PUF should be of the polyether type used for furniture upholstery, pillows, and mattresses. The PUF cylinders (plugs) should be slightly larger in diameter than the internal diameter of the cartridge. Sources of equipment are Tisch Environmental, Village of Cleves, OH; University Research Glassware, 116 S. Merritt Mill Road, Chapel Hill, NC; Thermo Environmental Instruments, Inc., 8 West Forge Parkway, Franklin, MA; Supelco, Supelco Park, Bellefonte, PA; and SKC Inc., 334 Valley View Road, Eighty Four, PA.

9.1.3 XAD-2® resin (optional). Supelco, Supelco Park, Bellefonte, PA.

9.1.4 Teflon® end caps (see Figure 5a). For sample cartridge; sources of equipment are Tisch Environmental, Village of Cleves, OH; and University Research Glassware, 116 S. Merritt Mill Road, Chapel Hill, NC.

9.1.5 Sample cartridge aluminum shipping containers (see Figure 5b). For sample cartridge shipping; sources of equipment are Tisch Environmental, Village of Cleves, OH; and University Research Glassware, 116 S. Merritt Mill Road, Chapel Hill, NC.

9.1.6 Glass sample cartridge (see Figure 5a). For sample collection; sources of equipment are Tisch Environmental, Village of Cleves, OH; Thermo Environmental Instruments, Inc., 8 West Forge Parkway, Franklin, MA; and University Research Glassware, 116 S. Merritt Mill Road, Chapel Hill, NC.

9.1.7 Aluminum foil. Best source.

9.1.8 Hexane, reagent grade. Best source.

9.2 Sample Clean-up and Concentration

9.2.1 Methylene chloride (extraction solvent for XAD-2®; optional). Chromatographic grade, glass-distilled, best source.

9.2.2 Sodium sulfate-anhydrous (ACS). Granular (purified by washing with methylene chloride followed by heating at 400°C for 4 hours in a shallow tray).

9.2.3 Boiling chips. Solvent extracted or heated in a muffle furnace at 450°C for 2 hours, approximately 10/40 mesh (silicon carbide or equivalent).

9.2.4 Nitrogen. High purity grade, best source.

9.2.5 Hexane. Chromatographic grade, glass-distilled, best source (extraction solvent for PUF).

9.2.6 Glass wool. Silanized, extracted with methylene chloride and hexane, and dried.

9.2.7 Diethyl ether. High purity, glass distilled (extraction solvent for PUF).

9.2.8 Pentane. High purity, glass distilled.

9.2.9 Silica gel. High purity, type 60, 70-230 mesh.

9.3 GC/MS Sample Analysis

9.3.1 Gas cylinder of helium. Ultra high purity, best source.

9.3.2 Chromatographic-grade stainless steel tubing and stainless steel fitting. For interconnections, Alltech Applied Science, 2051 Waukegan Road, Deerfield, IL 60015, 312-948-8600, or equivalent.

[Note: All such materials in contact with the sample, analyte, or support gases prior to analysis should be stainless steel or other inert metal. Do not use plastic or Teflon® tubing or fittings.]

9.3.3 Native and isotopically labeled PAH isomers for calibration and spiking standards. Cambridge Isotopes, 20 Commerce Way, Woburn, MA 01801 (617-547-1818). Suggested isotopically labeled PAH isomers are: D₁₀-fluoranthene, D₂-benzo(a)pyrene, D₁₀-fluorene, D₁₀-pyrene, D₁₂-perylene, D₁₀-acenaphthene, D₁₂-chrysene, D₈-naphthalene and D₁₀-phenanthrene.

9.3.4 Decafluorotriphenylphosphine (DFTPP). Used for tuning GC/MS, best source.

9.3.5 Native stock pure standard PAH analytes. For developing calibration curve for GC/MS analysis, best source.

10. Preparation of PUF Sampling Cartridge

[Note: This method was developed using the PS-1 sample cartridge provider by General Metal Works, Village of Cleves, OH as a guideline. EPA has experience in use of this equipment during various field monitoring program over the last several years. Other manufacturers' equipment should work as well; however, modifications to these procedures may be necessary if another commercially available sampler is selected.]

10.1 Summary of Method

10.1.1 This part of the procedure discusses pertinent information regarding the preparation and cleaning of the filter, sorbent, and filter/sorbent cartridge assembly. The separate batches of filters and sorbents are extracted with the appropriate solvent.

10.1.2 At least one PUF cartridge assembly and one filter from each batch, or 10 percent of the batch, whichever is greater, should be tested and certified before the batch is considered for field use.

10.1.3 Prior to sampling, the cartridges are spiked with field surrogate compounds.

10.2 Preparation of Sampling Cartridge

10.2.1 Bake the Whatman QMA-4 quartz filters at 400°C for 5 hours before use.

10.2.2 Set aside the filters in a clean container for shipment to the field or prior to combining with the PUF glass cartridge assembly for certification prior to field deployment.

10.2.3 The PUF plugs are 6.0-cm diameter cylindrical plugs cut from 3-inch sheet stock and should fit, with slight compression, in the glass cartridge, supported by the wire screen (see Figure 5a). During cutting, rotate the die at high speed (e.g., in a drill press) and continuously lubricate with deionized or distilled water. Pre-cleaned PUF plugs can be obtained from commercial sources (see Section 9.1.2).

10.2.4 For initial cleanup, place the PUF plugs in a Soxhlet apparatus and extract with acetone for 16 hours at approximately 4 cycles per hour. When cartridges are reused, use diethyl ether/hexane (5 to 10 percent volume/volume [v/v]) as the cleanup solvent.

[Note: A modified PUF cleanup procedure can be used to remove unknown interference components of the PUF blank. This method consists of rinsing 50 times with toluene, acetone, and diethyl ether/hexane (5 to 10 percent v/v), followed by Soxhlet extraction. The extracted PUF is placed in a vacuum oven connected to a water aspirator and dried at room temperature for approximately 2 to 4 hours (until no solvent odor is detected). The extract from the Soxhlet extraction procedure from each batch may be analyzed to determine initial cleanliness prior to certification.]

10.2.5 If using XAD-2® in the cartridge, initial cleanup of the resin is performed by placing approximately 50-60 grams in a Soxhlet apparatus and extracting with methylene chloride for 16 hours at approximately 4 cycles per hour. At the end of the initial Soxhlet extraction, the spent methylene chloride is discarded and replaced with a fresh reagent. The XAD-2® resin is once again extracted for 16 hours at approximately 4 cycles per hour. The XAD-2® resin is removed from the Soxhlet apparatus, placed in a vacuum oven connected to an ultra-pure nitrogen gas stream, and dried at room temperature for approximately 2-4 hours (until no solvent odor is detected).

10.2.6 Fit a nickel or stainless steel screen (mesh size 200/200) to the bottom of a hexane-rinsed glass sampling cartridge to retain the PUF or XAD-2® sorbents, as illustrated in Figure 5a. If using XAD-2® alone, then place a small diameter (~1/4") PUF plug on top of the nickel or stainless steel screen to retain the XAD-2® in the glass cartridge. Place the Soxhlet-extracted, vacuum-dried PUF (2.5-cm thick by 6.5-cm diameter) on top of the screen in the glass sampling cartridge using polyester gloves. Place ~200 g of the clean XAD-2® inside the glass sampling cartridge on top of the small diameter PUF plug.

10.2.7 Wrap the sampling cartridge with hexane-rinsed aluminum foil, cap with the Teflon® end caps (optional), place in a cleaned labeled aluminum shipping container, and seal with Teflon® tape. Analyze at least 1 cartridge from each batch of cartridges prepared using the procedure described in Section 10.3, before the batch is considered acceptable for field use.

The acceptance level of the cartridge is for each target PAH analyte to be less than or equal to the detection limit requirements to meet the project data quality objectives. It is generally not possible to eliminate the presence of naphthalene, but the amount detected on the cleaned PUF cartridge should be less than five times the concentration of the lowest calibration standard (~500 ng). This amount is insignificant compared to the amount collected from a typical air sample.

In general, the following guidelines are provided in determining whether a cartridge is clean for field use:

- Naphthalene <500 ng/cartridge
- Other PAHs <200 ng total/cartridge

10.3 Procedure for Certification of PUF Cartridge Assembly

[Note: The following procedure outlines the certification of a filter and PUF cartridge assembly. If using XAD-2® as the sorbent, the procedure remains the same, except the solvent is methylene chloride rather than 10 percent diethyl ether/hexane.]

10.3.1 Extract one filter and PUF sorbent cartridge by Soxhlet extraction and concentrate using a K-D evaporator for each lot of filters and cartridges sent to the field.

10.3.2 Assemble the Soxhlet apparatus. Charge the Soxhlet apparatus (see Figure 4a) with 700 mL of the extraction solvent (10 percent v/v diethyl ether/hexane) and reflux for 2 hours. Let the apparatus cool, disassemble it, and discard the used extraction solvent. Transfer the filter and PUF glass cartridge to the Soxhlet apparatus (the use of an extraction thimble is optional).

[Note: The filter and sorbent assembly are tested together in order to reach detection limits, to minimize cost and to prevent misinterpretation of the data. Separate analyses of the filter and PUF would not yield useful information about the physical state of most of the PAHs at the time of sampling due to evaporative losses from the filter during sampling.]

10.3.3 Add between 300 and 350 mL of diethyl ether/hexane (10 percent v/v) to the Soxhlet apparatus. Reflux the sample for 18 hours at a rate of at least 3 cycles per hour. Allow to cool, then disassemble the apparatus.

10.3.4 Assemble a K-D concentrator (see Figure 4b) by attaching a 10-mL concentrator tube to a 500-mL evaporative flask.

10.3.5 Transfer the extract by pouring it through a drying column containing about 10 cm of anhydrous granular sodium sulfate (see Figure 4c) and collect the extract in the K-D concentrator. Rinse the Erlenmeyer flask and column with 20 to 30 mL of 10 percent diethyl ether/hexane to complete the quantitative transfer.

10.3.6 Add one or two clean boiling chips and attach a 3-ball Snyder column to the evaporative flask. Pre-wet the Snyder column by adding about 1 mL of the extraction solvent to the top of the column. Place the K-D apparatus on a hot water bath (~50°C) so that the concentrator tube is partially immersed in the hot water, and the entire lower rounded surface of the flask is bathed with hot vapor. Adjust the vertical position of the apparatus and the water temperature as required to complete the concentration in 1 hour. At the proper rate of distillation, the balls of the column will actively chatter, but the chambers will not flood with condensed solvent. When the apparent volume of liquid reaches approximately 5 mL, remove the K-D apparatus from the water bath and allow it to drain and cool for at least 5 minutes. Remove the Snyder column and rinse the flask and its lower joint into the concentrator tube with 5 mL of cyclohexane. A 1-mL syringe is recommended for this operation.

10.3.7 Concentrate the extract to 5 mL and analyze using GC/MS.

10.3.8 The acceptance level of the cartridge is for each target PAH analyte to be less than or equal to the detection limit requirements to meet the project data quality objectives. It is generally not possible to eliminate the presence of naphthalene, but the amount detected on the cleaned PUF cartridge should be less than five times the concentration of the lowest calibration standard (~500 ng). This amount is insignificant compared to the amount collected from a typical air sample.

In general, the following guidelines are provided in determining whether a cartridge is clean for field use:

- Naphthalene <500 ng/cartridge
- Other PAHs <200 ng total/cartridge

Cartridges are considered clean for up to 30 days from date of certification when sealed in their containers.

10.4 Deployment of Cartridges for Field Sampling

10.4.1 Immediately prior to field deployment, add surrogate compounds (i.e., chemically inert compounds not expected to occur in an environmental sample) to the center of the PUF cartridge, using a microsyringe. Spike 20 μL of a 50 $\mu\text{g}/\text{mL}$ solution of the surrogates onto the center bed of the PUF trap to yield a final concentration of 1 μg . The surrogate compounds must be added to each cartridge assembly. The following field surrogate compounds should be added to each PUF cartridge prior to field deployment to monitor matrix effects, breakthrough, etc.

<u>Field Surrogate Compound</u>	<u>Total Spiked Amount (μg)</u>
D ₁₀ -Fluoranthene	1
D ₁₂ -Benzo(a)pyrene	1

Fill out a "chain-of-custody" indicating cartridge number, surrogate concentration, date of cartridge certification, etc. The chain-of-custody must accompany the cartridge to the field and return to the laboratory.

10.4.2 Use the recoveries of the surrogate compounds to monitor for unusual matrix effects and gross sample processing errors. Evaluate surrogate recovery for acceptance by determining whether the measured concentration falls within the acceptance limits of 60-120 percent.

10.4.3 Cartridges are placed in their shipping containers and shipped to the field. Blank cartridges do not need to be chilled when shipping to the field until after exposure to ambient air.

11. Assembly, Calibration, and Collection Using Sampling System

[Note: This method was developed using the PS-1 semi-volatile sampler provided by General Metal Works, Village of Cleves, OH as a guideline. EPA has experience in the use of this equipment during various field monitoring programs over the last several years. Other manufacturers' equipment should work as well; however, modifications to these procedures may be necessary if another commercially available sampler is selected.]

11.1 Sampling Apparatus

The entire sampling system is diagrammed in Figure 2. This apparatus was developed to operate at a rate of 4 to 10 scfm (0.114 to 0.285 std m³/min) and is used by EPA for high-volume sampling of ambient air. The method write-up presents the use of this device.

The sampling module (see Figure 3) consists of a filter and a glass sampling cartridge containing the PUF utilized to concentrate PAHs from the air. A field portable unit has been developed by EPA (see Figure 6).

11.2 Calibration of Sampling System

Each sampler should be calibrated (1) when new, (2) after major repairs or maintenance, (3) whenever any audit point deviates from the calibration curve by more than 7 percent, (4) before/after each sampling event, and (5) when a different sample collection medium, other than that which the sampler was originally calibrated to, will be used for sampling.

11.2.1 Calibration of Orifice Transfer Standard. Calibrate the modified high volume air sampler in the field using a calibrated orifice flow rate transfer standard. Certify the orifice transfer standard in the laboratory against a positive displacement rootsmeter (see Figure 7). Once certified, the recertification is performed rather infrequently if the orifice is protected from damage. Recertify the orifice transfer standard performed once per year utilizing a set of five multi-hole resistance plates.

[Note: The set of five multihole resistance plates is used to change the flow through the orifice so that several points can be obtained for the orifice calibration curve. The following procedure outlines the steps to calibrate the orifice transfer standard in the laboratory.]

11.2.1.1 Record the room temperature (T_1 in °C) and barometric pressure (P_b in mm Hg) on the Orifice Calibration Data Sheet (see Figure 8). Calculate the room temperature in K (absolute temperature) and record on Orifice Calibration Data Sheet.

$$T_1 \text{ in K} = 273^\circ + T_1 \text{ in } ^\circ\text{C}$$

11.2.1.2 Set up laboratory orifice calibration equipment as illustrated in Figure 7. Check the oil level of the rootsmeter prior to starting. There are three oil level indicators, one at the clear plastic end, and two sight glasses, one at each end of the measuring chamber.

11.2.1.3 Check for leaks by clamping both manometer lines, blocking the orifice with cellophane tape, turning on the high-volume motor, and noting any change in the rootsmeter's reading. If the rootsmeter's reading changes, there is a leak in the system. Eliminate the leak before proceeding. If the rootsmeter's reading remains constant, turn off the hi-vol motor, remove the cellophane tape, and unclamp both manometer lines.

11.2.1.4 Install the 5-hole resistance plate between the orifice and the filter adapter.

11.2.1.5 Turn manometer tubing connectors one turn counter-clockwise. Make sure all connectors are open.

11.2.1.6 Adjust both manometer midpoints by sliding their movable scales until the zero point corresponds with the meniscus. Gently shake or tap to remove any air bubbles and/or liquid remaining on tubing connectors. (If additional liquid is required for the water manometer, remove tubing connector and add clean water.)

11.2.1.7 Turn on the high-volume motor and let it run for 5 minutes to set the motor brushes. Turn the motor off. Ensure manometers are set to zero. Turn the high-volume motor on.

11.2.1.8 Record the time in minutes required to pass a known volume of air (approximately 5.6 to 8.4 m³ of air for each resistance plate) through the rootsmeter by using the rootsmeter's digital volume dial and a stopwatch.

11.2.1.9 Record both manometer readings [orifice water manometer (ΔH) and rootsmeter mercury manometer (ΔP)] on Orifice Calibration Data Sheet (see Figure 8).

[Note: ΔH is the sum of the difference from zero (0) of the two column heights.]

11.2.1.10 Turn off the high-volume motor.

11.2.1.11 Replace the 5-hole resistance plate with the 7-hole resistance plate.

11.2.1.12 Repeat Sections 11.2.1.3 through 11.2.1.11.

11.2.1.13 Repeat for each resistance plate. Note results on Orifice Calibration Data Sheet (see Figure 8). Only a minute is needed for warm-up of the motor. Be sure to tighten the orifice enough to eliminate any leaks. Also check the gaskets for cracks.

[Note: The placement of the orifice prior to the rootsmeter causes the pressure at the inlet of the rootsmeter to be reduced below atmospheric conditions, thus causing the measured volume to be incorrect. The volume measured by the rootsmeter must be corrected.]

11.2.1.14 Correct the measured volumes on the Orifice Calibration Data Sheet:

$$V_{\text{std}} = V_{\text{m}} \left(\frac{P_{\text{a}} - \Delta P}{P_{\text{std}}} \right) \left(\frac{T_{\text{std}}}{T_{\text{a}}} \right)$$

where:

V_{std} = standard volume, std m³

V_{m} = actual volume measured by the rootsmeter, m³

P_{a} = barometric pressure during calibration, mm Hg

ΔP = differential pressure at inlet to volume meter, mm Hg

P_{std} = 760 mm Hg

T_{std} = 298 K

T_{a} = ambient temperature during calibration, K.

11.2.1.15 Record standard volume on Orifice Calibration Data Sheet.

11.2.1.16 The standard flow rate as measured by the rootsmeter can now be calculated using the following formula:

$$Q_{\text{std}} = \frac{V_{\text{std}}}{\theta}$$

where:

Q_{std} = standard volumetric flow rate, std m³/min

θ = elapsed time, min

11.2.1.17 Record the standard flow rates to the nearest 0.01 std m³/min.

11.2.1.18 Calculate and record $\sqrt{\Delta H (P_1/P_{std})(298/T_1)}$ value for each standard flow rate.

11.2.1.19 Plot each $\sqrt{\Delta H (P_1/P_{std})(298/T_1)}$ value (y-axis) versus its associated standard flow rate (x-axis) on arithmetic graph paper and draw a line of best fit between the individual plotted points.

[*Note: This graph will be used in the field to determine standard flow rate.*]

11.2.2 Calibration of the High-Volume Sampling System Utilizing Calibrated Orifice Transfer Standard

For this calibration procedure, the following conditions are assumed in the field:

- The sampler is equipped with an valve to control sample flow rate.
- The sample flow rate is determined by measuring the orifice pressure differential using a Magnehelic gauge.
- The sampler is designed to operate at a standardized volumetric flow rate of 8 ft³/min (0.225 m³/min), with an acceptable flow rate range within 10 percent of this value.
- The transfer standard for the flow rate calibration is an orifice device. The flow rate through the orifice is determined by the pressure drop caused by the orifice and is measured using a "U" tube water manometer or equivalent.
- The sampler and the orifice transfer standard are calibrated to standard volumetric flow rate units (scfm or scmm).
- An orifice transfer standard with calibration traceable to NIST is used.
- A "U" tube water manometer or equivalent, with a 0- to 16-inch range and a maximum scale division of 0.1 inch, will be used to measure the pressure in the orifice transfer standard.
- A Magnehelic gauge or equivalent with a 9- to 100-inch range and a minimum scale division of 2 inches for measurements of the differential pressure across the sampler's orifice is used.
- A thermometer capable of measuring temperature over the range of 32° to 122°F (0° to 50°C) to ±2°F (±1°C) and referenced annually to a calibrated mercury thermometer is used.
- A portable aneroid barometer (or equivalent) capable of measuring ambient barometric pressure between 500 and 800 mm Hg (19.5 and 31.5 in. Hg) to the nearest mm Hg and referenced annually to a barometer of known accuracy is used.
- Miscellaneous handtools, calibration data sheets or station log book, and wide duct tape are available.

11.2.2.1 Set up the calibration system as illustrated in Figure 9. Monitor the airflow through the sampling system with a venturi/Magnehelic assembly, as illustrated in Figure 9. Audit the field sampling system once per quarter using a flow rate transfer standard, as described in the EPA *High-Volume Sampling Method, 40 CFR 50, Appendix B*. Perform a single-point calibration before and after each sample collection, using the procedures described in Section 11.2.3.

11.2.2.2 Prior to initial multi-point calibration, place an empty glass cartridge in the sampling head and activate the sampling motor. Fully open the flow control valve and adjust the voltage variator so that a sample flow rate corresponding to 110 percent of the desired flow rate (typically 0.20 to 0.28 m³/min) is indicated on the Magnehelic gauge (based on the previously obtained multipoint calibration curve). Allow the motor to warm up for 10 min and then adjust the flow control valve to achieve the desired flow rate. Turn off the sampler. Record the ambient temperature and barometric pressure on the Field Calibration Data Sheet (see Figure 10).

11.2.2.3 Place the orifice transfer standard on the sampling head and attach a manometer to the tap on the transfer standard, as illustrated in Figure 9. Properly align the retaining rings with the filter holder and secure by tightening the three screw clamps. Connect the orifice transfer standard by way of the pressure tap to a

manometer using a length of tubing. Set the zero level of the manometer or Magnehelic. Attach the Magnehelic gauge to the sampler venturi quick release connections. Adjust the zero (if needed) using the zero adjust screw on face of the gauge.

11.2.2.4 To leak test, block the orifice with a rubber stopper, wide duct tape, or other suitable means. Seal the pressure port with a rubber cap or similar device. Turn on the sampler.

Caution: Avoid running the sampler for too long a time with the orifice blocked. This precaution will reduce the chance that the motor will be overheated due to the lack of cooling air. Such overheating can shorten the life of the motor.

11.2.2.5 Gently rock the orifice transfer standard and listen for a whistling sound that would indicate a leak in the system. A leak-free system will not produce an upscale response on the sampler's magnehelic. Leaks are usually caused either by damaged or missing gaskets, by cross-threading, and/or not screwing sample cartridge together tightly. All leaks must be eliminated before proceeding with the calibration. When the sample is determined to be leak-free, turn off the sampler and unblock the orifice. Now remove the rubber stopper or plug from the calibrator orifice.

11.2.2.6 Turn the flow control valve to the fully open position and turn the sampler on. Adjust the flow control valve until a Magnehelic reading of approximately 70 in. is obtained. Allow the Magnehelic and manometer readings to stabilize and record these values on the orifice transfer Field Calibration Data Sheet (see Figure 10).

11.2.2.7 Record the manometer reading under Y1 and the Magnehelic reading under Y2 on the Field Calibration Data Sheet. For the first reading, the Magnehelic should still be at 70 inches as set above.

11.2.2.8 Set the Magnehelic to 60 inches by using the sampler's flow control valve. Record the manometer (Y1) and Magnehelic (Y2) readings on the Field Calibration Data Sheet (see Figure 10).

11.2.2.9 Repeat the above steps using Magnehelic settings of 50, 40, 30, 20, and 10 inches.

11.2.2.10 Turn the voltage variator to maximum power, open the flow control valve, and confirm that the Magnehelic reads at least 100 inches. Turn off the sampler and confirm that the Magnehelic reads zero.

11.2.2.11 Read and record the following parameters on the Field Calibration Data Sheet. Record the following on the calibration data sheet:

- Data, job number, and operator's signature.
- Sampler serial number.
- Ambient barometric pressure.
- Ambient temperature.

11.2.2.12 Remove the "dummy" cartridge and replace with a sample cartridge.

11.2.2.13 Obtain the manufacturer high volume orifice calibration certificate.

11.2.2.14 If not performed by the manufacturer, calculate values for each calibrator orifice static pressure (Column 6, inches of water) on the manufacturer's calibration certificate using the following equation:

$$\sqrt{\Delta H(P_a/760)[298/(T_a + 273)]}$$

where:

P_a = the barometric pressure (mm Hg) at time of manufacturer calibration, mm Hg

T_a = temperature at time of calibration, °C

11.2.2.15 Perform a linear regression analysis using the values in Column 7 of the manufacturer's High Volume Orifice Calibration Certificate for flow rate (Q_{std}) as the "X" values and the calculated values as the Y

values. From this relationship, determine the correlation (CC1), intercept (B1), and slope (M1) for the Orifice Transfer Standard.

11.2.2.16 Record these values on the Field Calibration Data Sheet (see Figure 10).

11.2.2.17 Using the Field Calibration Data Sheet values (see Figure 10), calculate the Orifice Manometer Calculated Values (Y3) for each orifice manometer reading using the following equation:

Y3 Calculation

$$Y3 = \{Y1(P_a/760)[298/(T_a + 273)]\}^{1/2}$$

11.2.2.18 Record the values obtained in Column Y3 on the Field Calibration Data Sheet (see Figure 10).

11.2.2.19 Calculate the Sampler Magnehelic Calculated Value (Y4) using the following equation:

Y4 Calculation

$$Y4 = \{Y2(P_a/760)[298/(T_a + 273)]\}^{1/2}$$

11.2.2.20 Record the value obtained in Column Y4 on the Field Calibration Data Sheet (see Figure 10).

11.2.2.21 Calculate the Orifice Flow Rate (X1) in scm using the following equation:

X1 Calculation

$$X1 = \frac{Y3 - B1}{M1}$$

11.2.2.22 Record the values obtained in Column X1 on the Field Calibration Data Sheet (see Figure 10).

11.2.2.23 Perform a linear regression of the values in Column X1 (as X) and the values in Column Y4 (as Y). Record the relationship for correlation (CC2), intercept (B2), and slope (M2) on the Field Calibration Data Sheet. The correlation coefficient must be 0.990 or greater.

11.2.2.24 Using the following equation, calculate a set point (SP) for the manometer to represent a desired flow rate:

Set Point

$$\text{Set point (SP)} = [(\text{Expected } P_a) / (\text{Expected } T_a) (T_{\text{std}} / P_{\text{std}})] [M2 (\text{Desired flow rate}) + B2]^2$$

where:

- P_a = Expected atmospheric pressure (P_a), mm Hg
- T_a = Expected atmospheric temperature (T_a), $273 + ^\circ\text{C}$
- M2 = Slope of developed relationship
- B2 = Intercept of developed relationship
- T_{std} = Temperature standard, $273 + 25^\circ\text{C}$
- P_{std} = Pressure standard, 760 mm Hg

11.2.2.25 During monitoring, calculate a flow rate from the observed Magnehelic reading using the following equations:

Flow Rate

$$Y5 = [\text{Average Magnehelic Reading } (\Delta H) (P_a/T_a)(T_{std}/P_{std})]^{1/2}$$

$$X2 = \frac{Y5 - B2}{M2}$$

where:

Y5 = Corrected average magnehelic reading
X2 = Instant calculated flow rate, scfm

11.2.2.26 The relationship in calibration of a sampling system between Orifice Transfer Standard and flow rate through the sampler is illustrated in Figure 11.

11.2.3 Single-Point Audit of the High Volume Sampling System Utilizing Calibrated Orifice Transfer Standard

Single point calibration checks are required as follows:

- Prior to the start of each 24-hour test period.
- After each 24-hour test period. The post-test calibration check may serve as the pre-test calibration check for the next sampling period if the sampler is not moved.
- Prior to sampling after a sample is moved.

For samplers, perform a calibration check for the operational flow rate before each 24-hour sampling event and when required as outlined in the user quality assurance program. The purpose of this check is to track the sampler's calibration stability. Maintain a control chart presenting the percentage difference between a sampler's indicated and measured flow rates. This chart provides a quick reference of sampler flow-rate drift problems and is useful for tracking the performance of the sampler. Either the sampler log book or a data sheet will be used to document flow-check information. This information includes, but is not limited to, sampler and orifice transfer standard serial number, ambient temperature, pressure conditions, and collected flow-check data.

In this subsection, the following is assumed:

- The flow rate through a sampler is indicated by the orifice differential pressure;
- Samplers are designed to operate at an actual flow rate of 8 scfm, with a maximum acceptable flow-rate fluctuation range of ± 10 percent of this value;
- The transfer standard will be an orifice device equipped with a pressure tap. The pressure is measured using a manometer; and
- The orifice transfer standard's calibration relationship is in terms of standard volumetric flow rate (Q_{std}).

11.2.3.1 Perform a single point flow audit check before and after each sampling period utilizing the Calibrated Orifice Transfer Standard (see Section 11.2.1).

11.2.3.2 Prior to single point audit, place a "dummy" glass cartridge in the sampling head and activate the sampling motor. Fully open the flow control valve and adjust the voltage variator so that a sample flow rate corresponding to 110 percent of the desired flow rate (typically 0.19 to 0.28 m³/min) is indicated on the Magnehelic gauge (based on the previously obtained multipoint calibration curve). Allow the motor to warm up for 10 minutes and then adjust the flow control valve to achieve the desired flow rate. Turn off the sampler. Record the ambient temperature and barometric pressure on the Field Test Data Sheet (see Figure 12).

11.2.3.3 Place the flow rate transfer standard on the sampling head.

11.2.3.4 Properly align the retaining rings with the filter holder and secure by tightening the three screw clamps. Connect the flow rate transfer standard to the manometer using a length of tubing.

11.2.3.5 Using tubing, attach one manometer connector to the pressure tap of the transfer standard. Leave the other connector open to the atmosphere.

11.2.3.6 Adjust the manometer midpoint by sliding the movable scale until the zero point corresponds with the water meniscus. Gently shake or tap to remove any air bubbles and/or liquid remaining on tubing connectors. (If additional liquid is required, remove tubing connector and add clean water.)

11.2.3.7 Turn on the high-volume motor and let run for 5 minutes.

11.2.3.8 Record the pressure differential indicated, ΔH , in inches of water, on the Field Test Data Sheet. Be sure a stable ΔH has been established.

11.2.3.9 Record the observed Magnehelic gauge reading in inches of water on the Field Test Data Sheet. Be sure stable ΔM has been established.

11.2.3.10 Using previous established Orifice Transfer Standard curve, calculate Q_{xs} (see Section 11.2.2.23).

11.2.3.11 This flow should be within ± 10 percent of the sampler set point, normally, 0.224 m³. If not, perform a new multipoint calibration of the sampler.

11.2.3.12 Remove flow rate transfer standard and dummy sorbent cartridge.

11.3 Sample Collection

11.3.1 General Requirements

11.3.1.1 The sampler should be located in an unobstructed area, at least 2 meters from any obstacle to air flow. The exhaust hose should be stretched out in the downwind direction to prevent recycling of air into the sample head.

11.3.1.2 All cleaning and sample module loading and unloading should be conducted in a controlled environment, to minimize any chance of potential contamination.

11.3.1.3 When new or when using the sampler at a different location, all sample contact areas need to be cleaned. Use triple rinses of reagent grade hexane or methylene chloride contained in Teflon® rinse bottles. Allow the solvents to evaporate before loading the PUF modules.

11.3.2 Preparing Cartridge for Sampling

11.3.2.1 Detach the lower chamber of the cleaned sample head. While wearing disposable, clean, lint-free nylon, or cotton gloves, remove a clean glass sorbent module from its shipping container. Remove the Teflon® end caps (if applicable). Replace the end caps in the sample container to be reused after the sample has been collected.

11.3.2.2 Insert the glass module into the lower chamber and tightly reattach the lower chambers to the module.

11.3.2.3 Using clean rinsed (with hexane) Teflon®-tipped forceps, carefully place a clean conditioned fiber filter atop the filter holder and secure in place by clamping the filter holder ring over the filter. Place the

aluminum protective cover on top of the cartridge head. Tighten the 3 screw clamps. Ensure that all module connections are tightly assembled. Place a small piece of aluminum foil on the ball-joint of the sample cartridge to protect from back-diffusion of semi-volatiles into the cartridge during transporting to the site.

[Note: Failure to do so could expose the cartridge to contamination during transport.]

11.3.2.4 Place the cartridge in a carrying bag to take to the sampler.

11.3.3 Collection

11.3.3.1 After the sampling system has been assembled, perform a single point flow check as described in Sections 11.2.3.

11.3.3.2 With the empty sample module removed from the sampler, rinse all sample contact areas using reagent grade hexane in a Teflon® squeeze bottle. Allow the hexane to evaporate from the module before loading the samples.

11.3.3.3 With the sample cartridge removed from the sampler and the flow control valve fully open, turn the pump on and allow it to warm-up for approximately 5 minutes.

11.3.3.4 Attach a "dummy" sampling cartridge loaded with the exact same type of filter and PUF media to be used for sample collection.

11.3.3.5 Turn the sampler on and adjust the flow control valve to the desired flow as indicated by the Magnehelic gauge reading determined in Section 11.2.2.4. Once the flow is properly adjusted, take extreme care not to inadvertently alter its setting.

11.3.3.6 Turn the sampler off and remove the "dummy" module. The sampler is now ready for field use.

11.3.3.7 Check the zero reading of the sampler Magnehelic. Record the ambient temperature, barometric pressure, elapsed time meter setting, sampler serial number, filter number, and PUF cartridge number on the Field Test Data Sheet (see Figure 12). Attach the loaded sampler cartridge assembly to the sampler.

11.3.3.8 Place the voltage variator and flow control valve at the settings used in Section 11.3.2, and the power switch. Activate the elapsed time meter and record the start time. Adjust the flow (Magnehelic setting), if necessary, using the flow control valve.

11.3.3.9 Record the Magnehelic reading every 6 hours during the sampling period. Use the calibration factors (see Section 11.2.2.4) to calculate the desired flow rate. Record the ambient temperature, barometric pressure, and Magnehelic reading at the beginning and during sampling period.

11.3.4 Sample Recovery

11.3.4.1 At the end of the desired sampling period, turn the power off. Carefully remove the sampling head containing the filter and sorbent cartridge. Place the protective "plate" over the filter to protect the cartridge during transport to a clean recovery area. Also, place a piece of aluminum foil around the bottom of the sampler cartridge assembly.

11.3.4.2 Perform a final calculated sampler flow check using the calibration orifice, assembly, as described in Section 11.3.2. If calibration deviates by more than 10 percent from initial reading, mark the flow data for that sample as suspect and inspect and/or remove from service, record results on Field Test Data Sheet, Figure 12.

11.3.4.3 Transport the sampler cartridge assembly to a clean recovery area.

11.3.4.4 While wearing white cotton gloves, remove the PUF glass cartridge from the lower module chamber and lay it on the retained aluminum foil in which the sample was originally wrapped.

11.3.4.5 Carefully remove the quartz fiber filter from the upper chamber using clean Teflon®-tipped forceps.

11.3.4.6 Fold the filter in half twice (sample side inward) and place it in the glass cartridge atop the PUF.

11.3.4.7 Wrap the combined samples in the original hexane-rinsed aluminum foil, attach Teflon® end caps (if applicable) and place them in their *original* aluminum shipping container. Complete a sample label and affix it to the aluminum shipping container.

11.3.4.8 Chain-of-custody should be maintained for all samples. Store the containers under blue ice or dry ice and protect from UV light to prevent possibly photo-decomposition of collected analytes. If the time span between sample collection and laboratory analysis is to exceed 24 hours, refrigerate sample at 4°C.

11.3.4.9 Return at least one field blank filter/PUF cartridge to the laboratory with each group of samples. Treat a field blank exactly as the sample except that air is not drawn through the filter/sorbent cartridge assembly.

11.3.4.10 Ship and store field samples chilled (<4°C) using blue ice until receipt at the analytical laboratory, after which samples should be refrigerated at less than or equal to 4°C for up to 7 days prior to extraction; extracts should be analyzed within 40 days of extraction.

12. Sample Extraction, Concentration, and Cleanup

[Note: The following sample extraction, concentration, solvent exchange and analysis procedures are outlined for user convenience in Figure 13.]

12.1 Sample Identification

12.1.1 The chilled (<4°C) samples are returned in the aluminum shipping container (containing the filter and sorbents) to the laboratory for analysis. The "chain-of-custody" should be completed.

12.1.2 The samples are logged in the laboratory logbook according to sample location, filter and sorbent cartridge number identification, and total air volume sampled (uncorrected).

12.1.3 If the time span between sample registration and analysis is greater than 24-hours, then the sample must be kept refrigerated at <4°C. Minimize exposure of samples to fluorescent light. All samples should be extracted within one week (7 days) after sampling.

12.2 Soxhlet Extraction and Concentration

[Note: If PUF is the sorbent, the extraction solvent is 10 percent diethyl ether in hexane. If XAD-2® resin is the sorbent, the extraction solvent is methylene chloride.]

12.2.1 Assemble the Soxhlet apparatus (see Figure 4a). Immediately before use, charge the Soxhlet apparatus with 700 to 750 mL of 10 percent diethyl ether in hexane and reflux for 2 hours. Let the apparatus cool, disassemble it, transfer the diethyl ether in hexane to a clean glass container, and retain it as a blank for later analysis, if required. Place the sorbent and filter together in the Soxhlet apparatus (the use of an extraction thimble is optional).

[Note: The filter and sorbent are analyzed together in order to reach detection limits, avoid questionable interpretation of the data, and minimize cost.]

12.2.1.1 Prior to extraction, add appropriate laboratory surrogate standards to the Soxhlet solvent. A surrogate standard (i.e., a chemically compound not expected to occur in an environmental sample) should be added to each sample, blank, and matrix spike sample just prior to extraction or processing. The recovery of the laboratory surrogate standard is used to monitor for unusual matrix effects, gross sample processing errors, etc. Surrogate recovery is evaluated for acceptance by determining whether the measure concentration falls within the acceptance limits. Spike 20 µL of a 50 µg/mL solution of the surrogates onto the PUF cartridge, prior to Soxhlet extraction, to yield a final concentration of 1 µg. The following laboratory surrogate standards have been

successfully utilized in determining Soxhlet extraction effects, sample process errors, etc., for GC/MS/DS analysis.

Laboratory Surrogate Standard	Total Spiked Amount (μg)
D ₁₀ -Fluorene	1
D ₁₀ -Pyrene	1

Section 13.2 outlines preparation of the laboratory surrogates. Add the laboratory surrogate compounds to the PUF cartridge. Add 700 mL of 10 percent diethyl ether in hexane to the apparatus and reflux for 18 hours at a rate of at least 3 cycles per hour. Allow to cool, then disassemble the apparatus.

12.2.1.2 Dry the extract from the Soxhlet extraction by passing it through a drying column containing about 10 grams of anhydrous sodium sulfate. Collect the dried extract in a K-D concentrator assembly. Wash the extractor flask and sodium sulfate column with 100-125 mL of 10 percent diethyl ether/hexane to complete the quantitative transfer.

12.2.2 Assemble a K-D concentrator (see Figure 4b) by attaching a 10 mL concentrator tube to a 500 mL evaporative flask.

[Note: Other concentration devices (vortex evaporator) or techniques may be used in place of the K-D as long as qualitative and quantitative recovery can be demonstrated.]

12.2.2.1 Add two boiling chips, attach a three-ball macro-Snyder column to the K-D flask, and concentrate the extract using a water bath at 60 to 65 °C. Place the K-D apparatus in the water bath so that the concentrator tube is about half immersed in the water and the entire rounded surface of the flask is bathed with water vapor. Adjust the vertical position of the apparatus and the water temperature as required to complete the concentration in one hour. At the proper rate of distillation, the balls of the column actively chatter but the chambers do not flood. When the liquid has reached an approximate volume of 5 mL, remove the K-D apparatus from the water bath and allow the solvent to drain for at least 5 minutes while cooling.

12.2.2.2 Remove the Snyder column and rinse the flask and its lower joint into the concentrator tube with 5 mL of cyclohexane. A 5 mL syringe is recommended for this operation. The extract is now ready for further concentration to 1.0 mL by nitrogen blowdown.

12.2.2.3 Place the 1 mL calibrated K-D concentrator tube with an open micro-Snyder attachment in a warm water bath (30 to 35 °C) and evaporate the solvent volume to just below 1 mL by blowing a gentle stream of clean, dry nitrogen (filtered through a column of activated carbon) above the extract.

12.2.2.4 The internal wall of the concentrator tube must be rinsed down several times with hexane during the operation.

12.2.2.5 During evaporation, the tube solvent level must be kept below the water level of the bath. The extract must never be allowed to become dry.

12.2.2.6 Bring the final volume back to 1.0 mL with hexane. Transfer the extract to a Teflon®-sealed screw-cap amber vial, label the vial, and store at 4 °C (± 2 °C).

[Note: It is not necessary to bring the volume to exactly 1.0 mL if the extract will be cleaned up by solid phase extraction cleanup methods. Final volume is brought to 1.0 mL after cleanup.]

12.3 Sample Cleanup

12.3.1 If the extract is cloudy, impurities may be removed from the extract by solid phase extraction using activated silica gel. Clean-up procedures may not be needed for relatively clean matrix samples.

12.3.2 Approximately 10 grams of silica gel, type 60 (70-230 mesh), are extracted in a Soxhlet extractor with 10 percent diethyl ether for 6 hours (minimum rate, 3 cycles/hr) and then activated by heating in a foil-covered glass container for 16 hours at 150°C.

12.3.3 Using a disposable Pasteur pipette (7.5-mm x 14.6-cm), place a small piece of glass wool in the neck of the pipette. Prepare a slurry of activated silica gel in 10 percent diethyl ether. Place 10 grams of the activated silica gel slurry into the column using additional 10 percent diethyl ether. Finally, 1 gram of anhydrous sodium sulfate is added to the top of the silica gel. Prior to use, the column is rinsed with 10 percent diethyl ether at 1 mL/min for 1 hour to remove any trace of contaminants. It is then pre-eluted with 40 mL of pentane and the eluate discarded.

12.3.4 While the pentane pre-elutant covers the top of the column, 1 mL of the sample extract is transferred to the column, and washed on with 2 mL of *n*-hexane to complete the transfer. Allow to elute through the column. Immediately prior to exposure of the sodium sulfate layer the air, add 25 mL of pentane and continue the elution process. The pentane eluate is discarded.

12.3.5 The column is finally eluted at 2 mL/min with 25 mL of 10 percent diethyl ether in pentane (4:6 v/v) and collected in a 50 mL K-D flask equipped with a 5 mL concentrator tube for concentration to less than 5 mL. The concentrate is further concentrated to 1.0 mL under a gentle stream of nitrogen as previously described.

12.3.6 The extract is now ready for GC/MS analysis. Spike the extract with internal standards (ISs) before analysis. The following internal standards (ISs) have been successfully used in PAH analysis by GC/MS.

Internal Standard (IS)	Total Spiked Amount (µg)
D ₈ -Naphthalene	0.5
D ₁₀ -Acenaphthene	0.5
D ₁₀ -Phenanthrene	0.5
D ₁₂ -Chrysene	0.5
D ₁₂ -Perylene	0.5

Section 13.2 outlines preparation of the ISs.

13. Gas Chromatography with Mass Spectrometry Detection

13.1 General

13.1.1 The analysis of the extracted sample for benzo[a]pyrene and other PAHs is accomplished by an electron ionization gas chromatograph/mass spectrometer (EI GC/MS) in the mode with a total cycle time (including voltage reset time) of 1 second or less. The GC is equipped with an DB-5 fused silica capillary column (30-m x 0.32-mm I.D.) with the helium carrier gas for analyte separation. The GC column is temperature controlled and interfaced directly to the MS ion source.

13.1.2 The laboratory must document that the EI GC/MS system is properly maintained through periodic calibration checks. The GC/MS system should be operated in accordance with specifications outlined in Table 2.

13.1.3 The GC/MS is tuned using a 50 ng/µL solution of decafluorotriphenylphosphine (DFTPP). The DFTPP permits the user to tune the mass spectrometer on a daily basis. If properly tuned, the DFTPP key ions and ion abundance criteria should be met as outlined in Table 3.

13.1.4 The GC/MS operating conditions are outlined in Table 2. The GC/MS system should be calibrated using the internal standard technique. Figure 14 outlines the following sequence involving the GC/MS calibration.

13.2 Calibration of GC/MS/DS

13.2.1 Standard Preparation

Stock PAH Standards Including Surrogate Compounds

13.2.1.1 Prepare stock standards of B[a]P and other PAHs. The stock standard solution of B[a]P (2.0 µg/µL) and other PAHs can be user prepared from pure standard materials or can be purchased commercially.

13.2.1.2 Place 0.2000 grams of native B[a]P and other PAHs on a tared aluminum weighing disk and weigh on a Mettler balance.

13.2.1.3 Quantitatively transfer the material to a 100 mL volumetric flask. Rinse the weighing disk with several small portions of 10 percent diethyl ether/hexane. Ensure all material has been transferred.

13.2.1.4 Dilute to mark with 10 percent diethyl ether/hexane.

13.2.1.5 The concentration of the stock standard solution of B[a]P or other PAHs in the flask is 2.0 µg/µL.

[Note: Commercially prepared stock PAH standards may be used at any concentration if they are certified by the manufacturer or by an independent source.]

13.2.1.6 Transfer the stock standard solutions into Teflon®-sealed screw-cap bottles. Store at 4°C and protect from light. Stock standard solutions should be checked frequently for signs of degradation or evaporation, especially just prior to preparing calibration standards from them.

13.2.1.7 Stock PAH standard solutions must be replaced after 1 year or sooner if comparison with quality control check samples indicates a problem.

Mix Internal Standard (IS) Solution

13.2.1.8 For PAH analysis, deuterated internal standards are selected that are similar in analytical behavior to the compound of interest. The following internal standards are suggested for PAH analysis:

D₁₂-Perylene

Benzo(e)pyrene
Benzo(a)pyrene
Benzo(k)fluoranthene

D₁₀-Acenaphthene

Acenaphthene (if using XAD-2® as the sorbent)
Acenaphthylene (if using XAD-2® as the sorbent)
Fluorene
Benzo(g,h,i)perylene
Dibenz(a,h)anthracene
Indeno(1,2,3-cd)pyrene
Perylene
Benzo(b)fluoranthene
Coronene

D₁₂-Chrysene

Benz(a)anthracene
Chrysene
Pyrene

D₈-Naphthalene

Naphthalene (if using XAD-2® as the sorbent)

D₁₀-Phenanthrene

Anthracene
Fluoranthene
Phenanthrene

13.2.1.9 Purchase a mix IS solution containing specific IS needed for quantitation at a concentration of 2,000 ng/ μ L.

Mixed Stock PAH Standard Including Surrogate Compounds

13.2.1.10 Prepare a mixed stock PAH standard by taking 125 μ L of the stock PAH standard(s) and diluting to mark with hexane in a 10-mL volumetric flask. The concentration of the mixed stock PAH standard(s) is 25 ng/ μ L.

Calibration PAH Standards Including Surrogate Compounds

13.2.1.11 Calibration PAH standards can be generated from the stock PAH standard using serial dilution utilizing the following equation:

$$C_1 V_1 = C_2 V_2$$

where:

C_1 = Concentration of stock PAH standards, ng/ μ L

V_1 = Volume of stock PAH standard solution taken to make calibration PAH standards, μ L

V_2 = Final volume diluted to generate calibration PAH standards, μ L

C_2 = Final concentration of calibration PAH standards, ng/ μ L

13.2.1.12 Using the above equation, prepare a series of calibration PAH standards which include the surrogate compounds (i.e., 2.50 ng/ μ L, 1.25 ng/ μ L, 0.50 ng/ μ L, 0.25 ng/ μ L, and 0.10 ng/ μ L) according to the scheme illustrated in Table 4 and described below.

- For CAL 5, transfer 1.00 mL of the mixed PAH stock standard in a 10-mL volumetric flask and dilute to 10.0 mL with hexane. The resulting concentration is 2.5 ng/ μ L for the PAH analytes.
- To prepare CAL 4, transfer 500 μ L of the mixed PAH stock standard solution to a 10-mL volumetric flask and dilute to 10.0 mL with hexane. The resulting concentration is 1.25 ng/ μ L for PAH analytes.
- To prepare CAL 3, transfer 200 μ L of the mixed PAH stock solution to a 10-mL volumetric flask and dilute to 10-mL with hexane. The resulting concentration is 0.50 ng/ μ L for PAH analytes.
- To prepare CAL 2, transfer 100 μ L of the mixed PAH stock solution to a 10-mL volumetric flask and dilute to 10-mL with hexane. The resulting concentration is 0.25 ng/ μ L for PAH analytes.
- To prepare CAL 1, transfer 40 μ L of the mixed PAH stock solution to a 10-mL volumetric flask and dilute to 10-mL with hexane. The resulting concentration is 0.10 ng/ μ L for PAH analytes.

13.2.2 Internal Standard Spiking

13.2.2.1 Prior to GC/MS analysis, each 1 mL aliquot of the five calibration standards is spiked with internal standard to a final concentration of 0.5 ng/ μ L. To do this, first prepare a 1:40 dilution of the 2,000 ng/ μ L mixed internal standard solution by diluting 250 μ L to a volume of 10 mL to yield a concentration of 50 ng/ μ L.

13.2.2.2 Each 1.0-mL portion of calibration standard and sample extract is then spiked with 10 μ L of the internal standard solution prior to analysis by GC/MS/DS operated in the SCAN mode.

13.2.3 Storage, Handling, and Retention of Standards

13.2.3.1 Store the stock and mixed standard solutions at 4 °C (± 2 °C) in Teflon®-lined screw-cap amber bottles. Store the working standard solutions at 4 °C (± 2 °C) in Teflon®-lined screw-cap amber bottles.

13.2.3.2 Protect all standards from light. Samples, sample extracts, and standards must be stored separately.

13.2.3.3 Stock standard solutions must be replaced every 12 months, or sooner, if comparison with quality control check samples indicates a problem. Diluted working standards are usable for 6 months. Analysis difficulties, which warrant investigation, may require preparation of new standards. All standards are securely stored at $\sim 4^{\circ}\text{C}$ ($\pm 2^{\circ}\text{C}$) but above freezing. The concentration, preparation and expiration date, and solvent are identified on standard vial labels. Each standard is uniquely identified with its laboratory notebook number and a prefix. This procedure helps provide traceability to standard preparation.

13.2.3.4 Take care to maintain the integrity of each standard. The solvent, hexane, is volatile and can easily evaporate. Make sure each vial is sealed after use, and mark the solvent level on the side of the vial. When retrieving a vial for use, if the solvent level does not match the mark, dispose of the standard and obtain a new one.

13.3 GC/MS Instrument Operating Conditions

13.3.1 Gas Chromatograph (GC). The following are the recommended GC analytical conditions, as also outlined in Table 3, to optimize conditions for compound separation and sensitivity.

Carrier Gas:	Helium
Linear Velocity:	28-29 cm ³ /sec
Injector Temperature:	250-300°C
Injector:	Grob-type, splitless, 2 μL
Temperature Program:	Initial Temperature: 70°C
Initial Hold Time:	4.0 \pm 0.1 min.
Ramp Rate:	10°C/min to 300°C, hold for 10 min
Final Temperature:	300°C
Final Hold Time:	10 min (or until all compounds of interest have eluted).
Analytical Time:	Approximately 50 min.

13.3.2 Mass Spectrometer. Following are the required mass spectrometer conditions for scan data acquisition:

Transfer Line Temperature:	290°C
Source Temperature:	According to manufacturer's specifications
Electron Energy:	70 volts (nominal)
Ionization Mode:	EI
Mass Range:	35 to 500 amu, SCAN data acquisition
Scan Time:	At least 5 scans per peak, not to exceed 1 second per scan

13.3.3 Instrument Performance Check for GC/MS.

13.3.3.1 Summary. It is necessary to establish that the GC/MS meet tuning and standard mass spectral abundance criteria prior to initiating any on-going data collection, as illustrated in Figure 14. This is accomplished through the analysis of decafluorotriphenylphosphine (DFTPP).

13.3.3.2 Frequency. The instrument performance check solution of DFTPP will be analyzed initially and once per 12-hour time period of operation. Also, whenever the laboratory takes corrective action which may change or affect the mass spectral criteria (e.g., ion source cleaning or repair, column replacement, etc.), the instrument performance check must be verified irrespective of the 12-hour laboratory requirement. The 12-hour

time period for GC/MS analysis begins at the injection of the DFTPP, which the laboratory submits as documentation of a compliance tune. The time period ends after 12 hours have elapsed. To meet instrument performance check requirements, samples, blanks, and standards must be injected within 12 hours of the DFTPP injection.

13.3.3.3 Procedure. Inject 50 ng of DFTPP into the GC/MS system. DFTPP may be analyzed separately or as part of the calibration standard.

13.3.3.4 Technical Acceptance Criteria. The following criteria have been established in order to generate accurate data:

- Prior to the analysis of any samples, blanks, or calibration standards, the laboratory must establish that the GC/MS system meets the mass spectral ion abundance criteria for the instrument performance check solution containing DFTPP.
- The GC/MS system must be tuned to meet the manufacturer's specifications, using a suitable calibrant. The mass calibration and resolution of the GC/MS system are verified by the analysis of the instrument performance check solution.
- The abundance criteria listed in Table 3 must be met for a 50 ng injection of DFTPP. The mass spectrum of DFTPP must be acquired by averaging three scans (the peak apex scan and the scans immediately preceding and following the apex). Background subtraction is required, and must be accomplished using a single scan prior to the elution of DFTPP.

*[Note: All ion abundance **MUST** be normalized to m/z 198, the nominal base peak, even though the ion abundances of m/z 442 may be up to 110 percent of m/z 198.]*

- The above criteria are based on adherence to the acquisition specifications identified in Table 4 and were developed for the specific target compound list associated with this document. The criteria are based on performance characteristics of instruments currently utilized in routine support of ambient air program activities. These specifications, in conjunction with relative response factor criteria for target analytes, are designed to control and monitor instrument performance associated with the requirements of this document. As they are performance-based criteria for these specific analytical requirements, they may not be optimal for additional target compounds.
- If the mass spectrometer has the ability for autotuning, then the user may utilize this function following manufacturer's specifications. Autotune automatically adjusts ion source parameters within the detector using FC-43 (Heptacos). Mass peaks at m/z 69, 219, and 502 are used for tuning. After the tuning is completed, the FC-43 abundances at m/z 50, 69, 131, 219, 414, 502, and 614 are further adjusted such that their relative intensities match the selected masses of DFTPP.

13.3.3.5 Corrective Action. If the DFTPP acceptance criteria are not met, the MS must be retuned. It may be necessary to clean the ion source, or quadrupoles, or take other actions to achieve the acceptance criteria. DFTPP acceptance criteria **MUST** be met before any standards, or required blanks, are analyzed. Any standards, field samples, or required blanks analyzed when tuning criteria have not been met will require reanalysis.

13.3.4 Initial Calibration for GC/MS.

13.3.4.1 Summary. Prior to the analysis of samples and required blanks, and after tuning criteria (instrument performance check) have been met, each GC/MS system will be initially calibrated at a minimum of five concentrations to determine instrument sensitivity and the linearity of GC/MS response for the analyte compounds and the surrogates.

13.3.4.2 Frequency. Each GC/MS system must be initially calibrated whenever the laboratory takes corrective action, which may change or affect the initial calibration criteria (e.g., ion source cleaning or repair,

column replacement, etc.), or if the continuing calibration acceptance criteria have not been met. If time still remains in the 12-hour time period after meeting the technical acceptance criteria for the initial calibration, samples may be analyzed. It is not necessary to analyze a continuing calibration standard within the 12-hour time period if the initial calibration standard (CAL 3) is the same concentration as the continuing calibration standard and both meet the continuing calibration technical acceptance criteria. Quantify all sample results using the mean of the relative response factors ($\overline{\text{RRFs}}$) from the initial calibration.

13.3.4.3 Procedure. Perform the following activities to generate quantitative data:

- Set up the GC/MS system.
- Warm all standard/spiking solutions, sample extracts, and blanks to ambient temperature (~1 hour) before analysis.
- Tune the GC/MS system to meet the technical acceptance criteria (see Section 13.3.3).
- Prepare five calibration standards containing the target compounds, internal standards, and surrogate compounds at the concentrations outlined in Table 4.
- Calibrate the GC/MS by injecting 2.0 μL of each standard. If a compound saturates when the CAL 5 standard is injected, and the system is calibrated to achieve a detection sensitivity of no less than the MDL for each compound, the laboratory must document it and attach a quantitation report and chromatogram. In this instance, the laboratory must calculate the results based on a four-point initial calibration for the *specific compound* that saturates. Secondary ion quantitation is only allowed when there are sample interferences with the primary quantitation ion. If secondary ion quantitation is used, calculate a relative response factor using the area response from the most intense secondary ion which is free of interferences and document the reasons for the use of the secondary ion.
- Record a mass spectrum of each target compound. Figure 15(a) through 15(q) documents the mass spectrum for each of the 16 target PAHs discussed in Compendium Method TO-13A. Judge the acceptability of recorded spectra by comparing them to spectra in libraries. If an acceptable spectrum of a calibration standard component is not acquired, take necessary actions to correct GC/MS performance. If performance cannot be corrected, report sample extract data for the particular compound(s), but document the affected compound(s) and the nature of the problem.

13.3.4.4 Calculations. Perform the following calculations to generate quantitative data:

[*Note: In the following calculations, the area response is that of the primary quantitation ion unless otherwise stated.*]

- **Relative Response Factors (RRFs).** Calculate RRFs for each analyte target compound and surrogate using the following equation with the appropriate internal standard. Table 5 outlines characteristic ions for the surrogate compounds and internal standards. Table 6 outlines primary quantitation ions for each PAH. Use the following equation for RRF calculation.

$$\text{RRF} = \frac{A_x C_{is}}{A_{is} C_x}$$

where:

A_x = area of the primary quantitation ion for the compound to be measured, counts

A_{is} = area of the primary quantitation ion for the internal standard, counts

C_{is} = concentration or amount of the internal standard, ng/ μL

C_x = concentration or amount of the compound to be measured, ng/ μ L

- **Percent Relative Standard Deviation (%RSD).** Using the RRFs from the initial calibration, calculate the %RSD for all target compounds and surrogates using the following equations:

$$\%RSD = \frac{SD_{RRF}}{\bar{x}} \times 100$$

and

$$SD_{RRF} = \sqrt{\sum_{i=1}^N \frac{(x_i - \bar{x})^2}{N - 1}}$$

where:

- SD_{RRF} = standard deviation of initial response factors (per compound)
- \bar{x} = mean of initial relative response factors (per compound)
- X_i = i th RRF
- N = number of determinations

- **Relative Retention Times (RRT).** Calculate the RRTs for each target compound and surrogate over the initial calibration range using the following equation:

$$RRT = \frac{RT_c}{RT_{is}}$$

where:

- RT_c = retention time of the target compound, minutes
- RT_{is} = retention time of the internal standard, minutes

- **Mean of the Relative Retention Times (\overline{RRT}).** Calculate the mean of the relative retention times (\overline{RRT}) for each analyte target compound and surrogate over the initial calibration range using the following equation:

$$\overline{RRT} = \sum_{i=1}^n \frac{RRT_i}{n}$$

where:

- \overline{RRT} = mean relative retention time for the target compound or surrogate for each initial calibration standard, minutes
- RRT = relative retention time for the target compound or surrogate for each initial calibration standard, minutes

- **Mean Area Response (\bar{Y}) for Internal Standard.** Calculate the area response (Y) mean for primary quantitation ion each internal standard compound over the initial calibration range using the following equation:

$$\bar{Y} = \sum_{i=1}^n \frac{Y_i}{n}$$

where:

\bar{Y} = mean area response, counts

Y_i = area response for the primary quantitation ion for the internal standard for each calibration standard, counts

- **Mean of the Retention Time (\bar{RT}) For Internal Standard.** Calculate the mean of the retention times (\bar{RT}) for each internal standard over the initial calibration range using the following equation:

$$\bar{RT} = \sum_{i=1}^n \frac{RT_i}{n}$$

where:

\bar{RT} = mean retention time, minutes

RT = retention time for the internal standard for each initial calibration standard, minutes

13.3.4.5 Technical Acceptance Criteria. All initial calibration standards must be analyzed at the concentration levels at the frequency described in Section 13.3.3 on a GC/MS system meeting the DFTPP instrument performance check criteria.

- The relative response factor (RRF) at each calibration concentration for each target compound and surrogate that has a required minimum response factor value must be greater than or equal to the minimum acceptable relative response factor (see Table 7) of the compound.
- The percent relative standard deviation (%RSD) over the initial calibration range for each target compound and surrogate that has a required maximum %RSD must be less than or equal to the required maximum value (see Table 7). For all the other target compounds, the value for %RSD must be less than or equal to 30 percent. When the value for %RSD exceeds 30 percent, analyze additional aliquots of appropriate CALs to obtain an acceptable %RSD of RRFs over the entire concentration range, or take action to improve GC/MS performance.
- The relative retention time for each of the target compounds and surrogates at each calibration level must be within ± 0.06 relative retention time units of the mean relative retention time for the compound.
- The retention time shift for each of the internal standards at each calibration level must be within ± 20.0 seconds compared to the mean retention time (\bar{RT}) over the initial calibration range for each internal standard.
- The compounds must meet the minimum RRF and maximum %RSD criteria for the initial calibration.

13.3.4.6 Corrective Action. If the technical acceptance criteria for initial calibration are not met, the system should be inspected for problems. It may be necessary to clean the ion source, change the column, or take other corrective actions to achieve the acceptance criteria. Initial calibration technical acceptance criteria MUST

be met before any samples or required blanks are analyzed in a 12-hour time period for an initial calibration analytical sequence.

13.3.5 Continuing Calibration.

13.3.5.1 Summary. Prior to the analysis of samples and required blanks and after tuning criteria have been met, the initial calibration of each GC/MS system must be routinely checked by analyzing a continuing calibration standard (see Table 4, CAL 3) to ensure that the instrument continues to meet the instrument sensitivity and linearity requirements of the method. The continuing calibration standard (CAL 3) shall contain the appropriate target compounds, surrogates, and internal standards.

13.3.5.2 Frequency. Each GC/MS used for analysis must be calibrated once every time period of operation. The 12-hour time period begins with injection of DFTPP. If time still remains in the 12-hour time period after meeting the technical acceptance criteria for the initial calibration, samples may be analyzed. It is not necessary to analyze a continuing calibration standard within this 12-hour time period, if the initial calibration standard that is the same concentration as the continuing calibration standard meets the continuing calibration technical acceptance criteria.

13.3.5.3 Procedure. The following activities should be performed for continuing calibration:

- Set up the GC/MS system as specified by the manufacturer.
- Tune the GC/MS system to meet the technical acceptance criteria (see Section 13.3.3).
- Analyze the CAL 3 standard solution containing all the target analytes, surrogate compounds, and internal standards using the procedure listed for the initial calibration.
- Allow all standard/spiking solutions and blanks to warm to ambient temperature (approximately 1 hour) before preparation or analysis.
- Start the analysis of the continuing calibration by injecting 2.0 μL of the CAL 3 standard solution.

13.3.5.4 Calculations. The following calculations should be performed:

- **Relative Response Factor (RRF).** Calculate a relative response factor (RRF) for each target compound and surrogate.
- **Percent Difference (%D).** Calculate the percent difference between the mean relative response factor ($\overline{\text{RRF}}$) from the most recent initial calibration and the continuing calibration RRF for each analyte target compound and surrogate using the following equation:

$$\%D_{\text{RRF}} = \frac{\text{RRF}_c - \overline{\text{RRF}}_i}{\overline{\text{RRF}}_i} \times 100$$

where:

$\%D_{\text{RRF}}$ = percent difference between relative response factors

$\overline{\text{RRF}}_i$ = average relative response factor from the most recent initial calibration

RRF_c = relative response factor from the continuing calibration standard

13.3.5.5 Technical Acceptance Criteria. The continuing calibration standard must be analyzed for the compounds listed in concentration levels at the frequency described and on a GC/MS system meeting the DFTPP instrument performance check and the initial calibration technical acceptance criteria. The relative response factor for each target analyte and surrogate that has a required minimum relative response factor value must be greater than or equal to the compound's minimum acceptable relative response factor. For an acceptable

continuing calibration, the %D between the measured RRF for each target/surrogate compound of the CAL 3 standard and the mean value calculated during initial calibration must be within ± 30 percent. If the criteria for %D are not met for the target or surrogate compounds, remedial action must be taken and recalibration may be necessary.

13.3.5.6 Corrective Action. If the continuing calibration technical acceptance criteria are not met, recalibrate the GC/MS instrument. It may be necessary to clean the ion source, change the column, or take other corrective actions to achieve the acceptance criteria. Continuing calibration technical acceptance criteria ***MUST*** be met before any samples or required blanks are analyzed in a 12-hour continuing calibration analytical sequence. Any samples or required blanks analyzed when continuing calibration criteria were not met will require reanalysis. Remedial actions, which include but are not limited to the following, must be taken if criteria are not met:

- Check and adjust GC and/or MS operating conditions.
- Clean or replace injector liner.
- Flush column with solvent according to manufacturers instructions.
- Break off a short portion (approximately 0.33 cm) of the column.
- Replace the GC column (performance of all initial calibration procedures are then required).
- Adjust MS for greater or lesser resolution.
- Calibrate MS mass scale.
- Prepare and analyze new continuing calibration.
- Prepare a new initial calibration curve.

13.3.6 Laboratory Method Blank (LMB).

13.3.6.1 Summary. The purpose of the LMB is to monitor for possible laboratory contamination. Perform all steps in the analytical procedure using all reagents, standards, surrogate compounds, equipment, apparatus, glassware, and solvents that would be used for a sample analysis. An LMB is an unused, certified filter/cartridge assembly which is carried through the same extraction procedure as a field sample. The LMB extract must contain the same amount of surrogate compounds and internal standards that is added to each sample. All field samples must be extracted and analyzed with an associated LMB.

13.3.6.2 Frequency. Analyze an LMB along with each batch of ≤ 20 samples through the entire extraction, concentration, and analysis process. The laboratory may also analyze a laboratory reagent blanks which is the same as an LMB except that no surrogate compounds or internal standards are added. This demonstrates that reagents contain no impurities producing an ion current above the level of background noise for quantitation ions for those compounds.

13.3.6.3 Procedure. Extract and analyze a clean, unused filter and glass cartridge assembly.

13.3.6.4 Technical Acceptance Criteria. Following are the technical criteria for the LMB:

- All blanks must be analyzed on a GC/MS system meeting the DFTPP instrument performance check and initial calibration or continuing calibration technical acceptance criteria.
- The percent recovery for each of the surrogates in the blank must be within the acceptance windows.
- The area response change for each of the internal standards for the blank must be within -50 percent and +100 percent compared to the internal standards in the most recent continuing calibration analysis.
- The retention time for each of the internal standards must be within ± 20.0 seconds between the blank and the most recent CAL 3 analysis.
- The LMB must not contain any target analyte at a concentration greater than the MDL and must not contain additional compounds with elution characteristics and mass spectral features that would interfere

with identification and measurement of a method analyte at its MDL. If the LMB that was extracted along with a batch of samples is contaminated, the entire batch of samples must be flagged.

13.3.6.5 Corrective Action. Perform the following if the LCBs exceed criteria:

- If the blanks do not meet the technical acceptance criteria, the analyst must consider the analytical system to be out of control. It is the analyst's responsibility to ensure that method interferences caused by contaminants in solvents, reagents, glassware, and other sample storage and processing hardware that lead to discrete artifacts and/or elevated baselines in gas chromatograms be eliminated. If contamination is a problem, the source of the contamination must be investigated and appropriate corrective measure *MUST* be taken and documented before further sample analysis proceeds.
- All samples processed with a method blank that is out of control (i.e., contaminated) will require data qualifiers to be attached to the analytical results.

13.3.7 Laboratory Control Spike (LCS).

13.3.7.1 Summary. The purpose of the LCS is to monitor the extraction efficiency of Compendium Method TO-13A target analytes from a clean, uncontaminated PUF cartridge. An LCS is an unused, certified PUF that is spiked with the target analytes (1 μg) and carried through the same extraction procedures as the field samples. The LCS must contain the same amount of surrogate compounds and internal standards that is added to each sample. All field samples must be extracted and analyzed with an associated LCS. All steps in the analytical procedure must use the same reagents, standards, surrogate compounds, equipment, apparatus, glassware, and solvents that would be used for a sample analysis.

13.3.7.2 Frequency. Analyze an LCS along with each of ≤ 20 samples through the entire extraction, concentration, and analysis. (The laboratory may also analyze a laboratory reagent blank which is the same as an LMB except that no surrogate compounds or internal standards are added. This demonstrates that reagents contain no impurities producing an ion current above the level of background noise for quantitation ions of those compounds.)

13.3.7.3 Procedure. Extract and analyze a clean, unused certified PUF cartridge assembly.

13.3.7.4 Technical Acceptance Criteria. Technical criteria for the LCS are:

- All LCSs must be analyzed on a GC/MS system meeting the DFTPP instrument performance check and initial calibration or continuing calibration technical acceptance criteria.
- The percent recovery for each of the surrogates in the LCS must be within the acceptance windows.
- The area response change for each of the internal standards for the LCS must be within -50 percent and +100 percent compared to the internal standards in the most recent continuing calibration analysis.
- The retention time for each of the internal standards must be within ± 20.0 seconds between the LCS and the most recent CAL 3 analysis.
- All target analytes spiked on the certified PUF cartridge must meet a percent recovery between 60-120 to be acceptable.

13.3.7.5 Corrective Action. Perform the following if the LCS exceed criteria:

- If the LCS do not meet the technical acceptance criteria, the analyst must consider the analytical system to be out of control. It is the analyst's responsibility to ensure that method interferences caused by contaminants in solvents, reagents, glassware, and other sample storage and processing hardware that lead to discrete artifacts and/or elevated baselines in gas chromatograms be eliminated. If contamination is a problem, the source of the contamination must be investigated and appropriate corrective measure *MUST* be taken and documented before further sample analysis proceeds.

- All samples processed with a LCS that is out of control (i.e., contaminated) will require re-analysis or data qualifiers to be attached to the analytical results.

13.4 Sample Analysis by GC/MS

13.4.1 Summary. The sample extract is analyzed by GC/MS and quantitated by the internal standard method.

13.4.2 Frequency. Before samples can be analyzed, the instrument must meet the GC/MS tuning and initial calibration or continuing calibration technical acceptance criteria. If there is time remaining in the 12-hour time period with a valid initial calibration or continuing calibration, samples may be analyzed in the GC/MS system that meet the instrument performance check criteria.

13.4.3 Procedure. For sample analysis, perform the following:

- Set up the GC/MS system.
- All sample extracts must be allowed to warm to ambient temperature (~1 hour) before analysis. All sample extracts must be analyzed under the same instrumental conditions as the calibration standards.
- Add the internal standard spiking solution to the 1.0 mL extract. For sample dilutions, add an appropriate amount of the internal standard spiking solution to maintain the concentration of the internal standards at 2 ng/ μ L in the diluted extract.
- Inject 2.0 μ L of sample extract into the GC/MS, and start data acquisition.
- When all semi-volatile target compounds have eluted from the GC, terminate the MS data acquisition and store data files on the data system storage device. Use appropriate data output software to display full range mass spectra and SICPs. The sample analysis using the GC/MS is based on a combination of retention times and relative abundances of selected ions (see Table 6). These qualifiers should be stored on the hard disk of the GC/MS data computer and are applied for identification of each chromatographic peak. The retention time qualifier is determined to be +0.10 minute of the library retention time of the compound. The acceptance level for relative abundance is determined to be $\pm 15\%$ of the expected abundance. Three ions are measured for most of the PAH compounds. When compound identification is made by the computer, any peak that fails any of the qualifying tests is flagged (e.g., with an *). The data should be manually examined by the analyst to determine the reason for the flag and whether the compound should be reported as found. Although this step adds some subjective judgment to the analysis, computer-generated identification problems can be clarified by an experienced operator. Manual inspection of the quantitative results should also be performed to verify concentrations outside the expected range.

13.4.4 Dilutions. The following section provides guidance when an analyte exceeds the calibration curve.

- When a sample extract is analyzed that has an analyte target compound concentration greater than the upper limit of the initial calibration range or saturated ions from a compound (excluding the compound peaks in the solvent front), the extract must be diluted and reanalyzed. Secondary ion quantitation is *only* allowed when there are sample interferences with the primary quantitation ion. If secondary ion quantitation is used, calculate a relative response factor using the area response for the most intense secondary ion which is free of sample interferences, and document the reasons for the use of the secondary ion.
- Calculate the sample dilution necessary to keep the semi-volatile target compounds that required dilution within the upper half of the initial calibration range so that no compound has saturated ions (excluding the compound peaks in the solvent front). Dilute the sample in hexane in a volumetric flask. Analyze the sample dilution.

- The dilution factor chosen should keep the response of the largest peak for a *target compound* in the upper half of the initial calibration range of the instrument.
- If the on-column concentration of any target compound in any sample exceeds the initial calibration range, that sample must be diluted, the internal standard concentration readjusted, and the sample extract reanalyzed.
- Use the results of the original analysis to determine the approximate dilution factor required to get the largest analyte peak within the initial calibration range.

13.4.5 Quantitation. This section provides guidance for quantitating PAH analytes.

- Target components identified shall be quantified by the internal standard method. The internal standards used for the target compounds are the ones nearest the retention time of a given analyte.
- The relative response factor (RRF) from the daily continuing calibration standard analysis (or RRF of CAL 3) if the sample is analyzed in the same 12-hour sequence as the initial calibration) is used to calculate the concentration in the sample. Secondary ion quantitation is allowed *only* when there are sample interferences with the primary ion. If secondary ion quantitation is performed, document the reasons. The area of a secondary ion cannot be substituted for the area of a primary ion unless a relative response factor is calculated using the secondary ion.
- A retention time window is calculated for each single component analyte and surrogate. Windows are established as ± 0.01 RRT units of the retention time for the analyte in CAL 3 of the initial calibration or the continuing calibration.

13.4.6 Calculations. Perform the following calculations:

13.4.6.1 Calculation of Concentration. Calculate target compound concentrations using the following equation:

$$\text{Concentration, (ng/std m}^3\text{)} = \frac{A_x I_s V_t D_f}{A_{is} V_i \overline{\text{RRF}}}$$

where:

A_x = area response for the compound to be measured, counts

A_{is} = area response for the internal standard, counts

I_s = amount of internal standard, ng/ μ L

$\overline{\text{RRF}}$ = the mean RRF from the most recent initial calibration, dimensionless

V_i = volume of air sampled, std m^3

V_t = volume of final extract, μ L

D_f = dilution factor for the extract. If there was no dilution, D_f equals 1. If the sample was diluted, the D_f is greater than 1.

The concentrations calculated can be converted to ppb_v for general reference. The analyte concentration can be converted to ppb_v using the following equation:

$$C_A(\text{ppb}_v) = C_A(\text{ng/m}^3) \times 24.4/\text{MW}_A$$

where:

- C_A = concentration of analyte calculated, ng/std. m³
 MW_A = molecular weight of analyte, g/g-mole
 24.4 = molar volume occupied by ideal gas at standard temperature and pressure (25°C and 760 mm Hg), L/mole.

13.4.6.2 Estimated Concentration. The equation in Section 13.4.6.1 is also used for calculating the concentrations of the non-target compounds. Total area counts (or peak heights) from the total ion chromatogram generated by the mass spectrometer for Compendium Method TO-13A PAHs (see Figure 16) are to be used for both the non-target compound to be measured (A_x) and the internal standard (A_{is}). Associate the nearest internal standard free of interferences with the non-target compound to be measured. A relative response factor (RRF) of one (1) is to be assumed. The value from this quantitation shall be qualified as estimated ("J") (estimated, due to lack of a compound-specific response factor) and "N" (presumptive evidence of presence), indicating the quantitative and qualitative uncertainties associated with this non-target component. An estimated concentration should be calculated for all tentatively identified compounds (TICs) as well as those identified as unknowns.

13.4.6.3 Surrogate Percent Recovery (%R). Calculate the surrogate percent recovery using the following equation:

$$\%R = \frac{Q_d}{Q_a} \times 100$$

where:

- Q_d = Quantity determined by analysis, ng
 Q_a = Quantity added to sample/blank, ng

The surrogate percent recovery must fall between 60-120% to be acceptable.

13.4.6.4 Percent Area Response Change (%ARC). Calculate the percent area response change (%ARC) for the sample/blank analysis compared to the most recent CAL 3 analysis for each of the internal standard compounds using the following equation:

$$\%ARC = \frac{A_s - A_x}{A_x} \times 100$$

where:

- $\%ARC$ = percent area response change, %
 A_s = area response of the internal standard in the sample/blank analysis, counts
 A_x = area response of the internal standard in the most recent CAL 3 analysis, counts

The area change for the internal standard must not exceed -50 to +100 percent.

13.4.6.5 Internal Standard Retention Time Shift (RTS). Calculate the retention time shift (RTS) between the sample/blank analysis and the most recent CAL 3 analysis for each of the internal standards using the following equation:

$$RTS = RT_s - RT_x$$

where:

RT_s = retention time of the IS in the sample

RT_x = retention time of the IS in the most recent CAL 3 analysis.

13.4.7 Technical Acceptance Criteria. The following guideline is provided as technical acceptance criteria.

13.4.7.1 All target compound concentrations must not exceed the upper limit of the initial calibration range and no compound ion (excluding the compound peaks in the solvent front) may saturate the detector.

13.4.7.2 Internal standard responses and retention times in all samples must be evaluated during or immediately after data acquisition. If the retention time for any internal standard changes by more than 20 seconds from the latest continuing calibration standard or CAL 3 if samples are analyzed in the same 12-hour sequence as the initial calibration, the chromatographic system must be inspected for malfunctions, and corrections made as required. The SICP of the internal standards must be monitored and evaluated for each field and QC sample. If the SICP area for any internal standard changes by more than a factor of -50 to +100 percent, the mass spectrometric system must be inspected for malfunction and corrections made as appropriate. If the analysis of a subsequent sample or standard indicates that the system is functioning properly, then corrections may not be required.

13.4.7.3 When target compounds are below the low standard, but the spectrum meets the identification criteria, report the concentration/amount with a "J." For example, if the low standard corresponds to $0.1\mu\text{g}$ and an amount of $0.05\mu\text{g}$ is calculated, report as "0.05J."

13.4.8 Corrective Action. The following section provides guidance if analyte exceeds the technical criteria.

- If the sample technical acceptance criteria for the surrogates and internal standards are not met, check calculations, surrogate and internal standard solutions, and instrument performance. It may be necessary to recalibrate the instrument or take other corrective action procedures to meet the surrogate and internal standard technical acceptance criteria.
- Sample analysis technical acceptance criteria *must* be met before data are reported. Samples contaminated from laboratory sources, or associated with a contaminated method blank, or any samples analyzed that are not meet the technical acceptance criteria will require reanalysis.
- The samples or standards with SICP areas outside the limits must be reanalyzed. If corrections are made, then the laboratory must demonstrate that the mass spectrometric system is functioning properly. This must be accomplished by the analysis of a standard or sample that meets the SICP criteria. After corrections are made, the reanalysis of samples analyzed while the system was malfunctioning is required.
- If after reanalysis, the SICP areas for all internal standards are inside the technical acceptance limits (-50 to +100 percent), then the problem with the first analysis is considered to have been within the control of the laboratory. Therefore, submit *only* data from the analysis with SICPs within the technical acceptance limits. This is considered the *initial* analysis and must be reported as such on all data deliverables.
- If the reanalysis of the sample does not solve the problem (i.e., the SICP areas are outside the technical acceptance limits for both analyses) then the laboratory must submit the SICP data and sample data from both analyses. Distinguish between the initial analysis and the reanalysis on all data deliverables, using the sample suffixes specified.
- Tentative identification of an analyte occurs when a peak from a sample extract falls within the daily retention time window.
- If sample peaks are not detected, or all are less than full-scale deflection, the undiluted extract is acceptable for GC/MS analysis. If any sample ions are greater than the 120 percent of the initial calibration curve range, calculate the dilution necessary to reduce the major ion to between half- and full-range response.

14. Quality Assurance/Quality Control (QA/QC)

14.1 General System QA/QC

14.1.1 Each laboratory that uses Compendium Method TO-13A must operate a formal quality control program. The minimum requirements of this program consist of an initial demonstration of laboratory capability and an ongoing analysis of spiked samples to evaluate and document quality data. The laboratory must maintain records to document the quality of the data generated. Ongoing data quality checks are compared with established performance criteria to determine if the results of analyses meet the performance characteristics of the method. When results of sample spikes indicate a typical method performance, a quality control check standard must be analyzed to confirm that the measurements were performed in an in-control mode of operation.

14.1.2 Before processing any samples, the analyst should demonstrate, through the analysis of a reagent solvent blank, that interferences from the analytical system, glassware, and reagents are under control. Each time a set of samples is extracted or there is a change in reagents, a reagent solvent blank should be processed as a safeguard against chronic laboratory contamination. The blank samples should be carried through all stages of the sample preparation and measurement steps.

14.1.3 For each analytical batch (up to 20 samples), a reagent blank, matrix spike, and deuterated/surrogate samples must be analyzed (the frequency of the spikes may be different for different monitoring programs). The blank and spiked samples must be carried through all stages of the sample preparation and measurement steps.

14.1.4 The experience of the analyst performing GC/MS is invaluable to the success of the methods. Each day that analysis is performed, the daily calibration sample should be evaluated to determine if the chromatographic system is operating properly. Questions that should be asked are: Do the peaks look normal? Are the response windows obtained comparable to the response from previous calibrations? Careful examination of the standard chromatogram can indicate whether the column is still good, the injector is leaking, the injector septum needs replacing, etc. If any changes are made to the system (e.g., column changed), recalibration of the system must take place.

14.2 Process, Field, and Solvent Blanks

14.2.1 One PUF cartridge and filter from each batch of approximately 20 should be analyzed without shipment to the field for the compounds of interest to serve as a process blank. A blank level specified in Section 10.2 for each cartridge/filter assembly is considered to be acceptable.

14.2.2 During each sampling episode, at least one cartridge and filter should be shipped to the field and returned, without drawing air through the sampler, to serve as a field blank.

14.2.3 During the analysis of each batch of samples at least one solvent process blank (all steps conducted but no cartridge or filter included) should be carried through the procedure and analyzed. Blank levels should be those specified in Section 10.2 for single components to be acceptable.

14.2.4 Because the sampling configuration (filter and backup sorbent) has been tested for targeted PAHs in the laboratory in relationship to collection efficiency and has been demonstrated to be greater than 95 percent for targeted PAHs (except naphthalene, acenaphthylene, and acenaphthene), no field recovery evaluation is required as part of the QA/QC program outlined in this section.

15. References

1. Dubois, L., Zdrojgowski, A., Baker, C., and Monknao, J.L., "Some Improvement in the Determination of Benzo[a]Pyrene in Air Samples," *J. Air Pollut. Contr. Assoc.*, 17:818-821, 1967.
2. Intersociety Committee, "Tentative Method of Analysis for Polynuclear Aromatic Hydrocarbon of Atmospheric Particulate Matter," *Health Laboratory Science*, 7(1):31-40, 1970.
3. Cautreels, W., and Van Cauwenberghe, K., "Experiments on the Distribution of Organic Pollutants Between Airborne Particulate Matter and Corresponding Gas Phase," *Atmos. Environ.*, 12:1133-1141, 1978.
4. "Tentative Method of Microanalysis for Benzo[a]Pyrene in Airborne Particles and Source Effluents," American Public Health Association, *Health Laboratory Science*, 7(1):56-59, 1970.
5. "Tentative Method of Chromatographic Analysis for Benzo[a]Pyrene and Benzo[k]Fluoranthene in Atmospheric Particulate Matter," American Public Health Association, *Health Laboratory Science*, 7(1):60-67, 1970.
6. "Tentative Method of Spectrophotometric Analysis for Benzo[a]Pyrene in Atmospheric Particulate Matter," American Public Health Association, *Health Laboratory Science*, 7(1):68-71, 1970.
7. Jones, P.W., Wilkinson, J.E., and Strup, P.E., *Measurement of Polycyclic Organic Materials and Other Hazardous Organic Compounds in Stack Gases: State-of-the-Art*, U. S. Environmental Protection Agency, Research Triangle Park, NC, U.S. EPA-600/2-77-202, 1977.
8. Walling, J.F., *Standard Operating Procedure for Ultrasonic Extraction and Analysis of Residual Benzo[a]Pyrene from Hi-Vol Filters via Thin-Layer Chromatography*, U.S. Environmental Protection Agency, Environmental Monitoring Systems Laboratory, Methods Development and Analysis Division, Research Triangle Park, NC, EMSL/RTP-SOP-MDAD-015, December, 1986.
9. Rasor, S., *Standard Operating Procedure for Polynuclear Aromatic Hydrocarbon Analysis by High Performance Liquid Chromatography Methods*, Acurex Corporation, Research Triangle Park, NC, 1978.
10. Rapport, S. W., Wang, Y. Y., Wei, E. T., Sawyer, R., Watkins, B. E., and Rapport, H., "Isolation and Identification of a Direct-Acting Mutagen in Diesel Exhaust Particulates," *Envir. Sci. Technol.*, 14:1505-1509, 1980.
11. Konlg, J., Balfanz, E., Funcke, W., and Romanowski, T., "Determination of Oxygenated Polycyclic Aromatic Hydrocarbons in Airborne Particulate Matter by Capillary Gas Chromatography and Gas Chromatography/Mass Spectrometry," *Anal. Chem.*, 55:599-603, 1983.
12. Chuang, J. C., Bresler, W. E., and Hannan, S. W., *Evaluation of Polyurethane Foam Cartridges for Measurement of Polynuclear Aromatic Hydrocarbons in Air*, U.S. Environmental Protection Agency, Environmental Monitoring Systems Laboratory, Methods Development and Analysis Division, Research Triangle Park, NC, EPA-600/4-85-055, September 1985.
13. Chuang, J. C., Hannan, S. W., and Koetz, J. R., *Stability of Polynuclear Aromatic Compounds Collected from Air on Quartz Fiber Filters and XAD-2 Resin*, U.S. Environmental Protection Agency, Environmental Monitoring Systems Laboratory, Methods Development and Analysis Division, Research Triangle Park, NC, EPA-600/4-86-029, September 1986.

14. Feng, Y., and Bidleman, T. F., "Influence of Volatility on the Collection of Polynuclear Aromatic Hydrocarbon Vapors with Polyurethane Foam," *Envir. Sci. Technol.*, 18:330-333, 1984.
15. Yamasaki, H., Kuwata, K., and Miyamoto, H., "Effects of Ambient Temperature on Aspects of Airborne Polycyclic Aromatic Hydrocarbons," *Envir. Sci. Technol.*, 16:89-194, 1982.
16. Galasyn, J. F., Hornig, J. F., and Soderberg, R. H., "The Loss of PAH from Quartz Fiber High Volume Filters," *J. Air Pollut. Contr. Assoc.*, 34:57-59, 1984.
17. You, F., and Bidleman, T. F., "Influence of Volatility on the Collection of Polynuclear Aromatic Hydrocarbon Vapors with Polyurethane Foam," *Envir. Sci. Technol.*, 18:330-333, 1984.
18. Chuang, J. C., Hannan, S. W., and Koetz, J. R., *Comparison of Polyurethane Foam and XAD-2 Resin as Collection Media for Polynuclear Aromatic Hydrocarbons in Air*, U.S. Environmental Protection Agency, Environmental Monitoring Systems Laboratory, Methods Development and Analysis Division, Research Triangle Park, NC, EPA-600/4-86-034, December 1986.
19. Chuang, J. C., Mack, G. A., Mondron, P. J., and Peterson, B. A., *Evaluation of Sampling and Analytical Methodology for Polynuclear Aromatic Compounds in Indoor Air*, U.S. Environmental Protection Agency, Environmental Monitoring Systems Laboratory, Methods Development and Analysis Division, Research Triangle Park, NC, EPA-600/4-85-065, January 1986.
20. Lewis, R. G., Brown, A. R., and Jackson, M. D., "Evaluation of Polyurethane Foam for High-Volume Air Sampling of Ambient Levels of Airborne Pesticides, Polychlorinated Biphenyls, and Polychlorinated Naphthalenes," *Anal. Chem.*, 49:1668-1672, 1977.
21. Lewis, R. G., and Jackson, M. D., "Modification and Evaluation of a High-Volume Air Sampler for Pesticides and Other Semi-volatile Industrial Organic Chemicals," *Anal. Chem.*, 54:592-594, 1982.
22. Winberry, W. T., and Murphy, N. T., *Supplement to Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air*, U.S. Environmental Protection Agency, Environmental Monitoring Systems Laboratory, Quality Assurance Division, Research Triangle Park, NC, EPA-600/4-87-006, September 1986.
23. Winberry, W. T., and Murphy, N. T., *Second Supplement to Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air*, U.S. Environmental Protection Agency, Environmental Monitoring Systems Laboratory, Quality Assurance Division, Research Triangle Park, NC, EPA 600/4-89-018, June 1989.
24. *Methods for Organic Chemical Analysis of Municipal and Industrial Wastewater*, U.S. Environmental Protection Agency, Environmental Monitoring and Support Laboratory, Cincinnati, OH, EPA-600/4-82-057, July 1982.
25. *ASTM Annual Book of Standards*, Part 31, D 3694, "Standard Practice for Preparation of Sample Containers and for Preservation," American Society for Testing and Materials, Philadelphia, PA, p. 679, 1980.
26. Burke, J. A., "Gas Chromatography for Pesticide Residue Analysis; Some Practical Aspects," *Journal of the Association of Official Analytical Chemists*, 48:1037, 1965.

27. Cole, T., Riggin, R., and Glaser, J., *Evaluation of Method Detection Limits: Analytical Curve for EPA Method 610 - PNAs*, 5th International Symposium on Polynuclear Aromatic Hydrocarbons, Battelle, Columbus, OH, 1980.
28. *Handbook of Analytical Quality Control in Water and Wastewater Laboratories*, U.S. Environmental Protection Agency, Environmental Monitoring and Support Laboratory, Cincinnati, OH, EPA-600/4-79-019, March 1979.
29. *ASTM Annual Book of Standards*, Part 31, D 3370, "Standard Practice for Sampling Water," American Society for Testing and Materials, Philadelphia, PA, p. 76, 1980.
30. *Protocol for the Collection and Analysis of Volatile POHC's (Principal Organic Hazardous Constituents) Using VOST (Volatile Organic Sampling Train)*, U. S. Environmental Protection Agency, Research Triangle Park, NC, EPA-600/8-84-007, March 1984.
31. *Sampling and Analysis Methods for Hazardous Waste Combustion - Methods 3500, 3540, 3610, 3630, 8100, 8270, and 8310; Test Methods for Evaluating Solid Waste (SW-846)*, U.S. Environmental Protection Agency, Office of Solid Waste, Washington, D.C.
32. Riggin, R. M., *Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air*, U.S. Environmental Protection Agency, Environmental Monitoring Systems Laboratory, Quality Assurance Division, Research Triangle Park, NC, EPA-600/4-84-041, April 1984.
33. Chuang, C. C., and Peterson, B. A., *Review of Sampling and Analysis Methodology for Polynuclear Aromatic Compounds in Air from Mobile Sources, Final Report*, U. S. Environmental Protection Agency, Research Triangle Park, NC, EPA-600/S4-85-045, August 1985.
34. *Measurement of Polycyclic Organic Matter for Environmental Assessment*, U.S. Environmental Protection Agency, Industrial Environmental Research Laboratory, Research Triangle Park, NC, EPA-600/7-79-191, August 1979.
35. Hudson, J. L., *Standard Operating Procedure No. FA 113C: Monitoring for Particulate and Vapor Phase Pollutants Using the Portable Particulate/Vapor Air Sampler*, U.S. Environmental Protection Agency, Region VII, Environmental Monitoring and Compliance Branch, Environmental Services Division, Kansas City, KS, March 1987.
36. Trane, K. E., and Mikalsen, A., "High-Volume Sampling of Airborne Polycyclic Aromatic Hydrocarbons Using Glass Fibre Filters and Polyurethane Foam," *Atmos. Environ.*, 15:909-918, 1981.
37. Keller, C. D., and Bidleman, T. F., "Collection of Airborne Polycyclic Hydrocarbons and Other Organics with a Glass Fiber Filter - Polyurethane Foam System," *Atmos. Environ.*, 18:837-845, 1984.
38. Hunt, G. T., and Pangaro, N., "Ambient Monitoring of Polynuclear Aromatic Hydrocarbons (PAHs) Employing High Volume Polyurethane Foam (PUF) Samplers," In *Polynuclear Aromatic Hydrocarbons*, Cooke, M., Dennis, and A. J., Eds, Battelle Press, Columbus, OH, pp. 583-608, 1985.

39. Alfeim, I., and Lindskog, A., "A Comparison Between Different High Volume Sampling Systems for Collecting Ambient Airborne Particles for Mutagenicity Testing and for Analysis of Organic Compounds," *Sci. Total Environ.*, 34:203-222, 1984.
40. Umlauf, G., and Kaupp, H., "A Sampling Device for Semivolatile Organic Compounds in Ambient Air," *Chemosphere*, 27:1293-1296, 1993.
41. Hippelein, M., Kaupp, H., Dorr, G., and McLachlan, M. S., "Testing of a Sampling System and Analytical Method for Determination of Semivolatile Organic Chemicals in Air," *Chemosphere*, 26:2255-2263, 1993.
42. Ligocki, M. P., and Pankow, J. F., "Assessment of Adsorption/Solvent Extraction with Polyurethane Foam and Adsorption/Thermal Desorption with Tenax-GC for Collection and Analysis of Ambient Organic Vapors," *Anal. Chem.*, 57:1138-1144, 1985.
43. Yamasaki, H., Kuwata, K., and Miyamoto, H., "Effects of Ambient Temperature on Aspects of Airborne Polycyclic Aromatic Hydrocarbons," *Envir. Sci. Technol.*, 16:189-194, 1982.
44. Hart, K. M., and Pankow, J. F., "High-Volume Air Sampler for Particle and Gas Sampling: Use of Backup Filters to Correct for the Adsorption of Gas-Phase Polycyclic Aromatic Hydrocarbons to the Front Filter," *Envir. Sci. Technol.*, 28:655-661, 1994.
45. Kaupp, H., and Umlauf, G., "Atmospheric Gas-Particle Partitioning of Organic Compounds: Comparison of Sampling Methods," *Atmos. Environ.*, 13:2259-2267, 1992.
46. Coutant, R. W., Brown, L., Chuang, J. C., Riggan, R. M., and Lewis, R. G., "Phase Distribution and Artifact Formation in Ambient Air Sampling for Polynuclear Aromatic Hydrocarbons," *Atmos. Environ.*, 22:403-409, 1988.
47. Coutant, R. W., Callahan, P. J., Kuhlman, M. R., and Lewis, R. G., "Design and Performance of a High-Volume Compound Annular Denuder," *Atmos. Environ.*, 23:2205-2211.
48. Lewis, R. G., Kelly, T. J., Chuang, J. C., Callahan, P. J., and Coutant, R. W., "Phase Distributions of Airborne Polycyclic Aromatic Hydrocarbons in Two U.S. Cities," In *Proceedings of the 9th World Clean Air Congress & Exhibition*, Montreal, Ontario, Canada, 1991, Vol., Paper IU-11E.02.
49. Kaupp, H., and Umlauf, G., "Atmospheric Gas-Particle Partitioning of Organic Compounds: Comparison of Sampling Methods," *Atmos. Environ.*, 26A:2259-2267, 1992.
50. Riggan, R. M., *Technical Assistance Document for Sampling and Analysis of Toxic Organic Compounds in Ambient Air*, U.S. Environmental Protection Agency, Environmental Monitoring Systems Laboratory, Quality Assurance Division, Research Triangle Park, NC, EPA-600/4-83-027, June 1983.
51. *ASTM Annual Book of Standards*, Part 31, D 3694, "Standard Practice for Preparation of Sample Containers and for Preservation," American Society for Testing and Materials, Philadelphia, PA, p. 679, 1980.
52. *Carcinogens - Working with Carcinogens*, Department of Health, Education, and Welfare, Public Health Service, Center for Disease Control, National Institute for Occupational Safety and Health, Publication No. 77-206, August 1977.

53. *OSHA Safety and Health Standards, General Industry*, (29CFR1910), Occupational Safety and Health Administration, OSHA, 2206, Revised, January 1976.
54. "Safety in Academic Chemistry Laboratories," American Chemical Society Publication, Committee on Chemical Safety, 3rd Edition, 1979.

TABLE 1. FORMULAE AND PHYSICAL PROPERTIES OF SELECTED PAHs

Compound	Formula	Molecular Weight	Melting Point, °C	Boiling Point, °C	Vapor Pressure, kPa	CAS RN #
Naphthalene	C ₁₀ H ₈	128.18	80.2	218	1.1x10	91-20-3
Acenaphthylene	C ₁₂ H ₈	152.20	92-93	265-280	3.9x10	208-96-8
Acenaphthene	C ₁₂ H ₁₀	154.20	90-96	278-279	2.1x10	83-32-9
Fluorene	C ₁₃ H ₁₀	166.23	116-118	293-295	8.7x10	86-73-7
Anthracene	C ₁₄ H ₁₀	178.24	216-219	340	36x10	120-12-7
Phenanthrene	C ₁₄ H ₁₀	178.24	96-101	339-340	2.3x10	85-01-8
Fluoranthene	C ₁₅ H ₁₀	202.26	107-111	375-393	6.5x10	206-44-0
Pyrene	C ₁₆ H ₁₀	202.26	150-156	360-404	3.1x10	129-00-0
Benz(a)anthracene	C ₁₈ H ₁₂	228.30	157-167	435	1.5x10	56-55-3
Chrysene	C ₁₈ H ₁₂	228.30	252-256	441-448	5.7x10	218-01-9
Benzo(b)fluoranthene	C ₂₀ H ₁₂	252.32	167-168	481	6.7x10	205-99-2
Benzo(k)fluoranthene	C ₂₀ H ₁₂	252.32	198-217	480-471	2.1x10	207-08-9
Perylene	C ₂₀ H ₁₂	252.32	273-278	500-503	7.0x10	198-55-8
Benzo(a)pyrene	C ₂₀ H ₁₂	252.32	177-179	493-496	7.3x10	50-32-8
Benzo(e)pyrene	C ₂₀ H ₁₂	252.32	178-179	493	7.4x10	192-92-2
Benzo(g,h,i)perylene	C ₂₂ H ₁₄	276.34	275-278	525	1.3x10	191-24-2
Indeno(1,2,3-cd)pyrene	C ₂₇ H ₁₈	276.34	162-163	--	ca.10	193-39-5
Dibenz(a,h)anthracene	C ₂₈ H ₁₈	278.35	266-270	524	1.3x10	53-70-3
Coronene	C ₂₄ H ₁₄	300.36	438-440	525	2.0x10	191-07-1

Many of these compounds sublime.

TABLE 2. GC-MS OPERATING CONDITIONS

Activity	Conditions
<u>Gas Chromatography</u>	
Column	J&W Scientific, DB-5 crosslinked 5% phenylmethyl silicone (30 m x 0.32 mm, 1.0 µm film thickness) or equivalent
Carrier Gas	Helium, velocity between 28-30 cm ³ /sec at 250°C
Injection Volume	2 µL, Grob-type, splitless
Injector Temperature	290°C
<u>Temperature Program</u>	
Initial Column Temperature	70°C
Initial Hold Time	4 ± 0.1 min.
Program	10°C/min to 300°C and hold 10 min.
Final Temperature	300°C
Final Hold Time	10 min. or until all compounds of interest have eluted
<u>Mass Spectrometer</u>	
Transfer Line Temperature	290°C or According to Manufacturer's Specification
Source Temperature	According to Manufacturer's Specifications
Electron Energy	70 volts (nominal)
Ionization Mode	EI
Mass Range	35 to 500 amu, full range data acquisition (SCAN) mode
Scan Time	At least 5 scans per peak, not to exceed 1 second per scan.

TABLE 3. DFTPP KEY IONS & ION ABUNDANCE CRITERIA

Mass	Ion Abundance Criteria
51	30 to 60% of mass 198
68 70	Less than 2% of mass 69 Less than 2% of mass 69
127	40 to 60% of mass 198
197 198 199	Less than 2% of mass 198 Base peak, 100% relative abundance 5 to 9% of mass 198
275	10 to 30% of mass 198
365	Greater than 1.0% of mass 198
441 442 443	Present but less than mass 443 40% of mass 198 17 to 23% of mass 442

TABLE 4. COMPOSITION AND APPROXIMATE CONCENTRATION OF CALIBRATION SOLUTIONS

Target Compound	Concentration, ng/ μ L				
	CAL 1	CAL 2	CAL 3	CAL 4	CAL 5
PAHs	0.10	0.25	0.50	1.25	2.50
Acenaphthene	0.10	0.25	0.50	1.25	2.50
Acenaphthylene	0.10	0.25	0.50	1.25	2.50
Anthracene	0.10	0.25	0.50	1.25	2.50
Benz(a)anthracene	0.10	0.25	0.50	1.25	2.50
Benzo(a)pyrene	0.10	0.25	0.50	1.25	2.50
Benzo(b)fluoranthene	0.10	0.25	0.50	1.25	2.50
Benzo(e)pyrene	0.10	0.25	0.50	1.25	2.50
Benzo(g,h,i)perylene	0.10	0.25	0.50	1.25	2.50
Benzo(k)fluoranthene	0.10	0.25	0.50	1.25	2.50
Chrysene	0.10	0.25	0.50	1.25	2.50
Perylene	0.10	0.25	0.50	1.25	2.50
Dibenz(a,h)anthracene	0.10	0.25	0.50	1.25	2.50
Fluoranthene	0.10	0.25	0.50	1.25	2.50
Fluorene	0.10	0.25	0.50	1.25	2.50
Indeno(1,2,3-c,d)pyrene	0.10	0.25	0.50	1.25	2.50
Naphthalene	0.10	0.25	0.50	1.25	2.50
Coronene	0.10	0.25	0.50	1.25	2.50
Phenanthrene	0.10	0.25	0.50	1.25	2.50
Pyrene	0.10	0.25	0.50	1.25	2.50

TABLE 4. (Continued)

Target Compound	Concentration, ng/ μ L				
	CAL 1	CAL 2	CAL 3	CAL 4	CAL 5
SUGGESTED INTERNAL STANDARDS					
D ₈ -Naphthalene	0.5	0.5	0.5	0.5	0.5
D ₁₀ -Acenaphthene	0.5	0.5	0.5	0.5	0.5
D ₁₀ -Phenanthrene	0.5	0.5	0.5	0.5	0.5
D ₁₂ -Chrysene	0.5	0.5	0.5	0.5	0.5
D ₁₂ -Perylene	0.5	0.5	0.5	0.5	0.5
SUGGESTED SURROGATE COMPOUNDS					
D ₁₀ -Fluoranthene (field)	0.10	0.25	0.50	1.25	2.50
D ₁₂ -Benzo[a]pyrene (field)	0.10	0.25	0.50	1.25	2.50
D ₁₀ -Fluorene (lab)	0.10	0.25	0.50	1.25	2.50
D ₁₀ -Pyrene (lab)	0.10	0.25	0.50	1.25	2.50

TABLE 5. CHARACTERISTIC IONS FOR SURROGATE SUGGESTED STANDARDS

Classification	Primary Ion	Secondary Ion
<u>Internal Standards</u>		
D ₈ -Naphthalene	136	68,137
D ₁₀ -Acenaphthene	164	162,165
D ₁₀ -Phenanthrene	188	94,189
D ₁₂ -Chrysene	240	120,241
D ₁₂ -Perylene	264	260,265
<u>Laboratory Surrogates</u>		
D ₁₀ -Fluorene	176	88,177
D ₁₀ -Pyrene	212	106,213
<u>Field Surrogates</u>		
D ₁₀ -Fluoranthene	212	106,213
D ₁₂ -Benzo(a)pyrene	264	132,265

TABLE 6. EXAMPLE OF CHARACTERISTIC IONS FOR COMMON PAHs

Analyte	Primary Ion	Secondary Ion(s)
Pyrene	202	101,203
Benz(a)anthracene	228	229,226
Chrysene	228	226,229
Benzo(a)pyrene	252	253,126
Benzo(b)fluoranthene	252	253,126
Benzo(k)fluoranthene	252	253,126
Benzo(g,h,i)perylene	276	138,277
Dibenz(a,h)anthracene	278	139,279
Anthracene	178	179,176
Phenanthrene	178	179,176
Acenaphthene	154	153,152
Acenaphthylene	152	151,153
Benzo(e)pyrene	252	253,126
Fluoranthene	202	101,203
Fluorene	166	165,167
Ideno(1,2,3-cd)pyrene	276	138,227
Naphthalene	128	129,127
Perylene	252	253,126
Coronene	300	150,301

TABLE 7. EXAMPLE OF RELATIVE RESPONSE FACTOR CRITERIA
FOR INITIAL AND CONTINUING CALIBRATION OF
COMMON SEMI-VOLATILE COMPOUNDS

Semi-volatile Compounds	Minimum RRF	Maximum %RSD	Maximum %Difference
Naphthalene	0.700	30	30
Acenaphthylene	1.300	30	30
Acenaphthene	0.800	30	30
Fluorene	0.900	30	30
Phenanthrene	0.700	30	30
Anthracene	0.700	30	30
Fluoranthene	0.600	30	30
Pyrene	0.600	30	30
Benzo(a)anthracene	0.800	30	30
Chrysene	0.700	30	30
Benzo(b)fluoranthene	0.700	30	30
Benzo(k)fluoranthene	0.700	30	30
Benzo(a)pyrene	0.700	30	30
Indeno(1,2,3-cd)pyrene	0.500	30	30
Dibenz(a,h)anthracene	0.400	30	30
Benzo(g,h,i)perylene	0.500	30	30
Perylene	0.500	30	30
Coronene	0.700	30	30

TABLE 8. MINIMUM SAMPLING EQUIPMENT CALIBRATION AND ACCURACY REQUIREMENTS

Equipment	Acceptance limits	Frequency and method of measurement	Action if requirements are not met
<u>Sampler</u>	Indicated flow rate = true flow rate, $\pm 10\%$.	Calibrate with certified transfer standard on receipt, after maintenance on sampler, and any time audits or flow checks deviate more than $\pm 10\%$ from the indicated flow rate or $\pm 10\%$ from the design flow rate.	Recalibrate
<u>Associated equipment</u>			
Sampler on/off timer	± 30 min/24 hour	Check at purchase and routinely on sample-recovery days	Adjust or replace
Elapsed-time meter	± 30 min/24 hour	Compare with a standard time-piece of known accuracy at receipt and at 6-month intervals	Adjust or replace
Flowrate transfer standard (orifice device)	Check at receipt for visual damage	Recalibrate annually against positive displacement standard volume meter	Adopt new calibration curve




Figure 1. Ring structure of common PAHs.




Figure 2. Typical high volume air sampler for PAHs.




Figure 3. Typical absorbent cartridge assembly for sampling PAHs.






Figure 4. Apparatus used for sample clean-up and extraction.

Glass PUF Cartridge with
Stainless Steel Screens



5a. Glass PUF cartridge, plug, and end caps.



5b. PUF shipping container.

Figure 5. Glass PUF cartridge (5a) and shipping container (5b) for use with Compendium Method TO-13A.




Figure 6. Example of a field portable high volume air sampler for sampling PAHs developed by EPA.




Figure 7. Positive displacement rootsmeter used to calibrate orifice transfer standard used in Compendium Method TO-13A.

COMPENDIUM METHOD TO-13A
ORIFICE CALIBRATION DATA SHEET

T₁ _____ Name _____
 P₁ _____ mmHg Date _____
 Orifice No. _____
 Rootsmeeter No. _____

Resistance Plants (No. of holes)	Air Volume Measured by Rootsmeeter V _m		Standard Volume, V _{std} (std m ³)	Time for Air Volume to Pass Through Rootsmeeter, θ (min)	Rootsmeeter Pressure Differential, ΔP (mm Hg)	Pressure Drop Across Orifice, ΔH (in. H ₂ O)	x-Axis Standard Flowrate, Q _{std} (std m ³ /min)
	(R ³)	(m ³)					
5	200	5.66					
7	200	5.66					
10	300	8.50					
13	300	8.50					
18	300	8.50					

Factors: $(R^3)(0.02832 \frac{m^3}{R^3}) = m^3$ and $(in. Hg) 25.4 (\frac{mm Hg}{in. Hg}) = mm Hg$

Calculation Equations:

$$1. V_{std} = V_m \left(\frac{P_1 - \Delta P}{P_{std}} \right) \left(\frac{T_{std}}{T_1} \right)$$

where:

$$T_{std} = 296^\circ K$$

$$P_{std} = 760.0 \text{ mm Hg}$$

$$2. Q_{std} = \frac{V_{std}}{\theta}$$

Figure 8. Example of a high-volume orifice calibration data sheet for Compendium Method TO-13A.




Figure 9. Typical field calibration configuration for Compendium Method TO-13A sampler.

FIELD CALIBRATION DATA SHEET FOR COMPENDIUM METHOD TO-13A PAH
SAMPLER CALIBRATION

Sampler ID: _____

Sampler Location: _____

Calibration Orifice ID: _____

Job No.: _____

High Volume Transfer Orifice Data:

Correlation Coefficient (CC1): _____

Slope (M1): _____

(CC2): _____

(M2): _____

Intercept (B1): _____

(B2): _____

Calibration Date: ____ Time: _____

Calibration Ambient Temperature: ____ °F ____ °C CALIBRATOR'S SIGNATURE _____

Calibration Ambient Barometric Pressure: ____ "Hg ____ mm Hg _____

Calibration set point (SP): _____

SAMPLER CALIBRATION

Actual values from calibration		Calibrated values		
Orifice manometer, inches (Y1)	Monitor magnehelic, inches (Y2)	Orifice manometer (Y3)	Monitor magnehelic (Y4)	Calculated value orifice flow, scm (X1)
	70			
	60			
	50			
	40			
	30			
	20			
	10			

Definitions

Y1 = Calibration orifice reading, in. H₂OY2 = Monitor magnehelic reading, in. H₂OP_a = Barometric pressure actual, mm Hg

B1 = Manufacturer's Calibration orifice Intercept

M1 = Manufacturer's Calibration orifice manometer slope

Y3 = Calculated value for orifice manometer
= {Y1(Pa/760)[298/(Ta + 273)]^{1/2}}

Y4 = Calculated value for magnehelic

= {Y2(Pa/760)[298/(Ta + 273)]^{1/2}}

X1 = Calculated value orifice flow, scm

= (Y3 - B1)/M1

P_{std} = Barometric pressure standard, 760 mm HgT_a = Temperature actual, °CT_{std} = Temperature standard, 25°C

Figure 10. Typical orifice transfer field calibration data sheet for Compendium Method TO-13A.




Figure 11. Example of relationship between orifice transfer standard and flow rate through Compendium Method TO-13A sampler.

**COMPENDIUM METHOD TO-13A
FIELD TEST DATA SHEET
GENERAL INFORMATION**

Sampler I.D. No.: _____
 Lab PUF Sample No.: _____
 Sample location: _____

Operator: _____
 Other: _____

PUF Cartridge Certification Date: _____
 Date/Time PUF Cartridge Installed: _____
 Elapsed Timer: _____
 Start _____
 Stop _____
 Diff. _____
 Sampling
 M1 _____ B1 _____
 M2 _____ B2 _____

	Start	Stop
Barometric pressure ("Hg)	_____	_____
Ambient Temperature (°F)	_____	_____
Rain	Yes _____	Yes _____
	No _____	No _____
Sampling time		
Start	_____	
Stop	_____	
Diff.	_____	

Audit flow check within ±10 of set point
 _____ Yes
 _____ No

TIME	TEMP	BAROMETRIC PRESSURE	MAGNEHELIC READING	CALCULATED FLOW RATE (std. m ³)	READ BY
Avg.					

Comments

Figure 12. Example of typical Compendium Method TO-13A field test data sheet (FTDS).




Figure 13. Sample clean-up, concentration, separation and analysis sequence for common PAHs.
 [Note: XAD-2 sequence is similar to PUF except methylene chloride is the solvent.]




Figure 14. Typical quality assurance specifications for GC/MS/DS operation.





Figure 15. Mass spectra of Compendium Method TO-13A compounds for (a) naphthalene and (b) acenaphthylene.





Figure 15 (Cont). Mass spectra of Compendium Method TO-13A compounds for (c) acenaphthene and (d) fluorene.





Figure 15 (Cont). Mass spectra of Compendium Method TO-13A compounds for (e) anthracene and (f) phenanthrene.





Figure 15 (Cont). Mass spectra of Compendium Method TO-13A compounds for (g) fluoranthene and (h) pyrene.





Figure 15 (Cont). Mass spectra of Compendium Method TO-13A compounds for (i) benz(a)anthracene and (j) chrysene.





Figure 15 (Cont). Mass spectra of Compendium Method TO-13A compounds for (k) benzo(b)fluoranthene and (l) benzo(k)fluoranthene.





Figure 15 (Cont). Mass spectra of Compendium Method TO-13A compounds for (m) benzo(a)pyrene and (n) benzo(e)pyrene.





Figure 15 (Cont). Mass spectra of Compendium Method TO-13A compounds for (o) benzo(g,h,i)perylene and (p) indeno(1,2,3-cd)pyrene.




Figure 15 (Cont). Mass spectra of Compendium Method TO-13A compounds for (q) dibenz(a,h)anthracene.




Figure 16. Total ion chromatogram (TIC) of Compendium Method TO-13A target PAHs.

This page intentionally left blank

Other Test Method 45 (OTM-45) Measurement of Selected Per- and Polyfluorinated Alkyl Substances from Stationary Sources

Background on OTM-45

The posting of a test method on the Other Test Methods portion of the EMC website is neither an endorsement by EPA regarding the validity of the test method nor a regulatory approval of the test method. The purpose of the Other Test Methods portion of the EMC website is to promote discussion of developing emission measurement methodologies and to provide regulatory agencies, the regulated community, and the public at large with potentially helpful tools. Other Test Methods are test methods which have not yet been subject to the Federal rulemaking process. Each of these methods, as well as the available technical documentation supporting them, have been reviewed by the EMC staff and have been found to be potentially useful to the emission measurement community. The types of technical information reviewed include field and laboratory validation studies; results of collaborative testing; articles from peer-reviewed journals; peer review comments; and quality assurance (QA) and quality control (QC) procedures in the method itself. The EPA strongly encourages the submission of additional supporting field and laboratory data as well as comments regarding these methods.

These methods may be considered for use in federally enforceable State and local programs [e.g., Title V permits, State Implementation Plans (SIP)] provided they are subject to an EPA Regional SIP approval process or permit veto opportunity and public notice with the opportunity for comment. The methods may also be candidates to be alternative methods to meet Federal requirements under 40 CFR Parts 60, 61, and 63. However, they must be approved as alternatives under Parts 60.8, 61.13, or 63.7(f) before a source may use them for this purpose. Consideration of a method's applicability for a particular purpose should be based on the stated applicability as well as the supporting technical information. The methods are available for application without EPA oversight for other non-EPA program uses including state permitting programs and scientific and engineering applications. As many of these methods are submitted by parties outside the Agency, the EPA staff may not necessarily be the technical experts on these methods. Therefore, technical support from EPA for these methods is limited, but the table at the end of this introduction contains contact information for the authors and developers so that you may contact them directly. Also, be aware that these methods are subject to change based on the review of additional validation studies or on public comment as a part of adoption as a Federal test method, the Title V permitting process, or inclusion in a SIP.

Validated measurement methods are limited and under development for reliably identifying and quantifying if per- and polyfluoroalkyl substances (PFAS) are released into the air from stationary sources. The current lack of standardized methods to measure PFAS emissions and the limited availability of data on the performance of methods to measure PFAS introduce uncertainty in the understanding of the release of PFAS into the air from these sources. The lack of validated stationary source measurement methods for PFAS also leads to inconsistent findings, incomparable measurements, and lack of coordination between policy makers, facilities and control technology development. This OTM recommends a consistent method for use by the facilities, stationary source test teams, research laboratories, and other stakeholders to measure a common list of PFAS compounds emitted from vents and stacks. This OTM includes

performance based PFAS measurement tools and performance criteria developed through field application of this method.

The analytical method imbedded in OTM 45 may support a variety of monitoring applications, which include the analysis of multiple short-chain PFAS that cannot be measured by EPA Method 537.1. This posting meets an agency commitment identified within the 2020 National Defense Authorization Act guidance for PFAS disposal and destruction. Posting this method, in and of itself, does not establish a requirement, although the use of this method may be specified by the EPA or a state through independent actions. Terms such as “must” or “required,” as used in this document, refer to procedures that are to be followed to conform with the method. References to specific brands and catalog numbers are included only as examples and do not imply endorsement of the products. Such reference does not preclude the use of equivalent products from other vendors or suppliers.

OTM 45 is a draft method under evaluation that will be updated as more data from stakeholders becomes available. Due to the urgent need for consistency, this method is being released as an “Other Test Method (OTM)” by EPA’s Emission Measurements Center to promote consistency with what we believe is the current best practices to sample and analyze the PFAS targets from stationary sources. We solicit any and all feedback, comments, and additional data coming from the application of this method as we work to adjust this method in anticipation of developing a reference method for PFAS from air emission sources. PFAS compounds encompass a wide variety of moiety's that can lead to different physical and chemical properties. As such this method may not be suitable for certain PFAS compounds with unique properties, requiring evaluation of this method for PFAS compounds that might be added later.

Note: Please submit a copy, either electronic or paper, of any test report from application of this OTM to EPA’s Measurement Technology Group.

- **Electronic copies should be submitted via email with the subject line “OTM-045” to: EMC@epa.gov**
- **Paper copies should be mailed to:
Measurement Technology Group
Office of Air Quality Planning and Standards
U.S. Environmental Protection Agency (Mail Code E143-02)
Research Triangle Park, NC 27711**

OTM-45 Authors and Developers		
Ray Merrill*	OAQPS/AQAD/MTG	Merrill.Raymond@epa.gov
Jeff Ryan*	ORD/CEMM/AMCD/CSB	Ryan.Jeff@epa.gov
Ken Krebs	ORD/CEMM/AMCD/CSB	Krebs.Ken@epa.gov
Stephen Jackson	ORD/CEMM/AMCD/CSB	Jackson.Stephen@epa.gov
Lindsay Wickersham	ORD/CEMM/AMCD/CSB	Wickersham.Lindsay@epa.gov
Bob Wright	ORD/CEMM/IOD	Wright.Bob@epa.gov

* **Primary contacts**

Other Test Method 45 - (OTM-45) Measurement of Selected Per- and Polyfluorinated Alkyl Substances from Stationary Sources)

1.0 Scope and Application

1.1 Applicability. OTM-45 is a performance-based method applicable to the collection and quantitative analysis of specific semivolatile (Boiling point > 100°C) and particulate-bound per- and polyfluorinated alkyl substances (PFAS) in air emissions from stationary sources. This method can also be used for the collection and recovery of ionic and covalent PFAS for non-targeted analysis (NTA) of PFAS compounds. Table 45-1 of this method lists the individual target analytes that have been evaluated for measurement by OTM-45.

1.2 Scope. This method describes the sampling and sample recovery procedures used to measure individual semivolatile PFAS from stationary source air emissions. OTM-45 incorporates by reference some of the specifications (e.g., equipment and supplies) and procedures (e.g., sampling and sample preparation) from other methods that are essential to conducting OTM-45. To obtain reliable samples, source sampling teams should be trained and experienced with the following additional EPA test methods: Method 1; Method 2; Method 3; Method 4; and Method 5 of Appendices A-1, A-2, and A-3 to 40 Code of Federal Regulations (CFR) Part 60.

Laboratory analysis teams should be trained and experienced in the use of liquid chromatography coupled with tandem mass spectrometry (LC-MS/MS) multiple reaction monitoring (MRM) as described in EPA Method 533 and Method 537.1

1.3 Branched and Linear PFAS Isomers

Both branched and linear PFAS isomers may be found in the environment. This method includes procedures for summing the contribution of multiple isomers to the final reported concentration.

In those cases where standard materials containing multiple isomers are commercially available, laboratories should obtain such standards for the method analytes.

1.4 Performance Based. This method provides some flexibility for analysis of PFAS by including the addition of isotopically labeled standards in various parts of the sampling system to assess and evaluate method performance against criteria for successful sampling and analysis procedures. The laboratory may select LC columns, LC conditions, and MS conditions different from those used to develop the method. At a minimum, the pre-sampling isotope spike standards and the pre-analysis isotope spike standards specified in the method must be used. Users may modify the method to overcome interferences or to substitute superior materials and equipment, provided they use LC-MS/MS as the basis for separation and quantitation of method target compounds and meet all performance criteria in this method. Section 9 of this method presents requirements for method performance.

2.0 Summary of Method

This method identifies and determines the concentration in mass per unit gas volume sampled of specific PFAS compounds in source emissions. Gaseous and particulate bound target pollutants are withdrawn from the gas stream isokinetically and collected in the sample probe, on a glass fiber or quartz filter, on a packed column of adsorbent material and in a series of impingers. The target compounds are extracted from the individual sample collection media. The OTM-45 train results in four (4) discrete sample extract fractions for analysis. The extracts are analyzed by LC-MS/MS in the MRM detection mode. Quantification of each analyte is calculated using the isotope dilution technique. For QC purposes, the percent recoveries of the **pre-extraction standards** are calculated using the integrated peak areas of **pre-analysis standard(s)**, which are

added to the final extract and function as traditional internal standards, exclusively applied to the pre-extraction standards. The use of **pre-sampling standards** added to XAD-2 collection media prior to sampling and analyzed in the same manner as targeted PFAS compounds serves as an indication of the method's quantitative capture efficiency. This method is not intended to differentiate between target compounds in particle or vapor fractions. This method uses isotopically labeled standards to improve method accuracy and precision.

3.0 Definitions

3.1 Alternate Recovery Standards. A group of isotopically labeled compounds that are not otherwise designated in this method for quality control purposes. Use alternative recovery standards to assess the recovery of a compound class relative to a step in the sampling and analysis procedure that is not already assessed as a mandatory part of this method.

3.2 Analysis Batch. A set of samples that are analyzed on the same instrument during a 24-hour period that begins and ends with the analysis of the appropriate Continuing Calibration Check (CCC) standards. Additional CCCs may be required depending on the length of the Analysis Batch and the number of field samples.

3.3 Batch Blank Sample. A laboratory blank sample composed of a precleaned filter and XAD-2 media processed and analyzed using the same procedures as a field sample.

3.4 Calibration Standard. A solution of the method analytes, pre-extraction standards, and pre-analysis standard(s) prepared from the Primary Dilution Standards and stock standards. The calibration standards are used to calibrate the instrument response with respect to analyte concentration.

3.5 Continuing Calibration Check Standard (CCC). A calibration standard that is analyzed periodically to verify the accuracy of the existing calibration. The mid-point calibration standard is typically used to verify calibration.

3.6 Congener. An individual compound with a common structure, only differing by the number of fluorine atoms attached to the structure.

3.7 Extraction Batch. A set of field samples (not including QC samples) extracted together using the same lot of extraction devices, solvents, and fortifying solutions.

3.8 Field Sample Media Blank (FSMB). Also called the field trip blank. The FSMB is intended to include and represent the sampling media (i.e., filter, XAD-2 adsorbent) and reagents (i.e., impinger contents, rinsing solvents) associated with the field sample collection and recovery, but is not actually used in the field. The FSMB is treated as a sample in all respects, including shipment to the sampling site, exposure to sampling site conditions, storage, and all analytical procedures. The purpose of the FSMB is to determine if method analytes or other interferences are introduced into the sample from shipping, storage, and the field environment and procedures.

3.9 Homolog. A compound belonging to a series of compounds with the same general molecular formula, differing from each other by the number of repeating units.

3.10 Pre-extraction Standards. Pre-extraction standards are isotopically labeled analogues of the method analytes that are added to the sample prior to extraction in a known amount.

Note: Not all target PFAS currently have an isotopically labeled analogue. In these cases, an alternate isotopically labelled analogue (isotopologue) is used as recommended in Table 45- 1.

3.11 Branched and linear isomers. Individual compounds with a common molecular formula, differing by the position of carbon and fluorine atoms attached to the structure.

3.12 Isotope Dilution Technique. An analytical technique for measuring analyte concentration using the ratio of the peak area of the native analyte to that of an isotopically labeled analogue, added to the original sample in a known amount and carried through the entire analytical procedure.

3.13 Isotopologue. An individual compound with an identical chemical formula and structure, differing only in isotopic composition.

3.14 Laboratory Fortified Media Blanks (LFMB). Also commonly referred to as Laboratory Control Samples. The LFMB includes and represents all the sampling media (i.e., filter, XAD-2 adsorbent) and reagents (i.e., impinger contents, rinsing solvents) associated with the field sample collection and recovery to which known quantities of the method analytes and isotope dilution analogues are added. The results of the LFMB verify method performance in the absence of sample matrix.

3.15 Laboratory Sample Media Blank (LSMB). The LSMB is intended to include and represent the sampling media (i.e., filter, XAD-2 adsorbent) and reagents (i.e., impinger contents, rinsing solvents) associated with the field sample collection and recovery, but is not actually shipped to the field and remains in the laboratory. The results of the LSMB verify that the sampling media and reagents are not introducing target analyte or interfering species.

3.16 Method Detection Limit (MDL). The minimum qualitatively recognizable signal in laboratory analyses above background for a target compound with 99 percent confidence. Procedures for determining MDL are provided in Section 9.2.1.3.

Note: MDL is typically determined following 40 CFR Part 136 appendix B and includes samples prepared from blank media, spiked within 5 times of the expected MDL and processed in a manner identical to field sample preparation.

3.17 Perfluorinated alkyl substances. Aliphatic substances where all of the hydrogen atoms attached to carbon atoms have been replaced by F atoms, except those H atoms whose substitution would modify the nature of any functional groups present.

3.18 Polyfluorinated alkyl substances. Aliphatic substances where all hydrogen atoms attached to at least one (but not all) carbon atoms have been replaced by F atoms.

3.19 Precursor Ion. The gas-phase species corresponding to the method analyte that is produced in the electrospray ionization interface of the LC-MS/MS. During tandem mass spectrometry, or MS/MS, the precursor ion is mass selected and fragmented by collision-activated dissociation to produce distinctive product ions of smaller mass to charge (m/z) ratio. For this method, the precursor ion is usually the deprotonated molecule ($[M - H]^-$) of the method analyte, except for HFPO-DA. For this analyte, the precursor ion is formed by decarboxylation of HFPO-DA.

3.20 Pre-analysis Standard(s). Pre-analysis standard(s) are quality control compounds that are added to all standard solutions and to extracts immediately prior to analysis in a known amount and used to measure the relative response of the isotopically labelled analogues that are

components of the same solution. For this method, the pre-analysis standard(s) are two isotopically labeled analogues of the method analytes. The pre-analysis standard(s) are indicators of instrument performance and are used to calculate the recovery of the pre-extraction standards through the extraction procedure. In this method, the pre-analysis standard(s) are not used in the calculation of the recovery of the native analytes. Pre-analysis standard(s) are added to every sample (including blank, quality control sample, and calibration solutions) at a known amount.

Note: Pre-analysis standard(s) is identical to Isotope Performance Standards in Method 533.

3.23 Pre-extraction Standard(s). A group of isotopically labeled analogues of the method analytes that are added in a known amount to all standard solutions, to each field sample fraction: (i.e., the primary and secondary XAD-2 adsorbent traps, filter, and impinger samples) and to laboratory blanks immediately before extraction. Used to correct the quantity of the native target compounds present in the sample for extraction, cleanup, and concentration recovery.

Note: Not all target PFAS currently have an isotopically labelled analogue. In these cases, an alternative isotopically labelled analogue is recommended in Table 45-1.

3.24 Pre-sampling Standard(s). A group of isotopically labeled compounds added in a known amount to the XAD-2 adsorbent prior to sampling used to indicate the sample collection and recovery efficiency of the method.

3.25 Pre-transport Standard(s). A group of isotopically labeled compounds added by the laboratory into the sample containers used in the field to contain and transport probe rinse and impinger samples. The measured concentration of the pre-transport recovery standard provides a

quality check on potential probe rinse sample spillage or mishandling after sample collection and during shipping.

3.26 Product Ions. One or more fragment ions that are produced in MS/MS by collision activated dissociation of the precursor ion.

3.27 Quality Control Standard (QCS). A calibration standard prepared independently from the primary calibration solutions. For this method, the QCS is a repeat of the entire dilution scheme starting with the same stock materials (neat compounds or purchased stock solutions) used to prepare the primary calibration solutions. Independent sources and separate lots of the starting materials are not required, provided the laboratory has obtained the purest form of the starting materials commercially available. The purpose of the QCS is to verify the integrity of the primary calibration standards.

3.28 Quantitative Reporting Limit (QRL): The minimum quantitative level that can reported. The QRL is based on the lowest concentration or level target PFAS compound used during calibration. Being sample specific, the QRL is affected by sample size, dilution, aliquots, etc.

3.29 Relative Response Factor (RRF). The response of the mass spectrometer to a known amount of an analyte relative to a known amount of an isotopically labeled standard.

3.30 Sampling Train Field Blank (STFB). The complete field assembly and recovery of an OTM-45 sampling train without actual sampling, including bringing the train to sampling location, heating and leak checks. The STFB is conducted using glassware that has been previously used for emissions sampling as part of the current field test. The purpose of the

STFB is to determine if method analytes or other interferences are introduced into the sample from previous sample runs using previously sampled sample train glassware.

3.31 Sampling Train Proof Blank (STPB). The complete field assembly and recovery of a clean OTM-45 sampling train without actual sampling, including bringing the train to sampling location, heating and leak checks. The STFB is conducted using clean glassware that has not been previously used for emissions sampling as part of the current field test. The purpose of the STPB is to determine if method analytes or other interferences are introduced into the sample from the clean, unused sample train glassware, train assembly, preparation and recovery, including the field environment.

3.32 Semivolatile and Condensable PFAS. Poly and perfluoro organic compounds with boiling points above 100°C.

3.33 Stack Detection Limit (SDL). The minimum qualitatively recognizable gaseous stack concentration above background for a target compound. The SDL is a mathematically derived from the method detection limit (MDL) for each sample fraction, the total gaseous stack sample volume collected, and the sample preparation steps in this method. Each sample fraction in this method has a distinct MDL based on the sample preparation, concentration, and aliquot splitting performed during the sample analysis procedures. Being sample specific, the SDL is affected by stack sample volume, sample extract volume, sample concentration, sample splits, and dilution, etc. The SDL is based on the sum of sample fractions 1-3 MDLs and the run-specific gaseous sample volume.

4.0 Interferences

4.1 Organic Compounds. Very high amounts of other organic compounds in the matrix may interfere with the analysis. This method provides examples of extraction and cleanup procedures to reduce, but not necessarily eliminate, matrix effects due to high concentrations of organic compounds.

4.2 Target compound contaminants or related organics in solvents, reagents, glassware, isotopically labeled spiking standards, and other sample processing hardware are potential method interferences. Routinely evaluate all these materials to demonstrate that they are either free from interferences under the conditions of the analysis, or that the interference does not compromise the quality of the analysis results. Evaluate chemical interference through the preparation and analysis of batch blank samples. Use high purity reagents, solvents, and standards to minimize interference problems in sample analysis.

Note: In this method, $^{13}\text{C}_3$ -PFBA is used as a pre-analysis standard(s) and $^{13}\text{C}_4$ -PFBA is used as an isotope dilution analogue. Both share the same product ion, m/z 172. Because the natural abundance of ^{13}C is 1.1%, there is a 1.1% contribution to the $^{13}\text{C}_4$ -PFBA area from the lone, unlabeled ^{12}C atom in $^{13}\text{C}_3$ -PFBA. Users of this method may consider this bias to the area of the PFBA isotope dilution analogue insignificant.

4.3 PTFE products including PTFE cap liners can be a source of PFAS contamination. The use of PTFE in this method must be avoided or the product must be tested and shown to be contaminant free before use. Polypropylene (PP) or polyethylene (PE, HDPE) products may be used in place of PTFE products to minimize PFAS contamination.

4.4 Labeled standards may include trace quantities of native PFAS and should be screened or certified pure before use.

4.5 Following extraction, aqueous samples should not come in contact with any glass containers or pipettes as PFAS analytes can potentially adsorb to glass surfaces. The eluate must be collected in a polypropylene tube prior to concentration to dryness. Concentration to dryness in glass tubes may cause poor recovery. Standards dissolved in organic solvent may be purchased in glass ampoules. These standards in organic solvent are acceptable and subsequent transfers may be performed using glass syringes and pipets.

4.6 Isotopic purity of $^{13}\text{C}_3$ -PFBA. In this method, $^{13}\text{C}_3$ -PFBA is used as a pre-analysis standard(s) and $^{13}\text{C}_4$ -PFBA is used as an isotope dilution analogue. Both share the same product ion, m/z 172. Ten nanograms per liter of $^{13}\text{C}_4$ -PFBA is added to the sample prior to extraction (10 ng/mL extract concentration assuming 100% recovery), and 10 ng/mL of $^{13}\text{C}_3$ -PFBA is added to the final extract. Because the natural abundance of ^{13}C is 1.1%, there is a 1.1% contribution to the $^{13}\text{C}_4$ -PFBA area from the lone, unlabeled ^{12}C atom in $^{13}\text{C}_3$ -PFBA. The authors confirmed this contribution empirically. Users of this method may consider this bias to the area of the PFBA isotope dilution analogue insignificant.

4.7 Isotopic purity of $^{13}\text{C}_4$ -PFBA. A trace amount of $^{13}\text{C}_3$ -PFBA was detected in the $^{13}\text{C}_4$ -PFBA. The contribution was no greater than 1%. The contribution of the pre-analysis standard(s) to the isotope dilution analogue is insignificant.

4.8 Telomer Sulfonates. Each of the three telomer sulfonates in the analyte list (4:2FTS, 6:2FTS, and 8:2FTS) are referenced to their $^{13}\text{C}_2$ isotope dilution analogue. The mass difference

between the telomer sulfonates and the pre-extraction standards is 2 mass units. The single sulfur atom in each of the unlabeled molecules has a naturally occurring M+2 isotope (^{34}S) at 4.25%. Thus, the precursor ions of the $^{13}\text{C}_2$ isotopically labeled analogues and the naturally occurring ^{34}S analogues present in the native analytes have the same nominal masses. The product ions of the telomer sulfonate pre-extraction standards listed in Table 45-2 would contain a small contribution from the ^{34}S analogue of the native telomer sulfonates. At the concentrations used in this study, the contribution of the ^{34}S analogue to the isotope dilution analogue was not greater than 2.7%. Alternate product ions may be used if there is sufficient abundance.

5.0 Safety

Note: Develop a strict safety program for the handling of PFAS samples to avoid contact with sampling media and sample recovery solutions.

5.1 Selected PFAS compounds, namely PFOS and PFOA have known aneugenic, endocrine, and teratogenic effects in laboratory animal studies. Be aware of the potential for inhalation, skin absorption and ingestion exposure to field personnel and laboratory analysts.

5.2 Commercial Standards. This method recommends that the laboratory purchase dilute standard solutions of the analytes required for this method. However, if preparing primary solutions, use a hood or glove box. Personnel handling primary solutions should wear personal protective equipment including nitrile gloves.

5.4 Toxicity. The toxicity or carcinogenicity of other reagents or chemicals used in this method is not precisely defined. However, treat each chemical as a potential health hazard and minimize exposure to these chemicals. The field and laboratory staff are responsible for maintaining a

current awareness file of Occupational Safety and Health Administration (OSHA) regulations regarding the safe handling of the chemicals specified in this method. Ensure that a reference file or list of internet sites that contain safety data sheets (SDS) is available to all personnel involved in the sampling and chemical analysis of samples known or suspected to contain PFAS.

6.0 Equipment and Supplies

Note: Brand names, suppliers, and part numbers are for illustration purposes only and no endorsement is implied. Apparatus and materials other than those specified in this method may achieve equivalent performance. Meeting the performance requirements of this method is the responsibility of the source testing team and laboratory team.

6.1 Sampling Train. Figure OTM 45-1 of this method shows a schematic of the OTM 45 sampling train. This sampling train configuration is adapted from EPA Method 5 procedures, and, as such, the majority of the required equipment is identical to that used in EPA Method 5 determinations. The OTM 45 sampling train is very similar, but not identical to the SW-846 Method 0010 sampling train. The specific OTM 45 adaptations are: the use of condenser and XAD-2 adsorbent module, which are used for the primary capture of PFAS compounds that pass through the glass or quartz-fiber filter in the gas phase; a series of impingers for additional capture of PFAS compounds that pass through the primary XAD-2 adsorbent module; and a secondary XAD-2 adsorbent module which is to determine the breakthrough of PFAS compounds not captured by the primary XAD-2 adsorbent module and impingers. The train is identical to that described in section 6.1.1 of Method 5 of appendix A-3 to 40 CFR part 60 with the following additions:

6.1.1 Nozzle. The nozzle must be made of quartz or borosilicate glass or titanium. Stainless steel nozzles should not be used.

6.1.2 Probe Liner. Use either borosilicate, or quartz glass probe liners with a heating system capable of maintaining a probe gas temperature of 120 ± 14 °C (248 ± 25 °F) during sampling. Use a PTFE ferrule or single-use PTFE coated O-ring to achieve the seal at the nozzle end of the probe for stack temperatures up to about 300 °C (572 °F). Use a quartz glass liner and integrated quartz nozzle for stack temperatures between 300 and 1,200 °C (572 and 2,192 °F).

6.1.3 Filter Holder. Use a filter holder of borosilicate glass with a PTFE frit or PTFE-coated wire filter support. The holder design should provide a positive seal against leakage from the outside or around the filter. The holder should be durable, easy to load, leak-free in normal applications, and positioned immediately following the probe and cyclone bypass (or cyclone, if used) with the active side of the filter perpendicular to the source of the flow. The filter support must be evaluated for PFAS contamination.

6.1.4 Filter Heating System. Use any heating system capable of monitoring and maintaining the temperature around the filter to ensure that the sample gas temperature exiting the filter is 120 ± 14 °C (248 ± 25 °F) during sampling or such other temperature as specified by an applicable subpart of the standards or approved by the Administrator for a particular application.

6.1.5 Filter Temperature Sensor. Install a temperature sensor capable of measuring temperature to within ± 3 °C (5.4 °F) so that the sensing tip protrudes at least 1.3 centimeters (cm) (1/2 in.) into the sample gas exiting the filter. Encase the sensing tip of the sensor in glass or PTFE if needed.

6.1.6 Sample Transfer Line. The sample transfer line transports gaseous emissions from the heated filter holder to the condenser and must be heat traced and constructed of glass or PTFE with connecting fittings that form leak-free, vacuum-tight connections without using sealing greases or tapes. PFA tubing may also be used for the sample transfer line. Keep the sample transfer lines as short as possible and maintain the lines at a temperature of $120\text{ }^{\circ}\text{C} \pm 14\text{ }^{\circ}\text{C}$ ($248\text{ }^{\circ}\text{F} \pm 25\text{ }^{\circ}\text{F}$) using active heating when necessary. Orient the sample transfer lines with the downstream end lower than the upstream end so that any condensate will flow away from the filter and into the condenser.

Note: The use of a sample transfer line should be avoided if possible. If a sample transfer line is used, it must be evaluated for PFAS contamination.

6.1.7 Condenser. Glass, water-jacketed, coil-type with compatible fittings. Orient the condenser to cause moisture to flow down to the adsorbent module to facilitate condensate drainage. Figure OTM 45-2 of this method shows a schematic diagram of the condenser.

6.1.8 Water Circulating Bath. Use a bath pump circulating system capable of providing chilled water flow to the condenser and adsorbent module water jackets. Typically, a submersible pump is placed in the impinger ice water bath to circulate the ice water contained in the bath. Verify the function of this system by measuring the gas temperature at the entrance to the adsorbent module. Maintain this temperature at $< 20\text{ }^{\circ}\text{C}$ ($68\text{ }^{\circ}\text{F}$).

6.1.9 Primary and Secondary Adsorbent Module. Use a water-jacketed glass container to hold up to 40 grams (g) of the solid adsorbent. Figure OTM 45-2 of this method shows a schematic diagram of the adsorbent module. Other physical configurations of the adsorbent resin

module/condenser assembly are acceptable if the configuration contains the requisite amount of solid adsorbent and maintains the minimum length-to-width adsorbent bed ratio of two-to-one. Orient the adsorbent module vertically to facilitate condensate drainage. The connecting fittings must form leak-free, vacuum-tight seals. Include a coarse glass frit in the adsorbent module to retain the adsorbent.

6.1.10 Impingers. Use four or five impingers connected in series with leak-free ground glass fittings or any similar leak-free noncontaminating fittings. The first impinger is an optional water knockout impinger and must be a short-stem (water-dropout) design or equivalent. The second, fourth, and fifth impingers must be of the Greenburg-Smith design, modified by replacing the tip with a 1.3 cm (1/2 in.) inside diameter (ID) glass tube extending to approximately 1.3 cm (1/2 in.) from the bottom of the flask. The third impinger must be of the Greenburg-Smith design with the standard tip. The second, third, and fourth impingers must contain known quantities of water, the fifth impinger is used to contain a known weight of silica gel or equivalent desiccant.

6.2 Sample Recovery Equipment.

6.2.1 Fitting Caps. Use leak-free ground glass fittings or any similar leak-free non-contaminating fitting to cap the sections of the sampling train exposed to the sample gas.

6.2.2 Wash Bottles. Use high density polyethylene (HDPE) bottles.

6.2.3 Probe-Liner, Probe-Nozzle, and Filter-Holder Brushes. Use inert bristle brushes with precleaned stainless steel handles. Extensions of the probe brush must be made of stainless steel and be at least as long as the probe. Use brushes that are properly sized and shaped to remove accumulated material from the nozzle and probe liner if used.

6.2.4 Filter Storage Container. Use a sealed filter holder, wide-mouth amber glass jar with HDPE-lined cap, or glass petri dish sealed with HDPE tape or encased in a resealable polyethylene bag. Purchase precleaned amber glass jars and petri dishes or clean according to the glassware cleaning procedures listed in Section 8.1.1.1 of this method.

6.2.5 Field Balance. Use a weighing device capable of measurements to an accuracy of 0.5g.

6.2.6 Aluminum Foil. Use heavy duty aluminum foil cleaned by rinsing three times with 5% ammonium hydroxide in methanol and stored in a pre-cleaned glass petri dish or glass jar. Do not use aluminum foil to wrap or contact filter samples due to the possibility of reaction between the sample and the aluminum.

6.2.7 Silica Gel Storage Containers. Use an air-tight container to store silica gel.

6.2.8 Sample Storage Containers. Recover samples in high density polyethylene (HDPE) bottles, 125, 250, 500- or 1000-milliliters (mL) with leak-free polyethylene-lined caps. Either purchase precleaned bottles or clean containers according to glassware cleaning procedures listed in Section 8.1.1.1 of this method.

6.3 Sample Extraction Equipment.

6.3.1 Sample Containers. Use 125- and 250-mL HDPE bottles with polypropylene or polyethylene-lined caps.

6.3.2 Test Tubes. Use polypropylene test tubes or small (e.g., 5 to 15 mL) polypropylene vials with polypropylene screw caps.

6.3.3 Nitrogen Evaporative Concentrator. Use a nitrogen evaporative concentrator equipped with a water bath with the temperature controlled in the range of 30 to 60 °C (86 to 140 °F) (N-Evap Organomation Associates, Inc., South Berlin, MA, or equivalent).

6.3.4 Shaker table. Use a shaker table (Eberbach Shaker Model E6013, Eberbach Corporation, Belleville, MI, or equivalent) capable of holding samples securely and operating uninterrupted for at least 18 hours.

6.3.5 Filter paper (0.45 µm pore size). Filter extract from “front half” probe filter before concentration step.

6.3.6 Digestion block (“hot block”). Use a digestion block capable of reaching 55-60 °C and securely holding digestion vessels. Used for concentration of extracts.

6.3.7 Digestion Vessels. Use polypropylene digestion vessels capable of holding 70 mL.

6.3.8 Watch Glass. Use ribbed polypropylene watch glass capable of covering digestion vessels.

6.4.1 Pasteur Pipettes. Use disposable pipettes, or glass serological pipettes typically 150 mm long x 6 mm ID.

6.5 Solid Phase Extraction (SPE)

6.5.1 SPE Cartridge. (Waters OASIS WAX 500mg/6 cc, Waters Corporation, Milford, MA or equivalent) SPE cartridges containing weak anion exchange, mixed-mode polymeric sorbent (polymeric backbone and a diamino ligand), particle size approximately 33 µm. The SPE sorbent

must have a pKa above 8 so that it remains positively charged during extraction. SPE cartridges containing 500 mg sorbent.

6.5.1.1 SPE Cartridge Interferences. Solid phase extraction cartridges may be a source of interferences. The analysis of LSMBs provides important information regarding the presence or absence of such interferences. Each brand and lot of SPE devices must be tested as part of the LSMB to ensure that contamination does not preclude analyte identification and quantitation. SPE cartridges should be sealed while in storage to prevent ambient contamination of the SPE sorbent.

6.5.2 Vacuum Extraction Manifold. Equipped with flow and vacuum control [Supelco Cat. No. 57030-U, UCT Cat. No. VMF016GL (the latter requires UCT Cat. No. VMF02116 control valves), or equivalent systems]. Automated devices designed for use with SPE cartridges may be used; however, all extraction and elution steps must be the same as in the manual procedure. Care must be taken with automated SPE systems to ensure that Teflon tubing and other PTFE components commonly used in these systems, do not contribute to unacceptable analyte concentrations in LSMBs.

6.5.3 Sample Delivery System. Use of large volume sampling lines, constructed with polyethylene tubing, are recommended, but not mandatory. Large volume sample transfer lines, constructed with PTFE tubing, are commercially available for standard extraction manifolds (Supelco Cat. No. 57275 or equivalent). The PTFE tubing can be replaced with 1/8" o.d. x 1/16" i.d. polyethylene tubing [Freelin-Wade (McMinnville, Oregon) LLDPE or equivalent] cut to an appropriate length. This prevents potential contamination from PTFE transfer lines. Other types of non-PTFE tubing may be used provided it meets the LSMB and LFMB QC requirements.

PTFE tubing may be used, but an LSMB must be run on each individual transfer line and the QC requirements in Section 9.2.2.1 must be met. In the case of automated SPE, the removal of PTFE lines may not be feasible; therefore, acceptable performance for the LSMB must be met for each port during the IDC (Section 9.2.1). LSMBs must be rotated among the ports during routine analyses thereafter. Plastic reservoirs are difficult to rinse during elution and their use may lead to lower recovery.

6.6 LC-MS/MS System

6.6.1 LC System. The LC system must provide consistent sample injection volumes and be capable of performing binary linear gradients at a constant flow rate.

Note: On some LC systems, PFAS may build up in PTFE transfer lines when the system is idle for more than one day. To prevent long delays in purging high levels of PFAS from the LC solvent lines, it may be useful to replace PTFE tubing with polyetheretherketone (PEEK) tubing and the PTFE solvent frits with stainless steel frits. These modifications were not used on the LC system used for method development. However, a delay column, HLB Direct Connect 2.1 x 30 mm (Waters 186005231), was placed in the mobile phase flow path immediately before the injection valve. This direct connect column may have reduced the co-elution of PFAS originating from sources prior to the sample loop from the PFAS injected in the sample. It may not be possible to remove all PFAS background contamination.

6.6.2 Chromatography Column. C18 liquid chromatography column (2 x 50 mm) packed with 3 µm C18 solid phase particles (Phenomenex Part Number 00B-4439-B0 or equivalent).

6.6.3 Electrospray Ionization Tandem Mass Spectrometer (ESI-MS/MS). The mass spectrometer must be capable of electrospray ionization in the negative ion mode. The system must be capable of performing MS/MS to produce unique product ions for the method analytes within specified retention time segments. A minimum of 10 scans across the chromatographic peak is needed to ensure adequate precision. Some ESI-MS/MS instruments may not be suitable for PFAS analysis. For this method, the m/z 80 product ion must be used for PFOS and PFHxS to minimize this problem and promote comparability between laboratories. Some MS/MS instruments may not be able to scan a product ion with such a wide mass difference from the precursor ion. These instruments may not be used for this method if PFOS or PFHxS analysis is to be conducted.

6.6.4 MS/MS Data System. An interfaced data system is required to acquire, store, and output MS data. The computer software must have the capability of processing stored data by recognizing a chromatographic peak within a given retention time window. The software must allow integration of the abundance of any specific ion between specified time or scan number limits. The software must be able to reproducibly integrate analyte and internal standard ion abundances in order to construct calibration curves and calculate analyte concentrations using the internal standard technique.

7.0 Reagents, Media, and Standards

Unless otherwise indicated, all reagents must conform to the *Specifications and Procedures for Reagents and Standard-Grade Reference Materials* (see <https://pubs.acs.org/isbn/9780841230460>) of the American Chemical Society (ACS) Committee on Analytical Reagents where such specifications are available. Other grades may be used if the

reagent is demonstrated to be free of analytes and interferences and all requirements in Section 13 are met when using these reagents media and standards.

7.1 Sampling Media

7.1.1 Filter. Glass fiber filters, without organic binder, exhibiting at least 99.95 percent efficiency (<0.05 percent penetration) on 0.3-micron dioctyl phthalate smoke particles.

7.1.1.1 Filter quality control check. Conduct a filter lot blank evaluation prior to the field test to demonstrate that filters are free from contamination or interference. Perform extraction and analysis using the same procedures used to process field samples as outlined in Section 11 of this method on a minimum of three filters from the lot. The blank filter check analysis must meet the performance requirements in Section 9.2.2.1 of this method.

7.1.2 Adsorbent Resin. Amberlite® XAD-2 resin. All adsorbent resin must meet the cleanliness criteria in Section 9.2.2.1 of this method for all target compounds on the analysis list following the same extraction, concentration, cleanup, and analysis steps as field samples. This method recommends using the procedures provided in the Appendix to this method to clean the resin before use, if needed. However, this method allows alternative cleanup procedures that use automated extraction equipment if the adsorbent meets the required performance criteria in Section 9.2.2.1 of this method.

7.1.3 Conduct a quality control check on the cleaned adsorbent. Perform extraction and analysis using the same procedures used to process field samples as outlined in Section 11 of this method on a quantity of the sorbent representative of the amount typically packed into a sampling

module. The cleaned adsorbent must meet the criteria in Section 9.2.2.1 of this method. A batch blank conducted on a filter and adsorbent lot combination used for a test can serve this purpose.

7.1.3.1 Storage. Store adsorbent in its original purchase container, a clean wide-mouth HDPE container with a polypropylene or polyethylene-lined cap, or in glass adsorbent modules tightly sealed with glass caps.

7.1.4 Glass Wool. Clean the glass wool to meet the specifications in Section 9.2.2.1 of this method. Using sequential immersion in three clean aliquots of 5% ammonium hydroxide in methanol, drying in a 110 °C (230 °F) oven, and storing in a 5% ammonium hydroxide in methanol rinsed glass jar with a polypropylene or polyethylene-lined screw cap can meet these requirements.

7.1.5 Water. Use deionized or distilled PFAS free water meeting requirements in Section 9.2.2.1 of this method and store in its original container or in a 5% ammonium hydroxide in methanol-rinsed glass container with a polypropylene or polyethylene-lined screw cap.

7.1.6 Silica Gel. Indicating type, 6–16 mesh. If previously used, dry at 175 °C (347 °F) for two hours. Use new silica gel as received. As an alternative, use other types of desiccants (equivalent or better), subject to the approval of the Administrator.

7.2 Sample Recovery Reagents.

7.2.1 Methanol. CH₃OH, CASRN 67-56-1, LC grade (Fisher Scientific, Cat. No. A456 or equivalent).

7.2.2 Ammonium hydroxide solution. NH_4OH , CASRN 1336-21-6, approximately 56.6% in water as ammonium hydroxide (w/w), approximately 28% in water as ammonia, approximately 14.5 N (Fisher Scientific, Cat. No. A669, Certified ACS Plus grade, or equivalent).

7.2.3 Ammonium hydroxide, 5% in methanol rinsing solution. Prepared by diluting of 50 mL ammonium hydroxide solution into 1 L final volume in methanol.

7.3 Sample Extraction and Cleanup Reagents.

7.3.1 Ammonium hydroxide 5% in methanol extraction solution. (methanol / 5% NH_4OH).

Prepared by diluting of 50 mL ammonium hydroxide solution into 1 L final volume in methanol (see Section 7.2). This reagent is added to methanolic solutions of PFAS to prevent esterification.

7.3.2 Ammonium hydroxide, 0.3%, in methanol (SPE extraction solution). Prepared by diluting 12 mL of ammonium hydroxide into 4L of methanol.

7.3.3 Sodium Hydroxide (NaOH), 0.1 N, in water. Prepared by diluting 400 mL of 1N NaOH in 3.6L of water for a total volume of 4L.

7.3.4 Hexane. Reagent grade.

7.4 Standard Solutions

7.4.1 Stability of Methanolic Solutions

Fluorinated carboxylic acids will esterify in anhydrous acidic methanol. To prevent esterification, standards must be stored under basic conditions. If base is not already present, this may be accomplished by the addition of ammonium hydroxide (approximately 4 mole

equivalents) when standards are diluted in methanol. When calculating molarity for solutions containing multiple PFAS, the molecular weight can be estimated as 250 atomic mass units (amu). It is necessary to include ammonium hydroxide in solutions of both isotopically labeled and native analytes. The amount of ammonium hydroxide needed may be calculated using the following equation:

$$\frac{\text{Total PFAS mass}(g) \times 160\left(\frac{g}{mol}\right)}{250\left(\frac{g}{mol}\right)} = \text{Mass of } NH_4OH \text{ Required } (g) \quad \text{Eq. 45-1}$$

7.4.2 Preparation of Standards

When a compound purity is assayed to be 96% or greater, the weight can be used without correction to calculate the concentration of the stock standard. PFAS analyte and isotopically labeled analogues commercially purchased in glass ampoules are acceptable; however, all subsequent transfers or dilutions performed by the analyst must be stored in polypropylene containers.

Solution concentrations listed in this section are included as examples. Alternate concentrations may be used as necessary depending on instrument sensitivity and the calibration range used. Standards for sample fortification should be prepared in the smallest volume that can be accurately measured to minimize the addition of excess organic solvent to aqueous samples. The analyte supplier's guidelines are used to determine when standards need to be replaced.

7.4.3 Storage Temperatures for Standard Solutions. Store stock standards at less than 4 °C unless the vendor recommends otherwise. The Primary Dilution Standards (PDS) may be stored at any

temperature, but cold storage is recommended to prevent solvent evaporation. PDS stored at -20°C showed no change in analyte concentrations over a period of 6 months.

7.4.4 Pre-analysis standard(s). Obtain the pre-analysis standard(s) as certified standard solutions, if available, or as the neat compounds. Note that Chemical Abstracts Registry Numbers are not currently available for these compounds. All the pre-analysis standard(s) listed in this section must be used. Additional pre-analysis standard(s) may be used provided they are isotopically labeled analytes or labeled analytes with similar functional groups as the method analytes. Linear isomers are recommended to simplify peak integration. Method modification QC requirements must be met (Sect. 9.3) whenever additional pre-analysis standard(s) are used. The final sample extracts were fortified with $10\ \mu\text{L}$ of the pre-analysis standard(s) to yield a concentration of $10\ \text{ng/mL}$ for $^{13}\text{C}_3\text{-PFBA}$ and $^{13}\text{C}_2\text{-PFOA}$.

7.4.5 Pre-analysis primary dilution standard (pre-analysis standard(s)-PDS). Prepare the pre-analysis standard(s)-PDS in methanol and add ammonium hydroxide if not already present to prevent esterification. The PDS concentrations for the method are listed in the Table 45-3

7.4.6 Pre-extraction Standards. Obtain the isotopically labeled analogues listed in Table 45-1 as individual certified standard solutions or as certified standard mixes. All listed isotope dilution analogues must be used, if available. Linear isomers are recommended to simplify peak integration.

Note: Chemical Abstracts Registry Numbers are not currently available for these isotopically labeled analogues.

Note: Pre-extraction standard(s) is identical to Isotope Dilution Analogues in Method 533

7.4.7 Pre-extraction Primary Dilution Standard (pre-extraction standards -PDS). Prepare the pre-extraction standards - PDS in methanol and add ammonium hydroxide if not already present to prevent esterification as described in Section 7.4.1. The final extracts are fortified with 10 µL of the PDS to yield concentrations of 40 – 160 ng/sample fraction. Note that the concentrations of sulfonates in the pre-extraction standards PDS is based on the weight of the salt. It is not necessary to account for difference in the formula weight of the salt compared to the free acid for sample quantitation.

7.4.8 Analyte Standard Materials. Analyte standards may be purchased as certified standard solutions or prepared from neat materials of assayed purity. If available, the method analytes should be purchased as technical-grade to ensure that linear and branched isomers are represented. Standards or neat materials that contain only the linear isomer can be substituted if technical-grade analytes are not available as quantitative standards. Stock standards are made by dilution in methanol containing 4 mole equivalents of ammonium hydroxide, as described in Section 7.4.2. Technical grade, quantitative PFHxS and PFOS standards containing branched and linear isomers must be used when available.

7.4.9 Correction for Analytes Obtained in the Salt Form. This method measures all forms of the analytes as anions while the identity of the counterion is inconsequential. Analytes may be commercially available as neat materials or as certified stock standards as their corresponding ammonium, sodium, or potassium salts. These salts are acceptable standards provided the measured mass, or concentration, is corrected for the salt content. The equation for this correction is provided below.

$$mass(acid\ form) = mass(salt\ form) \times \frac{MW\ acid}{MW\ salt} \quad \text{Eq. 45-2}$$

7.4.10 Analyte PDS. The analyte PDS is used to prepare the calibration standards and to fortify the laboratory fortified blanks, laboratory fortified sample matrix spikes, laboratory fortified sample matrix spike duplicates (LFMB s, LFSMs and LFSMDs) with the method analytes. Prepare the analyte PDS by combining and diluting the analyte stock standards in 100% methanol and add ammonium hydroxide if not already present to prevent esterification. Select nominal analyte concentrations for the PDS such that between 5 and 100 μL of the PDS is used to fortify samples and prepare standard solutions. More than one PDS concentration may be necessary to meet this requirement. Nominal analyte PDS may be prepared at an identical concentration for all analytes of 0.5 ng/ μL . The user may modify the concentrations of the individual analytes based on the confirmed QRLs and the desired measurement range. Premixed standards containing most of the target analytes in this method may be used as a PDS. For those compounds not available in this mixture, a second source standard from the same vendor as the ICAL may be used to complete the target list compounds in the PDS. If the PDS is stored cold, warm the vials to room temperature and vortex mix prior to use.

7.4.11 Calibration Standards. Prepare a series of calibration standards of at least five levels by diluting the analyte PDS into methanol containing 20% reagent water. The lowest calibration standard must be at or below the QRL for each analyte. The calibration standards may also be used as Continuing Calibration Checks (CCCs). Using the PDS solutions, add a constant amount of the pre-analysis standard(s) and to each calibration standard. The concentration of the pre-extraction standards should match the concentration of the analogues in sample extracts, assuming 100% recovery through the extraction process. The nominal concentrations of the pre-extraction standards can be 40 ng/mL in the extract for 4:2FTS, 6:2FTS and 8:2FTS, and 10

ng/mL for all others. The analyte calibration ranges from approximately 0.50 ng/mL to 25 ng/mL extract concentration.

7.5 Nitrogen. 99.999 percent (ultra-high) purity used to concentrate sample extracts.

7.6 Argon. Used as collision gas in MS/MS instruments. Argon should meet or exceed instrument manufacturer's specifications. Nitrogen may be used as the collision gas if recommended by the instrument manufacturer.

8.0 Sample Collection, Recovery, Preservation, and Storage

8.1 Sampling. This method involves collection and recovery of trace concentrations of semivolatile organic compounds. Therefore, field sampling and recovery staff must be trained in the best practices for handling and using organic solvents in field environments to recover and protect samples from contamination.

8.1.1 Pretest Preparation.

8.1.1.1 Cleaning Glassware. Clean glassware thoroughly before using. This section provides a recommended procedure, but any protocol that consistently results in contamination-free glassware meeting the batch blank criteria in Section 13.2 of this method is acceptable.

8.1.1.1.1 Soak all glassware in hot soapy water (Alconox® or equivalent) at 50 °C or higher.

8.1.1.1.2 Rinse three times with hot tap water.

8.1.1.1.3 Rinse three times with deionized/distilled water.

8.1.1.1.4 Rinse three times each with Acetone, dichloromethane, and methanol.

8.1.1.1.5 Bake glassware at 300 °C (572 °F) for a minimum of 2 hours

Note: Repeated baking of glassware may cause active sites on the glass surface that may irreversibly absorb target compounds.

8.1.1.1.6 Cover glassware openings with clean glass fitting caps or cleaned aluminum (see Section 6.2.6 of this method).

8.1.1.1.7 Rinse glassware immediately before use with 5% ammonium hydroxide in methanol.

Note: To prepare heavily soiled glassware, remove surface residuals from the glassware by soaking in hot soapy water, rinsing with hot water, then soaking with a non-chromic acid oxidizing cleaning reagent in a strong acid (e.g., NOCHROMIX® prepared according to manufacturer's directions). After the acid soak, rinse with hot water and repeat the cleaning procedures in Section 8.1.1.1 of this method.

8.1.1.2 Adsorbent Module. Load the modules in a clean area to avoid contamination. Spike modules before the sampling event, but do not spike the modules in the field. Fill a module with 20 to 40 g of XAD-2. Add the pre-sampling standard spike to the top quarter of the adsorbent bed. Add sufficient spike (picograms (pg)/module) to result in the final theoretical concentrations specified in Table 45-3 of this method. For samples with known or anticipated target compound concentration significantly higher or lower than the specified amount in these tables, add a pre-sampling spike amount appropriate to the expected native compound concentration, but no less

than 10 times the detection limit. Follow the XAD-2 with cleaned glass wool and tightly cap both ends of the module.

8.1.1.3 Sampling Train. Figure OTM 45-1 of this method shows the complete sampling train.

8.1.1.4 Silica Gel. Weigh several 200 to 300 g portions of silica gel in an air-tight container to the nearest 0.5 g. Record the total weight of the silica gel plus container on the outside of each container. As an alternative, directly weigh the silica gel in its impinger or sampling holder just prior to sampling.

8.1.1.5 Filter. Check each filter against light for irregularities and flaws or pinhole leaks. Pack the filters flat in a clean glass container. Do not mark filters with ink or any other contaminating substance.

8.1.2 Preliminary Determinations. Use the procedures specified in Section 8.2 of Method 5 of appendix A-3 to 40 CFR part 60.

8.1.2.1 Sample Volume. This method recommends sampling enough gas volume to reach a DL sufficient to meet test objectives. Unless otherwise specified in an applicable rule, permit, or other requirement, collect a minimum of 3.0 dry standard cubic meters of source gas.

8.1.2.2 For continuously operating processes, use the same sampling time at each traverse point. To avoid timekeeping errors, use an integer, or an integer plus one-half minute, for each traverse point.

8.1.2.3 For batch processes, determine the minimum operating cycle duration, dividing the sampling time evenly between the required numbers of traverse points. After sampling all

traverse points once, sample each point again for the same duration of time per sampling point in reverse order until the operating cycle is completed. Sample all traverse points at least once during each test run.

8.1.3 Preparation of Sampling Train. Do not use sealing greases or brominated flame retardant-coated tape in assembling the train.

8.1.3.1 During field preparation and assembly of the sampling train, keep all train openings sealed where contamination can enter until just prior to assembly or until sampling is about to begin. To protect the adsorbent module from radiant heat and sunlight, you must wrap the module with aluminum foil or other suitable material capable of shielding the module from light. The XAD-2 adsorbent resin temperature must never exceed 50 °C (122 °F) because thermal decomposition will occur. Clean and prepare a complete set of sampling train components that will contact the sample for each sampling run. Include at least one complete field test proof blank and at least one field test field blank, as described in Sections 9.1.3 and 9.1.4 of this method.

8.1.3.2 Place approximately 100 mL of water in each of the second, third and fourth impingers but leave the first (condensate trap) impinger empty. Transfer approximately 200 g or more of silica gel from its container to the fifth impinger. Weigh each impinger and the adsorbent module, including the fitting caps, to the nearest 0.5 g using the field balance and record the weight for moisture determination. Remove the aluminum foil from the adsorbent module before weighing. Keep the module out of direct sunlight and rewrap the module with foil immediately after recording the module weight.

8.1.3.3 Using tweezers or clean disposable surgical gloves, place a filter in the filter holder. Be sure that the filter is properly centered, and the gasket properly placed, to prevent the sample gas stream from circumventing the filter. Check the filter for tears after completing the assembly.

8.1.3.4 Prepare the inside of the sampling probe and nozzle by brushing each component while rinsing three times each with methanol. Install the selected nozzle. You may use connecting systems described in Section 6.1.2 of this method. Mark the probe with heat resistant tape or by some other method to denote the proper distance into the stack or duct for each sampling point. Assemble the train as shown in Figure 45–1 of this method. Orient the adsorbent module vertically so condensed moisture drains into the first impinger. See APTD-0576 Maintenance, Calibration, and Operation of Isokinetic Source-sampling Equipment (U.S. EPA 1972) for details.

8.1.3.5 Turn on the recirculation pump to the adsorbent module and condenser coil and begin monitoring the temperature of the gas entering the primary adsorbent module. Ensure proper temperature of the gas entering the adsorbent module before proceeding.

8.1.4 Leak-Check Procedure. Same as Section 8.4 of Method 5 of appendix A-3 to 40 CFR part 60.

8.1.5 Sampling Train Operation. Same as Sections 8.5.1 through 8.5.9 of Method 5 of appendix A-3 to 40 CFR part 60 with the exception that the probe and filter holder (and heated sample transfer line, if used) temperature are limited to minimize the potential thermal degradation of thermally labile PFAS compounds such as HFPO-DA which are known to decompose at temperatures below the standard Method 5 probe and filter operating temperatures.

8.1.5.1 Probe and Filter Operating Temperatures. For stack temperatures below 120 °C (248 °F), limit the probe and filter (and heated sample transfer line, if used) temperature to approximately 10 °C (20 °F) above the sampling location stack temperature. Ensure the operating temperature is sufficient to avoid moisture condensation in the probe and filter holder. For stack temperatures at or above 120 °C, operate the probe and filter at 120 °C ± 14 °C (248 °F ± 25 °F). Monitor the filter temperature sensor and record the filter temperature during sampling. A nominal filter exit temperature of 120 °C ± 14 °C (248 °F ± 25 °F) should not be exceeded.

8.1.5.2 XAD-2 Adsorbent Module Temperatures. During testing, you must record the temperature of the gas entering the XAD-2 adsorbent modules. The temperature of the gas must not exceed 20 °C (68 °F) for efficient capture of the target compounds.

8.2 Sample Recovery. Begin the cleanup procedure as soon as the probe is removed from the stack at the end of the sampling period.

8.2.1 Preparation. Allow the probe to cool. Do not cap the probe tip tightly while the sampling train is cooling down because this will create a vacuum in the filter holder, drawing water from the impingers into the sorbent module. When the probe can be safely handled, wipe off all external particulate matter near the tip of the probe. Conduct a post-test leak check. Remove the probe from the train and close off both ends. Seal off the inlet to the filter. Remove the umbilical cord from the last impinger and cap the impinger. If a flexible line is used between the primary sorbent module and the filter holder, disconnect the line at the filter holder and let any condensed water or liquid drain into the organic module. Cap the filter-holder outlet and the inlet to the organic module. Separate the sorbent trap section of the organic module from the condensate knockout trap and the gas-conditioning section. Cap all sorbent module openings. Disconnect the

sorbent module knockout trap from the impinger train inlet and cap both of these openings. Ground-glass stoppers, Teflon tape, or other inert materials such as cleaned aluminum foil (e.g., rinsed with 5% ammonium hydroxide in methanol rinsing solution) may be used to seal all openings.

8.2.2 Transfer and Inspection. Transfer the sampling train components to the cleanup area. This method recommends cleaning and enclosing this area to minimize the chances of losing or contaminating the sample. To avoid sample contamination and unnecessary exposure to toxic chemicals, smoking or eating in the sample recovery area shall not be allowed. Inspect the train prior to and during disassembly. Note and record any abnormal conditions (e.g., broken filters, colored impinger liquid). Recover and prepare samples for shipping as follows in Sections 8.2.4 through 8.2.12 of this method.

8.2.3 Moisture Weight. Weigh the adsorbent module, impingers, and silica gel impinger to within ± 0.5 g using the field balance and record the weights. This information is required to calculate the moisture content of the effluent gas.

Note: Moisture measurement in the field using the OTM 45 train requires weighing the primary adsorbent module before the sampling run described in 8.1.3.2 and after sampling as part of the sample recovery for stack moisture determination.

8.2.4 Container No. 1 – Filter. Either seal the filter holder or carefully remove the filter from the filter holder and place it in its identified container. If it is necessary to remove the filter, use a pair of cleaned tweezers to handle the filter. If necessary, fold the filter such that the particulate cake is inside the fold. Carefully transfer to the container any particulate matter and filter fibers

that adhere to the filter holder gasket by using a dry inert bristle brush and a sharp-edged blade. Seal the container and store in a thermally insulated container for transport to the laboratory.

8.2.5 Container No. 2 – Front Half Rinse. Quantitatively recover material deposited in the nozzle, the front half of the filter holder, and the cyclone, if used, by brushing while rinsing three times with the 5% Ammonium hydroxide in methanol rinsing solution. Collect all the rinses in the HDPE sample bottle and label as Container No. 2. Mark the level of the liquid on the container. Store the sample container refrigerated or on ice until laboratory shipment.

8.2.6 Container No. 3 – Primary Adsorbent Module Sample. Remove the module from the train and tightly cover both ends. Replace the retaining clips around the glass joint. Remove the foil, drain the recirculating water from the module, weigh and record the module weight. The adsorbent trap module should be used as a sample transport container. Both ends should be sealed with tightly fitting ground-glass stoppers followed by Teflon tape around the glass joint. The sorbent trap should then be labeled, re-covered with aluminum foil, and packaged on ice for transport to the laboratory.

Note: The XAD-2 resin modules (primary and breakthrough) are shipped back from the field as separate fractions for analysis. As more data is collected from the use of this method the requirement of analysis of the XAD-2 module as a separate fraction may change.

8.2.7 Container No. 4 – Back Half Rinse. All sampling train components located between the back half of the filter holder and the inlet of the primary adsorbent module, including the condenser if a separate condenser and adsorbent module are used, and the heated sample transfer line connecting the filter outlet to the condenser (if used) shall be triple rinsed with 5%

ammonium hydroxide in methanol rinsing solution. Collect all the rinses in the HDPE sample bottle and label as Container No. 4. Mark the level of the liquid on the container. Store the sample container refrigerated or on ice until laboratory shipment.

8.2.9 Container No. 5 – Condensate and Impinger Water. After weighing the impingers, quantitatively recover the impinger water samples, including the contents of the knockout impinger (if used), in the HDPE sample bottle and label as Container No. 5. Mark the level of the liquid on the container. Store the sample container refrigerated or on ice until laboratory shipment.

Note: Make sure that no ammonium hydroxide in methanol rinsing solution is transferred to the container No. 5 sample. Doing so may compromise the sample for analysis.

8.2.10 Container 6 – Impingers Rinse. Rinse impingers 1-4 three times with the 5% ammonium hydroxide in methanol rinsing solution. Collect all the rinses in the HDPE sample bottle and label as Container No. 6. Mark the level of the liquid on the container. Store the sample container refrigerated or on ice until laboratory shipment. If impingers are used in a subsequent sampling run they must be rinsed three times with reagent water to remove residual ammonium hydroxide in methanol.

8.2.11 Container 7 – Secondary Adsorbent Module Sample. Remove the module from the train and tightly cover both ends. Replace the retaining clips around the glass joint. Remove the foil, drain the recirculating water from the module, weigh and record the module weight. The adsorbent trap module should be used as a sample transport container. Both ends should be sealed with tightly fitting ground-glass stoppers followed by Teflon tape around the glass joint.

The sorbent trap should then be labeled, re-covered with aluminum foil, and packaged on ice for transport to the laboratory.

8.2.10 Silica Gel. Note the color of the indicating silica gel to determine if it has been completely spent and report its condition on the field data sheet.

8.2.11 Field Sample Handling, Preservation, Storage, and Transport. Store all field samples, with the exception of the particulate filter, temporarily on ice (approximately 4°C) and dark conditions prior to transport to the laboratory. The particulate filters should be shipped in a thermally insulated container and may remain unrefrigerated. Ship samples on ice, shielded from ultraviolet light. The particulate filters should be shipped unrefrigerated. In addition, follow the procedures in ASTM D6911-15 (Guide for Packaging and Shipping Environmental Samples for Laboratory Analysis) for all samples, where appropriate. To avoid contamination of the samples, pay special attention to cleanliness during transport, field handling, sampling, recovery, and laboratory analysis, as well as during preparation of the adsorbent cartridges.

Prior to shipment, recheck all sample containers to ensure that the caps are well secured. Seal the lids of all containers around the circumference with Teflon tape. Ship all liquid samples upright on ice and all particulate filters with the particulate catch facing upward.

8.2.12 Sample Custody. Proper procedures and documentation for sample chain of custody are critical to ensuring data integrity. Follow the chain of custody procedures in ASTM D4840-99 (2018)e1 (Standard Guide for Sampling Chain-of-Custody Procedures) for all samples (including field samples and blanks).

8.3 Laboratory Sample Storage Conditions and Laboratory Hold Times.

8.3.1 Table 45-4 of this method summarizes the sample storage conditions and laboratory hold times.

8.3.2 Store sampling train rinses and filter samples in the dark at 6 °C (43 °F) or less from the time the laboratory receives the samples until extraction but not longer than 28 days from date of collection

8.3.3 You may store adsorbent samples prior to extraction in the dark at 6 °C (43 °F) or less for up to one year from the time the laboratory receives the samples.

8.3.4 Samples must be extracted within 28 days of collection. Extracts are stored at room temperature and must be analyzed within 28 days of extraction.

8.3.6 You may store archived extracted samples in the dark at refrigerator temperature of approximately 6 °C (43 °F) for up to one year.

9.0 Quality Control

It is the testing team's responsibility to establish the conditions for optimum sample collection, extraction, cleanup, and concentration to meet the performance criteria in this method. However, you may not change the fundamental procedures of isokinetic sampling with an adsorbent collection media followed by sample extraction, and LC-MS/MS with isotopic dilution quantification. This method requires performing a media blank (i.e., batch blank) assessment to evaluate an individual laboratory's performance against the performance criteria in this method.

In recognition of advances that are occurring in sampling and analytical technology for PFAS measurement, and to allow the test team to overcome analyte sensitivity and matrix

interferences, this method allows certain options to increase sample collection volume and to improve separations and the quality of the analysis results for target analytes.

QC procedures include the Initial Demonstration of Capability (IDC) and ongoing quality control (QC) requirements. This section describes each QC parameter, its required frequency, and the performance criteria that must be met in order to satisfy method objectives. The analysis QC criteria discussed in the following sections are summarized in Table 45-5. These QC requirements are considered the minimum for an acceptable QC program. Laboratories are encouraged to institute additional QC practices to meet their specific needs. At a minimum, laboratories must evaluate changes within the alternatives allowed in this method using a media blank sample to re-demonstrate that the performance criteria are achieved.

9.1 Sampling Quality Control.

9.1.1 Sampling System. Same as Sections 8.4 and 9.2 of Method 5 of appendix A-3 to 40 CFR part 60.

9.1.2 Field Sample Media Blank (FSMB). Also called the field trip blank. The FSMB is intended to include and represent the sampling media (i.e., filter, XAD-2 adsorbent) and reagents (i.e., impinger contents, rinsing solvents) associated with the field sample collection and recovery, but is not actually used in the field. The FSMB is treated as a sample in all respects, including shipment to the sampling site, exposure to sampling site conditions, storage, and all analytical procedures. The purpose of the FSMB is to determine if method analytes or other interferences are introduced into the sample from shipping, storage, and the field environment. Analysis of the FSMBs should be compared to the background level criteria in 9.2.2.1. Failure to meet these

levels does not invalidate data. However, the measured target compound mass in each fraction will need to be reported and used to interpret sample results.

9.1.3 Sample Train Proof Blank (STPB). A STPB must be submitted with the samples collected at each sampling site. At a minimum, conduct at least one sample train proof blank for each test series at a single facility. A sample train proof blank train consists of a fully assembled train at the sampling site using glassware that has been cleaned, but not yet used for sampling. Prepare and assemble the proof blank train in a manner identical to that described in Section 8.1.3 and 8.1.4 of this method. The STPB is taken to the sampling area, and leak checked at the beginning and end of the testing (or for the same total number of times as the actual test train). The filter housing and probe of the blank train will be heated during the sample test. No gaseous sample will be passed through the sampling train. Recover the proof blank train as an actual train described in Section 8.2 of this method. Follow all subsequent steps used for actual field train samples including sample preparation, analysis and data reporting. Table 45-5 of this method includes the performance criteria for field train proof blank assessment samples. Failure to meet these levels does not invalidate data. However, the measured target compound mass in each fraction will need to be reported and used to interpret sample results

9.1.4 Sample Train Field Blank. STFBs must be submitted with the samples collected at each sampling site. At a minimum, conduct at least one sample train field blank for each test series at a single facility. A sample train field blank train consists of a fully assembled train at the sampling site using glassware that has been previously used for sampling. Prepare and assemble the proof blank train in a manner identical to that described in Section 8.1.3 and 8.1.4 of this method. The STFB is taken to the sampling area, and leak checked at the beginning and end of

the testing (or for the same total number of times as the actual test train). The filter housing and probe of the blank train will be heated during the sample test. No gaseous sample will be passed through the sampling train. The train will be recovered as if it were an actual test sample.

Recover the proof blank train as an actual train described in Section 8.2 of this method. Follow all subsequent steps for blank train sample preparation and analysis used for field train samples including data reporting. Table 45-5 of this method includes the performance criteria for field train proof blank assessment samples. Failure to meet these levels does not invalidate data.

However, the measured target compound mass in each fraction will need to be reported and used to interpret sample results.

9.1.5 Pre-sampling Standard Recoveries. Pre-sampling standard XAD-2 adsorbent spike recoveries must demonstrate on a per sample basis that recovery of the labeled standards achieve the requirements in Table 45-5 of this method. Recoveries below the acceptable range of 70-130% for the pre-sampling standard spikes may require a root cause evaluation of the cause for poor recovery. If the recovery of all the pre-sampling standard adsorbent spikes is below 70%, but, greater than 50%, the results have not met the recoveries experienced during method development but may still be acceptable. Flag recoveries that are between 50 and 70% and describe their potential impact on results. If the pre-sampling standard recoveries are less than 50%, the data for that train are not considered valid.

Note: The Pre-sampling standard recoveries include the sum of the analytical results of Fractions 2 and 3. (The pre-sampling standard may have migrated from the XAD-2 to the impingers).

9.1.6 Secondary XAD-2 Breakthrough. Determine the relative breakthrough (BT) of PFAS through the OTM-45 train. For each PFAS target compound, calculate breakthrough.

$$BT(\%) = \frac{(Fraction\ 4\ mass)}{(Fracton\ 1+2+3\ mass)} \times 100\% \quad \text{Eq. 45-3}$$

BT should be less than 30%. For any BT greater than 10%, add the Fraction 4 mass to the total sample mass for emissions calculations. Failure to meet breakthrough requirements may invalidate reported results and require repeat sampling.

9.2 Analysis Quality Control

9.2.1 Initial Demonstration of Capability (IDC). The IDC must be successfully performed prior to analyzing field samples by meeting the QC requirements in Table 45-6. The IDC must be repeated if changes are made to analytical parameters not previously validated during the IDC. This may include, for example, changing the sample volume, selecting alternate quantitation ions, extending the calibration range, adding additional pre-analysis standard(s), or adding additional pre-extraction standards. Prior to conducting the IDC, the analyst must meet the calibration requirements outlined in Section 10. The same calibration range used during the IDC must be used for the analysis of field samples.

9.2.1.1 Perform initial calibration following the procedures in Section 10.4 The lowest calibration standard used to establish the initial calibration (as well as the low-level CCC) must be within two to ten times the estimated detection limit.

9.2.1.2 Demonstration of Low System Background. Analyze an LSMB immediately after injecting the highest calibration standard in the selected calibration range. Confirm that the blank is free from contamination as defined in Section 9.2.2.1. If an automated extraction system is used, an LSMB must be extracted on each port to fulfil this requirement.

9.2.1.3 Initial MDL Determination. Perform an MDL determination for each sample media fraction (filter, XAD-2, and impinger solution) following the requirements in 40 CFR Part 136 Appendix B. The MDL determination includes seven LSMB and seven LFMB that are prepared from blank media, spiked within 2 to 10 times of the expected MDL, and processed in a manner identical to field sample preparation. The MDL study establishes the lowest detectable concentrations for each sampling train fraction. Sample specific MDLs are reported inclusive of sample-specific dilutions, final volumes, aliquots, etc.

9.2.1.4 MDL Confirmation. Prepare a LSMB for each sampling media by spiking each media with native target compounds at the MDL and pre-extraction isotopic labeled standards at the concentration used to analyze field samples. Prepare and analyze the spiked LSMB and confirm target compounds meet the qualitative identification criteria in Section 12.3.2 of this method.

9.2.1.5 Demonstration of Precision. Prepare, extract, and analyze seven replicate LFMBs in a valid Extraction Batch (seven LFMBs and an LSMB). Fortify the LFMBs near the midpoint of the initial calibration curve. The percent relative standard deviation (%RSD) of the concentrations from the replicate analyses must be less than 20% for all method analytes.

9.2.1.6 Demonstration of Accuracy. Using the same set of replicate data generated for Section 9.2.1.5, calculate the average percent recovery. The average recovery for each analyte must be within a range of 70–130%.

9.2.1.7 Lowest Calibration Concentration Confirmation. Establish a target concentration for the lowest calibration standard based on the intended use of the method. The lowest calibration concentration may be established by a laboratory for their specific purpose or may be set by a

regulatory agency. If there is a regulatory or programmatic lowest quantitative reporting requirement, the laboratory calibration curve must be set at or below this level. In doing so, one should consider that establishing the lowest calibration concentration too low may cause repeated failure of ongoing QC requirements.

9.2.1.7.1 Prepare and Analyze LFMB Samples. Fortify, extract, and analyze seven replicate LFMBs at the proposed lowest calibration concentration.

9.2.1.7.2 Calculate Lowest Calibration Statistics. Calculate the mean and standard deviation for each analyte in these replicates. Determine the Half Range for the Prediction Interval of Results (HR_{PIR}) using the following equation:

$$HR_{PIR} = 3.963S \quad \text{Eq. 45-4}$$

Where,

S = the standard deviation

3.963 is a constant value for seven replicates¹

9.2.1.7.3 Calculate the Upper and Lower Limits for the Prediction Interval of Results ($PIR = \text{Mean} \pm HR_{PIR}$) as shown below. These equations are only defined for seven replicate samples.

$$U_{PIR} = \left[\frac{(\text{Mean} + HR_{pir})}{\text{Fortified Concentration}} \right] 100\% \quad \text{Eq. 45-5}$$

$$L_{PIR} = \left[\frac{(\text{Mean} - HR_{pir})}{\text{Fortified Concentration}} \right] 100\% \quad \text{Eq. 45-6}$$

9.2.1.7.4 Lowest Calibration Point Acceptance Criteria. The laboratory's ability to measure analyte concentrations down to the lowest calibration point is confirmed if the Upper PIR Limit is less than, or equal to, 150%; and the Lower PIR Limit is greater than, or equal to, 50%. If

these criteria are not met, the lowest calibration point has been set too low and must be confirmed again at a higher concentration.

9.2.1.8 Calibration Verification. Analyze a QCS (Sect. 9.2.2.12) to confirm the accuracy of the primary calibration standards.

9.2.2 Ongoing QC Requirements. This section describes the ongoing QC elements that must be included when processing and analyzing field samples.

9.2.2.1 Blanks and Background Levels. The quantitative measurement of various blanks and sampling media background levels is required. Unless otherwise stated, estimated quantitative measurement, and therefore acceptable blank and background levels for these samples should be at or below the MDLs established in Section 9.2.1.3. Ideally, estimated quantitative limits, and therefore acceptable blank and background levels, should be below the established MDLs, but no greater than the MDL.

9.2.2.1.1 Sampling Media Background Level Checks. When performing cleanliness checks prior to field sampling on the sampling media (i.e., filters, XAD-2, reagents, solvents, etc), acceptable levels should be below one third of the established MDLs, but no greater than the MDL. If levels are found above the MDL, further clean the sampling media until levels are below MDL.

9.2.2.1.2 Laboratory Sample Media Blank (LSMB). Analyze at least one LSMB during an analytical sequence or every 24 hours, whichever is shorter. Sampling media and reagents are fortified with the pre-extraction standards and processed identically to a field sample of the same media. LSMBs for each sampling media and reagent (i.e., filter, XAD-2, water, rinsing solutions) are included in each Extraction Batch to determine if the method analytes or other interferences

are introduced from the laboratory environment, the reagents, glassware, or extraction apparatus. Acceptable levels should be below the established MDLs, but no greater than the MDL. If method analytes are detected in the LSMB at concentrations greater than or equal to this level the results should be flagged accordingly. Resolve the source of contamination before proceeding with additional analyses.

9.2.2.2 Estimating Background Concentrations.

Although data below the lowest calibration concentration may not be accurate enough for quantitative data reporting, such data are useful in determining the magnitude of background interference. Therefore, the analyte concentrations in the LSMB may be estimated by extrapolation when results are below the lowest calibration concentration.

9.2.2.2.1 Influence of Background on Selection of MDLs. Because background contamination can be a significant problem, some MDLs may be background limited.

9.2.2.2.2 Evaluation of Background when Analytes Exceed the Calibration Range. After analysis of a sample in which method analytes exceed the calibration range, one or more LSMBs must be analyzed (to detect potential carryover) until the system meets the LSMB acceptance criteria. If this occurs during an automated sequence, examine the results of samples analyzed following the sample that exceeded the calibration range. If the analytes that exceeded the calibration range in the previous sample are detected at, or above, the MDL, these samples are invalid. If the affected analytes do not exceed the MDL, these subsequent samples may be reported.

9.2.2.3 Calibration Acceptance Criteria. Evaluate the initial calibration by calculating the concentration of each analyte as an unknown against its regression equation. (See Section 10.4)

For calibration levels, the result for each analyte should be within 90 – 110% of their true value. If these criteria cannot be met, the analyst could have difficulty meeting ongoing QC criteria. In this case, corrective action is recommended such as reanalyzing the calibration standards, restricting the range of calibration, or performing instrument maintenance. If the cause for failure to meet the criteria is due to contamination or standard degradation, prepare fresh calibration standards and repeat the initial calibration.

9.2.2.3.1 Continuing Calibration Check (CCC). Analyze CCC standards at the beginning of each Analysis Batch, after every tenth field sample, and at the end of the Analysis Batch. (See Section 10.5) CCCs must be within 70-130% of true value. If the CCC fails because concentration is >130% (150% for low-level CCC) and field sample extracts show no concentrations above the MDL for that analyte, non-detects can be reported without re-analysis.

9.2.2.4 Laboratory Fortified Media Blanks. Duplicate low level and high level LFMBs are required with each extraction batch for each media fraction (i.e., filter, XAD-2, water).

9.2.2.4.1 LFMB Concentration Requirements. Fortify the low concentration LFMB no more than two times the lowest calibration point. Fortify the high level LFMBs at a concentration between the mid and high-level calibration points.

9.2.2.4.2 Evaluate Analyte Recovery. Calculate the percent recovery (%R) using the equation:

$$\%R = \frac{(A-B)}{C} \times 100 \quad \text{Eq. 45-7}$$

Where,

A = measured concentration in the fortified sample,

B = measured concentration in the unfortified sample, and

C = fortification concentration.

Note: In order to obtain meaningful percent recovery results, correct the measured values in the LFSM and LFSMD for the native levels in the unfortified samples, even if the native values are less than the lowest calibration concentration.

Results for analytes fortified at concentrations near or at the lowest calibration point (within a factor of two times the lowest calibration concentration) must be within 50–150% of the true value. Results for analytes fortified at higher concentrations must be within 70–130% of the true value. If the LFMB results do not meet these criteria, then report all data for the problem analytes in the Extraction Batch with a note that the LFMB accuracy criteria were not met. The laboratory must investigate the root cause for this failure and report their findings and corrective action.

9.2.2.5 Pre-analysis Standard(s) Areas. The analyst must monitor the peak areas of the pre-analysis standard(s) in all injections of the Analysis Batch. The pre-analysis standard(s) (as indicated by peak area) in any chromatographic run must be within 50–150% of the average area measured during the initial calibration. Random evaporation losses have been observed with the polypropylene caps causing high-biased pre-analysis standard(s) areas. If a pre-analysis standard(s) standard area for a single sample in an analysis batch does not meet these criteria, reanalyze the extract in a subsequent analysis batch. If the pre-analysis standard(s) area fails to meet the acceptance criteria in the repeat analysis, or if multiple samples in a batch fail to meet

the pre-analysis spike criteria, perform corrective action and reanalyze the failed samples.
extract.

9.2.2.6 Pre-extraction Standard Recoveries. Pre-extraction standard recoveries determined during the analysis of samples must demonstrate on a per sample basis that recovery of the labeled standard achieved the 20-130% recovery requirements summarized in Table 45-5. Calculate the percent recovery (%R) for each analogue as follows:

$$\%R = \frac{\text{Measured Concentration of the Pre-extraction Isotope analogue}}{\text{Fortification concentration of the Pre-extractoin Isotope analogue}} \times 100 \quad \text{Eq. 45-8}$$

Recoveries below the acceptable range for pre-extraction spikes are an indication that sample preparation procedures did not adequately address sample and or sample matrix processing to recover native target compounds. Compounds that fail this criterion should be flagged and reported as not quantitative because of QC failure. If this failure involves target compounds that are critical to the test objectives, this is a failure that requires root cause investigation and may require a repeat field sampling effort.

9.2.2.7 Calibration Verification using QCS. A QCS must be analyzed during the IDC, and then quarterly thereafter. For this method, the laboratory is not required to obtain standards from a source independent of the primary calibration standards. Instead, the laboratory should acquire the best available quantitative standards and use these to prepare both the primary calibration standards and the QCS. The QCS must be an independent dilution beginning with the common starting materials. Preparation by a second analyst is recommended. The acceptance criterion for the QCS is 70–130% of the true value. If the accuracy for any analyte fails the recovery criterion, prepare fresh standard dilutions and repeat the Calibration Verification.

9.3 Method Modification QC Requirements

The analyst is permitted to modify the chromatographic and MS/MS conditions. Examples of permissible method modifications include alternate LC columns, MRM transitions, and additional QC analytes proposed for use with the method. Any method modifications must be within the scope of the established method flexibility and must retain the basic chromatographic elements of this method. The following are required after a method modification.

9.3.1 Repeat the IDC. Establish an acceptable initial calibration using the modified conditions.

Repeat the procedures of the IDC.

9.3.2 Document Performance in Representative Sample Matrices. The analyst is also required to evaluate and document method performance for the modifications in an archived field sample treated as a matrix spike.

9.4 Record and Report Requirements

Record and report data and information that will allow an independent reviewer to validate the determination of each target compound concentration. At a minimum, record and report the data as described in Sections 9.4.1 through 9.4.7 of this method.

9.4.1 Sample numbers and other sample identifiers. Each sample must have a unique identifier.

9.4.2 Field sample volume.

9.4.3 Field sampling date.

9.4.4 Extraction dates.

9.4.5 Analysis dates and times.

9.4.6 Analysis sequence/run chronology.

9.4.7 Quantitation Reports.

10.0 Calibration and Standardization

10.1 Sampling System. Same as Sections 6.1 and 10.1 through 10.7 of Method 5 of appendix A-3 to 40 CFR part 60.

10.2 MS/MS Optimization.

10.2.1 MS Parameters. Instrumental parameters are optimized for the precursor and product ions listed in Table 45-2. Product ions other than those listed may be selected; however, the analyst should avoid using ions with lower mass or common ions that may not provide enough discrimination between the analytes of interest and co-eluting interferences.

10.2.2 Precursor Ion. Optimize the response of the precursor ion ($[M - H]^-$ or $[M - CO_2 - H]^-$) for each analyte following instrument manufacturer's guidance. Optimization may be performed at analyte concentrations of 1.0 $\mu\text{g/mL}$. Vary the MS parameters (source voltages, source and desolvation temperatures, gas flows, etc.) until optimal analyte responses are determined. The analytes may have different optimal parameters, requiring some compromise on the final operating conditions.

10.2.3 Product Ion. Optimize the product ion for each analyte following the instrument manufacturer's guidance. Typically, the carboxylic acids have similar MS/MS conditions and the sulfonic acids have similar MS/MS conditions.

10.3 Chromatographic Conditions. Establish LC operating parameters that optimize resolution and peak shape. Example LC starting conditions can be found in Table 45-7. Modifying the solvent composition of the standard or extract by increasing the aqueous content to better focus early eluting compounds on the column is not permitted. A decrease in methanol concentration could lead to lower or imprecise recovery of the more hydrophobic method analytes, while higher methanol concentration could lead to the precipitation of salts in some extracts. The peak shape of the early eluting compounds may be improved by increasing the volume of the injection loop or increasing the aqueous content of the initial mobile phase composition.

10.3.1 Minimizing PFAS Background. LC system components, as well as the mobile phase constituents, may contain many of the analytes in this method. These PFAS may build up on the head of the LC column during mobile phase equilibration. To minimize the background PFAS peaks and to keep baseline levels constant, the time the LC column sits at initial conditions must be kept constant and as short as possible (while ensuring reproducible retention times). In addition, priming the mobile phase and flushing the column with at least 90% methanol before initiating a sequence may reduce background contamination.

10.3.2 Establishing Branched vs. Linear Isomer Profiles. Prepare and analyze the technical-grade standard of PFOA at a mid- to high-level concentration. Identify the retention times of the branched isomers of PFOA present in the technical-grade PFOA standard. When PFOA is chromatographed on a reversed-phase column, the branched isomers elute prior to the linear

isomer. Repeat the procedure in this section for technical grade PFHxS and PFOS, and any other analytes for which technical-grade standards have been acquired. The branched isomer identification checks must be repeated any time chromatographic changes occur that alter analyte retention times.

10.3.3 Establish LC-MS/MS Retention Times and MRM Segments. Inject a mid- to high-level calibration standard under optimized LC-MS/MS conditions to obtain the retention times of each method analyte. Divide the chromatogram into segments that contain one or more chromatographic peaks. For maximum sensitivity, minimize the number of MRM transitions that are simultaneously monitored within each segment. Both primary and secondary product ions may need to be monitored to quantitate selected PFAS. Ensure that the retention time window used to collect data for each analyte is of sufficient width to detect earlier eluting branched isomers.

10.4 Initial Calibration. This method has two pre-analysis standard(s) that are used as reference compounds for the internal standard quantitation of the pre-extraction standards. The pre-extraction standards are used as reference compounds to quantitate the native analyte concentrations. The pre-extraction standard references for the native analytes are listed in Table 45-1.

10.4.1 Calibration Standards. Prepare a set of at least five calibration standards as described in Section 7.4.11. The analyte concentrations in the lowest calibration standard must not be within two to ten times the MDL. Suggested calibration standard concentrations for native target compounds are shown in Table 45-8. Suggested calibration standard concentrations for pre-extraction isotopic labeled compounds are shown in Table 45-9.

10.4.2 Calibration Curves of Native Analytes. Quantitate the native analytes using the internal standard calibration technique. The internal standard technique calculates concentration based on the ratio of the peak area of the native analyte to that of the isotope dilution analogue. Calibrate the LC-MS/MS and fit the calibration points with either a linear or quadratic regression.

Weighting may be used.

Note: Forcing the calibration curve through the origin may improve the estimate of the background levels of method analytes which are an important consideration for this method.

Note: The MS/MS instrument used during method development was calibrated using weighted (1/x) quadratic regression with forced zero.

10.4.3 Calibration of pre-extraction standards. The pre-extraction standards are quantified using the pre-analysis calibration technique. Because these isotopes are added at a single concentration to the calibration standards, calibrate for each of these using an average response factor.

10.4.4 Calibration of pre-analysis standard(s). Because pre-analysis standard(s) are added at a single concentration to the calibration standards, calibrate for each of these using an average response factor.

10.4.5 Calibration Acceptance Criteria. Evaluate the initial calibration by calculating the concentration of each analyte as an unknown against its regression equation. All calibration points should be within 90 – 110% of their true value. If these criteria cannot be met, the analyst could have difficulty meeting ongoing QC criteria. In this case, corrective action is recommended such as reanalyzing the calibration standards, restricting the range of calibration, or performing instrument maintenance. If the cause for failure to meet the criteria is due to

contamination or standard degradation, prepare fresh calibration standards and repeat the initial calibration.

10.5 Continuing Calibration. Analyze a CCC to verify the initial calibration at the beginning of each Analysis Batch, after every tenth field sample, and at the end of each Analysis Batch. The beginning CCC for each Analysis Batch must be at, or below, the MDL for each analyte. This CCC verifies instrument sensitivity prior to the analysis of samples. If standards have been prepared such that all low calibration levels are not in the same solution, it may be necessary to analyze two standards to meet this requirement. Alternatively, the nominal analyte concentrations in the analyte PDS may be customized to meet these criteria. Alternate subsequent CCCs between the mid and high calibration levels. Verify that the CCC meets the criteria in the following sections.

10.5.1 CCC- Pre-analysis standard(s) Responses. The absolute area of the quantitation ion for each of the two pre-analysis standard(s) must be within 50–150% of the average area measured during the initial calibration. If these limits are exceeded, corrective action is necessary (Section 10.5).

10.5.2 CCC Isotope Dilution Analogue Recovery. Using the average response factor determined during the initial calibration and the internal standard calibration technique, calculate the percent recovery of each isotope dilution analogue in the CCC. The recovery for each analogue must be within a range of 70–130%. If these limits are exceeded, corrective action is necessary (Section 10.5).

10.5.3 CCC Analyte Responses. Calculate the concentration of each method analyte in the CCC. Each analyte fortified at a level less than or equal to the QRL must be within 50–150% of the true value. The concentration of the analytes in CCCs fortified at all other levels must be within 70–130%. If these limits are exceeded, then all data for the failed analytes must be considered invalid. Any field samples analyzed since the last acceptable CCC that are still within holding time must be reanalyzed after an acceptable calibration has been restored.

10.5.3.1 Exception for High Recovery. If the CCC fails because the calculated concentration is greater than 130% (150% for the low-level CCC) for a method analyte, and field sample extracts show no concentrations above the MDL for that analyte, non-detects may be reported without re-analysis.

10.6 Corrective Action. Failure to meet the CCC QC performance criteria requires corrective action. Following a minor remedial action, such as servicing the autosampler or flushing the column, check the calibration with a mid-level CCC and a CCC at the QRL, or recalibrate according to Section 10.4 of this method. If pre-analysis standard(s) and calibration failures persist, maintenance may be required, such as servicing the LC-MS/MS system or replacing the LC column. These latter measures constitute major maintenance and the analyst must return to the initial calibration step.

10.7 Calibration Range Flexibility. The calibration ranges for native PFAS target compounds in Tables 45-8 are provided as an example. Calibration solutions are prepared by diluting the appropriate amounts of calibration stock solutions in 80 percent methanol/water. The actual ICAL concentration used for each sample batch will depend upon the quantitation limit requirements of the program. The concentration of pre-extraction isotopically labeled

compounds used for isotopic dilution quantitation are kept constant and will depend upon the quantitation limit requirements of the test program.

11.0 Analysis Procedure

11.1 Sample Extraction. The OTM-45 sampling train (Figure 45-1) currently results in seven (7) containers that are extracted and analyzed as four (4) discrete analytical fractions. Figure OTM 45-4 provides a flow chart showing how the respective sample containers are combined, extracted, and concentrated for analysis.

11.1.1 Fraction 1 Particulate Filter and FH Rinse (Containers 1 and 2). Place the particulate filter into an appropriate size HDPE bottle. Spike the filter with the appropriate pre-extraction recovery standard solution. Add front-half sample rinsate (Container 2) to each respective filter container. If there is less than 50 mL of rinsate, add extraction solvent for a total of 50 mL. Extract samples on a shaker table for a minimum of 18 hours. Measure and record the final volume. After extraction, the solvent is divided: half of the volume used in the extraction is measured for concentration; the remaining portion is archived. This is achieved by filtering into separate, appropriately sized HDPE containers by placing a filter paper in a disposable polypropylene funnel and pouring the sample extracts through the funnels and into the HDPE containers.

11.1.2 Fraction 2 Primary XAD-2 Adsorbent and Back-Half (BH) Rinse (Containers 3 and 4).

11.1.2.1 Spike the XAD-2 in the module with the appropriate pre-extraction recovery standard solution (Table 45-3). Empty the module contents into an HDPE wide mouth bottle. Rinse the inside of the module with 5% ammonium hydroxide in methanol and add to the HDPE bottle.

11.1.2.2 Add up to 180 mL of the back-half sample rinsate (Container 4) to the respective XAD container. If there is less than 180 mL of rinsate, add 5% ammonium hydroxide in methanol for a total of 180 mL. If there is excess rinsate, you may add up to half of the volume and save the remaining half for the second extraction described in 11.1.2.5.

11.1.2.3 Extract the samples on a shaker table for a minimum of 18 hours.

11.1.2.4 Decant the extraction solvent into a new HDPE container leaving the XAD-2 in the extraction container.

11.1.2.5 Add any unused rinsate to the XAD- HDPE extraction container. If there is less than 180 mL of sample rinsate remaining, add 5% ammonium hydroxide in methanol for a total of 180 mL. Extract the sample a second time on a shaker table for a minimum of 18 hours. After the second extraction is completed, the extraction solvent is decanted into the container with the first extract leaving the XAD-2 in the extraction container. Measure and record the final volume. Divide extract with half of the volume going for concentration (Section 11.2) and archive the remaining half.

11.1.3 Fraction 3 Condensate/Impinger Water and Impinger Rinses (Containers 5 and 6). The sample is brought up to 500 mL in a Nalgene bottle with DI water. If the amount of sample is more than 500 mL, the entire sample is prepared (no additional DI water is added). Spike each sample with 1.0 mL of the IDA solution. The water sample is divided: 250 mL (or half the sample if more than 500 mL was received) is measured for analysis in an appropriately sized Nalgene bottle; the remaining half is archived in an appropriately sized Nalgene bottle.

11.1.3.1 Solid Phase Extraction (SPE) of Aqueous Samples. Wash the SPE cartridges with 5.0 mL 0.3% NH₄OH/methanol. Wash with 5.0 mL of 0.1 N NaOH/water. Close valve when ~ 200 uL remains on top to keep columns wet. After this step, ensure the columns do not go dry until the completion of loading and rinsing samples. Label and add the reservoir to the column. Add sample to the column. Using a vacuum, pull the entire sample through the cartridge at a rate of 2-5 drops per second. After the entire sample has been loaded onto the column, rinse the Nalgene bottle twice with 5 mL of reagent water and pour onto the column reservoir. After the final loading of the sample but before completely passed through the cartridge, allow the column to dry well with vacuum for 15 minutes.

11.1.3.2 SPE Column Wash with Hexane. Load 5 mL of hexane to the column and soak for five minutes. Elute to waste. Load an additional 5 mL of hexane and elute to waste without a soaking period. Allow the column to dry completely with vacuum for 5 to 10 minutes.

11.1.3.3 SPE Elution. Rinse Nalgene bottle with 4 mL of 0.3% NH₄OH/methanol. Pour rinse into the column reservoir onto the cartridge. Allow the solution to soak for 5 minutes and then elute into a polypropylene container large enough to contain volume from 11.1.3.4. Repeat rinse with an additional 4 mL of 0.3% NH₄OH/methanol. The total collection should be approximately 8 mL. Measure and record the final volume.

11.1.3.4 Combine with Container 6 (Impinger Rinse). Measure and record the Container 6 rinse volume. Transfer half to a HDPE container for sample archive. Combine with half of the SPE eluent. Cap, seal with Teflon tape, mark the level and store refrigerated at approximately 4 °C. Combine the remaining Container 6 rinse with the remaining SPE eluent and concentrate as described in 11.2.

11.1.3.4 Final volume for extract. Add 0.5 mL of IS 50 ng/mL concentration and 2 mL of water to the extract. This will create an extract with a final solvent composition of 80:20 methanol:water. Transfer a portion of the extract to a 300 uL polypropylene autosampler vial. Archive the rest of the extracts for re-injection and dilution. Seal the vial with a polypropylene screw cap.

11.1.4 Fraction 4 Breakthrough XAD-2 Adsorbent (Container 7).

11.1.4.1 Spike the XAD-2 in the module with the appropriate pre-extraction recovery standard solution. Empty the module contents into an HDPE wide mouth bottle. Rinse the module with 5% ammonium hydroxide and add to HDPE bottle.

11.1.4.2 Add 5% ammonium hydroxide in methanol for a total volume of 180 mL

11.1.4.3 Extract the samples on a shaker table for a minimum of 18 hours.

11.1.4.4 Decant the extraction solvent into an HDPE container leaving the XAD-2 in the extraction container.

11.1.4.5 Add 180 mL of 5% ammonium hydroxide in methanol to the XAD extraction container. Extract the sample a second time on a shaker table for a minimum of 18 hours. After the second extraction is completed, the extraction solvent is decanted into the container with the first extract leaving the XAD-2 in the extraction container. Divide extract with half of the volume going for concentration (Section 11.2) and archive the remaining half.

11.2 Sample Concentration

Revision 0 (1/13/2021)

11.2.1 Pour extract into a polypropylene digestion vessel and insert into a digestion block heated between 55-60 °C. Rinse container three times with small portions of 5% ammonium hydroxide in methanol and add to digestion vessel. Cover each vessel with a ribbed polypropylene watch glass.

11.2.1.1 To accommodate large volumes, the extract and rinsate can either be concentrated in multiple digestion vessels until concentrated to a level that allows recombination (with proper rinsing) OR additional extract can be added to a digestion vessel as volume is reduced during the concentration step. Concentrate the sample to below 10 mL, but not to dryness. Transfer to a 10 mL polypropylene tube, rinsing the digestion vessel adequately.

11.2.2 Blow down remaining volume to below 2 mL, but not to dryness, with an N-EVAP at 55-60 °C. Bring up volume to 2 mL with DI water or 5% ammonium hydroxide in methanol. Cap snugly and ensure there are no leaks. Store refrigerated at approximately 4 °C until analysis. Analyze within 28 days of extracting.

11.3 Sample Analysis.

Note: Sample analysis for this method is fashioned after EPA's Method 533 with modifications to accommodate the various fractions recovered from the stationary source sampling train.

11.3.1 Establish LC-MS/MS Operating Conditions. Establish MS/MS operating conditions per the procedures in Section 10.2 and chromatographic conditions per Section 10.3. Establish a valid initial calibration following the procedures in Section 10.3 or confirm that the existing calibration is still valid by analyzing a low-level CCC. If establishing an initial calibration for the

first time, complete the IDC prior to analyzing field samples. Analyze field and QC samples in a properly sequenced Analysis Batch as described in Section 11.4.

11.3.2 Verify Retention Time Windows. The analyst must ensure that each method analyte elutes entirely within the assigned window during each Analysis Batch. Make this observation by viewing the quantitation ion for each analyte in the CCCs analyzed during an Analysis Batch. If an analyte peak drifts out of the assigned window, then data for that analyte is invalid in all injections acquired since the last valid CCC. In addition, all peaks representing multiple isomers of an analyte must elute entirely within the same MRM window.

11.4 Analysis Batch Sequence. An Analysis Batch is a sequence of samples, analyzed within a 24-hour period, of no more than 20 field samples and includes all required QC samples (LSMB, CCCs, the LFSM and LFSMD (or FD)). The required QC samples are not included in counting the maximum field sample total of 20. LC-MS/MS conditions for the Analysis Batch must be the same as those used during calibration.

11.4.1 Analyze Initial CCC. After a valid calibration is established, begin every Analysis Batch by analyzing an initial low-level CCC at or below the QRL. This initial CCC must be within 50–150% of the true value for each method analyte and must pass both the pre-analysis standard(s) area response criterion (Section 9.2.2.4) and the pre-extraction isotope recovery criterion (Section 9.2.2.6). The initial CCC confirms that the calibration is still valid. Failure to meet the QC criteria may indicate that recalibration is required prior to analyzing samples.

11.4.2 Analyze Field and QC Samples. After the initial CCC, continue the Analysis Batch by analyzing an LSMB, followed by the field samples and QC samples. Analyze a mid- or high-

level CCC after every ten field samples and at the end of each Analysis Batch. Do not count QC samples (LSMBs, FDs, LFSMs, LFSMDs) when calculating the required frequency of CCCs.

11.4.3 Analyze Final CCC. The last injection of the Analysis Batch must be a mid- or high-level CCC. The acquisition start time of the final CCC must be within 24 hours of the acquisition start time of the low-level CCC at the beginning of the Analysis Batch. More than one Analysis Batch within a 24-hour period is permitted. An Analysis Batch may contain field and QC samples from multiple extraction batches.

11.4.4 Initial Calibration Frequency. A full calibration curve is not required before starting a new Analysis Batch. A previous calibration can be confirmed by running an initial, low-level CCC followed by an LSMB. If a new calibration curve is analyzed, an Analysis Batch run immediately thereafter must begin with a low-level CCC and an LSMB.

12.0 Data Analysis and Calculations

12.1 Calculation Nomenclature

A_s = the area of the characteristic mass for the compound in the continuing calibration verification sample.

A_{is} = the area of the characteristic mass of the pre-extraction isotopically labeled standard in the continuing calibration verification sample.

C_{is} = the concentration of the pre-extraction isotopically labeled standard (pg/ μ L).

C_s = the concentration of the native compound in the continuing calibration standard (pg/ μ L).

A_i = Integrated ion current for the isotopically labeled compound.

A_n = Integrated ion current for the target native compound.

C_i = The concentration of the labeled compound used to perform isotope recovery correction, pg/ μ L.

C_n = The concentration of the target native compound, pg/ μ L.

C_{ndscm} = Concentration of target native compound i in the emission gas, pg/dscm.

C_{next} = Concentration of target native compound i in the extract, pg.

D = Difference in the RRF of the continuing calibration verification compared to the average RRF of the initial calibration, percent (%).

dscm = Dry standard cubic meters of gas volume sample measured by the dry gas meter, corrected to standard conditions.

R^* = Recovery of labeled compound standards, %.

RSD = Relative standard deviation, in this case, of RRFs over the five calibration levels, %.

SD_{RRF} = Standard deviation of initial calibration RRFs.

V_{ext} = Extract volume, μ L.

12.2 Source gas volume calculations. Carry out calculations for stack gas velocity, volumetric flow rate, sampling volume, moisture, isokinetic variation following the procedures in Section 12 of Method 5 of appendix A-3 to 40 CFR part 60, with the following additions.

12.3 Qualitative Identification of Target Compounds

12.3.1 Qualitative Identity of Peaks by Retention Times. At the conclusion of analysis data acquisition, use the same software settings established during the calibration procedure to qualitatively identify analyte peaks in the predetermined retention time windows. Confirm the identity of each analyte by comparison of its retention time with that of the corresponding analyte peak in an initial calibration standard or CCC.

12.3.2 Qualitative Identity by Confirming the Parent and Product ions. Confirm that the native and pre-extraction isotope parent and product ions are consistent with the continuing calibration results. The signals for all characteristic masses shall be present and shall maximize within the same two consecutive scans. The retention time difference between the native analyte and its labeled analog shall agree within ± 6 scans or ± 6 seconds (whichever is greater) of this difference in the continuing calibration standard.

12.4 Recovery of Labeled Compound Standards. Use the following equation to determine the recovery of any labeled compounds, including pre-sampling spikes, pre-extraction filter spike, pre-extraction spikes, pre-analysis spikes. Verify and report the percent recovery of the pre-extraction isotopic labeled spike. The recovery performance criteria and actions if performance criteria are not met for these spikes is in Section 9. of this method.

$$R^* = \frac{\text{conc. found}}{\text{conc. spiked}} \times 100\% \quad \text{Eq. 45-9}$$

12.5 Quantitative Determination of Target Compounds by Isotope Dilution. Isotope dilution in this method is performed by adding a known amount of a labeled compound to every sample prior to extraction. Correction for recovery of the pollutant can be made because the pollutant and its labeled analog exhibit the same effects upon extraction, concentration, and chromatography.

Because quantitative native standards and isotopically labeled analogs are not currently available for all of the PFAS that may be target compounds in this procedure as originally written or in the future, integration and quantitation of the PFAS is dependent on the type of isotopically labeled

standard and native compound materials available. Procedures in this section must be followed to identify and report the identity and quantity of PFAS found in source gas samples.

12.5.1 PFOA Quantification. For PFOA, identify the branched and linear isomers by analyzing a technical-grade standard that includes both linear and branched isomers as directed and ensure that all isomers elute within the same acquisition segment. Quantitate field samples and fortified matrix samples by integrating the total response, accounting for peaks that are identified as linear and branched isomers. Quantitate based on the quantitative PFOA standard containing just the linear isomer.

12.5.2 PFHxS, PFOS, and other Analytes with Technical-Grade Standards. Multiple chromatographic peaks, representing branched and linear isomers, have been observed for current standards of PFHxS and PFOS using the LC conditions in this method. For PFHxS and PFOS, all the chromatographic peaks observed in the standard must be integrated and the areas summed. Chromatographic peaks in all field samples and QC samples must be integrated in the same way as the calibration standard for analytes with quantitative standards containing the branched and linear isomers.

12.5.3 Calculate the concentration of each of the target compounds in each of the sample fractions using the multipoint calibration and the measured sample volume including compensation for archived splits of the sample following the procedures in this section. Calculations must use all available digits of precision, but final reported concentrations should be rounded to an appropriate number of significant figures (one digit of uncertainty), typically two, and not more than three significant figures. Results must be reported individually for the filter, sorbent, impinger, and backup sorbent fractions. Report all results as total mass per media

fraction corrected for sample aliquot, sample dilution, and sample concentration as described in the following sections of this method.

12.5.3.1 Sample Concentrations Below the Lowest Calibration Standard. It is important to estimate the concentrations of target PFAS compounds below the lowest calibration standard concentration as well as below the MDL whenever possible. Samples may be concentrated and reanalyzed, or calibration curves may be extended to lower concentrations as necessary if initial concentration determination shows results below the lowest calibration standard. Concentrations may be estimated using the calibration equation or by other accepted approaches (e.g., using the RRF for the lowest calibration standard. The lowest calibration point may not be lower than three times the MDL.

12.5.3.1.1 Concentrations Between the QRL and MDL. Report concentrations between the detection limit (MDL) and the QRL for target compounds in each fraction with a J flag and explain the meaning of the J flag in your laboratory report, including the approach for estimating concentrations below the lowest calibration standard.

12.5.3.1.2 Concentrations below the MDL. Report results below the detection limit (MDL) with a BDL flag, and explain the meaning of the BDL flag in your laboratory report, including the approach for estimating concentrations below the lowest calibration standard.

12.5.4 Sample Concentrations Above the Highest Calibration Standard. Report concentrations of analytes that exceeded the calibration range in the original sample based on measurement in a diluted sample. Incorporate the dilution factor into final concentration calculations.

12.6 Quantitation Formulas.

Revision 0 (1/13/2021)

12.6.1 Response Factors. Calculate daily response factors for each of the target analytes in the continuing calibration verification sample using the following equation:

$$RF = \frac{(A_s)(C_{is})}{(A_{is})(C_s)} \quad \text{Eq. 45-10}$$

Where,

A_{is} = the area of the characteristic mass of the pre-extraction isotopically labeled standard in the continuing calibration verification sample.

A_s = the area of the characteristic mass for the native compound in the continuing calibration verification sample.

C_{is} = the concentration of the pre-extraction isotopically labeled analog in the standard (pg/ μ L).

C_s = the concentration of the native compound in the continuing calibration standard (pg/ μ L).

Confirm that the daily response factors for each target analyte meet the QC criteria in Section 9 of this method.

12.6.2 Isotope Dilution Quantification of Target Compounds in Sample Extracts

12.6.2.1 Proceed with target compound quantitation based on the type of isotopic labeled standard available used for each method analyte. If standards containing the branched and linear isomers cannot be purchased (i.e., only the linear isomer is available), only the linear isomer can be identified and quantitated in field samples and QC samples. Target analytes and branched isomers that do not have corresponding isotopic labels must be reported separately showing which native and isotopic labels were used to generate semiquantitative results for the analytes and isomers.

12.6.2.2 The response of each target compound (RR) relative to its labeled analog is determined using the area responses for each calibration standard, as follows:

$$RR = \frac{(A_n)(C_i)}{(A_i)(C_n)} \quad \text{Eq. 45-11}$$

Where,

A_n = The areas of the characteristic m/z for the target compound(s).

A_i = The areas of the peak representing the primary m/z for the labeled compound.

C_i = The concentration of the labeled compound in the calibration standard.

C_n = The corresponding concentration of the native compound in the calibration standard.

12.6.2.3 Calculate the Concentration of Individual Target Compound in the Extract by Isotope Dilution (pg/ μ L). This equation corrects for the target native compound recovery by its labeled pre-extraction spike analog. To accomplish this the pre-extraction spike, labeled compound concentrations must remain constant. Compute the concentration in the extract using the following equation:

$$C_{ex} \left(\frac{\text{pg}}{\mu\text{L}} \right) = \frac{(A_n)(C_i)}{(A_i)(RR)} \quad \text{Eq. 45-12}$$

Where:

C_{ex} = the concentration of the target compound in the extract in pg/ μ L.

C_i = The concentration of the labeled compound in the calibration standard in pg/ μ L.

Revision 0 (1/13/2021)

C_n = The concentration of the native compound in the calibration standard in pg/ μ L.

12.6.2.4 Total Mass of the Individual Target Compounds in the Sample Extract (pg). Calculate the mass for each target compound in each sample fraction using the concentration of the compound in the extract and the volume of extract, including any dilution, aliquots and/or archiving.

$$M_{total} = \frac{(C_{ex})(V_{ext})}{f} \quad \text{Eq. 45-13}$$

Where,

f = the fraction of the original sample extract before concentration taken for analysis (volume extracted/volume concentrated).

C_{ex} = The concentration of the compound in the extract in pg/ μ L.

V_{ext} = The final extract volume after concentration and/or dilution in mL

12.6.3 Concentration of the Individual Target Compound or in the Emission Gas (pg/dscm).

Calculate the gaseous emission concentration for each target compound based on the sum of the measured mass for each sample train fraction where PFAS were detected.

$$C_{dscm} = \frac{M_{sum\ total}}{dscm} \quad \text{Eq. 45-14}$$

Where,

$dscm$ = the standard dry cubic meters of gas collected in the sampling run.

12.6.3.1 In-stack Detectable Limit (DL). Calculate the gaseous emission, in-stack DL (pg/dscm) for each target compound based the sum of the sample-specific MDL masses for sample train fractions 1, 2 and 3, divided by the standard dry cubic meters of gas collected in the sampling run.]

$$DL = \frac{M_{MDLSum}}{dscm} \quad \text{Eq. 45-15}$$

Where,

M_{MDLSum} = sum of the sample-specific MDL masses for sample train fractions 1, 2 and 3.

12.7 Data Reporting. Report the following data in the emissions test reports.

12.7.1 Analytical Data. Include the following data and information in analytical data reports and emissions data test report.

12.7.1.1 Calibration Data. Include the number of calibration points and associated concentrations, including the lowest point in the calibration curve. Provide the calibration information used to derive the quantitative relationship and the approach used. Report the agreement between the calculated value and known value for each calibration point (See Section 10.4.5 calibration criteria).

12.7.1.2 MDL Study Data. Report the results for the MDL study identified and described in Section 9.2.1.3.

12.7.1.2.1 MDL Concentration and Mass for Each Fraction. Report the study's individual MDLs established for Fractions 1 – 3. Include the overall mass for each analyte as well as the associated volumetric concentration (pg/ μ L) used for determining the MDL mass specific to a collected emissions sample.

12.7.1.2.2 MDL Accuracy Data. Report the accuracy data from the MDL study and Section 9.2.1.6.

12.7.1.2.3 MDL Precision Data. Report the precision data from the MDL study and Section 9.2.1.5.

12.7.1.3 Laboratory Fortified Media Blank Samples Data. Report the results of the dual level LFMB samples from Section 9.2.2.4.

12.7.1.4 Laboratory Blanks. Report the target PFAS compound masses for all laboratory blanks.

12.7.1.5 Sample Train Analytical Data. Report all field sample analytical data, including blanks.

12.7.1.5.1 Sample-Specific Concentration and Mass for Each Fraction. Report the sample-specific analytical mass/volume (liquid) concentration and associated mass for each PFAS target compound and each fraction, inclusive of sample-specific final extract volumes, archive volumes, dilutions, etc.

12.7.1.4.2 Sample-Specific MDL Mass for Each Fraction. Report the sample-specific analytical MDL mass/volume (liquid) concentration and associated mass for each PFAS target compound and each fraction, inclusive of total sample-specific final extract volumes, archive volumes, dilutions, etc.

12.7.1.4.3 Sample-Specific QRL Mass for Each Fraction. Report the sample-specific analytical QRL mass/volume (liquid) concentration and associated mass for each PFAS target compound and each fraction, inclusive of total sample-specific final extract volumes, archive volumes, dilutions, etc.

12.7.1.4.4 Pre-extraction Standard Recoveries for Each Fraction. Report the pre-extraction standard recoveries for each standard, for each fraction.

12.7.1.4.5 Pre-Sampling Standard Mass and Recoveries. Report PSS mass for Fractions 2 (primary XAD-2 adsorbent and rinses) and 3 (condensate/impingers and rinse) and individual and combined recoveries.

Note: The Pre-sampling standard recoveries include both Fractions 2 and 3 and the pre-sampling standard may have migrated from the XAD-2 to the impingers.

12.7.2 Sampling and Emissions Data. Include the following data and information in the emissions data test report.

12.7.2.1 Emissions Data. For each run, report the target compound masses measured for each fraction. Indicate whether reported masses are in the quantitative range, estimated, or non-detect. Report the associated gaseous concentration for that run, based on the sum of detected fractions and the gaseous sample volume collected for that run.

12.7.2.2 In-stacks DLs. For each run, report the sample-specific, in-stack DLs, based on run-specific QRL and MDL masses (sum of fractions 1-3) and the gaseous sample volume collected for that run.

13.0 Method Performance

Data to support OTM-45 Method Performance to date is limited. This method has been developed based on a combination of empirical knowledge and data as well as incorporation of procedures and concepts adapted from other methods and measurement practices. EPA/ORD research, including field evaluation testing, has resulted in data that support the specified OTM-45 method performance criteria, including pre-extraction standard and pre-sampling standard isotopic label recovery criteria. Moreover, these performance criteria are further confirmed by

communications with and data shared by commercial analytical laboratories. Tables 45-9 and 45-10 present the pre-sampling standard recoveries and pre-extraction standard recoveries achieved as part of an ORD OTM-45 field evaluation. An average pre-sampling standard recovery of 100.3% was observed for PFOA, while an average pre-sampling standard recovery of 73.0% was observed for PFOS. Additional pre-sampling standard candidate compounds are needed to better represent overall measurement performance. Candidates are limited to PFOA and PFOS at this time due to the lack of additional PFAS compounds with necessary multiple isotopologues. The pre-extraction standard recoveries presented in Table 45-10 also support OTM-45 performance criteria. In general, pre-extraction standard recoveries for Fractions 1 (Filter) and 3 (Impingers) were greater than for Fraction 2 (XAD-2).

Table 45-11 presents the calculated in-stack detection levels, based on both QRLs and MDLs and a nominal 3 m³ sample volume. The values represent the sum of Fractions 1, 2, and 3. The in-stack detection levels reported indicate that measurements in the pg/m³ range are possible for the majority of OTM-45 targeted PFAS compounds.

The preparation and release of this method is intended to further establish data to support refinement of these initial performance criteria.

The collection of blanks, in replicate and representing multiple forms, is particularly important to data interpretation and validation. Managing PFAS contamination from a variety of sources, particularly XAD-2, has been a critical and acknowledged factor to collecting emissions data of known and acceptable quality. The recent application of PFAS emissions testing has resulted in advancements to minimizing PFAS contamination. However, the comprehensive collection of

laboratory and field sampling blank samples remains critical to identifying sources of contamination and an inherent component to interpret results from this method.

14.0 Pollution Prevention

The target compounds used as standards in this method are prepared in extremely small amounts and pose little threat to the environment when managed properly. Prepare standards in volumes consistent with laboratory use to minimize the disposal of excess volumes of expired standards.

15.0 Waste Management

15.1 The laboratory is responsible for complying with all federal, state, and local regulations governing waste management, particularly the hazardous waste identification rules and land disposal restrictions, and for protecting the air, water, and land by minimizing and controlling all releases from fume hoods and bench operations. The laboratory must also comply with any sewage discharge permits and regulations. The EPA's *Environmental Management Guide for Small Laboratories* (EPA 233-B-98-001) provides an overview of requirements.

15.2 For further information on waste management, consult *The Waste Management Manual for Laboratory Personnel* and *Less is Better-Laboratory Chemical Management for Waste Reduction*, available from the American Chemical Society's Department of Government Relations and Science Policy, 1155 16th Street N.W., Washington, D.C. 20036.

16.0 Bibliography

- 1) EPA Method 533
<https://www.epa.gov/sites/production/files/2019-12/documents/method-533-815b19020.pdf>
- 2) EPA Method 537.1
https://cfpub.epa.gov/si/si_public_record_Report.cfm?dirEntryId=343042&Lab=NERL
- 3) EPA Method 5. 40 CFR Part 60 Appendix A-3
- 4) SW-846 Method 0010
<https://www.epa.gov/hw-sw846/sw-846-test-method-0010-modified-method-5-sampling-train>.
- 5) Shoemaker, J. and Dan Tettenhorst. Method 537.1: Determination of Selected Per- and Polyfluorinated Alkyl Substances in Drinking Water by Solid Phase Extraction and Liquid Chromatography/Tandem Mass Spectrometry (LC/MS/MS). Version 2.0, U.S. Environmental Protection Agency, Office of Research and Development, Center for Environmental Solutions and Emergency Response, Cincinnati, OH, EPA Document #: EPA/600/R-20/006, 2020.
- 6) Rosenblum, L. and Wendelken, S. C. Method 533: Determination of Per- and Polyfluorinated Alkyl Substances in Drinking Water by Isotope Dilution Anion Exchange Solid Phase Extraction and Liquid Chromatography/Tandem Mass Spectrometry. U.S. Environmental Protection Agency, Office of Ground Water and Drinking Water, Standards and Risk Management Division, Cincinnati, OH, EPA Document No. 815-B-19-020, 2019.
- 7) Tyner, T, and Francis, J. ACS reagent chemicals: specifications and procedures for reagents and standard-grade reference materials. American Chemical Society, <http://pubs.acs.org/isbn/9780841230460>, 2017.
- 8) U.S. Environmental Protection Agency. Office of Air Programs Publication No. APTD-0576: Maintenance, Calibration, and Operation of Isokinetic Source Sampling Equipment. Research Triangle Park, NC. March 1972.
- 9) US EPA. *Statistical Protocol for the Determination of the Single-Laboratory Lowest Concentration Minimum Reporting Level (LCMRL) and Validation of Laboratory Performance at or Below the Minimum Reporting Level (MRL)*; EPA 815-R-05-006; Office of Water: Cincinnati, OH, November 2004.
- 10) US EPA. *Technical Basis for the Lowest Concentration Minimum Reporting Level (LCMRL) Calculator*; EPA 815-R-11-001; Office of Water: Cincinnati, OH, December 2010.

17.0 Tables, Diagrams, Flowcharts, and Validation Data

Table 45-1 PFAS Target Analytes

Common Name ^a	Abbreviated Name	CAS ^b Registry Number	Isotopic Pre-Extraction Pair
Perfluoroalkylcarboxylic acids (PFCAs)			
Perfluorobutanoic acid ^{1,3,4}	PFBA	375-22-4	¹³ C ₄ -PFBA
Perfluoropentanoic acid ^{1,3,4}	PFPeA	2706-90-3	¹³ C ₅ -PFPeA
Perfluorohexanoic acid ^{1,2,3,4}	PFHxA	307-24-4	¹³ C ₂ -PFHxA
Perfluoroheptanoic acid ^{1,2,3,4}	PFHpA	375-85-9	¹³ C ₄ -PFHpA
Perfluorooctanoic acid ^{1,2,3,4}	PFOA	335-67-1	¹³ C ₄ -PFOA
Perfluorononanoic acid ^{1,2,3,4}	PFNA	375-95-1	¹³ C ₅ -PFNA
Perfluorodecanoic acid ^{1,2,3,4}	PFDA	335-76-2	¹³ C ₂ -PFDA
Perfluoroundecanoic acid ^{1,2,3,4}	PFUnDA	2058-94-8	¹³ C ₂ -PFUnDA
Perfluorododecanoic acid ^{1,2,3,4}	PFDoA	307-55-1	¹³ C ₂ -PFDoA
Perfluorotridecanoic acid ^{2,3,4}	PFTTrDA	72629-94-8	¹³ C ₂ -PFDoA
Perfluorotetradecanoic acid ^{2,3,4}	PFTeDA	376-06-7	¹³ C ₂ -PFTeDA
Perfluoro-n-hexadecanoic acid	PFHxDA	67905-19-5	¹³ C ₂ -PFHxDA
Perfluoro-n-octadecanoic acid	PFODA	16517-11-6	¹³ C ₂ -PFDoA
Perfluorinated sulfonic acids (PFSAs)			
Perfluoro-1-butanedisulfonic acid ^{1,2,3,4}	PFBS	375-73-5	¹³ C ₃ -PFBS
Perfluoro-1-pentadisulfonic acid ^{1,3}	PFPeS	2706-91-4	¹³ C ₃ -PFHxS Or ¹³ C ₃ -PFBS
Perfluoro-1-hexadisulfonic acid ^{1,2,3,4}	PFHxS	355-46-4	¹⁸ O ₂ -PFHxS or ¹³ C ₃ -PFHxS
Perfluoro-1-heptadisulfonic acid ^{1,3}	PFHpS	375-92-8	¹³ C ₄ -PFHpA
Perfluoro-1-octadisulfonic acid ^{1,2,3,4}	PFOS	1763-23-1	¹³ C ₄ -PFOS
Perfluoro-1-nonadisulfonic acid ³	PFNS	68259-12-1	¹³ C ₄ -PFOS
Perfluoro-1-decadisulfonic acid ³	PFDS	335-77-3	¹³ C ₄ -PFOS
Perfluorododecane sulfonate	PFDoS	79780-39-5	¹³ C ₄ -PFOS
Perfluorinated sulfonamides (FOSAs)			
Perfluoro-1-octanesulfonamide ^{3,5}	FOSA	754-91-6	¹³ C ₈ -FOSA
N-Methylperfluorooctanesulfonamide ⁵	MeFOSA	31506-32-8	d3-MeFOSA
N-ethylperfluorooctanesulfonamide ⁵	EtFOSA	4151-50-2	d5-EtFOSA
Perfluorinated sulfonamide ethanols (FOSEs)			
2-(N-methylperfluoro-1-octanesulfonamido)-ethanol ⁵	N-MeFOSE	24448-09-7	d7-N-MeFOSE
2-(N-ethylperfluoro-1-octanesulfonamido)-ethanol ⁵	N-EtFOSE	1691-99-2	d9-N-EtFOSE
Perfluorinated sulfonamidoacetic acids (FOSAAAs)			
N-methyl perfluorooctanesulfonamidoacetic acid ^{2,3}	MeFOSAA	2355-31-9	d3-MeFOSAA
N-ethyl perfluorooctanesulfonamidoacetic acid ^{2,3}	EtFOSAA	2991-50-6	d5-EtFOSAA

Fluorotelomer sulfonates (FTS)			
1H,1H,2H,2H-Perfluorohexane sulfonic acid ^{1,3}	4:2 FTS	757124-72-4	M2-4:2 FTS
1H,1H,2H,2H -Perfluorooctane sulfonic acid ^{1,3}	6:2 FTS	27619-97-2	M2-6:2 FTS
1H,1H,2H,2H -Perfluorodecane sulfonic acid ^{1,3}	8:2 FTS	39108-34-4	M2-8:2 FTS
1H,1H,2H,2H-perfluorododecane sulfonate (10:2)	10:2 FTS	120226-60-0	M2-10:2 FTS
Fluorinated Replacement Chemicals			
4,8-Dioxa-3H-perfluorononanoic acid	ADONA ¹	919005-14-4	¹³ C ₄ -PFOS
Hexafluoropropylene Oxide Dimer Acid	HFPO-DA (GenX) ¹	13252-13-6	¹³ C ₃ -HFPO-DA
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	9Cl-PF3ONS (F-53B Major) ¹	756426-58-1	¹³ C ₄ -PFOS
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid OR 11-Chloroeicosafluoro-3-oxaundecane-1-sulfonate ^a	11Cl-PF3OUdS (F-53B Minor) ¹	763051-92-9 83329-89-9	¹³ C ₄ -PFOS
Additional Targets			
Nonafluoro-3,6-dioxaheptanoic acid ^{1,5}	NFDHA	151772-58-6	¹³ C ₅ -PFHxA
Perfluoro(2-ethoxyethane)sulfonic acid ^{1,5}	PFEESA	113507-82-7	¹³ C ₃ -PFBS
Sodium perfluoro-1-dodecanesulfonate ⁵	PFDoS	1260224-54-1	¹³ C ₄ -PFOS
Perfluoro-4-methoxybutanoic acid ^{1,5}	PFMBA	863090-89-5	¹³ C ₅ -PFPeA
Perfluoro-3-methoxypropanoic acid ^{1,5}	PFMPA	377-73-1	¹³ C ₄ -PFBA
Decafluoro-4-(pentafluoroethyl)cyclohexanesulfonate ⁴	PFecHS	67584-42-3	¹⁸ O ₂ -PFHxS
2H-perfluoro-2-decenoic acid ⁴	8:2 FTUCA or FOUEA	70887-84-2	¹³ C ₂ -FOUEA
2-perfluorodecyl ethanoic acid ⁴	10:2 FDEA	53826-13-4	¹³ C ₂ -FDEA
2-perfluorooctyl ethanoic acid ⁴	8:2 FTA or FOEA	27854-31-5	¹³ C ₂ -FOEA
2H-perfluoro-2-octenoic acid ⁴	6:2 FHUEA	70887-88-6	¹³ C ₂ -FHUEA
2-perfluorohexyl ethanoic acid ⁴	6:2FTCA or 6:2 FHEA	53826-12-3	¹³ C ₂ -FHEA
3:3 Fluorotelomer carboxylic acid ⁵	3:3 FTCA	356-02-5	¹³ C ₂ -FHEA
5:3 Fluorotelomer carboxylic acid ⁵	5:3 FTCA	914637-49-3	¹³ C ₂ -FHEA
7:3 Fluorotelomer carboxylic acid or 3-perfluoropheptyl propanoic acid ^{4,5}	7:3 FTCA or FHpPA	812-70-4	¹³ C ₂ -FOEA

^a Some PFAS are commercially available as ammonium, sodium, and potassium salts. This method measures all forms of the analytes as anions while the identity of the counterion is inconsequential. Analytes may be purchased as acids or as any of the corresponding salts

^b Chemical Abstract Service.

¹ Compound targeted in EPA Method 533

² Compound targeted in EPA Method 537.1

³ Compound targeted in EPA Method 8327

⁴ Compound targeted in ASTM Method D7968

⁵ Compound targeted in DoD Isotope Dilution Method

Table 45-2 MS/MS Conditions and Characteristic Ions

Analyte	Precursor Ion (m/z)	Primary Product Ion (m/z)	Secondary Product Ion (m/z)
PFBA	213	169	
¹³ C ₃ -PFBA	216	172	
¹³ C ₄ -PFBA	217	172	
PFPeA	263	219	
¹³ C ₅ -PFPeA	268	223	
PFH _x A	313	269	119
¹³ C ₅ -PFH _x A	318	273	
PFHpA	363	319	169
¹³ C ₄ -PFHpA	367	322	
PFOA	413	369	169
¹³ C ₂ -PFOA	415	370	
¹³ C ₈ -PFOA	421	376	
PFNA	463	419	169
¹³ C ₉ -PFNA	472	427	
PFDA	513	469	169
¹³ C ₆ -PFDA	519	474	
PFUnA	563	519	169
¹³ C ₇ -PFUnA	570	525	
PFDoA	613	569	169
¹³ C ₂ -PFDoA	615	570	
PFT _r DA	663	619	169
PFTeDA	713	169	219
PFH _x DA	813	769	169
PFODA	913	869	169
PFBS	299	80	99
¹³ C ₃ -PFBS	302	80	83
PFPeS	349	80	99
PFH _x S	399	80	99
¹³ C ₃ -PFH _x S	402	80	
PFHpS	449	80	99
PFOS	499	80	99
¹³ C ₄ -PFOS	503	80	
¹³ C ₈ -PFOS	507	80	
PFNS	549	80	99
PFDS	599	80	99
PFDoS	699	80	99
FOSA	498	78	
MeFOSA	512	169	
EtFOSA	526	169	
N-MeFOSE	616	59	
N-EtFOSE	630	59	
MeFOSAA	570	419	512
EtFOSAA	584	419	526
4:2FTS	327	307	

¹³ C ₂ -4:2FTS	329	309	81
6:2FTS	427	407	
¹³ C ₂ -6:2FTS	429	409	81
8:2FTS	527	507	
¹³ C ₂ -8:2FTS	529	509	81
10:2FTS	627	607	80
ADONA	377	251	
HFPO-DA	285	169	
¹³ C ₃ -HFPO-DA	287	169	
9Cl-PF3ONS	531	351	
11Cl-PF3OUdS	631	451	
NFDHA	295	201	
PFEESA	315	135	
PFDoS	699	80	99
PFMBA	279	85	
PFMPA	229	85	
PFecHS	461	381	99
8:2 FTUCA or FOUEA	457	393	
10:2 FDEA	577	493	
8:2 FTA or FOEA	477	393	
6:2 FHUEA	357	293	
6:2FTCA or 6:2 FHEA	377	243	
3:3 FTCA	241	177	117
5:3 FTCA	341	237	217
7:3 FTCA or FHpPA	441	337	317

Table 45-3 Fortification and Recovery Solutions for Poly and Perfluorinated Alkyl Compounds^a

Compound	Amount (µg/µL of final extract) ^b	Spike Recovery (percent)
Pre-sampling Adsorbent Standards		
¹³ C ₈ PFOA	75	70-130
¹³ C ₈ PFOS	75	70-130
Pre-extraction Filter Recovery Spike Standards		
Needs to be identified	75	70-130
Pre-extraction Standards		
	50	20-130
¹³ C ₄ PFBA	50	20-130
¹³ C ₅ PFPeA	50	20-130
¹³ C ₂ PFHxA	50	20-130
¹³ C ₄ PFHpA	50	20-130
¹³ C ₄ PFOA	50	20-130
¹³ C ₅ PFNA	50	20-130
¹³ C ₂ PFDA	50	20-130
¹³ C ₂ PFUnA	50	20-130
¹³ C ₂ PFDoA	50	20-130
¹³ C ₂ PFTeDA	50	20-130
¹³ C ₂ -PFHxDA	50	20-130
¹³ C ₂ -PFDoA	50	20-130
¹³ C ₃ - PFBS	50	20-130
¹³ C ₃ -PFHxS	50	20-130
¹⁸ O ₂ PFHxS	50	20-130
¹³ C ₄ PFOS	50	20-130
¹³ C ₈ FOSA	50	20-130
d ₃ -NMeFOSAA	50	20-130
d ₅ -NEtFOSAA	50	20-130
d ₇ -N-MeFOSE	50	20-130
d ₉ -N-EtFOSE	50	20-130
M2-6:2 FTS	50	20-130
M2-8:2 FTS	50	20-130
M2-4:2 FTS	50	20-130
M2-10:2 FTS (D4 labeled)	50	20-130
¹³ C ₃ HFPO-DA	50	20-130
¹³ C ₂ -FOUEA	50	20-130
¹³ C ₂ -FDEA	50	20-130
¹³ C ₂ -FOEA	50	20-130
¹³ C ₂ -FHUEA	50	20-130
¹³ C ₂ -FHEA	50	20-130
Pre-analysis Standards		
¹³ C ₂ PFOA	50	S/N≥10
¹³ C ₃ -PFBA	50	S/N≥10

^a Changes in the amounts of spike standards added to the sample or its representative extract will necessitate an adjustment of the calibration solutions to prevent the introduction of inconsistencies. Spike concentration assumes 1µL sample injection volume for analysis.

^b Spike levels assume half of the extract will be archived before cleanup. Spike levels may be adjusted for different split levels.

Table 45- 4 Sample Storage Conditions^a and Laboratory Hold Times

Stage	Type	Storage Conditions	Laboratory Holding time
Field collection	Particulate Filter	$\leq 20 \pm 3$ °C, 68 ± 5 °F Store cool but NOT on ice	N/A
	All other Field samples	Store on Ice (4°C)	
Shipping/Transport	Particulate Filter	Ship unrefrigerated.	N/A
	All other Field samples	Ship on Ice (4°C). Follow procedures in ASTM D6911-15	
Laboratory Storage: Before Extraction	All Sampling train Rinses and Particulate Filter Samples	≤ 6 °C (43°F)	≤ 28 days from date of collection; Extract within 28 days of collection
	Adsorbent samples (XAD-2)	≤ 6 °C (43°F)	≤ 1 year from received date; Extract within 28 days of collection
Laboratory Storage: After Extraction (archived)	All archived extracted samples	at ≤ 6 °C (43°F)	≤ 1 year

^a All samples must be stored in the dark after collection.

Table 45-5. General QA/QC Requirements for OTM-45

Section	Requirement	Specification and Frequency	Acceptance Criteria	Consequences and Corrective Actions
Sampling Quality Controls (Section 9.1)				
9.1.2	Field Sampling Media Blank (FSMB)	Represents the sampling media and reagents associated with field sample collection. One per each test series.	Levels should be \leq compound MDL	Analysis of the FSMBs should be compared to the background level criteria in Section 9.2.2.1. Failure to meet these levels does not invalidate the sampling run. If $>$ MDL, flag data. The measured target compound mass in each fraction will need to be reported and used to interpret sample assess impact on results.
9.1.3	Sample Train Proof Blank (STPB)	At least one STPB per each test series.	Levels \leq 10% of actual samples	If $>$ 10%, flag data and assess impact on results
9.1.4	Sample Train Field Blank (STFB)	At least one STPB per each test series	Levels \leq 10% of actual samples	If $>$ 10%, flag data and assess impact on results
9.1.5	Pre-Sampling Standards	Added to each XAD-2 adsorbent cartridge prior to sampling. Indicates sample collection and recovery efficiency.	\geq 70% and \leq 130% recovery of all spike standards as the sum of the recovery of sampling train fractions 2 and 3.	Recoveries below the acceptable range of 70-130% for the pre-sampling standard spikes may require a root cause evaluation. If the recovery of all the pre-sampling standard adsorbent spikes is below 70%, but, greater than 50%, the results have not met the recoveries experienced during method development but may still be acceptable. Flag recoveries that are between 50 and 70% and describe their potential impact on results. If the pre-sampling standard recoveries are less than 50%, the data for that train are not considered valid.

9.1.6	Secondary XAD-2 Breakthrough	Determines the relative breakthrough (BT) of each target through the OTM-45 train.	≤30% BT for each target compound.	For any BT ≥ 10%, add the fraction 4 mass to the total sample mass for emissions calculations If >30%, flag data and assess impact on results. Failure may invalidate reported results and require repeat sampling.
Ongoing Quality Control Requirements (Section 9.2.2)				
9.2.2.1.1	Sampling Media Background Level Checks	Confirm sample media background before use for sampling	Levels should be ≤ compound MDL	If >MDL, further clean sampling media until levels are ≤ MDL
9.2.2.1.2	Laboratory Sample Media Blank (LSMB)	Analyze a LSMB for each sampling medium and reagent with each extraction batch and ≥1 LSMB when method analytes exceed the calibration range.	Levels should be ≤ compound MDL	If >MDL, flag data and assess impact on results Resolve source of contamination before proceeding to additional analyses.
9.2.2.3	Calibration Acceptance Criteria	Evaluate the initial concentration of each analyte as an unknown against its regression equation (Section 10.4)	Between 90-110% of each analyte true value.	Reanalyze the calibration standards, restrict the range of calibration, or perform instrument maintenance. If failure is due to contamination or standard degradation, prepare fresh calibration standards and repeat initial calibration
9.2.2.3.1	Continuing Calibration Check (CCC)	Analyze CCC at the beginning of each analysis batch, after every tenth field sample, and at the end of the analysis batch.	Beginning CCC must be ≤ MDL for each analyte. Must be within 70-130% of true value	If the CCC fails because concentration is >130% (150% for low-level CCC) and field sample extracts show no concentrations above the MDL for that analyte, non-detects can be reported without re-analysis See Section 10.6 for Corrective Action.

9.2.2.4	Laboratory Fortified Media Blanks (LFMB)	Duplicate low and high LFMBs are required with each extraction batch for each fraction.	Analytes fortified near or at the lowest calibration point must be within 50-150% of the true value. Analytes fortified at all other concentrations must be within 70-130% of the true value.	If the LFMB results do not meet these criteria, the laboratory must investigate the cause for this failure, report their findings and corrective action. Then report all data for the problem analytes with a note that LFMB accuracy criteria were not met.
9.2.2.5	Pre-Analysis Standard(s) Areas	The analyst must monitor the peak areas of the pre-analysis standards in all injections of the analysis batch.	The pre-analysis standards (as indicated by peak area) in any chromatographic run must be within 50-150% of the average area measured during in the initial calibration.	If criteria is not met, reanalyze the extract in a subsequent analysis batch. If the pre-analysis standard(s) area fails to meet the acceptance criteria in the repeat analysis, or if multiple samples in a batch fail to meet the pre-analysis spike criteria, perform corrective action and reanalyze the failed samples extract.
9.2.2.6	Pre-Extraction Isotope Dilution	For each sample fraction, calculate the concentration and percent recovery of each isotope dilution analogue in field and QC samples using the average area in the initial calibration and internal standard	Percent recovery must be within a range of 20-130%	Recoveries below the acceptable range for pre-extraction spikes are an indication that sample preparation procedures did not adequately address sample and or sample matrix processing to recover native target compounds. Compounds that fail this criterion should be flagged and reported as not quantitative because of QC failure. If this failure involves target compounds that are critical to the test objectives, this is a failure that requires root cause investigation and may require a repeat field sampling effort.
9.2.2.78	Calibration Verification using Quality Control Standards (QCS)	Perform a calibration verification during the IDC and at least quarterly after	Results must be within 70-130% of the true value	If accuracy fails, prepare fresh standard dilutions and repeat the calibration verification
9.3	Method Modification QC Requirements	Perform after modifying chromatographic and MS/MS conditions	Must pass IDC criteria. Must evaluate and document method performance in an archived field sample	Repeat until IDC is passed.

Table 45-6 Initial Demonstration of Capability QC Requirements

Section	Requirement	Specification and Frequency	Acceptance Criteria
10.3.2	Establish retention times for branched isomers	Each time chromatographic conditions change	All isomers of each analyte must elute within the same MRM window.
9.2.1.2	Demonstration of low system background	Analyze a Laboratory Sample media Blank (LSMB) after the highest standard in the calibration range. If an automated extraction system is used, an LSMB must be extracted on each port.	Confirm that the LSMB is free from contamination as defined in section 9.2.2.1
9.2.1.3	Determination of MDL	7 LSMB and LFMB that are spiked within 2 to 10 times of the expected MDL. See 40 CFR Part 136 Appendix B	Establishes detectable concentrations for each sampling train fraction. Sample Specific MDL's are reported inclusive of sample-specific dilutions, final volumes, aliquots, etc.
9.2.1.4	MDL confirmation	Prepare a LSMB for each sampling media by spiking each media with native targets at the MDL and pre-extraction isotopic labeled standards at the concentration used to analyze field samples. Prepare and analyze the spiked LSMB Fortify and analyze 7 replicate LFMBs at the proposed MDL concentration. Confirm that the Upper Prediction Interval of Results (PIR) and Lower PIR meet the recovery criteria.	Upper PIR \leq 150% Lower PIR \geq 50% If not met, the lowest calibration point has been set too low and must be confirmed again at a higher concentration.
9.2.1.5	Demonstration of precision	Extract and analyze 7 replicate Laboratory Fortified Media Blanks (LFMBs) near the mid-range concentration.	Percent relative standard deviation must be \leq 20%.
9.2.1.6	Demonstration of accuracy	Calculate mean recovery for replicates used in Section 9.2.1.5	Mean recovery within 70–130% of the true value.
9.2.1.7	Lowest Calibration Concentration Confirmation	Establish a target concentration for the lowest calibration based on the intended use of the method. See subsections 9.1.2.7.1 - 9.2.1.7.2 for details	Upper PIR \leq 150% Lower PIR \geq 50% If not met, the lowest calibration point has been set too low and must be confirmed again at a higher concentration.
9.2.1.8	Calibration Verification	Analyze mid-level QCS.	Results must be within 70–130% of the true value.

Note: All aspects of the IDC must be successfully performed prior to analyzing field samples. If any of the above fail, you must repeat until successful.

Table 45-7 Example HPLC Method Conditions

Time (min)	% 20 mM Ammonium acetate	% Methanol
Initial	95.0	5.0
0.5	95.0	5.0
3.0	60.0	40.0
16.0	20.0	80.0
18.0	20.0	80.0
20.0	5.0	95.0
22.0	5.0	95.0
25.0	95.0	5.0
35.0	95.0	5.0

Table 45-8 Recommended Initial Calibration (ICAL) (pg/uL)

Compound	CS-1	CS-2	CS-3	CS-4	CS-5	CS-6	CS-7
Perfluoroalkylcarboxylic acids (PFCAs)							
PFBA	0.25	0.50	1	2.5	5	20	100
PFPeA	0.25	0.50	1	2.5	5	20	100
PFHxA	0.25	0.50	1	2.5	5	20	100
PFHpA	0.25	0.50	1	2.5	5	20	100
PFOA	0.25	0.50	1	2.5	5	20	100
PFNA	0.25	0.50	1	2.5	5	20	100
PFDA	0.25	0.50	1	2.5	5	20	100
PFUnDA	0.25	0.50	1	2.5	5	20	100
PFDoA	0.25	0.50	1	2.5	5	20	100
PFTTrDA	0.25	0.50	1	2.5	5	20	100
PFTeDA	0.25	0.50	1	2.5	5	20	100
PFHxDA	0.25	0.50	1	2.5	5	20	100
PFODA	0.25	0.50	1	2.5	5	20	100
Perfluorinated sulfonic acids (PFSAs)							
PFBS	0.25	0.50	1	2.5	5	20	100
PFPeS	0.25	0.50	1	2.5	5	20	100
PFHxS	0.25	0.50	1	2.5	5	20	100
PFHpA	0.25	0.50	1	2.5	5	20	100
PFOS	0.25	0.50	1	2.5	5	20	100
PFNS	0.25	0.50	1	2.5	5	20	100
PFDS	0.25	0.50	1	2.5	5	20	100
PFDoS	0.25	0.50	1	2.5	5	20	100
Perfluorinated sulfonamides (FOSAs)							
FOSA	0.25	0.50	1	2.5	5	20	100
EtFOSA	0.25	0.50	1	2.5	5	20	100
MeFOSA	0.25	0.50	1	2.5	5	20	100
Perfluorinated sulfonamide ethanols (FOSEs)							
MeFOSE	0.25	0.50	1	2.5	5	20	100
EtFOSE	0.25	0.50	1	2.5	5	20	100
Perfluorinated sulfonamidoacetic acids (FOSAAs)							
EtFOSAA	0.25	0.50	1	2.5	5	20	100
MeFOSAA	0.25	0.50	1	2.5	5	20	100
Fluorotelomer sulfonates (FTS)							
4:2 FTS	0.25	0.50	1	2.5	5	20	100
6:2 FTS	0.25	0.50	1	2.5	5	20	100
8:2 FTS	0.25	0.50	1	2.5	5	20	100
10:2 FTS	0.25	0.50	1	2.5	5	20	100
Fluorinated Replacement Chemicals							
ADONA ¹	0.25	0.50	1	2.5	5	20	100
HFPO-DA (GenX) ¹	0.25	0.50	1	2.5	5	20	100
⁹ Cl-PF ₃ ONS (F-53B Major) ¹	0.25	0.50	1	2.5	5	20	100
F-53B Minor (¹¹ Cl-PF ₃ OUdS) ¹	0.25	0.50	1	2.5	5	20	100
F5B minor (¹¹ Cl-PF ₃ OUdS)	0.25	0.50	1	2.5	5	20	100
Additional Targets							
NFDHA	0.25	0.50	1	2.5	5	20	100
PFEESA	0.25	0.50	1	2.5	5	20	100
PFDoS	0.25	0.50	1	2.5	5	20	100
PFMBA	0.25	0.50	1	2.5	5	20	100

PFMPA	0.25	0.50	1	2.5	5	20	100
<i>PFecHS</i>	0.25	0.50	1	2.5	5	20	100
8:2 FTUCA or FOUEA	0.25	0.50	1	2.5	5	20	100
10:2 FDEA	0.25	0.50	1	2.5	5	20	100
8:2 FTA or FOEA	0.25	0.50	1	2.5	5	20	100
6:2 FHUEA	0.25	0.50	1	2.5	5	20	100
6:2 FTCA or 6:2 FHEA	0.25	0.50	1	2.5	5	20	100
3:3 FTCA	0.25	0.50	1	2.5	5	20	100
5:3 FTCA	0.25	0.50	1	2.5	5	20	100
7:3 FTCA Or FHpPA	0.25	0.50	1	2.5	5	20	100

Table 45-8 Isotopic Dilution Pairs Initial Calibration (ICAL) Concentration (pg/uL)

Compound	CS-1	CS-2	CS-3	CS-4	CS-5	CS-6	CS-7
Perfluoroalkylcarboxylic acids (PFCAs)							
¹³ C ₄ -PFBA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₅ -PFPeA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₂ -PFHxA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₄ -PFHpA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₄ -PFOA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₅ -PFNA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₂ -PFDA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₂ -PFUDa	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₂ -PFDoA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₂ -PFTeDA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
Perfluorinated sulfonic acids (PFSAs)							
¹³ C ₃ - PFBS	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₃ -PFHxS (533) Or ¹³ C ₃ -PFBS	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹⁸ O ₂ -PFHxS	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₄ -PFHpA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₄ -PFOS	2.5	2.5	2.5	2.5	2.5	2.5	2.5
Perfluorinated sulfonamides (FOSAs)							
¹³ C ₈ -FOSA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
d5-EtFOSA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
d3-MeFOSA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
Perfluorinated sulfonamide ethanols (FOSEs)							
d7-N-MeFOSE	2.5	2.5	2.5	2.5	2.5	2.5	2.5
d9-N-EtFOSE	2.5	2.5	2.5	2.5	2.5	2.5	2.5
Perfluorinated sulfonamidoacetic acids (FOSAAs)							
d5-EtFOSAA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
d3-MeFOSAA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
Fluorotelomer sulfonates (FTS)							
M2-4:2 FTS	2.5	2.5	2.5	2.5	2.5	2.5	2.5
M2-6:2 FTS	2.5	2.5	2.5	2.5	2.5	2.5	2.5
M2-8:2 FTS	2.5	2.5	2.5	2.5	2.5	2.5	2.5
	2.5	2.5	2.5	2.5	2.5	2.5	2.5
Fluorinated Replacement Chemicals							
¹³ C ₄ -PFOS	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₃ -HFPO-DA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₄ -PFOS	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₄ -PFOS	2.5	2.5	2.5	2.5	2.5	2.5	2.5
	2.5	2.5	2.5	2.5	2.5	2.5	2.5
Pre-analysis Spiking Standard							
¹³ C ₂ -PFOA	2.5	2.5	2.5	2.5	2.5	2.5	2.5

Table 45-8 Continued Isotopic Dilution Pairs Initial Calibration (ICAL) Concentration (pg/uL)

¹³ C ₅ -PFHxA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₃ -PFBS	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₄ -PFOS	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₅ -PFPeA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₄ -PFBA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₂ -FOUEA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₂ -FDEA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₂ -FOEA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₂ -FHUEA	2.5	2.5	2.5	2.5	2.5	2.5	2.5
¹³ C ₂ -FHEA	2.5	2.5	2.5	2.5	2.5	2.5	2.5

Table 45-9 Pre-Sampling Standard Recoveries

Pre-Sampling Standard	Percent (%) Recovery			
	Avg	RSD	Max	Min
13C8 PFOA	100.3	9.2%	111.0	89.0
13C8 PFOS	73.0	18.6%	85.0	58.0

Table 45-10 Pre-Extraction Standard Recoveries

Pre-Extraction Standard	Fraction 1 (Filter) Percent (%) Recovery				Fraction 2 (XAD-2) Percent (%) Recovery				Fraction 3 (Impingers) Percent (%) Recovery			
	Avg	RSD	Max	Min	Avg	RSD	Max	Min	Avg	RSD	Max	Min
13C2 PFDA	98	6.8%	107	92	72	16.9%	83	55	112	8.2%	124	102
13C2 PFDoA	98	5.3%	102	90	84	25.7%	105	54	114	12.6%	132	98
13C2 PFHxA	98	7.1%	108	92	85	7.0%	93	80	106	7.4%	115	97
13C2 PFTeDA	93	9.9%	104	82	75	31.8%	107	50	113	12.2%	127	94
13C2 PFUnA	101	7.9%	110	91	80	21.4%	94	55	112	9.7%	126	100
13C3 HFPO-DA	79	6.3%	84	74	72	12.0%	79	60	104	11.3%	121	96
13C4 PFBA	94	6.7%	103	89	56	12.5%	65	49	102	8.4%	114	94
13C4 PFHpA	100	4.8%	106	95	76	5.6%	81	71	111	9.4%	124	99
13C4 PFOA	99	6.9%	108	92	88	10.3%	99	78	108	10.0%	120	94
13C4 PFOS	92	4.7%	98	89	82	19.2%	94	59	108	9.1%	120	96
13C5 PFNA	99	6.1%	106	92	81	14.5%	93	65	114	10.0%	128	100
13C5 PFPeA	93	6.9%	101	86	64	6.5%	68	60	101	7.5%	111	93
13C8 FOSA	89	8.7%	99	80	55	48.0%	89	25	108	8.1%	120	99
18O2 PFHxS	105	4.1%	111	101	81	9.8%	86	69	119	10.4%	135	105
d3-NMeFOSAA	83	13.1%	93	68	77	25.7%	100	52	94	9.4%	106	85
d5-NEtFOSAA	104	8.2%	111	93	84	28.6%	108	51	97	16.1%	118	82
M2-4:2 FTS	110	13.8%	129	92	156	11.8%	174	131	110	21.1%	131	80
M2-6:2 FTS	118	18.8%	144	91	143	25.6%	186	97	130	28.4%	176	86
M2-8:2 FTS	121	29.3%	167	82	89	25.0%	105	57	113	17.5%	135	87

Table 45-11. In-Stack Detection Limits

Analyte Description	CAS Number	QRL (ng/train)	MDL (ng/train)	QRL (ng/m3)	MDL (ng/m3)	QRL (PPQ)	MDL (PPQ)
Perfluorobutanoic acid (PFBA)	375-22-4	12.60	6.25	4.20	2.08	471.75	234.00
Perfluoropentanoic acid (PFPeA)	2706-90-3	1.80	0.59	0.60	0.20	54.63	17.88
Perfluorohexanoic acid (PFHxA)	307-24-4	2.40	0.92	0.80	0.31	61.24	23.42
Perfluoroheptanoic acid (PFHpA)	375-85-9	1.70	0.62	0.57	0.21	37.42	13.63
Perfluorooctanoic acid (PFOA)	335-67-1	2.50	1.28	0.83	0.43	48.38	24.69
Perfluorononanoic acid (PFNA)	375-95-1	1.50	0.46	0.50	0.15	25.90	7.87
Perfluorodecanoic acid (PFDA)	335-76-2	1.50	0.39	0.50	0.13	23.38	6.02
Perfluoroundecanoic acid (PFUnA)	2058-94-8	2.00	0.99	0.67	0.33	28.41	14.11
Perfluorododecanoic acid (PFDoA)	307-55-1	1.50	0.36	0.50	0.12	19.57	4.67
Perfluorotridecanoic acid (PFTriA)	72629-94-8	1.50	0.35	0.50	0.12	18.10	4.20
Perfluorotetradecanoic acid (PFTeA)	376-06-7	1.50	0.58	0.50	0.19	16.83	6.49
Perfluorobutanesulfonic acid (PFBS)	375-73-5	1.50	0.51	0.50	0.17	40.06	13.62
Perfluoropentanesulfonic acid (PFPeS)	2706-91-4	1.50	0.43	0.50	0.14	34.33	9.82
Perfluorohexanesulfonic acid (PFHxS)	355-46-4	1.70	0.82	0.57	0.27	34.05	16.50
Perfluoroheptanesulfonic Acid (PFHpS)	375-92-8	1.50	0.25	0.50	0.08	26.71	4.52
Perfluorooctanesulfonic acid (PFOS)	1763-23-1	2.40	1.06	0.80	0.35	38.46	16.98
Perfluorononanesulfonic acid (PFNS)	68259-12-1	1.50	0.42	0.50	0.14	21.85	6.16
Perfluorodecanesulfonic acid (PFDS)	335-77-3	1.50	0.51	0.50	0.17	20.03	6.86
Perfluorooctanesulfonamide (FOSA)	754-91-6	2.00	0.82	0.67	0.27	32.11	13.23
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2355-31-9	2.40	1.21	0.80	0.40	33.67	16.92
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2991-50-6	2.40	1.18	0.80	0.39	32.86	16.14
4:2 FTS	757124-72-4	1.50	0.59	0.50	0.20	36.63	14.43
6:2 FTS	27619-97-2	2.00	0.87	0.67	0.29	37.43	16.21
8:2 FTS	39108-34-4	1.70	0.81	0.57	0.27	25.79	12.23
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	919005-14-4	1.60	0.41	0.53	0.14	33.91	8.61
HFPO-DA (GenX)	13252-13-6	16.60	8.30	5.53	2.77	403.05	201.62
F-53B Major	756426-58-1	1.50	0.51	0.50	0.17	22.57	7.61
F-53B Minor	763051-92-9	1.50	0.55	0.50	0.18	19.00	6.93

RL: Reporting Limit

MDL: Method Detection Limit

PPQ: Parts Per Quadrillion




Figure OTM-45-1. Sampling Train




Figure OTM 45-2. Condenser and Adsorbent Module




Figure OTM 45-3. Sample Preparation Flow Chart

APPENDIX

PREPARATION OF XAD-2 ADSORBENT RESIN

1.0 Scope and Application

XAD-2® resin, as supplied by the original manufacturer, is impregnated with a bicarbonate solution to inhibit microbial growth during storage. Remove both the salt solution and any residual extractable chemicals used in the polymerization process before use. Prepare the resin by a series of water and organic extractions, followed by careful drying.

2.0 Extraction

2.1 You may perform the extraction using a Soxhlet extractor or other apparatus that generates resin meeting the requirements in Section 13.14 of OTM 45. Use an all-glass thimble containing an extra-coarse frit for extraction of the resin. The frit is recessed 10-15 mm above a crenellated ring at the bottom of the thimble to facilitate drainage. Because the resin floats on methylene chloride, carefully retain the resin in the extractor cup with a glass wool plug and stainless-steel screen. This process involves sequential extraction with the following recommended solvents in the listed order.

- Water initial rinse: Place resin in a suitable container, soak for approximately 5 min with Type II water, remove fine floating resin particles and discard the water. Fill with Type II water a second time, let stand overnight, remove fine floating resin particles and discard the water.
- Hot water: Extract with water for 8 hr.
- Methyl alcohol: Extract for 22 hr.
- Methylene chloride: Extract for 22 hr.
- 5% ammonium hydroxide in methanol: Extract for 22 hr.

Note: You may store the resin in a sealed glass container filled with 5% ammonium hydroxide in methanol prior to the final extraction.

2.2 You may use alternative extraction procedures to clean large batches of resin. Any size extractor may be constructed; the choice depends on the needs of the sampling programs. The resin is held in a glass or stainless-steel cylinder between a pair of coarse and fine screens. Spacers placed under the bottom screen allow for even distribution of clean solvent. Clean solvent is circulated through the resin for extraction. A flow rate is maintained upward through the resin to allow maximum solvent contact and prevent channeling.

2.2.1 Experience has shown that 1 mL/g of resin extracted is the minimum necessary to extract and clean the resin. The aqueous rinse is critical to the subsequent organic rinses and may be accomplished by simply flushing the canister with about 1 liter of distilled water for every 25 g

of resin. A small pump may be useful for pumping the water through the canister. You should perform the water extraction at the rate of about 20 to 40 mL/min.

2.2.2 All materials of construction are glass, PTFE, or stainless steel. Pumps, if used, should not contain extractable materials.

3.0 Drying

3.1 Dry the adsorbent of extraction solvent before use. This section provides a recommended procedure to dry adsorbent that is wet with solvent. However, you may use other procedures if the cleanliness requirements in Sections 13.2 and 13.14 are met.

3.2 Drying Column. A simple column with suitable retainers, as shown in Figure A–2, will hold all the XAD-2 from the extractor shown in Figure A–1 or the Soxhlet extractor, with sufficient space for drying the bed while generating a minimum backpressure in the column.

3.3 Drying Procedure: Dry the adsorbent using clean inert gas. Liquid nitrogen from a standard commercial liquid nitrogen cylinder has proven to be a reliable source of large volumes of gas free from organic contaminants. You may use high-purity tank nitrogen to dry the resin. However, you should pass the high-purity nitrogen through a bed of activated charcoal approximately 150 mL in volume prior to entering the drying apparatus.

3.3.1 Connect the gas vent of a liquid nitrogen cylinder or the exit of the activated carbon scrubber to the column by a length of precleaned copper tubing (e.g., 0.95 cm ID) coiled to pass through a heat source. A convenient heat source is a water bath heated from a steam line. The final nitrogen temperature should only be warm to the touch and not over 40 °C.

3.3.2 Allow the solvent to drain from the resin prior to placing the resin in the drying apparatus.

3.3.3 Flow nitrogen through the drying apparatus at a rate that does not fluidize or agitate the resin. Continue the nitrogen flow until the residual solvent is removed.

Note: Experience has shown that about 500 g of resin may be dried overnight by consuming a full 160-L cylinder of liquid nitrogen.

4.0 Quality Control Procedures

4.1 Report quality control results for the batch. Re-extract the batch if the residual extractable organics fail the criteria in Section 9.

4.2 Residual Quality Check. If adsorbent resin is cleaned or recleaned by the laboratory, perform a quality control check for residual PFAS. Analyze a portion of each batch of cleaned XAD-2 as you would a laboratory sample matrix blank (LSMB) in OTM 45. If the adsorbent exceeds the QC criteria the batch must be re-extracted with 4 % ammonium hydroxide in methanol, dried and reanalyzed.




Figure A-1. XAD-2 fluidized-bed drying apparatus