



Suggested Revisions to Primary and Secondary Compliance Testing Methods and Protocols

David M. Harmon
Technical Manager
Advanced Testing Services


23 August 2011

Background


- CARB Website lists 916 line items for certified mills (6/22/11) and 99 line items for NAF/ULEF Exemptions (6/24/11).
 - There are significant questions about the overall validity of the testing behind these certifications and exemptions. (See Slide 3)
 - Mills and downstream customers concerned with the spread of results among labs.
- The formaldehyde emissions classifications and limitations for the covered composite wood products are performance-based.
 - Most critical at determination of NAF/ULEF Exempt (See Slide 4)
- All labs performing compliance testing must be able to deliver accurate and reproducible results.
 - Differentiate between 0.05 and 0.04 ppm.
 - Should be able to determine concentration to ± 0.005 ppm.
 - Higher quality lab performance must be defined in the regulation, and enforceable.

ILCS 1 – Plots of Lab Average/Test Type

Normal Quantile Plot - Lg. Chamber Data




Normal Quantile Plot - Sm. Ch Data



23 August 2011

Title 17 CCR § 93120 *et.seq.* – Critical Limits



23 August 2011

Recommended Test Method Changes – (1)



Changes compared to existing test methods. Blanks indicate no change. The "TPC Changes" are for compliance testing by qualified laboratories.		E1333		D6007		
		Existing Standard	Proposed Changes	Existing Standard	QC Changes	TPC Changes
Conditioning:	Units					
Time to Condition Target	Day	7		----		7
Time to Condition Target	Hour	----		2		----
Condition Time Range Allowance	Hour	± 3		----		± 3
Condition Time Range Allowance	Min.	----		± 15		----
Temperature Target	°F	75	77	75	77	77
Temperature Range Allowance	°F	± 5	± 3	± 5	± 3	± 3
RH Target	%	50		50		
RH Range Allowance	%	± 5	± 4	± 5	± 4	± 4
Background Formaldehyde	ppm	≤ 0.1	≤ 0.020	≤ 0.1	≤ 0.020	≤ 0.020
Air Changes per Hour past sample	ACH	N/S	≥ 1.00	N/S	≥ 1.00	≥ 1.00
Test Chamber:						
Minimum Chamber Volume	m ³	22		0.02		0.060
Maximum Chamber Volume	m ³	N/S		1		1.000
Air Changes per Hour, Minimum	ACH	0.5	0.500	Q/A based		0.500 OR
Air Changes per Hour, Maximum	ACH	0.5	0.500	Q/A based		1.000 ←
Air Change Allowance	ACH	± 0.05	± 0.010	N/S	± 0.010	± 0.010
Pre-test time in chamber	Hour	18 ± 2		Steady State	MIN 3 air changes	18 ± 2
Temperature Target	°F	77		77		
Temperature Range Allowance	°F	± 2		± 2		
RH Target	%	50		50		
RH Range Allowance	%	± 4		± 4		
Background Formaldehyde	ppm	≤ 0.02	≤ 0.010	≤ 0.02	≤ 0.010	≤ 0.010
Positive pressure required	Pa	≥ 1		N/S	≥ 1	≥ 1

Recommended Test Method Changes – (2)



Changes compared to existing test methods. Blanks indicate no change. The "TPC Changes" are for compliance testing by qualified laboratories.		E1333		D6007			
		Existing	Proposed	Existing	QC	TPC	
Product Loading Ratios ± 2%						@ 0.5 ACH	@ 1.0 ACH
Hardwood Plywood Wall Paneling	m ² /m ³	0.95		N/A		0.95	1.90
Particleboard Flooring Materials	m ² /m ³	0.43		N/A		0.43	0.86
Industrial Particleboard Panels							
Industrial Hardwood Plywood Panels							
Medium Density Fiberboard (MDF)	m ² /m ³	0.26		N/A		0.26	0.52
Particleboard Door Core	m ² /m ³	N/S	0.13	N/A		0.13	0.26
Product Q/A Ratios ± 2%							
Hardwood Plywood Wall Paneling	m/h	N/A		0.526		N/A	
Particleboard Flooring Materials	m/h	N/A		1.173		N/A	
Industrial Particleboard Panels							
Industrial Hardwood Plywood Panels							
Medium Density Fiberboard (MDF)	m/h	N/A		1.905		N/A	
Particleboard Door Core	m/h	N/A		3.846		N/A	
Air Sampling & Testing							
Pre-sampling line purge time	min	5		1			
Collection Rate (Liquid Trap MAX)	L/min	1	1.000	1	1.000	1.000	
Collection Rate - DNPH (up to)	L/min	N/A	1.500	N/A	1.500	1.500	
Collection Rate Allowance	L/min	± 0.05	± 0.010	± 0.05	± 0.010	± 0.010	
Collection Time	min	≥ 60		30	≥ 30	≥ 60	
Collection Time Allowance	sec	± 5		± 5			
Simultaneous Samples Collected	number	2		N/A			
Reagent blank vs DI max difference	Absorbance	0.040	0.005	0.100	0.005	0.005	
Required agreement between matched air samples	ppm	≤ 0.03	0.005				
Report							
Emissions, round to	ppm	0.01	0.001	0.01	0.001	0.001	
Emission Factor, round to	mg/(m ² •h)	0.001		N/S		0.001	

Recommendations for TPC labs using E1333 and D6007 methods –(3)



- All laboratories should participate in a proficiency demonstration under the “Tedlar Bag Protocol”, and show that they are capable of achieving results that are within ± 0.005 ppm of the standard air concentration.
 - Base solutions for standard air concentrations to be supplied as unknowns suitable for the above protocol, by the controlling agency/designee.
- Require that all test results (E1333 and D6007): [Also see slide 12.]
 - Be reported to the third decimal place.
 - Have an in-lab standard deviation of ≤ 0.010 ppm, and
 - There also needs to be a reasonable accuracy requirement for interlaboratory comparison study results. Perhaps this may also be in terms of a standard deviation of ≤ 0.010 ppm.

[Note: More clarity and detail will be provided separately.]


Recommendations for TPC labs using E1333 and D6007 methods –(4)



- All flow measurement by devices calibrated to NIST standards.
 - Use of mass flow controllers recommended.
 - Eliminates need for barometric pressure correction.
 - Easily data logged to corrected (STP) volume collection.
 - Significant improvement over conventional dry gas meter measurement for air change per hour/ air volume.
 - Removes ambiguity from allowing use of 'bubble tubes' as calibration device for sample volume collection equipment.
 - With data logging, eliminates need to use calibrated stopwatch to calculate collected volume. Also help eliminate calculation errors.

23 August 2011

Is the “Tedlar Protocol Feasible? Collection/Analytical Methods Study



23 August 2011

Large and Small Scale Test - Equality

- Current Equivalency demonstration requirement allows too much variation, and does not allow for alternative small scale testing that is statistically equal to large scale results.
- Requirement to test in two different ranges is becoming very difficult, as most products today are right around 0.07 ppm and lower.
- See recap of 53 sets of equivalency tests below, plus statistical analyses.

Combined Low and Intermediate Ranges

E1333 LSCT ppm	D6007 3 Avg ppm	3 Avg Diff. ppm	D6007 Single ppm	Single Diff. ppm
Average		0.005		0.000
Std. Dev.		0.008		0.010
Equiv. Value		0.012		0.009
Max Equiv. Value		≤ 0.026		≤ 0.026

Regression Analysis: Large Chamber E1333 ppm versus Small Chamber D6007 Single Sample (S1) ppm

The regression equation is
 $LC\text{-}ppm = 0.00499 + 0.964 \text{ SC}(1)\text{-}ppm$

Predictor	Coef	SE Coef	T	P
Constant	0.004994	0.002811	1.78	0.082
SC(1)-ppm	0.96399	0.03965	24.31	0.000

S = 0.0100127 R-Sq = 92.1% R-Sq(adj) = 91.9%

Analysis of Variance

Source	DF	SS	MS	F	P
Regression	1	0.059261	0.059261	591.11	0.000
Residual Error	51	0.005113	0.000100		
Total	52	0.064374			

23 August 2011

D6007 Equivalency versus Equality

- Current equivalency demonstration requirement allows too much variation, and the CARB regulation does not allow for alternative small scale testing that is statistically equal to large scale chamber results.
 - Requirement to test in two different ranges is becoming very difficult, as most products today are right around 0.07 ppm and lower.
 - Appropriate single sample small chamber tests are statistically suitable as an alternate secondary method, in place of the “three sets of three” currently required by CARB.
- Repeatability and reproducibility of results should be limited to a standard deviation ≤ 0.010 ppm.
 - For both E1333 and D6007 testing of matched samples.

Are In-lab and Interlab Requirements Feasible?

Individual Test Results					Inter-laboratory Deviation				
Test	Lab 1	Lab 2	Lab 3		Test	Lab 1	Lab 2	Lab 3	
1	0.100	0.122	0.110		1	-0.015	0.007	-0.005	
2	0.098	0.121	0.110		2	-0.017	0.006	-0.005	
3	0.119	0.119	0.120		3	0.004	0.004	0.005	
4	0.110	0.121	0.120		4	-0.005	0.006	0.005	
5	0.117	0.118	0.110		5	0.002	0.003	-0.005	
6	0.117	0.117	0.120		6	0.002	0.002	0.005	
7	0.117	0.117	0.120		7	0.002	0.002	0.005	
8	0.116	0.116	0.120		8	0.001	0.001	0.005	
9	0.109	0.111	0.110		9	-0.006	-0.004	-0.005	
10	0.111	0.113	0.110		10	-0.004	-0.002	-0.005	
				GRAND					GRAND
Average	0.111	0.118	0.115	0.115	Average	-0.003	0.003	0.000	0.000
S. D.	0.007	0.004	0.005	0.006	S. D.	0.007	0.004	0.005	0.006
In-Lab Deviation									
Test	Lab 1	Lab 2	Lab 3						
1	-0.011	0.004	-0.005						
2	-0.013	0.003	-0.005						
3	0.008	0.001	0.005						
4	-0.001	0.003	0.005						
5	0.006	0.000	-0.005						
6	0.006	-0.001	0.005						
7	0.006	-0.001	0.005						
8	0.005	-0.002	0.005						
9	-0.002	-0.007	-0.005						
10	0.000	-0.005	-0.005						
				GRAND					
Average	0.000	0.000	0.000	0.000					
S. D.	0.007	0.004	0.005	0.005					

23 August 2011

Safe Harbor

- Incent importers to self-test the goods they are bringing in, by offering a safe harbor for their appropriate due diligence.
 - Testing to be done by an approved laboratory that meets suggested accuracy and precision requirements.
 - Structure so that it lowers financial risk of importers.
 - Importers using this can then effectively enforce the regulation through their purchasing power, which transcends regulatory jurisdictional boundaries on a global basis.
 - Reduce burden on regulatory enforcement agencies.
- Details will require much additional input from importers, and others along the chain of commerce.
- Deconstructive testing method must be finalized and shared with all concerned parties.
 - Protocol for deconstruction
 - Clear definition of what range of results are acceptable to demonstrate compliance.

23 August 2011